

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM010524\
 Data File : BM043759.D
 Acq On : 05 Jan 2024 14:24
 Operator : MA/JU
 Sample : PB158241BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK241

Quant Time: Jan 06 01:18:59 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM121523.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 29 22:31:57 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.952	152	305409	20.000	ng/u1	0.00
20) Naphthalene-d8	10.763	136	1201255	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.598	164	652082	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.345	188	1293150	20.000	ng/u1	0.00
79) Chrysene-d12	21.533	240	1189766	20.000	ng/u1	0.01
88) Perylene-d12	23.950	264	1400196	20.000	ng/u1	0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.364	96	47967	5.105	ng/uL	0.00
4) Pyridine-d5	3.787	84	664373	24.565	ng/u1	0.00
7) Phenol-d5	7.128	99	760986	21.903	ng/u1	0.01
9) Bis-(2-Chloroethyl)eth...	7.293	67	491072	22.318	ng/u1	0.00
11) 2-Chlorophenol-d4	7.481	132	617334	23.272	ng/u1	0.00
15) 4-Methylphenol-d8	8.675	113	574010	21.291	ng/u1	0.00
21) Nitrobenzene-d5	9.134	128	290870	25.027	ng/u1	0.00
24) 2-Nitrophenol-d4	9.851	143	319788	26.630	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.393	165	528165	24.915	ng/u1	0.01
31) 4-Chloroaniline-d4	10.916	131	788612	25.286	ng/u1	0.00
46) Dimethylphthalate-d6	14.016	166	1494827	25.465	ng/u1	0.00
49) Acenaphthylene-d8	14.292	160	1804854	25.919	ng/u1	0.00
54) 4-Nitrophenol-d4	14.828	143	258684	23.283	ng/u1	0.04
60) Fluorene-d10	15.592	176	1277438	26.317	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.727	200	176620	20.651	ng/u1	0.02
73) Anthracene-d10	17.445	188	1939039	26.737	ng/u1	0.00
81) Pyrene-d10	19.739	212	2182709	23.237	ng/u1	0.01
92) Benzo(a)pyrene-d12	23.792	264	2294063	26.550	ng/u1	0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM010524\
Data File : BM043759.D
Acq On : 05 Jan 2024 14:24
Operator : MA/JU
Sample : PB158241BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK241

Quant Time: Jan 06 01:18:59 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM121523.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Dec 29 22:31:57 2023
Response via : Initial Calibration

