


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM010719\
 Data File : BM018435.D
 Acq On : 07 Jan 2019 09:33
 Operator : JU/SJ
 Sample : DFTPP01
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP01

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM010719MA.M
 Title : SVOA CALIBRATION
 Last Update : Mon Jan 07 23:55:53 2019



AutoFind: Scans 2499, 2500, 2501; Background Corrected with Scan 2492

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.5	541822	PASS
68	69	0.00	2	2.0	11555	PASS
69	198	0.00	100	33.9	584677	PASS
70	69	0.00	2	0.5	2823	PASS
127	198	10	80	45.8	789226	PASS
197	198	0.00	2	0.5	8461	PASS
198	198	100	100	100.0	1722538	PASS
199	198	5	9	6.7	114554	PASS
275	198	10	60	28.3	487680	PASS
365	198	1	100	4.2	72656	PASS
441	198	0.01	100	14.3	246122	PASS
442	198	50	100	93.9	1616725	PASS
443	442	15	24	19.1	308714	PASS

m/z	Abundance
39.00	587.0
40.00	1648.0
41.00	1674.0
42.00	526.0
43.00	1675.0
44.00	18256.0
45.10	785.0
49.00	550.0
51.10	343.0
55.00	1115.0
56.00	548.0
57.10	929.0
67.10	448.0
69.10	539.0
70.00	448.0
71.00	1066.0
73.10	2623.0
77.20	717.0
77.90	311.0
78.90	467.0
81.10	355.0
83.20	525.0
84.00	723.0
85.00	373.0
85.90	376.0
91.00	371.0
94.90	540.0
95.90	449.0
97.00	575.0
99.20	352.0
105.20	665.0
111.00	407.0
133.00	418.0
135.20	566.0
147.10	1482.0
148.80	367.0
149.90	303.0
157.00	332.0
177.10	467.0
191.10	378.0
193.00	952.0
207.00	3038.0
208.00	599.0
209.00	585.0
221.00	779.0
221.80	300.0
267.00	374.0
281.10	1127.0
282.10	489.0
341.00	493.0
355.00	693.0

Instrument :
BNA_M
ClientSampleId :
DFTPP01

m/z	Abundance
36.10	426.0
37.00	2884.0
38.10	9757.0
39.10	57280.0
40.00	4160.0
41.00	2219.0
41.90	650.0
43.10	1921.0
44.00	18488.0
45.10	1921.0
46.00	437.0
49.10	5708.0
50.10	179392.0
51.00	696576.0
52.10	33624.0
53.00	1910.0
55.00	4735.0
56.00	21576.0
57.00	48056.0
58.00	2125.0
58.90	644.0
60.20	635.0
61.00	8808.0
62.00	9689.0
63.00	29136.0
64.00	4860.0
65.10	13637.0
66.00	1348.0
67.00	1323.0
68.00	14853.0
69.00	734400.0
70.00	4594.0
71.00	1020.0
73.10	8105.0
74.00	76848.0
75.00	119128.0
76.10	37888.0
77.10	814400.0
78.10	53896.0
79.00	59200.0
80.00	46336.0
81.00	59584.0
82.00	15456.0
83.00	13937.0
84.00	1730.0
85.00	8715.0
86.00	17184.0
87.00	8845.0
88.00	3105.0
89.00	1757.0
91.00	12774.0
92.00	17168.0
93.00	101064.0
94.00	7515.0
95.00	1899.0
96.10	4644.0
97.10	2505.0
98.00	77376.0
99.00	60944.0
100.00	4394.0
101.00	36152.0
102.00	2038.0
103.00	10818.0
104.00	23056.0
105.00	21712.0
106.00	5978.0
107.00	271808.0
108.00	40088.0
109.10	7981.0
110.00	475008.0
111.00	75936.0
112.00	8354.0
113.00	2741.0
113.80	452.0
114.30	503.0
115.00	1405.0
116.10	16352.0
117.00	234944.0
118.00	18488.0
118.90	2222.0
120.10	2917.0
120.90	1282.0
122.00	19048.0
123.00	27448.0
124.00	12821.0
125.00	13258.0
127.00	959040.0
128.00	71168.0
129.00	373184.0
130.00	31688.0
131.00	5882.0

Instrument :
BNA_M
ClientSampleId :
DFTPP01

132.20	2805.0
133.00	1670.0
134.00	11601.0
135.00	30888.0
136.00	13045.0
137.00	16696.0
138.00	3969.0
138.90	2218.0
139.90	6085.0
141.00	52736.0
142.00	15393.0
143.00	10076.0
144.00	3173.0
145.10	2618.0
146.00	10070.0
147.00	27416.0
148.00	59504.0
149.00	11962.0
150.00	2923.0
151.10	6680.0
151.90	3608.0
153.00	16816.0
154.00	12416.0
155.00	26528.0
156.00	42440.0
157.00	7435.0
158.00	8562.0
159.00	7478.0
160.00	17304.0
161.00	23712.0
162.00	6059.0
163.10	2236.0
164.00	1798.0
165.00	21368.0
166.00	14297.0
167.00	103168.0
168.00	50488.0
169.00	9308.0
170.10	3905.0
171.10	5021.0
172.00	8932.0
173.00	13956.0
174.00	20296.0
175.00	41712.0
176.00	9643.0
177.00	19680.0
178.00	6632.0
179.00	79896.0
180.00	54696.0
181.00	24360.0
182.00	4569.0
183.00	1935.0
184.00	7343.0
185.00	36912.0
186.00	292992.0
187.00	85496.0
188.10	8593.0
189.00	19064.0
190.00	3493.0
191.00	9335.0
192.00	25520.0
193.00	27208.0
194.00	6361.0
194.90	2771.0
196.00	54704.0
197.00	9476.0
198.00	1953280.0
199.00	135296.0
200.00	10612.0
201.40	7785.0
203.00	14985.0
204.00	79832.0
205.00	128648.0
206.00	556672.0
207.00	72800.0
208.00	20184.0
209.00	6398.0
210.20	9521.0
211.10	20048.0
212.10	3647.0
213.00	1393.0
214.00	829.0
215.00	6800.0
216.00	13701.0
217.00	150592.0
218.00	19168.0
219.00	2321.0
221.00	77296.0
221.80	20144.0
223.00	32504.0
224.00	305984.0
225.00	81696.0
226.10	7890.0
227.00	132800.0
228.00	19376.0

Instrument :
BNA_M
ClientSampleId :
DFTPP01

229.00	28056.0
229.90	4620.0
231.00	10306.0
232.00	1670.0
233.00	2386.0
234.00	8352.0
235.00	10722.0
236.00	6344.0
237.00	11668.0
238.00	2375.0
239.00	5614.0
240.00	3966.0
241.00	6926.0
242.00	14867.0
243.10	15987.0
244.00	220032.0
245.10	30456.0
246.00	51200.0
247.00	10007.0
248.00	1793.0
249.00	7499.0
249.90	1620.0
251.00	2282.0
252.00	1812.0
253.00	4913.0
254.10	8189.0
255.00	1208832.0
256.00	179776.0
257.00	15716.0
258.00	74376.0
259.00	11268.0
259.90	1967.0
261.00	1507.0
262.00	308.0
263.10	470.0
264.00	1968.0
265.00	30656.0
266.10	3932.0
267.10	916.0
267.90	456.0
269.90	1180.0
271.00	3008.0
272.10	3336.0
273.00	35096.0
274.00	92264.0
275.00	519296.0
276.00	71192.0
277.00	49104.0
278.00	7525.0
278.90	1931.0
281.10	1424.0
282.00	1360.0
283.00	5107.0
284.00	3317.0
285.00	8008.0
286.10	1873.0
288.00	399.0
289.10	1402.0
290.00	1757.0
291.10	1297.0
292.00	2078.0
293.00	9741.0
294.00	3162.0
295.10	3455.0
296.00	155456.0
297.00	21648.0
298.00	1652.0
299.00	594.0
300.90	1667.0
302.00	2357.0
303.00	16992.0
304.00	5144.0
304.80	817.0
306.00	304.0
308.00	2499.0
308.90	1203.0
310.00	1777.0
311.10	502.0
312.10	800.0
312.90	1324.0
314.00	7429.0
315.00	17640.0
316.00	8588.0
317.00	1999.0
320.20	766.0
321.00	4623.0
322.00	2678.0
323.10	45408.0
324.00	8942.0
324.90	946.0
326.00	1296.0
327.00	9071.0
328.00	4877.0
329.10	674.0
332.00	3379.0

Instrument :
BNA_M
ClientSampleId :
DFTPP01

333.10	3855.0
334.00	32120.0
335.10	7773.0
336.00	1307.0
338.80	836.0
340.00	458.0
341.10	5449.0
342.00	1429.0
343.00	438.0
346.00	10354.0
347.00	1851.0
347.90	382.0
351.00	1135.0
352.00	14118.0
353.00	9703.0
354.10	14822.0
355.10	2897.0
356.20	303.0
357.10	461.0
358.20	342.0
359.00	1133.0
360.90	520.0
364.10	755.0
365.00	68968.0
366.00	9422.0
367.00	784.0
370.00	1360.0
371.10	3656.0
372.00	21712.0
373.00	6703.0
374.00	776.0
376.90	478.0
383.00	6538.0
384.00	2212.0
385.10	525.0
390.00	3058.0
391.00	1839.0
392.10	1574.0
393.10	490.0
400.90	1569.0
402.00	9918.0
403.00	13780.0
404.10	5091.0
405.00	986.0
409.90	351.0
415.00	935.0
421.00	11724.0
422.00	11895.0
423.10	76840.0
424.00	18776.0
425.10	2232.0
437.90	392.0
439.00	648.0
439.70	366.0
440.20	418.0
441.00	218304.0
442.10	1398784.0
443.10	273536.0
444.10	26064.0
445.00	1309.0

Instrument :
BNA_M
ClientSampleId :
DFTPP01

m/z	Abundance
35.90	464.0
37.10	2871.0
38.10	8462.0
39.10	51264.0
40.00	3887.0
41.10	2374.0
42.10	525.0
43.10	1730.0
44.00	17224.0
45.00	1782.0
47.90	623.0
49.10	5176.0
50.10	164416.0
51.10	622080.0
52.10	32456.0
53.00	1423.0
55.00	4238.0
56.00	18752.0
57.00	44808.0
58.00	2598.0
59.00	832.0
60.00	755.0
61.00	8184.0
62.00	9402.0
63.00	26392.0
64.10	2975.0
65.10	13257.0
66.00	890.0
67.00	1165.0
68.10	13431.0
69.00	676160.0
70.00	3447.0
71.10	1068.0
72.10	446.0
73.00	8032.0
74.00	73160.0
75.00	112928.0
76.10	36688.0
77.10	744384.0
78.10	46680.0
79.00	54768.0
80.00	42288.0
81.00	54584.0
82.00	13283.0
83.00	12357.0
84.00	1964.0
85.00	9683.0
86.10	15415.0
87.00	7040.0
88.00	2666.0
89.10	1893.0
90.00	552.0
91.00	12359.0
92.00	14561.0
93.00	93304.0
94.00	6817.0
95.20	1723.0
96.00	4010.0
97.10	2146.0
98.00	76624.0
99.00	58208.0
100.10	4621.0
101.00	32344.0
102.00	1994.0
103.00	11274.0
104.00	23224.0
105.00	19504.0
106.10	7122.0
107.00	256064.0
108.00	38904.0
109.10	8618.0
110.00	452544.0
111.00	71288.0
112.00	7749.0
113.00	3556.0
114.20	758.0
115.00	1082.0
116.00	13955.0
117.00	223168.0
118.00	16383.0
119.00	1652.0
120.00	3564.0
120.90	1464.0
122.00	19848.0
123.00	25840.0
124.00	11762.0
125.00	11921.0
126.20	5256.0
127.00	918336.0
128.00	67560.0
129.00	370752.0

Instrument :
BNA_M
ClientSampleId :
DFTPP01

130.00	31192.0
131.00	7047.0
132.20	3011.0
132.90	1918.0
134.00	11071.0
135.00	30008.0
136.00	11241.0
137.00	14976.0
137.90	3053.0
138.90	2092.0
140.10	6160.0
141.00	50360.0
142.00	16544.0
143.00	11737.0
144.00	2782.0
145.00	2932.0
146.00	9378.0
147.00	25928.0
148.00	58400.0
149.00	12707.0
150.00	3109.0
151.20	6097.0
152.00	3440.0
153.00	17128.0
154.00	12629.0
155.10	28392.0
156.00	40576.0
157.10	9157.0
158.00	9100.0
159.00	8086.0
160.00	16760.0
161.00	23808.0
161.90	6517.0
163.00	1870.0
164.10	2189.0
165.00	18496.0
166.00	15761.0
167.00	103088.0
168.00	51680.0
169.10	9910.0
170.10	2666.0
170.90	4871.0
172.00	10984.0
173.00	11298.0
174.00	23248.0
175.00	44208.0
176.00	11600.0
177.00	19384.0
178.00	7517.0
179.00	80096.0
180.00	54664.0
181.00	25120.0
182.10	3967.0
182.90	2034.0
184.00	4974.0
185.00	39640.0
186.00	301568.0
187.00	85144.0
188.00	9687.0
189.00	20416.0
189.90	2699.0
191.00	11304.0
192.00	24336.0
193.00	29392.0
194.00	6331.0
195.00	3065.0
196.00	56104.0
197.00	9827.0
198.00	2038784.0
199.00	133440.0
200.00	12079.0
201.40	8638.0
203.00	16228.0
204.00	82408.0
205.00	142784.0
206.00	592896.0
207.00	80984.0
208.00	19992.0
209.00	6241.0
210.20	9147.0
211.00	22904.0
213.00	1162.0
214.10	779.0
215.00	7143.0
216.00	13112.0
217.00	163840.0
218.00	21616.0
219.00	1373.0
221.10	76864.0
221.90	18632.0
223.00	37376.0
224.00	327680.0
225.00	83440.0
226.00	9840.0
227.00	143488.0

Instrument :
BNA_M
ClientSampleId :
DFTPP01

228.00	21312.0
229.00	28688.0
230.00	4472.0
231.00	12209.0
231.90	1690.0
233.10	2829.0
234.00	9234.0
235.00	10668.0
236.00	6067.0
237.00	11215.0
238.00	1912.0
239.00	5635.0
240.00	4553.0
241.00	7855.0
242.00	17552.0
243.10	18736.0
244.00	239808.0
245.00	29872.0
246.00	53552.0
247.00	10948.0
248.00	2410.0
249.00	9000.0
249.90	2065.0
250.90	2290.0
252.00	2243.0
253.00	5486.0
254.10	9587.0
255.00	1361920.0
256.00	198336.0
257.00	15243.0
258.00	81488.0
259.00	13504.0
260.00	2175.0
261.00	1944.0
263.00	826.0
263.90	2488.0
265.00	34536.0
266.00	4507.0
267.10	754.0
268.00	489.0
270.00	1960.0
271.00	3258.0
272.00	3682.0
273.00	39704.0
274.00	99848.0
275.00	575552.0
276.00	74856.0
277.00	53408.0
278.00	8456.0
278.90	1890.0
280.00	453.0
281.00	1856.0
282.00	1470.0
283.10	5258.0
284.00	3480.0
285.10	8875.0
286.00	1467.0
288.00	886.0
289.00	1831.0
290.00	1693.0
291.00	1118.0
292.00	2358.0
293.00	11115.0
294.00	3597.0
295.10	3723.0
296.00	180352.0
297.00	23536.0
298.00	1684.0
298.90	419.0
301.00	2558.0
302.10	4104.0
303.00	20360.0
304.00	5396.0
305.10	577.0
306.00	449.0
307.10	466.0
308.00	2704.0
309.00	2083.0
310.00	1949.0
311.00	833.0
312.00	664.0
313.00	1579.0
314.00	9618.0
315.00	20200.0
316.00	10998.0
317.00	1736.0
319.90	696.0
321.00	5239.0
322.10	2731.0
323.10	54928.0
324.00	10182.0
325.00	1063.0
326.00	1350.0
327.00	10996.0
328.00	5354.0

Instrument :
BNA_M
ClientSampleId :
DFTPP01

328.90	1236.0
330.00	374.0
332.10	4644.0
333.00	5833.0
334.00	35528.0
335.00	10460.0
336.10	1001.0
339.10	951.0
340.10	899.0
341.00	6503.0
342.10	1710.0
342.90	358.0
344.80	465.0
346.00	12379.0
347.00	3410.0
349.90	522.0
350.90	1133.0
352.00	19608.0
353.00	12385.0
354.00	17104.0
355.00	3201.0
356.20	499.0
357.10	330.0
358.90	1150.0
360.00	346.0
360.80	311.0
362.90	347.0
363.90	633.0
364.20	724.0
365.00	85936.0
366.00	12694.0
366.90	714.0
370.00	1885.0
371.00	4777.0
372.10	28456.0
373.00	7135.0
374.10	845.0
377.10	588.0
383.00	8312.0
384.00	2177.0
384.90	660.0
390.00	3921.0
391.00	2296.0
392.00	1827.0
393.00	403.0
397.00	315.0
401.00	1230.0
402.00	12760.0
403.00	17384.0
404.00	6111.0
405.20	882.0
410.10	437.0
415.10	962.0
421.00	14898.0
422.00	15336.0
423.10	103968.0
424.10	22616.0
425.00	2683.0
429.00	364.0
438.90	699.0
440.20	1048.0
441.10	299648.0
442.10	1979392.0
443.10	377792.0
444.10	35248.0
445.00	2006.0
475.10	307.0

Instrument :

BNA_M

ClientSampleId :

DFTPP01

m/z	Abundance
35.90	382.0
37.10	1543.0
38.00	4031.0
39.10	24592.0
40.00	3072.0
41.00	1359.0
42.00	381.0
43.00	1958.0
44.00	18600.0
45.00	1091.0
49.10	2368.0
50.10	79376.0
51.10	307840.0
52.10	14399.0
53.00	1070.0
55.00	2512.0
56.00	9512.0
57.00	23128.0
58.00	1278.0
59.10	369.0
60.00	533.0
61.00	3917.0
62.00	5047.0
63.10	12770.0
64.10	1957.0
65.10	6861.0
66.10	524.0
66.90	989.0
68.00	6382.0
69.00	345088.0
70.10	1774.0
71.00	758.0
73.00	5010.0
74.00	36224.0
75.00	54056.0
76.10	18520.0
77.10	372800.0
78.10	24712.0
79.00	27280.0
80.00	20544.0
81.00	30064.0
82.00	6391.0
83.00	6911.0
84.00	1534.0
85.10	4837.0
86.00	9616.0
87.10	3722.0
88.10	2006.0
89.00	864.0
91.00	7902.0
92.00	7411.0
93.00	52608.0
94.10	3150.0
95.10	1462.0
96.00	2165.0
97.20	1195.0
98.00	39416.0
99.00	30688.0
100.00	2617.0
101.00	16792.0
101.90	765.0
103.00	5526.0
104.00	10594.0
105.00	10288.0
106.00	3729.0
107.00	137088.0
108.00	21016.0
109.00	3665.0
110.00	233856.0
111.00	36040.0
112.00	4571.0
113.00	1185.0
114.20	553.0
115.00	804.0
116.00	7629.0
117.00	130448.0
118.00	8485.0
119.00	1121.0
120.10	1726.0
120.90	542.0
122.00	10539.0
123.00	15507.0
124.00	7145.0
125.00	5460.0
126.10	2740.0
127.00	490304.0
128.00	36608.0
129.00	201472.0
130.00	16920.0
131.10	3186.0
132.10	1676.0

Instrument :
BNA_M
ClientSampleId :
DFTPP01

132.90	1310.0
133.90	5806.0
135.00	19696.0
136.00	6615.0
137.00	8591.0
137.80	2086.0
139.00	1224.0
139.90	2776.0
141.00	28584.0
142.00	8991.0
143.00	6320.0
144.00	1832.0
145.00	1052.0
146.00	4638.0
147.00	14578.0
148.00	35472.0
149.00	7135.0
149.90	1895.0
151.10	3816.0
151.90	2095.0
153.00	9696.0
154.00	6365.0
155.00	15908.0
156.00	22968.0
157.00	4239.0
158.00	5579.0
158.90	3724.0
160.00	8585.0
161.00	13761.0
162.10	3797.0
163.10	1270.0
164.00	1740.0
165.00	10860.0
166.00	9015.0
167.00	59112.0
168.00	30960.0
169.10	4404.0
170.10	2452.0
171.00	2309.0
172.00	5798.0
172.90	7394.0
174.00	12024.0
175.00	24304.0
176.00	5772.0
177.00	11317.0
178.00	3782.0
179.00	47224.0
180.00	30456.0
181.00	15873.0
182.00	2201.0
182.90	1102.0
184.00	3995.0
185.00	21848.0
186.00	172608.0
187.00	48120.0
188.00	5185.0
189.00	12393.0
190.10	1617.0
190.90	5513.0
192.00	16448.0
193.00	16672.0
194.00	2832.0
195.00	1940.0
196.00	34848.0
197.00	6081.0
198.00	1175552.0
198.90	74928.0
200.00	6063.0
201.50	5760.0
203.00	9526.0
204.00	46944.0
205.00	81384.0
206.00	342208.0
207.00	45464.0
208.00	12956.0
209.00	4415.0
210.00	4790.0
211.00	13651.0
211.90	2257.0
213.00	815.0
213.90	511.0
214.90	3566.0
216.10	6966.0
217.00	97184.0
218.00	10797.0
219.10	1039.0
221.00	44864.0
221.90	11539.0
223.00	19456.0
224.00	196160.0
225.00	48824.0
226.00	4819.0
227.00	90440.0
228.10	12376.0
229.00	18416.0

Instrument :
BNA_M
ClientSampleId :
DFTPP01

229.90	2213.0
231.10	7029.0
232.00	743.0
233.00	1759.0
234.00	6208.0
235.00	6577.0
236.00	4437.0
237.00	7259.0
238.00	614.0
239.00	2833.0
240.00	3009.0
240.90	5699.0
242.10	10067.0
243.00	10461.0
244.00	153920.0
245.10	19792.0
246.00	35640.0
247.00	6890.0
248.00	1843.0
249.00	6010.0
250.00	741.0
250.90	1336.0
252.00	1540.0
253.00	3522.0
254.10	4877.0
255.00	830784.0
256.00	116888.0
257.00	8119.0
258.00	51000.0
259.00	8599.0
260.00	1307.0
261.00	1162.0
263.10	521.0
264.10	916.0
265.00	20360.0
266.10	2957.0
266.90	572.0
267.90	516.0
269.10	409.0
269.90	1243.0
271.00	1884.0
271.90	2450.0
273.00	25840.0
274.00	61992.0
275.00	368192.0
276.00	48392.0
277.00	32032.0
278.00	5402.0
278.90	1347.0
281.00	1602.0
282.00	1168.0
283.00	4315.0
284.00	2400.0
285.10	6584.0
286.00	1170.0
288.90	1299.0
290.00	1208.0
291.10	808.0
292.10	1816.0
293.00	7505.0
293.90	1885.0
295.00	2093.0
296.00	122640.0
297.00	16070.0
298.00	932.0
299.00	360.0
300.90	1112.0
302.00	2088.0
303.00	13718.0
304.00	2920.0
305.00	574.0
308.00	2036.0
309.00	1196.0
310.10	1293.0
311.00	399.0
313.00	1233.0
314.00	5354.0
315.00	13583.0
316.00	8084.0
316.90	1378.0
320.00	504.0
321.00	3921.0
322.00	2470.0
323.00	39312.0
324.00	6722.0
324.90	489.0
326.00	977.0
327.00	8236.0
327.90	3818.0
328.90	914.0
332.00	2876.0
333.10	4005.0
334.00	27312.0
335.00	5842.0
336.10	759.0

Instrument :
BNA_M
ClientSampleId :
DFTPP01

339.00	958.0
340.10	444.0
341.00	4911.0
342.10	1422.0
346.00	8924.0
347.00	1628.0
351.00	474.0
352.00	11577.0
353.00	7261.0
354.00	12211.0
355.10	2811.0
356.10	406.0
359.00	913.0
364.00	383.0
365.00	63064.0
366.00	8554.0
366.90	353.0
370.00	1118.0
371.00	3219.0
372.10	19584.0
373.00	5268.0
374.10	664.0
377.00	521.0
383.00	6364.0
383.90	1799.0
385.10	625.0
390.00	2658.0
391.10	2072.0
392.00	1656.0
393.10	352.0
396.90	331.0
401.00	1290.0
402.00	8670.0
403.00	13128.0
404.10	4474.0
404.90	457.0
415.20	675.0
415.90	404.0
421.10	11262.0
422.00	10414.0
423.00	79104.0
424.10	15265.0
425.00	1896.0
438.90	349.0
441.10	220416.0
442.10	1472000.0
443.10	274816.0
444.10	24648.0
445.00	1419.0

Instrument :
BNA_M
ClientSampleId :
DFTPP01