

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM010815\
 Data File : BM000145.D
 Acq On : 08 Jan 2015 15:20
 Operator : TP/IZ
 Sample : SSTD0.422
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Quant Time: Jan 08 16:34:16 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM010715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 08 10:41:47 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	98	-0.01
3	Naphthalene	1.072	1.043	2.7	98	-0.01
4 SURR	2-Methylnaphthalene-d10	0.532	0.520	2.3	98	-0.01
5	2-Methylnaphthalene	0.710	0.694	2.3	98	-0.01
6 I	Acenaphthene-d10	1.000	1.000	0.0	97	0.00
7	Acenaphthylene	1.996	1.976	1.0	100	-0.01
8 C	Acenaphthene	1.247	1.228	1.5	100	-0.02
9	Fluorene	1.433	1.449	-1.1	100	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
11	Pentachlorophenol	0.063	0.047	25.4#	97	-0.02
12	Phenanthrene	1.235	1.173	5.0	98	0.00
13	Anthracene	1.159	1.136	2.0	103	0.00
14 SURR	Fluoranthene-d10	0.862	0.858	0.5	102	0.00
15 C	Fluoranthene	1.355	1.343	0.9	102	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	103	-0.01
17	Pyrene	1.669	1.572	5.8	101	0.00
18	Benzo(a)anthracene	1.324	1.279	3.4	101	-0.01
19	Chrysene	1.444	1.404	2.8	104	-0.01
20 I	Perylene-d12	1.000	1.000	0.0	105	-0.01
21	Benzo(b)fluoranthene	1.574	1.571	0.2	115	-0.01
22	Benzo(k)fluoranthene	1.503	1.377	8.4	94	-0.01
23 C	Benzo(a)pyrene	1.425	1.373	3.6	104	-0.01
24	Indeno(1,2,3-cd)pyrene	1.626	1.560	4.1	105	-0.01
25	Dibenzo(a,h)anthracene	1.293	1.256	2.9	106	-0.01
26	Benzo(g,h,i)perylene	1.444	1.380	4.4	103	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0