


Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM010820\
 Data File : BM024412.D
 Acq On : 08 Jan 2020 09:26
 Operator : JU
 Sample : DFTPP09
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP09

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM010820MA.M
 Title : SVOA CALIBRATION
 Last Update : Thu Jan 09 01:07:25 2020



AutoFind: Scans 2466, 2467, 2468; Background Corrected with Scan 2460

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.1	157884	PASS
68	69	0.00	2	1.9	3078	PASS
69	198	0.00	100	40.0	165504	PASS
70	69	0.00	2	0.3	509	PASS
127	198	10	80	46.7	193402	PASS
197	198	0.00	2	0.6	2682	PASS
198	198	100	100	100.0	413952	PASS
199	198	5	9	6.7	27584	PASS
275	198	10	60	30.0	123994	PASS
365	198	1	100	4.0	16702	PASS
441	198	0.01	100	14.3	59090	PASS
442	198	50	100	95.3	394304	PASS
443	442	15	24	19.2	75720	PASS

m/z	Abundance
35.90	524.0
36.80	312.0
38.10	528.0
39.00	2142.0
40.10	2101.0
41.10	6635.0
42.00	1639.0
43.10	8991.0
44.00	28056.0
45.00	2097.0
46.90	342.0
49.00	320.0
50.10	695.0
50.90	673.0
52.00	497.0
53.00	1075.0
54.00	1448.0
55.10	5328.0
56.20	1561.0
57.10	6359.0
58.20	739.0
59.00	927.0
60.10	1155.0
63.90	365.0
64.90	449.0
65.90	356.0
67.10	2047.0
68.10	884.0
69.00	3138.0
70.00	1858.0
71.10	4098.0
71.90	338.0
73.10	6512.0
74.00	804.0
75.00	938.0
76.00	604.0
77.00	1860.0
78.00	526.0
79.00	2136.0
79.90	425.0
80.90	1766.0
82.00	1589.0
83.00	2482.0
84.00	1308.0
85.10	2330.0
85.90	315.0
86.90	490.0
88.00	416.0
91.00	1527.0
92.00	782.0
93.00	810.0
93.90	311.0
95.10	1353.0
96.10	1341.0
97.10	2688.0
98.10	2437.0
99.10	1286.0
101.00	432.0
102.00	411.0
103.20	874.0
104.00	646.0
105.20	1413.0
106.10	526.0
107.20	1236.0
108.00	429.0
109.00	807.0
110.20	838.0
111.10	1345.0
112.10	802.0
112.90	888.0
115.00	859.0
116.00	405.0
117.10	523.0
119.00	823.0
119.90	336.0
121.10	821.0
121.90	628.0
123.10	315.0
124.90	438.0
125.10	422.0
125.90	456.0
127.00	816.0
128.10	460.0
128.90	827.0
130.90	732.0
133.00	980.0
134.10	337.0
135.00	921.0
137.10	406.0
138.90	449.0
141.20	521.0
142.10	424.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

143.10	439.0
145.20	486.0
146.20	416.0
147.10	2503.0
148.90	797.0
150.00	506.0
151.90	628.0
152.90	683.0
155.00	458.0
156.90	352.0
160.90	411.0
162.00	308.0
163.00	489.0
165.10	525.0
166.00	814.0
167.00	316.0
168.20	379.0
168.90	373.0
170.00	385.0
170.90	362.0
177.90	517.0
178.90	612.0
182.90	323.0
186.00	330.0
189.10	319.0
190.90	1443.0
192.00	378.0
193.00	2725.0
194.00	551.0
194.80	431.0
196.10	335.0
207.00	9369.0
208.00	1480.0
209.10	1509.0
210.90	392.0
213.30	337.0
221.00	1886.0
222.10	1061.0
222.90	872.0
239.20	500.0
267.00	1052.0
268.00	332.0
281.10	1708.0
282.10	592.0
283.00	520.0
295.00	1110.0
296.00	386.0
327.00	417.0
341.10	773.0
342.00	377.0
355.00	938.0
356.20	322.0
429.00	305.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

m/z	Abundance
35.90	541.0
37.10	1395.0
38.00	3589.0
39.00	18168.0
40.00	3347.0
41.10	7102.0
42.00	2177.0
43.10	8138.0
44.00	32328.0
45.00	1681.0
48.20	355.0
49.00	1444.0
50.10	48480.0
51.10	187712.0
52.00	9120.0
53.00	1400.0
54.00	1120.0
55.00	6851.0
56.00	8655.0
57.00	18576.0
58.00	1514.0
59.00	713.0
60.00	1293.0
61.10	2772.0
62.10	2503.0
63.00	9581.0
64.00	1159.0
65.00	3888.0
66.00	755.0
67.00	2529.0
68.10	4087.0
69.00	193024.0
70.00	2770.0
71.10	3956.0
72.20	361.0
73.00	6607.0
74.00	19920.0
75.00	30040.0
76.00	10483.0
77.00	198016.0
78.00	15074.0
79.00	17792.0
80.00	10305.0
81.00	15704.0
82.00	5151.0
83.00	5184.0
84.10	1826.0
85.10	5784.0
86.00	4372.0
87.00	2079.0
88.00	781.0
89.00	730.0
91.00	5293.0
92.00	3717.0
93.00	25712.0
94.00	2482.0
95.10	2201.0
96.10	2842.0
97.10	3489.0
98.00	21320.0
99.00	13613.0
100.00	1451.0
101.00	8173.0
101.90	645.0
102.90	3330.0
103.90	5990.0
105.00	5966.0
106.10	2678.0
107.00	65168.0
108.00	10204.0
109.00	3084.0
110.00	111600.0
111.00	17208.0
112.10	2886.0
113.10	1547.0
113.90	338.0
114.90	809.0
115.20	871.0
116.00	4117.0
117.00	52408.0
118.00	3953.0
119.10	1306.0
120.00	965.0
120.90	1036.0
122.00	5004.0
123.00	7936.0
124.00	2695.0
125.00	2798.0
127.00	212736.0
128.00	18688.0
129.00	83624.0
130.00	6490.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

130.90	2153.0
132.00	1034.0
133.10	1542.0
134.00	2538.0
135.00	7354.0
135.90	3327.0
137.00	4007.0
137.90	1412.0
138.90	610.0
140.00	1967.0
141.00	13177.0
142.00	3729.0
143.00	3139.0
144.00	910.0
145.00	504.0
146.00	2269.0
147.00	7330.0
148.00	14046.0
149.00	2986.0
150.10	1063.0
151.10	1595.0
152.00	1863.0
153.10	3849.0
154.00	3482.0
155.00	7236.0
156.00	9679.0
157.00	1964.0
158.00	2498.0
159.00	1623.0
159.90	3561.0
161.00	5517.0
162.00	1715.0
162.90	688.0
163.80	469.0
164.00	443.0
165.00	4969.0
166.00	3662.0
167.00	20080.0
168.00	9666.0
169.00	1962.0
169.90	933.0
171.00	1894.0
172.00	2041.0
172.90	2779.0
174.00	5499.0
175.10	8744.0
176.00	2526.0
176.90	4487.0
178.00	1809.0
179.00	17680.0
180.00	12141.0
181.00	5497.0
181.90	1394.0
183.00	636.0
183.90	1916.0
185.00	9060.0
186.00	62048.0
187.00	19328.0
188.00	2171.0
189.00	3951.0
190.00	1088.0
191.00	3257.0
192.00	5415.0
193.00	7608.0
194.10	1789.0
194.90	1562.0
196.00	12088.0
197.10	2604.0
197.90	430016.0
198.90	29920.0
200.00	2487.0
201.40	2428.0
203.00	3654.0
204.00	16159.0
205.00	26704.0
206.00	115056.0
207.00	24408.0
208.00	6635.0
209.00	3344.0
210.00	1965.0
211.00	4317.0
211.80	829.0
212.20	704.0
212.80	525.0
214.90	1292.0
216.00	2904.0
217.00	31584.0
218.00	3705.0
218.80	463.0
219.20	538.0
221.00	18984.0
221.90	5613.0
223.00	8097.0
224.00	64848.0
225.00	16688.0
226.00	1983.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

227.00	27568.0
228.00	4178.0
229.00	5829.0
230.00	725.0
231.00	3029.0
232.00	783.0
232.80	878.0
233.90	1641.0
234.90	2577.0
235.90	1439.0
237.00	2184.0
238.20	510.0
239.00	1288.0
240.10	1311.0
240.90	2199.0
242.00	3364.0
243.00	3054.0
244.00	50960.0
245.00	6208.0
246.00	10729.0
246.90	2302.0
248.10	727.0
249.00	2275.0
249.90	526.0
250.90	936.0
252.00	792.0
252.90	1236.0
254.20	1806.0
255.00	260032.0
256.00	36008.0
257.00	2292.0
257.90	15966.0
259.00	2936.0
260.00	558.0
260.80	425.0
264.10	541.0
265.00	6478.0
266.10	897.0
267.20	1176.0
268.10	348.0
269.10	371.0
270.00	410.0
270.90	663.0
272.00	860.0
272.90	8039.0
274.00	19032.0
275.00	118376.0
276.00	16360.0
277.00	9916.0
278.00	1804.0
279.00	401.0
281.00	2717.0
282.10	1288.0
283.10	1216.0
283.90	1114.0
284.90	2234.0
286.00	525.0
289.00	543.0
289.90	437.0
292.10	354.0
292.90	2646.0
294.00	474.0
295.10	1682.0
296.00	32664.0
297.00	5053.0
297.90	343.0
301.10	825.0
301.90	876.0
303.00	4323.0
304.00	919.0
307.90	705.0
309.10	444.0
310.10	510.0
313.00	612.0
314.00	1704.0
314.90	3821.0
315.90	2181.0
317.00	452.0
321.00	844.0
321.90	320.0
323.00	9730.0
324.00	2566.0
325.80	529.0
326.90	2435.0
328.00	937.0
331.80	686.0
333.00	1581.0
334.00	7370.0
335.00	1932.0
335.90	339.0
338.90	348.0
341.00	1685.0
342.00	594.0
346.00	2690.0
347.00	584.0
351.00	446.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

352.00	3279.0
353.10	2226.0
354.00	3243.0
355.00	1274.0
365.00	14690.0
366.00	2282.0
369.80	301.0
371.10	952.0
372.10	4800.0
373.00	1563.0
383.10	1457.0
383.80	588.0
390.00	728.0
390.90	536.0
401.10	774.0
402.00	2166.0
403.00	3061.0
404.10	987.0
405.00	410.0
414.90	304.0
420.90	2433.0
422.00	2272.0
423.00	16424.0
424.00	3566.0
425.00	306.0
441.00	46568.0
442.00	316736.0
443.00	62952.0
444.10	5705.0
445.10	606.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

m/z	Abundance
37.10	737.0
38.00	2358.0
39.00	17176.0
40.00	2891.0
41.10	6539.0
42.10	2010.0
43.10	9859.0
44.00	29376.0
45.00	1439.0
47.00	453.0
48.90	1773.0
50.00	49608.0
51.10	187904.0
52.00	9441.0
53.00	1640.0
54.00	1376.0
55.10	7747.0
56.10	7411.0
57.00	19024.0
58.10	1532.0
58.90	556.0
60.10	1636.0
61.00	2697.0
62.00	3381.0
63.00	7222.0
64.00	1509.0
65.10	4143.0
66.00	866.0
67.10	2449.0
68.00	4558.0
69.00	195200.0
70.00	2497.0
71.00	3261.0
72.10	510.0
73.00	6731.0
74.00	21880.0
75.00	31120.0
76.10	10643.0
77.00	205504.0
78.10	14115.0
79.00	18512.0
80.00	11906.0
81.00	17440.0
82.00	4751.0
83.00	6578.0
84.00	1807.0
85.00	5682.0
86.00	5147.0
87.00	2471.0
88.10	613.0
89.00	430.0
91.10	5083.0
92.00	4009.0
93.00	27224.0
94.00	2006.0
94.90	1785.0
96.00	3193.0
97.10	3017.0
98.00	21192.0
99.00	15293.0
99.90	1900.0
101.00	7958.0
102.00	918.0
103.00	3951.0
103.90	6037.0
105.00	6415.0
106.10	1901.0
107.00	65888.0
108.00	9768.0
109.10	3567.0
110.00	116272.0
111.00	18584.0
112.00	3032.0
113.00	1263.0
113.70	318.0
114.30	420.0
115.00	1178.0
116.00	4086.0
117.00	55784.0
118.00	4172.0
119.00	1381.0
120.00	1181.0
120.90	1137.0
122.00	4837.0
123.00	8604.0
123.90	3210.0
125.10	3634.0
126.20	2114.0
127.00	232832.0
128.00	15882.0
129.00	94528.0
130.00	7656.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

131.00	2274.0
132.00	774.0
132.90	1676.0
134.00	3001.0
135.00	7660.0
136.00	3028.0
137.00	4446.0
138.10	689.0
138.90	1012.0
140.00	1303.0
141.00	14106.0
141.90	4322.0
143.00	3005.0
144.00	801.0
145.00	1293.0
146.00	3197.0
147.00	8795.0
148.00	15073.0
149.00	3414.0
150.10	1050.0
151.10	1892.0
152.00	1753.0
153.00	4286.0
154.00	3470.0
155.00	7805.0
156.00	10776.0
157.00	1795.0
157.90	2694.0
158.90	1864.0
159.90	3652.0
161.00	6112.0
162.00	2057.0
162.90	693.0
164.10	801.0
165.00	5703.0
166.00	4121.0
167.00	24008.0
168.00	11349.0
169.00	2501.0
170.00	1169.0
171.00	1399.0
171.90	2539.0
173.00	2759.0
174.00	5943.0
175.00	10639.0
176.00	2496.0
177.00	5658.0
178.00	1902.0
178.90	21216.0
180.00	13424.0
181.00	5628.0
182.10	827.0
183.00	1324.0
184.00	2135.0
185.00	10343.0
186.00	72088.0
187.00	22264.0
188.10	2168.0
189.00	5153.0
190.10	1071.0
191.00	3324.0
192.00	6849.0
193.00	9163.0
194.00	2463.0
194.90	1508.0
196.00	14091.0
197.10	3551.0
197.90	495680.0
198.90	32592.0
200.00	3447.0
200.80	406.0
201.40	2368.0
201.80	1580.0
202.90	4227.0
204.00	19328.0
205.00	33936.0
206.00	138816.0
207.00	25624.0
208.00	6119.0
209.00	3010.0
210.20	3249.0
211.00	6231.0
211.80	1111.0
212.90	452.0
215.00	1831.0
216.00	3183.0
217.00	36968.0
218.00	4067.0
218.80	541.0
220.10	994.0
221.00	20856.0
221.80	4366.0
223.00	8568.0
224.00	78328.0
225.00	20856.0
226.10	2604.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

227.00	34048.0
228.00	4407.0
228.90	6368.0
230.00	403.0
231.00	3225.0
231.90	747.0
232.90	659.0
234.00	1877.0
234.90	1950.0
235.90	1715.0
236.90	2637.0
238.20	801.0
239.00	1762.0
240.10	1162.0
241.00	1607.0
242.00	4302.0
243.00	4111.0
244.00	59504.0
245.00	8068.0
246.00	11546.0
247.00	3268.0
247.90	1269.0
249.00	2238.0
250.00	532.0
251.00	1002.0
251.90	1122.0
253.10	1538.0
254.10	2881.0
255.00	318848.0
256.00	48848.0
257.00	4381.0
258.00	17792.0
259.10	3188.0
260.00	809.0
260.90	706.0
263.10	362.0
264.10	1061.0
265.00	8379.0
266.00	1049.0
266.90	804.0
267.90	435.0
269.10	428.0
269.90	470.0
270.80	860.0
272.00	789.0
273.00	9406.0
274.00	25360.0
275.00	149952.0
276.00	20720.0
277.00	12874.0
278.00	2073.0
278.90	746.0
280.90	2565.0
282.10	1440.0
282.90	1685.0
284.00	1297.0
285.00	2485.0
287.80	319.0
288.90	771.0
290.00	445.0
291.10	325.0
291.90	692.0
293.00	2807.0
293.90	523.0
294.90	1502.0
296.00	40464.0
296.90	5472.0
297.90	424.0
301.00	614.0
302.10	790.0
303.00	4781.0
304.10	1664.0
307.90	569.0
309.10	623.0
310.10	414.0
310.70	337.0
312.90	420.0
314.00	2443.0
315.00	4483.0
316.00	2368.0
317.00	693.0
320.90	1224.0
321.90	877.0
323.00	12737.0
323.90	1866.0
324.80	678.0
325.90	481.0
327.00	2511.0
328.00	1854.0
329.10	417.0
331.90	1307.0
333.10	1174.0
334.00	9921.0
335.00	2904.0
335.90	339.0
339.90	485.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

341.00	1915.0
342.10	497.0
346.00	2964.0
347.00	629.0
352.00	4695.0
353.00	3024.0
354.00	4166.0
355.00	1758.0
356.00	433.0
357.10	315.0
365.00	20664.0
366.00	3006.0
370.00	615.0
371.00	886.0
372.10	6871.0
373.00	1957.0
383.10	1845.0
384.00	691.0
384.90	309.0
389.90	861.0
391.10	750.0
392.00	981.0
401.00	603.0
402.10	2733.0
403.00	4618.0
404.00	1648.0
414.90	310.0
421.00	3792.0
422.00	3777.0
423.00	27184.0
424.00	5743.0
425.00	665.0
440.10	415.0
441.00	70216.0
442.00	468288.0
443.00	90504.0
444.00	8512.0
445.00	562.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

m/z	Abundance
35.00	538.0
37.00	504.0
38.00	2276.0
39.10	11270.0
40.00	2771.0
41.10	7654.0
42.10	2839.0
43.10	9313.0
44.00	27752.0
45.00	1809.0
46.10	402.0
48.00	363.0
49.20	1056.0
50.00	29880.0
51.00	100056.0
52.10	4930.0
53.00	1131.0
54.00	965.0
55.00	6204.0
56.00	5259.0
57.00	13539.0
58.00	991.0
59.00	923.0
60.00	918.0
61.00	1771.0
62.00	2441.0
63.00	4350.0
64.00	1052.0
65.00	2444.0
65.90	521.0
67.00	1690.0
68.20	3242.0
69.00	117704.0
70.10	1835.0
71.10	4825.0
71.90	457.0
73.10	6359.0
74.00	12188.0
75.00	18304.0
76.10	7018.0
77.00	118888.0
78.10	7846.0
79.00	11098.0
80.00	7148.0
81.00	11448.0
82.00	3491.0
83.10	4587.0
84.10	1548.0
85.10	4686.0
86.00	3033.0
87.00	1390.0
88.00	602.0
89.00	673.0
91.00	3820.0
92.10	2596.0
93.00	15112.0
93.90	1073.0
95.00	1190.0
96.00	1981.0
97.10	2979.0
98.00	14487.0
99.00	9463.0
100.10	792.0
101.00	5659.0
101.90	469.0
103.00	2186.0
104.00	4023.0
105.00	5414.0
106.00	1439.0
107.00	37984.0
108.00	6885.0
109.10	2887.0
110.00	67872.0
111.00	10989.0
111.90	2036.0
113.00	1193.0
114.00	404.0
114.90	1723.0
116.10	2799.0
117.00	35360.0
118.00	2761.0
119.00	1183.0
120.00	945.0
121.00	543.0
122.00	3354.0
123.00	3989.0
124.00	1968.0
125.00	2310.0
126.20	1164.0
127.00	137088.0
128.00	11017.0
129.00	55816.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

130.00	4295.0
131.00	1460.0
132.10	830.0
133.00	1802.0
134.00	2217.0
135.00	5405.0
136.10	2310.0
137.10	3071.0
138.10	790.0
139.20	770.0
139.90	1200.0
141.00	8007.0
141.90	3085.0
143.00	1934.0
144.00	1007.0
145.00	883.0
145.90	1752.0
147.00	5844.0
148.00	10713.0
149.00	2941.0
150.10	549.0
150.90	1053.0
151.80	938.0
153.10	3415.0
154.00	2722.0
155.00	4910.0
156.00	6265.0
157.10	1852.0
158.10	1587.0
158.90	1914.0
160.00	2827.0
161.00	3976.0
162.10	1151.0
163.00	792.0
163.90	665.0
165.00	3686.0
166.00	2522.0
167.00	15484.0
168.00	7339.0
168.90	1525.0
170.00	810.0
171.00	937.0
172.00	1629.0
173.00	2455.0
174.00	3442.0
175.00	7410.0
175.90	2439.0
176.90	3410.0
178.00	1523.0
179.00	12063.0
180.00	8086.0
181.00	4899.0
181.90	823.0
183.10	568.0
184.00	1522.0
185.00	6679.0
186.00	47744.0
187.00	12272.0
187.90	1531.0
188.90	3806.0
190.00	551.0
191.00	2244.0
192.00	3896.0
193.00	5354.0
194.00	1115.0
195.10	1268.0
195.90	9718.0
197.10	1893.0
197.90	316160.0
198.90	20240.0
200.00	1396.0
200.80	349.0
201.30	1716.0
202.10	1033.0
202.90	2921.0
204.00	11952.0
205.00	20800.0
206.00	89616.0
207.00	21920.0
208.00	5737.0
209.00	2660.0
210.20	2040.0
210.90	4044.0
211.70	866.0
213.20	646.0
213.90	361.0
214.90	1519.0
216.00	2103.0
217.00	24888.0
218.00	3357.0
219.10	663.0
221.00	14672.0
221.90	3761.0
223.00	5860.0
224.00	46848.0
225.00	12804.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

225.90	1535.0
227.00	20368.0
228.00	2547.0
229.00	5458.0
230.00	721.0
231.00	1537.0
231.90	515.0
233.00	622.0
234.00	1748.0
234.90	1991.0
235.80	1129.0
237.10	2036.0
237.90	396.0
238.90	1875.0
240.10	719.0
241.00	1548.0
242.00	2731.0
243.10	3233.0
244.00	39672.0
245.00	5606.0
245.90	8982.0
247.00	1625.0
247.90	330.0
248.90	1680.0
249.80	414.0
250.90	912.0
252.00	595.0
252.90	971.0
253.90	1224.0
255.00	225024.0
256.00	31936.0
256.90	2285.0
258.00	13575.0
259.00	2057.0
260.00	518.0
264.10	517.0
264.90	5120.0
266.00	845.0
267.10	1254.0
267.90	519.0
270.90	512.0
272.10	926.0
273.00	6963.0
274.00	18040.0
275.00	103656.0
276.00	14062.0
277.00	9146.0
277.90	1482.0
281.00	2794.0
282.10	655.0
283.00	1676.0
284.00	1174.0
285.10	1477.0
286.00	317.0
288.80	583.0
290.00	630.0
292.00	638.0
292.90	2152.0
293.90	574.0
295.00	1117.0
296.00	32224.0
296.90	3850.0
298.10	657.0
301.10	682.0
302.00	583.0
303.00	3626.0
303.90	1176.0
307.90	575.0
308.80	460.0
310.10	461.0
312.90	445.0
314.00	1424.0
315.00	3241.0
315.90	1719.0
317.10	379.0
321.00	701.0
322.00	426.0
323.00	11031.0
324.20	1812.0
325.70	403.0
326.90	2951.0
328.00	1528.0
328.90	438.0
332.00	651.0
332.90	1251.0
334.00	7086.0
335.00	1566.0
341.00	1884.0
341.90	490.0
346.00	2318.0
346.80	502.0
350.80	416.0
352.00	2560.0
353.00	2473.0
354.00	3600.0
355.00	1453.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09

356.10	354.0
359.10	302.0
365.00	14753.0
366.00	2053.0
369.20	326.0
370.90	899.0
372.00	6296.0
373.00	1592.0
382.90	1497.0
383.90	749.0
390.00	717.0
391.00	775.0
392.10	429.0
400.90	709.0
402.00	2135.0
402.90	3105.0
404.10	1187.0
404.90	339.0
415.10	599.0
421.10	3147.0
421.90	2380.0
423.00	21032.0
424.00	5010.0
425.20	321.0
440.00	384.0
441.00	60488.0
442.00	397888.0
443.00	73704.0
444.10	7327.0
445.10	504.0

Instrument :
BNA_M
ClientSampleId :
DFTPP09