

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM010919\  
 Data File : BM018498.D  
 Acq On : 10 Jan 2019 07:33  
 Operator : JU/SJ  
 Sample : K1082-04  
 Misc :  
 ALS Vial : 26 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 SP-1-2

Quant Time: Jan 10 08:26:46 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM010919.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Jan 10 05:57:41 2019  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 7.76  | 152  | 617586   | 20.00  | ng    | 0.00     |
| 21) Naphthalene-d8          | 10.56 | 136  | 2530981  | 20.00  | ng    | 0.00     |
| 39) Acenaphthene-d10        | 14.41 | 164  | 1421702  | 20.00  | ng    | 0.00     |
| 64) Phenanthrene-d10        | 17.16 | 188  | 2955670  | 20.00  | ng    | 0.00     |
| 76) Chrysene-d12            | 21.35 | 240  | 2272124  | 20.00  | ng    | 0.00     |
| 87) Perylene-d12            | 23.64 | 264  | 2258910  | 20.00  | ng    | 0.00     |
| System Monitoring Compounds |       |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.36  | 112  | 3733109  | 111.61 | ng    | 0.00     |
| 7) Phenol-d6                | 6.95  | 99   | 4304455  | 101.33 | ng    | 0.01     |
| 23) Nitrobenzene-d5         | 8.93  | 82   | 3431478  | 78.60  | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol    | 15.91 | 330  | 2271798  | 125.50 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 13.03 | 172  | 7971156  | 73.16  | ng    | 0.00     |
| 79) Terphenyl-d14           | 19.79 | 244  | 10620284 | 98.53  | ng    | 0.00     |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM010919\  
Data File : BM018498.D  
Acq On : 10 Jan 2019 07:33  
Operator : JU/SJ  
Sample : K1082-04  
Misc :  
ALS Vial : 26 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
SP-1-2

Quant Time: Jan 10 08:26:46 2019  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM010919.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Jan 10 05:57:41 2019  
Response via : Initial Calibration

