

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM011019\
 Data File : BM018527.D
 Acq On : 11 Jan 2019 02:17
 Operator : JU/SJ
 Sample : K1099-21
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 TP-21

Manual Integrations
 APPROVED

Sohil
 1/11/2019 4:41:14 PM

Quant Time: Jan 11 04:22:06 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM010919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 10 05:57:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.76	152	710774	20.00	ng	0.00
21) Naphthalene-d8	10.55	136	2851763	20.00	ng	0.00
39) Acenaphthene-d10	14.41	164	1512129	20.00	ng	0.00
64) Phenanthrene-d10	17.16	188	2863633	20.00	ng	0.00
76) Chrysene-d12	21.35	240	2224595	20.00	ng	0.00
87) Perylene-d12	23.65	264	2294348	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.35	112	5460872	141.87	ng	0.00
7) Phenol-d6	6.95	99	7128846	145.81	ng	0.01
23) Nitrobenzene-d5	8.93	82	4429964	90.05	ng	0.00
42) 2,4,6-Tribromophenol	15.90	330	2465955	128.08	ng	0.00
45) 2-Fluorobiphenyl	13.03	172	8826735	76.17	ng	0.00
79) Terphenyl-d14	19.79	244	8126209	77.00	ng	0.00
Target Compounds						
10) Phenol	6.97	94	123768	2.491	ng	96
50) Dimethylphthalate	13.87	163	2733156	22.008	ng	99
71) Phenanthrene	17.20	178	362071	2.377	ng	99
75) Fluoranthene	19.22	202	938537	5.344	ng	96
78) Pyrene	19.59	202	1092949	8.289	ng	98
81) Benzo(a)anthracene	21.33	228	652169	4.976	ng	# 89
83) Chrysene	21.39	228	677703	5.355	ng	# 92
86) Indeno(1,2,3-cd)pyrene	25.97	276	862904	5.913	ng	# 90
88) Benzo(b)fluoranthene	22.96	252	1096780	8.420	ng	# 88
89) Benzo(k)fluoranthene	23.00	252	335876m	2.559	ng	
90) Benzo(a)pyrene	23.54	252	805909	6.463	ng	# 88
92) Benzo(a,h,i)perylene	26.69	276	870707	7.081	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM011019\
Data File : BM018527.D
Acq On : 11 Jan 2019 02:17
Operator : JU/SJ
Sample : K1099-21
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
TP-21

Manual Integrations
APPROVED

Sohil
1/11/2019 4:41:14 PM

Quant Time: Jan 11 04:22:06 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM010919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 10 05:57:41 2019
Response via : Initial Calibration

