




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011322\
 Data File : BM033711.D
 Acq On : 13 Jan 2022 09:55
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 13 21:28:42 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM010522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 13 07:49:01 2022
 Response via : Initial Calibration






TIC: BM033711.D\data.ms

(77) Benzidine**19.565min (-0.006) 103798.91 ng****response 1445503**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.20	14.18
183.00	12.10	11.99
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011322\
Data File : BM033711.D
Acq On : 13 Jan 2022 09:55
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 13 21:28:42 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM010522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 13 07:49:01 2022
Response via : Initial Calibration



TIC: BM033711.D\data.ms

(70) Pentachlorophenol (C)

16.989min (-0.006) 0.00 ng

response **395622**

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.60	63.87
264.00	61.40	62.45
0.00	0.00	0.00