


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011322\
 Data File : BM033717.D
 Acq On : 13 Jan 2022 15:40
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM010522.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jan 13 07:49:01 2022






AutoFind: Scans 2473, 2474, 2475; Background Corrected with Scan 2466

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.8	81695	PASS
68	69	0.00	2	2.0	1648	PASS
69	198	0.00	100	36.5	83307	PASS
70	69	0.00	2	0.5	414	PASS
127	198	10	80	46.4	106016	PASS
197	198	0.00	2	0.7	1621	PASS
198	198	100	100	100.0	228501	PASS
199	198	5	9	6.6	15133	PASS
275	198	10	60	28.8	65781	PASS
365	198	1	100	4.2	9527	PASS
441	198	0.01	100	13.5	30773	PASS
442	442	50	100	100.0	198336	PASS
443	442	15	24	19.5	38715	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011322\
 Data File : BM033717.D
 Acq On : 13 Jan 2022 15:40
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP

Quant Time: Jan 13 21:31:30 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM010522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 13 07:49:01 2022
 Response via : Initial Calibration



TIC: BM033717.D\data.ms

(77) Benzidine

19.566min (-0.005) 101836.00 ng




response 1274517

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.20	14.31
183.00	12.10	12.19
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011322\
 Data File : BM033717.D
 Acq On : 13 Jan 2022 15:40
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP

Quant Time: Jan 13 21:31:30 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM010522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 13 07:49:01 2022
 Response via : Initial Calibration



TIC: BM033717.D\data.ms

(70) Pentachlorophenol (C)

16.989min (-0.005) 0.00 ng

response 266116

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.60	64.55
264.00	61.40	62.90
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
1/13/2022	BNA_M	<u>BM033717.D</u>
Compound Name	Response	Retention Time
DDT	632045	20.801
DDD	46082	20.366
DDE	607	19.86
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
46689	678734	6.88

Instrument :
BNA_M
ClientSampleId :
DFTPP