

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM011720\
 Data File : BM024499.D
 Acq On : 17 Jan 2020 17:24
 Operator : JU
 Sample : PB126136BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK36

Quant Time: Jan 17 17:57:10 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM011520MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jan 17 12:17:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.46	152	161221	20.00	ng/ul	0.00
18) Naphthalene-d8	10.22	136	701298	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.10	164	527291	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.86	188	1249798	20.00	ng/ul	0.00
78) Chrysene-d12	21.06	240	1276534	20.00	ng/ul	0.00
86) Perylene-d12	23.18	264	1354858	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.06	96	25029	7.59	ng/uL	0.00
5) Phenol-d5	6.65	99	383544	32.09	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.81	67	231498	34.87	ng/ul	0.00
9) 2-Chlorophenol-d4	7.00	132	360472	35.32	ng/ul	0.00
13) 4-Methylphenol-d8	8.17	113	350727	31.83	ng/ul	0.00
19) Nitrobenzene-d5	8.60	128	190285	38.62	ng/ul	0.00
22) 2-Nitrophenol-d4	9.31	143	214150	40.13	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.85	165	404974	33.72	ng/ul	0.00
29) 4-Chloroaniline-d4	10.36	131	565086	42.32	ng/ul	0.00
44) Dimethylphthalate-d6	13.53	166	1568487	37.96	ng/ul	0.00
47) Acenaphthylene-d8	13.79	160	1770185	37.23	ng/ul	0.00
52) 4-Nitrophenol-d4	14.32	143	218867	31.74	ng/ul	0.00
58) Fluorene-d10	15.11	176	1352285	37.18	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.23	200	218126	30.57	ng/ul	0.00
71) Anthracene-d10	16.96	188	2143526	36.88	ng/ul	0.00
79) Pyrene-d10	19.26	212	2524633	37.00	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.05	264	2597764	36.97	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM011720\
Data File : BM024499.D
Acq On : 17 Jan 2020 17:24
Operator : JU
Sample : PB126136BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK36

Quant Time: Jan 17 17:57:10 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM011520MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jan 17 12:17:39 2020
Response via : Initial Calibration

