

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011722\
 Data File : BM033751.D
 Acq On : 17 Jan 2022 16:05
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM011722

Quant Time: Jan 17 17:31:45 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM011722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 17 16:30:56 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	107	0.00
2	1,4-Dioxane	40.000	36.688	8.3	109	0.00
3	Pyridine	40.000	36.880	7.8	106	0.00
4	n-Nitrosodimethylamine	40.000	36.729	8.2	104	0.00
5 S	2-Fluorophenol	80.000	73.172	8.5	106	0.00
6	Aniline	40.000	36.881	7.8	105	0.00
7 S	Phenol-d6	80.000	73.615	8.0	105	0.00
8	2-Chlorophenol	40.000	36.614	8.5	105	0.00
9	Benzaldehyde	40.000	38.666	3.3	104	0.00
10 C	Phenol	40.000	37.000	7.5	104	0.00
11	bis(2-Chloroethyl)ether	40.000	36.213	9.5	105	0.00
12	1,3-Dichlorobenzene	40.000	36.588	8.5	106	0.00
13 C	1,4-Dichlorobenzene	40.000	36.469	8.8	106	0.00
14	1,2-Dichlorobenzene	40.000	36.562	8.6	106	0.00
15	Benzyl Alcohol	40.000	36.966	7.6	105	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.090	9.8	105	0.00
17	2-Methylphenol	40.000	36.665	8.3	103	0.00
18	Hexachloroethane	40.000	36.801	8.0	108	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	35.815	10.5	103	0.00
20	3+4-Methylphenols	40.000	37.178	7.1	105	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	105	0.00
22	Acetophenone	40.000	35.702	10.7	104	0.00
23 S	Nitrobenzene-d5	80.000	72.148	9.8	105	0.00
24	Nitrobenzene	40.000	35.945	10.1	105	0.00
25	Isophorone	40.000	35.570	11.1	104	0.00
26 C	2-Nitrophenol	40.000	36.764	8.1	104	0.00
27	2,4-Dimethylphenol	40.000	36.060	9.8	104	0.00
28	bis(2-Chloroethoxy)methane	40.000	35.283	11.8	104	0.00
29 C	2,4-Dichlorophenol	40.000	36.351	9.1	105	0.00
30	1,2,4-Trichlorobenzene	40.000	35.644	10.9	104	0.00
31	Naphthalene	40.000	35.665	10.8	104	0.00
32	Benzoic acid	40.000	39.469	1.3	107	0.00
33	4-Chloroaniline	40.000	36.341	9.1	105	0.00
34 C	Hexachlorobutadiene	40.000	35.892	10.3	105	0.00
35	Caprolactam	40.000	36.499	8.8	105	0.00
36 C	4-Chloro-3-methylphenol	40.000	35.970	10.1	104	0.00
37	2-Methylnaphthalene	40.000	35.251	11.9	104	0.00
38	1-Methylnaphthalene	40.000	35.394	11.5	104	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	104	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	35.948	10.1	104	0.00
41 P	Hexachlorocyclopentadiene	40.000	35.769	10.6	101	0.00
42 S	2,4,6-Tribromophenol	80.000	73.712	7.9	106	0.00
43 C	2,4,6-Trichlorophenol	40.000	36.364	9.1	104	0.00
44	2,4,5-Trichlorophenol	40.000	36.516	8.7	105	0.00
45 S	2-Fluorobiphenyl	80.000	70.715	11.6	103	0.00
46	1,1'-Biphenyl	40.000	35.718	10.7	104	0.00
47	2-Chloronaphthalene	40.000	35.647	10.9	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011722\
 Data File : BM033751.D
 Acq On : 17 Jan 2022 16:05
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM011722

Quant Time: Jan 17 17:31:45 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM011722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 17 16:30:56 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	37.429	6.4	104	0.00
49	Acenaphthylene	40.000	36.004	10.0	105	0.00
50	Dimethylphthalate	40.000	35.844	10.4	105	0.00
51	2,6-Dinitrotoluene	40.000	36.617	8.5	105	0.00
52 C	Acenaphthene	40.000	36.196	9.5	104	0.00
53	3-Nitroaniline	40.000	37.396	6.5	106	0.00
54 P	2,4-Dinitrophenol	40.000	39.425	1.4	107	0.00
55	Dibenzofuran	40.000	35.720	10.7	105	0.00
56 P	4-Nitrophenol	40.000	38.283	4.3	106	0.00
57	2,4-Dinitrotoluene	40.000	37.471	6.3	106	0.00
58	Fluorene	40.000	35.789	10.5	104	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	36.457	8.9	105	0.00
60	Diethylphthalate	40.000	35.905	10.2	104	0.00
61	4-Chlorophenyl-phenylether	40.000	35.482	11.3	105	0.00
62	4-Nitroaniline	40.000	37.850	5.4	106	0.00
63	Azobenzene	40.000	36.122	9.7	104	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	106	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.462	3.8	106	0.00
66 c	n-Nitrosodiphenylamine	40.000	35.540	11.2	105	0.00
67	4-Bromophenyl-phenylether	40.000	35.240	11.9	105	0.00
68	Hexachlorobenzene	40.000	35.112	12.2	106	0.00
69	Atrazine	40.000	35.693	10.8	105	0.00
70 C	Pentachlorophenol	40.000	37.433	6.4	106	0.00
71	Phenanthrene	40.000	35.302	11.7	105	0.00
72	Anthracene	40.000	35.731	10.7	105	0.00
73	Carbazole	40.000	36.033	9.9	106	0.00
74	Di-n-butylphthalate	40.000	36.455	8.9	106	0.00
75 C	Fluoranthene	40.000	36.703	8.2	109	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	119	0.00
77	Benzydine	40.000	35.300	11.8	119	0.00
78	Pyrene	40.000	33.269	16.8	109	0.00
79 S	Terphenyl-d14	80.000	66.555	16.8	109	0.00
80	Butylbenzylphthalate	40.000	35.434	11.4	114	0.00
81	Benzo(a)anthracene	40.000	35.987	10.0	118	0.00
82	3,3'-Dichlorobenzidine	40.000	36.503	8.7	121	0.00
83	Chrysene	40.000	35.958	10.1	120	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	35.833	10.4	117	0.00
85 c	Di-n-octyl phthalate	40.000	38.085	4.8	126	0.00
86 I	Perylene-d12	20.000	20.000	0.0	129	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	33.478	16.3	121	0.00
88	Benzo(b)fluoranthene	40.000	36.334	9.2	128	0.00
89	Benzo(k)fluoranthene	40.000	36.938	7.7	130	0.00
90 C	Benzo(a)pyrene	40.000	36.427	8.9	128	0.00
91	Dibenzo(a,h)anthracene	40.000	33.692	15.8	121	0.00
92	Benzo(g,h,i)perylene	40.000	32.732	18.2	118	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011722\
Data File : BM033751.D
Acq On : 17 Jan 2022 16:05
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ICVBM011722

Quant Time: Jan 17 17:31:45 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM011722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 17 16:30:56 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0