


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011722\
 Data File : BM033743.D
 Acq On : 17 Jan 2022 11:18
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM011722.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jan 17 16:30:56 2022






AutoFind: Scans 2471, 2472, 2473; Background Corrected with Scan 2464

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.0	96448	PASS
68	69	0.00	2	2.0	1933	PASS
69	198	0.00	100	36.2	96832	PASS
70	69	0.00	2	0.6	554	PASS
127	198	10	80	46.9	125603	PASS
197	198	0.00	2	0.7	2007	PASS
198	198	100	100	100.0	267648	PASS
199	198	5	9	6.8	18301	PASS
275	198	10	60	28.5	76229	PASS
365	198	1	100	4.1	10882	PASS
441	198	0.01	100	13.5	36107	PASS
442	442	50	100	100.0	232683	PASS
443	442	15	24	19.2	44597	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011722\
 Data File : BM033743.D
 Acq On : 17 Jan 2022 11:18
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP

Quant Time: Jan 17 15:32:34 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM010522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 17 15:29:26 2022
 Response via : Initial Calibration



TIC: BM033743.D\data.ms

(70) Pentachlorophenol (C)

16.977min (-0.018) 0.00 ng




response 298161

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.60	62.78
264.00	61.40	62.67
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011722\
 Data File : BM033743.D
 Acq On : 17 Jan 2022 11:18
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP

Quant Time: Jan 17 15:32:34 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM010522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 17 15:29:26 2022
 Response via : Initial Calibration



TIC: BM033743.D\data.ms

(77) Benzidine

19.559min (-0.012) 821236.98 ng

response 1577454

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.20	13.79
183.00	12.10	12.19
0.00	0.00	0.00

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP

DDT Breakdown

Date	Instrument Name	DFTPP Data File
1/17/2022	BNA_M	<u>BM033743.D</u>
Compound Name	Response	Retention Time
DDT	883660	20.79
DDD	3758	20.4
DDE	0	19.96
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
3758	887418	0.42