


Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM011821\
 Data File : BM028416.D
 Acq On : 18 Jan 2021 08:27
 Operator : CG/JU
 Sample : DFTPP20
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP20

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM011621MA.M
 Title : SVOA CALIBRATION
 Last Update : Mon Jan 18 10:13:28 2021



AutoFind: Scans 2495, 2496, 2497; Background Corrected with Scan 2490

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	49.0	31010	PASS
68	69	0.00	2	1.9	551	PASS
69	198	0.00	100	44.8	28397	PASS
70	69	0.00	2	0.2	60	PASS
127	198	10	80	51.5	32584	PASS
197	198	0.00	2	0.5	326	PASS
198	198	100	100	100.0	63317	PASS
199	198	5	9	6.8	4275	PASS
275	198	10	60	28.0	17753	PASS
365	198	1	100	3.8	2403	PASS
441	198	0.01	100	13.6	8632	PASS
442	198	50	100	92.2	58384	PASS
443	442	15	24	19.8	11564	PASS

Scan 2490 (17.833 min): BM028416.D
DFTPP20

m/z	Abundance
44.00	542.0
73.00	188.0
207.00	445.0
281.00	160.0

Instrument :
BNA_M
ClientSampleId :
DFTPP20

m/z	Abundance
37.00	166.0
38.00	445.0
39.10	2684.0
40.00	196.0
44.00	637.0
49.10	317.0
50.10	9034.0
51.10	32712.0
52.00	1725.0
55.00	245.0
56.00	1123.0
57.00	2154.0
61.00	477.0
62.00	460.0
63.10	1164.0
65.10	540.0
68.00	588.0
69.00	29848.0
73.10	433.0
74.00	3283.0
75.00	4438.0
76.10	1630.0
77.10	32248.0
78.10	2398.0
79.00	2301.0
80.00	1836.0
81.00	2376.0
82.00	583.0
83.00	527.0
84.90	350.0
86.00	658.0
87.00	334.0
91.00	499.0
92.00	576.0
93.00	4220.0
93.90	229.0
96.10	234.0
98.00	3472.0
99.00	2184.0
100.00	207.0
101.00	1216.0
103.00	419.0
104.00	926.0
105.00	749.0
106.00	268.0
107.00	9087.0
108.00	1426.0
109.00	285.0
110.00	16073.0
111.00	2594.0
112.00	349.0
116.00	556.0
117.00	9114.0
118.00	644.0
121.90	588.0
123.00	1004.0
124.00	465.0
125.00	410.0
127.00	32136.0
128.00	2601.0
129.00	13213.0
130.00	1150.0
130.90	196.0
134.00	404.0
135.00	1126.0
136.00	472.0
137.00	582.0
140.00	163.0
141.00	1767.0
142.00	477.0
142.90	340.0
146.00	353.0
147.00	873.0
148.00	2133.0
149.00	503.0
151.20	167.0
151.80	167.0
153.00	495.0
154.00	435.0
155.00	878.0
156.00	1330.0
157.00	204.0
158.00	337.0
158.90	216.0
160.00	487.0
161.00	708.0
162.00	189.0
165.00	611.0
166.00	407.0
167.00	3296.0
168.00	1896.0
168.90	240.0

Instrument :
BNA_M
ClientSampleId :
DFTPP20

172.00	345.0
172.90	300.0
174.00	638.0
175.00	1247.0
175.90	347.0
177.00	519.0
178.00	248.0
179.00	2205.0
180.00	1788.0
181.00	808.0
184.00	184.0
185.00	1034.0
186.00	8260.0
187.00	2491.0
188.10	281.0
189.00	614.0
191.00	258.0
191.90	763.0
193.00	777.0
194.00	158.0
196.00	1738.0
196.70	331.0
198.00	58456.0
199.00	4111.0
200.00	379.0
201.40	321.0
203.00	370.0
204.00	2264.0
205.00	3816.0
206.10	15518.0
207.00	2604.0
208.00	669.0
209.10	225.0
210.00	236.0
211.00	566.0
215.00	203.0
216.00	331.0
217.00	4137.0
217.90	525.0
221.00	2651.0
221.80	505.0
223.00	1019.0
224.00	8430.0
225.00	2124.0
226.00	314.0
227.00	3607.0
228.00	448.0
229.00	773.0
230.90	358.0
234.00	205.0
235.00	231.0
235.90	219.0
237.00	379.0
241.00	309.0
242.00	427.0
243.00	444.0
244.00	6344.0
245.00	803.0
246.00	1386.0
246.90	279.0
249.00	246.0
254.20	328.0
255.00	33424.0
256.00	5047.0
257.00	466.0
258.00	1832.0
259.00	304.0
265.00	838.0
273.00	961.0
274.00	2629.0
275.00	15363.0
276.00	2204.0
277.00	1309.0
277.90	210.0
285.00	199.0
292.90	292.0
296.00	4138.0
297.00	697.0
303.00	491.0
313.90	228.0
314.90	381.0
316.00	232.0
323.00	1249.0
324.10	258.0
326.90	250.0
328.00	150.0
334.00	777.0
335.00	259.0
341.00	195.0
345.90	295.0
352.00	406.0
352.90	246.0
354.00	413.0
365.00	1949.0
366.00	286.0
372.00	611.0

Instrument :
BNA_M
ClientSampleId :
DFTPP20

372.90	193.0
382.90	184.0
401.90	187.0
403.00	375.0
403.90	183.0
420.90	292.0
422.00	386.0
423.00	2350.0
424.00	714.0
441.00	5971.0
442.00	41984.0
443.00	9120.0
444.00	721.0

Instrument :
BNA_M
ClientSampleId :
DFTPP20

m/z	Abundance
36.90	159.0
38.00	489.0
39.10	2938.0
39.90	228.0
44.00	481.0
49.00	296.0
50.10	10052.0
51.10	36936.0
52.10	1774.0
55.00	258.0
56.00	1222.0
57.00	2667.0
61.00	477.0
62.00	527.0
63.00	1540.0
64.00	178.0
65.00	670.0
68.10	647.0
69.00	33640.0
70.00	182.0
73.00	550.0
74.00	3561.0
75.00	5562.0
76.10	1873.0
77.10	37616.0
78.10	2440.0
79.00	2657.0
80.00	2111.0
81.00	2760.0
82.00	657.0
83.00	658.0
85.00	424.0
86.00	845.0
87.00	335.0
91.00	569.0
92.00	734.0
93.00	4732.0
94.00	360.0
96.00	210.0
98.00	3674.0
99.00	2712.0
100.00	272.0
101.00	1566.0
102.90	515.0
104.00	1010.0
105.00	967.0
105.90	314.0
107.00	11054.0
108.00	1729.0
109.10	402.0
110.00	19768.0
111.00	3216.0
112.10	351.0
116.00	729.0
117.00	10708.0
118.00	763.0
122.00	744.0
123.00	1237.0
124.00	584.0
125.00	540.0
127.00	39376.0
128.00	2931.0
129.00	15662.0
130.00	1392.0
131.00	260.0
134.00	495.0
135.00	1281.0
136.00	535.0
137.00	685.0
139.90	189.0
141.00	2152.0
141.90	614.0
143.00	334.0
145.90	398.0
147.00	1093.0
148.00	2611.0
149.00	566.0
151.10	248.0
151.90	185.0
153.00	595.0
153.90	390.0
155.00	1209.0
156.00	1751.0
157.00	309.0
157.90	397.0
158.90	285.0
159.90	622.0
161.00	835.0
162.00	261.0
165.00	788.0
166.00	588.0
167.00	4493.0

Instrument :
BNA_M
ClientSampleId :
DFTPP20

168.00	2151.0
169.00	314.0
171.90	390.0
172.90	441.0
174.00	915.0
175.00	1658.0
176.00	493.0
177.00	645.0
178.00	295.0
179.00	3207.0
180.00	2148.0
181.00	953.0
182.00	159.0
184.00	234.0
185.00	1270.0
186.00	10582.0
187.00	2926.0
188.00	333.0
189.00	739.0
191.00	327.0
192.00	996.0
193.00	1029.0
194.00	232.0
196.00	2410.0
197.10	533.0
198.00	75568.0
199.00	5075.0
200.00	399.0
201.60	308.0
202.90	633.0
204.00	2675.0
205.10	4449.0
206.10	20320.0
207.10	3003.0
208.00	848.0
209.00	308.0
210.10	290.0
211.00	895.0
211.70	162.0
215.00	235.0
216.00	496.0
217.00	5577.0
218.00	861.0
221.00	3342.0
221.80	744.0
223.00	1128.0
224.00	11458.0
225.00	2912.0
226.10	332.0
227.00	4802.0
228.00	670.0
229.00	1029.0
231.00	397.0
234.00	381.0
234.90	294.0
236.00	271.0
237.00	410.0
238.90	196.0
240.00	159.0
241.00	270.0
242.00	650.0
243.10	566.0
244.10	8653.0
245.00	1263.0
246.00	1668.0
246.90	368.0
249.00	305.0
252.90	257.0
254.20	362.0
255.00	45192.0
256.00	6601.0
257.00	613.0
258.00	2498.0
259.00	369.0
265.00	1144.0
265.90	162.0
273.00	1461.0
274.00	3643.0
275.00	20496.0
276.00	2597.0
277.00	1812.0
277.90	322.0
283.00	211.0
285.00	332.0
293.00	455.0
296.00	5700.0
297.00	880.0
303.00	677.0
304.00	150.0
313.90	347.0
315.00	566.0
316.00	280.0
323.00	1723.0
324.10	375.0
327.00	398.0
327.90	209.0

Instrument :
BNA_M
ClientSampleId :
DFTPP20

333.00	157.0
334.00	1301.0
335.00	306.0
341.00	254.0
345.90	371.0
352.00	596.0
353.00	348.0
354.00	700.0
365.00	2679.0
366.00	341.0
372.00	1244.0
373.00	217.0
383.00	272.0
402.00	358.0
403.00	587.0
404.00	235.0
420.90	448.0
422.00	566.0
423.00	3531.0
424.00	962.0
441.00	10280.0
442.00	67736.0
443.00	13369.0
444.00	1167.0

Instrument :
BNA_M
ClientSampleId :
DFTPP20

m/z	Abundance
38.00	300.0
39.00	1778.0
40.00	215.0
44.00	513.0
49.00	197.0
50.10	6367.0
51.10	23384.0
52.00	1225.0
55.00	219.0
56.00	743.0
57.00	1578.0
61.00	279.0
62.00	329.0
63.00	849.0
65.10	431.0
68.00	418.0
69.00	21704.0
73.00	500.0
74.00	2360.0
75.00	3601.0
76.10	1219.0
77.10	24240.0
78.00	1543.0
79.00	1802.0
80.00	1307.0
81.00	1866.0
82.00	393.0
83.00	337.0
85.00	332.0
86.00	512.0
87.00	231.0
90.90	458.0
92.00	434.0
93.00	3190.0
93.90	255.0
96.10	194.0
98.00	2456.0
99.00	1795.0
100.00	181.0
101.00	926.0
102.90	341.0
104.00	680.0
105.00	679.0
106.10	198.0
107.00	7647.0
108.00	1146.0
109.00	264.0
110.00	12875.0
111.00	2060.0
111.90	228.0
116.00	412.0
117.00	7232.0
118.00	524.0
122.00	590.0
123.00	844.0
123.90	370.0
124.90	323.0
127.00	26240.0
128.00	2007.0
129.00	10918.0
130.00	928.0
131.00	245.0
134.00	342.0
135.00	932.0
136.00	262.0
137.00	472.0
141.00	1431.0
142.00	480.0
142.90	306.0
146.00	319.0
147.00	810.0
148.00	1766.0
149.00	406.0
151.00	196.0
153.00	429.0
154.00	272.0
155.00	718.0
156.00	1190.0
157.00	223.0
157.90	280.0
158.90	253.0
160.00	494.0
161.00	645.0
161.90	186.0
165.00	567.0
165.90	397.0
167.00	3146.0
168.00	1453.0
168.90	306.0
172.00	277.0
172.90	350.0
174.00	543.0

Instrument :
BNA_M
ClientSampleId :
DFTPP20

175.00	1095.0
176.00	282.0
177.00	624.0
178.00	164.0
179.00	2198.0
180.00	1626.0
181.00	645.0
183.90	160.0
185.00	1002.0
186.00	7814.0
187.10	2257.0
188.00	271.0
188.90	537.0
191.00	325.0
192.00	715.0
193.00	831.0
194.00	195.0
196.00	1857.0
197.20	447.0
198.00	55928.0
199.00	3640.0
199.90	233.0
201.40	331.0
203.00	389.0
204.00	2194.0
205.00	3462.0
206.10	15263.0
207.00	2309.0
208.00	583.0
209.00	201.0
210.20	313.0
211.10	633.0
215.10	205.0
216.00	359.0
217.00	3981.0
217.90	535.0
221.00	2645.0
223.00	887.0
224.00	8555.0
225.00	2199.0
226.00	266.0
227.00	3651.0
228.00	550.0
228.90	701.0
231.00	302.0
233.90	231.0
234.90	255.0
235.90	181.0
237.00	313.0
238.90	155.0
241.00	262.0
242.00	513.0
243.10	574.0
244.00	6989.0
245.00	934.0
246.00	1451.0
247.00	256.0
249.00	321.0
252.90	156.0
254.10	356.0
255.00	35136.0
256.00	5143.0
257.00	345.0
258.00	2216.0
259.00	297.0
264.90	977.0
273.00	1078.0
274.00	2978.0
275.00	17400.0
276.00	2229.0
277.00	1413.0
278.00	277.0
283.00	206.0
285.00	222.0
293.00	336.0
296.00	4718.0
297.00	672.0
303.00	508.0
304.00	151.0
314.10	198.0
315.00	593.0
315.90	285.0
321.00	190.0
323.00	1639.0
324.00	240.0
327.00	263.0
333.00	156.0
334.00	982.0
334.90	251.0
341.00	219.0
346.00	354.0
352.00	498.0
353.00	264.0
354.00	637.0
355.10	163.0
365.00	2582.0

Instrument :
BNA_M
ClientSampleId :
DFTPP20

365.90	369.0
370.90	155.0
372.00	943.0
373.00	222.0
382.90	231.0
402.00	431.0
402.90	577.0
404.00	283.0
421.00	445.0
422.00	523.0
423.00	3548.0
424.00	773.0
441.00	9647.0
442.00	65432.0
443.00	12205.0
444.00	1140.0

Instrument :
BNA_M
ClientSampleId :
DFTPP20