

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011922\
 Data File : BM033785.D
 Acq On : 19 Jan 2022 16:38
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 19 17:09:00 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM011722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 17 16:30:56 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	129	0.00
2	1,4-Dioxane	40.000	36.567	8.6	132	0.00
3	Pyridine	40.000	36.447	8.9	126	0.00
4	n-Nitrosodimethylamine	40.000	35.411	11.5	121	0.00
5 S	2-Fluorophenol	80.000	74.622	6.7	131	0.00
6	Aniline	40.000	36.235	9.4	125	0.00
7 S	Phenol-d6	80.000	72.108	9.9	124	0.00
8	2-Chlorophenol	40.000	36.326	9.2	126	0.00
9	Benzaldehyde	40.000	37.524	6.2	123	0.00
10 C	Phenol	40.000	36.507	8.7	125	0.00
11	bis(2-Chloroethyl)ether	40.000	35.766	10.6	126	0.00
12	1,3-Dichlorobenzene	40.000	36.243	9.4	127	0.00
13 C	1,4-Dichlorobenzene	40.000	36.309	9.2	128	0.00
14	1,2-Dichlorobenzene	40.000	36.030	9.9	127	0.00
15	Benzyl Alcohol	40.000	35.344	11.6	121	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.465	8.8	128	-0.01
17	2-Methylphenol	40.000	35.472	11.3	121	0.00
18	Hexachloroethane	40.000	36.339	9.2	129	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	34.386	14.0	120	0.00
20	3+4-Methylphenols	40.000	35.470	11.3	121	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	120	0.00
22	Acetophenone	40.000	36.019	10.0	119	0.00
23 S	Nitrobenzene-d5	80.000	72.514	9.4	119	0.00
24	Nitrobenzene	40.000	36.287	9.3	120	0.00
25	Isophorone	40.000	35.803	10.5	119	0.00
26 C	2-Nitrophenol	40.000	37.500	6.3	120	0.00
27	2,4-Dimethylphenol	40.000	36.384	9.0	119	0.00
28	bis(2-Chloroethoxy)methane	40.000	35.931	10.2	121	0.00
29 C	2,4-Dichlorophenol	40.000	36.268	9.3	119	0.00
30	1,2,4-Trichlorobenzene	40.000	36.130	9.7	120	0.00
31	Naphthalene	40.000	35.961	10.1	120	0.00
32	Benzoic acid	40.000	40.128	-0.3	123	0.00
33	4-Chloroaniline	40.000	35.915	10.2	118	0.00
34 C	Hexachlorobutadiene	40.000	35.777	10.6	119	0.00
35	Caprolactam	40.000	36.699	8.3	120	0.00
36 C	4-Chloro-3-methylphenol	40.000	35.438	11.4	116	0.00
37	2-Methylnaphthalene	40.000	34.888	12.8	117	0.00
38	1-Methylnaphthalene	40.000	34.972	12.6	117	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	116	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	35.454	11.4	115	0.00
41 P	Hexachlorocyclopentadiene	40.000	33.819	15.5	107	0.00
42 S	2,4,6-Tribromophenol	80.000	70.242	12.2	113	0.00
43 C	2,4,6-Trichlorophenol	40.000	36.110	9.7	116	0.00
44	2,4,5-Trichlorophenol	40.000	36.176	9.6	116	0.00
45 S	2-Fluorobiphenyl	80.000	70.142	12.3	115	0.00
46	1,1'-Biphenyl	40.000	35.259	11.9	115	0.00
47	2-Chloronaphthalene	40.000	35.433	11.4	115	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011922\
 Data File : BM033785.D
 Acq On : 19 Jan 2022 16:38
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 19 17:09:00 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM011722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 17 16:30:56 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	36.780	8.0	114	0.00
49	Acenaphthylene	40.000	35.680	10.8	116	0.00
50	Dimethylphthalate	40.000	35.339	11.7	116	0.00
51	2,6-Dinitrotoluene	40.000	36.761	8.1	118	0.00
52 C	Acenaphthene	40.000	35.857	10.4	115	0.00
53	3-Nitroaniline	40.000	37.066	7.3	117	0.00
54 P	2,4-Dinitrophenol	40.000	38.869	2.8	118	0.00
55	Dibenzofuran	40.000	34.939	12.7	115	0.00
56 P	4-Nitrophenol	40.000	38.297	4.3	119	0.00
57	2,4-Dinitrotoluene	40.000	37.163	7.1	118	0.00
58	Fluorene	40.000	34.675	13.3	113	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	35.576	11.1	115	0.00
60	Diethylphthalate	40.000	35.515	11.2	115	0.00
61	4-Chlorophenyl-phenylether	40.000	34.500	13.8	114	0.00
62	4-Nitroaniline	40.000	37.273	6.8	117	0.00
63	Azobenzene	40.000	36.074	9.8	117	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	115	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.568	3.6	116	0.00
66 c	n-Nitrosodiphenylamine	40.000	35.492	11.3	114	0.00
67	4-Bromophenyl-phenylether	40.000	35.059	12.4	114	0.00
68	Hexachlorobenzene	40.000	34.551	13.6	113	0.00
69	Atrazine	40.000	35.835	10.4	115	0.00
70 C	Pentachlorophenol	40.000	37.567	6.1	116	0.00
71	Phenanthrene	40.000	35.242	11.9	115	0.00
72	Anthracene	40.000	35.670	10.8	115	0.00
73	Carbazole	40.000	35.911	10.2	115	0.00
74	Di-n-butylphthalate	40.000	37.239	6.9	118	0.00
75 C	Fluoranthene	40.000	36.207	9.5	117	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	120	0.00
77	Benzydine	40.000	39.238	1.9	133	0.00
78	Pyrene	40.000	35.220	12.0	115	0.00
79 S	Terphenyl-d14	80.000	69.282	13.4	114	0.00
80	Butylbenzylphthalate	40.000	38.217	4.5	124	0.00
81	Benzo(a)anthracene	40.000	36.017	10.0	119	0.00
82	3,3'-Dichlorobenzidine	40.000	36.433	8.9	121	0.00
83	Chrysene	40.000	35.758	10.6	120	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	37.416	6.5	123	0.00
85 c	Di-n-octyl phthalate	40.000	38.691	3.3	129	0.00
86 I	Perylene-d12	20.000	20.000	0.0	128	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	35.566	11.1	128	-0.01
88	Benzo(b)fluoranthene	40.000	36.504	8.7	128	0.00
89	Benzo(k)fluoranthene	40.000	34.981	12.5	122	0.00
90 C	Benzo(a)pyrene	40.000	36.147	9.6	127	0.00
91	Dibenzo(a,h)anthracene	40.000	35.959	10.1	128	0.00
92	Benzo(g,h,i)perylene	40.000	35.657	10.9	127	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011922\
Data File : BM033785.D
Acq On : 19 Jan 2022 16:38
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Jan 19 17:09:00 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM011722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 17 16:30:56 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0