

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012023\
 Data File : BM038504.D
 Acq On : 20 Jan 2023 17:51
 Operator : CG/JU
 Sample : PB150235BL
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB150235BL

Quant Time: Jan 21 01:24:24 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM012023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 21 01:21:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.969	152	67058	20.000	ng	0.00
21) Naphthalene-d8	10.780	136	218490	20.000	ng	0.00
39) Acenaphthene-d10	14.604	164	124526	20.000	ng	0.00
64) Phenanthrene-d10	17.345	188	262266	20.000	ng	# 0.00
76) Chrysene-d12	21.515	240	232588	20.000	ng	0.00
86) Perylene-d12	23.932	264	286896	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.522	112	449493	110.810	ng	0.00
7) Phenol-d6	7.134	99	729243	140.902	ng	0.00
23) Nitrobenzene-d5	9.145	82	447026	88.847	ng	0.00
42) 2,4,6-Tribromophenol	16.092	330	266914	148.847	ng	0.00
45) 2-Fluorobiphenyl	13.233	172	1001343	99.053	ng	0.00
79) Terphenyl-d14	19.956	244	1335744	111.960	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.351	88	570	0.300	ng	# 48
4) n-Nitrosodimethylamine	3.716	42	243	0.102	ng	# 1
9) Benzaldehyde	7.139	77	1896	0.618	ng	# 5
15) Benzyl Alcohol	7.975	79	2105	0.544	ng	# 32
24) Nitrobenzene	9.139	77	1993	0.391	ng	# 52
25) Isophorone	9.592	82	58	0.007	ng	# 67
46) 1,1'-Biphenyl	13.227	154	1947	0.195	ng	# 1
52) Acenaphthene	14.604	154	261	0.035	ng	# 4
60) Diethylphthalate	15.421	149	90	0.010	ng	# 58
63) Azobenzene	15.821	77	79	0.008	ng	# 51
66) n-Nitrosodiphenylamine	16.092	169	8112	1.030	ng	# 47
71) Phenanthrene	17.386	178	318	0.022	ng	# 60
72) Anthracene	17.486	178	125	0.008	ng	# 61
74) Di-n-butylphthalate	18.303	149	685	0.048	ng	# 78
75) Fluoranthene	19.415	202	231	0.013	ng	65
78) Pyrene	19.774	202	303	0.019	ng	# 57
80) Butylbenzylphthalate	20.644	149	430	0.077	ng	# 23
81) Benzo(a)anthracene	21.521	228	768	0.047	ng	# 51
82) 3,3'-Dichlorobenzidine	21.450	252	163	0.031	ng	# 25
83) Chrysene	21.556	228	149	0.009	ng	# 49
84) Bis(2-ethylhexyl)phtha...	21.415	149	181	0.022	ng	# 53
85) Di-n-octyl phthalate	22.350	149	183	0.013	ng	# 80
87) Indeno(1,2,3-cd)pyrene	26.468	276	133	0.006	ng	# 53
88) Benzo(b)fluoranthene	23.191	252	451	0.026	ng	# 1
89) Benzo(k)fluoranthene	23.274	252	728	0.041	ng	# 1
90) Benzo(a)pyrene	23.803	252	779	0.045	ng	# 1
92) Benzo(g,h,i)perylene	27.232	276	56	0.003	ng	# 56


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012023\
 Data File : BM038504.D
 Acq On : 20 Jan 2023 17:51
 Operator : CG/JU
 Sample : PB150235BL
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB150235BL

Quant Time: Jan 21 01:24:24 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM012023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 21 01:21:49 2023
 Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.969 min Scan# 813
Delta R.T. -0.000 min
Lab File: BM038504.D
Acq: 20 Jan 2023 17:51




Instrument :
BNA_M
ClientSampleId :
PB150235BL


Tgt Ion:152 Resp: 67058
Ion Ratio Lower Upper
152 100
150 155.9 124.2 186.4
115 56.4 44.8 67.2



#2
1,4-Dioxane
Concen: 0.300 ng
RT: 3.351 min Scan# 28
Delta R.T. -0.024 min
Lab File: BM038504.D
Acq: 20 Jan 2023 17:51

Tgt Ion: 88 Resp: 570
Ion Ratio Lower Upper
88 100
58 23.5 50.5 75.7#
43 0.0 23.9 35.9#





#4

n-Nitrosodimethylamine

Concen: 0.102 ng

RT: 3.716 min Scan# 90

Delta R.T. 0.012 min

Lab File: BM038504.D

Acq: 20 Jan 2023 17:51

Instrument :

BNA_M

ClientSampleId :

PB150235BL





Tgt Ion: 42 Resp: 243

Ion Ratio Lower Upper

42 100

74 0.0 105.0 157.4#

44 115.8 4.1 6.1#



#5

2-Fluorophenol

Concen: 110.810 ng

RT: 5.522 min Scan# 397

Delta R.T. 0.006 min

Lab File: BM038504.D

Acq: 20 Jan 2023 17:51




Tgt Ion: 112 Resp: 449493


Ion Ratio Lower Upper

112 100

64 50.7 47.3 70.9

63 31.6 28.8 43.2









#7
Phenol-d6
Concen: 140.902 ng
RT: 7.134 min Scan# 61
Delta R.T. -0.000 min
Lab File: BM038504.D
Acq: 20 Jan 2023 17:51




Instrument :
BNA_M
ClientSampleId :
PB150235BL


Tgt Ion: 99 Resp: 729243
Ion Ratio Lower Upper
99 100
42 18.2 16.0 24.0
71 40.7 33.8 50.6



#9
Benzaldehyde
Concen: 0.618 ng
RT: 7.139 min Scan# 672
Delta R.T. 0.029 min
Lab File: BM038504.D
Acq: 20 Jan 2023 17:51

Tgt Ion: 77 Resp: 1896
Ion Ratio Lower Upper
77 100
105 0.0 72.9 112.9#
106 0.0 63.8 103.8#









#15
Benzyl Alcohol
Concen: 0.544 ng
RT: 7.975 min Scan# 814
Delta R.T. -0.241 min
Lab File: BM038504.D
Acq: 20 Jan 2023 17:51




Instrument :
BNA_M
ClientSampleId :
PB150235BL


Tgt Ion: 79 Resp: 2105
Ion Ratio Lower Upper
79 100
108 18.6 55.0 82.6#
77 128.5 55.0 82.4#



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.780 min Scan# 1291
Delta R.T. -0.000 min
Lab File: BM038504.D
Acq: 20 Jan 2023 17:51

Tgt Ion: 136 Resp: 218490
Ion Ratio Lower Upper
136 100
137 11.0 8.4 12.6
54 6.9 6.0 9.0
68 5.0 4.5 6.7









#23
Nitrobenzene-d5
Concen: 88.847 ng
RT: 9.145 min Scan# 1013
Delta R.T. -0.000 min
Lab File: BM038504.D
Acq: 20 Jan 2023 17:51




Instrument :
BNA_M
ClientSampleId :
PB150235BL


Tgt Ion: 82 Resp: 447026
Ion Ratio Lower Upper
82 100
128 39.2 29.3 43.9
54 44.8 36.9 55.3



#24
Nitrobenzene
Concen: 0.391 ng
RT: 9.139 min Scan# 1012
Delta R.T. -0.047 min
Lab File: BM038504.D
Acq: 20 Jan 2023 17:51

Tgt Ion: 77 Resp: 1993
Ion Ratio Lower Upper
77 100
123 0.0 31.5 47.3#
65 13.2 11.2 16.8









#25
Isophorone
Concen: 0.007 ng
RT: 9.592 min Scan# 1089
Delta R.T. -0.118 min
Lab File: BM038504.D
Acq: 20 Jan 2023 17:51




Instrument :
BNA_M
ClientSampleId :
PB150235BL


Tgt Ion: 82 Resp: 58
Ion Ratio Lower Upper
82 100
95 0.0 5.6 8.4#
138 0.0 13.2 19.8#



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.604 min Scan# 1941
Delta R.T. -0.000 min
Lab File: BM038504.D
Acq: 20 Jan 2023 17:51


Tgt Ion: 164 Resp: 124526
Ion Ratio Lower Upper
164 100
162 105.2 82.0 123.0
160 46.6 35.8 53.6








#42
2,4,6-Tribromophenol
Concen: 148.847 ng
RT: 16.092 min Scan# 2194
Delta R.T. -0.000 min
Lab File: BM038504.D
Acq: 20 Jan 2023 17:51

Instrument :
BNA_M
ClientSampleId :
PB150235BL







Tgt Ion:330 Resp: 266914
Ion Ratio Lower Upper
330 100
332 96.2 77.5 116.3
141 33.9 29.0 43.4



#45
2-Fluorobiphenyl
Concen: 99.053 ng
RT: 13.233 min Scan# 1708
Delta R.T. -0.000 min
Lab File: BM038504.D
Acq: 20 Jan 2023 17:51

Tgt Ion:172 Resp: 1001343
Ion Ratio Lower Upper
172 100
171 34.5 28.0 42.0
170 22.6 19.5 29.3









#46
1,1'-Biphenyl
Concen: 0.195 ng
RT: 13.227 min Scan# 11
Delta R.T. -0.218 min
Lab File: BM038504.D
Acq: 20 Jan 2023 17:51




Instrument :
BNA_M
ClientSampleId :
PB150235BL


Tgt Ion:154 Resp: 1947
Ion Ratio Lower Upper
154 100
153 728.3 20.1 60.1#
76 1295.4 0.0 31.3#



#52
Acenaphthene
Concen: 0.035 ng
RT: 14.604 min Scan# 1941
Delta R.T. -0.065 min
Lab File: BM038504.D
Acq: 20 Jan 2023 17:51

Tgt Ion:154 Resp: 261
Ion Ratio Lower Upper
154 100
153 0.0 90.7 136.1#
152 0.0 43.8 65.8#









#60
Diethylphthalate
Concen: 0.010 ng
RT: 15.421 min Scan# 2080
Delta R.T. -0.000 min
Lab File: BM038504.D
Acq: 20 Jan 2023 17:51




Instrument :
BNA_M
ClientSampleId :
PB150235BL


Tgt Ion:149 Resp: 90
Ion Ratio Lower Upper
149 100
177 0.0 18.6 27.8#
150 0.0 9.5 14.3#



#63
Azobenzene
Concen: 0.008 ng
RT: 15.821 min Scan# 2148
Delta R.T. -0.112 min
Lab File: BM038504.D
Acq: 20 Jan 2023 17:51

Tgt Ion: 77 Resp: 79
Ion Ratio Lower Upper
77 100
182 0.0 4.7 44.7#
105 0.0 0.0 36.7
51 0.0 9.4 49.4#





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.345 min Scan# 24

Delta R.T. -0.000 min

Lab File: BM038504.D

Acq: 20 Jan 2023 17:51

Instrument :

BNA_M

ClientSampleId :

PB150235BL





Tgt Ion:188 Resp: 262266

Ion Ratio Lower Upper

188 100

94 5.8 6.0 9.0#

80 7.2 6.9 10.3



#66

n-Nitrosodiphenylamine

Concen: 1.030 ng

RT: 16.092 min Scan# 2194

Delta R.T. 0.235 min

Lab File: BM038504.D

Acq: 20 Jan 2023 17:51




Tgt Ion:169 Resp: 8112


Ion Ratio Lower Upper

169 100

168 6.6 55.8 83.8#

167 32.2 30.3 45.5









#71
Phenanthrene
Concen: 0.022 ng
RT: 17.386 min Scan# 2415
Delta R.T. -0.006 min
Lab File: BM038504.D
Acq: 20 Jan 2023 17:51




Instrument :
BNA_M
ClientSampleId :
PB150235BL


Tgt Ion:178 Resp: 318
Ion Ratio Lower Upper
178 100
176 0.0 15.4 23.2#
179 0.0 12.2 18.4#



#72
Anthracene
Concen: 0.008 ng
RT: 17.486 min Scan# 2431
Delta R.T. 0.006 min
Lab File: BM038504.D
Acq: 20 Jan 2023 17:51

Tgt Ion:178 Resp: 125
Ion Ratio Lower Upper
178 100
176 0.0 15.0 22.6#
179 0.0 12.2 18.4#





#74

Di-n-butylphthalate

Concen: 0.048 ng

RT: 18.303 min Scan# 21

Delta R.T. -0.000 min

Lab File: BM038504.D

Acq: 20 Jan 2023 17:51

Instrument :

BNA_M

ClientSampleId :

PB150235BL





Tgt Ion:149 Resp: 685

Ion Ratio Lower Upper

149 100

150 0.0 7.4 11.0#

104 0.0 4.6 6.8#



#75

Fluoranthene

Concen: 0.013 ng

RT: 19.415 min Scan# 2759

Delta R.T. 0.018 min

Lab File: BM038504.D

Acq: 20 Jan 2023 17:51




Tgt Ion:202 Resp: 231


Ion Ratio Lower Upper

202 100

101 0.0 0.0 28.7

203 0.0 0.0 37.6









#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.515 min Scan# 3117
Delta R.T. -0.006 min
Lab File: BM038504.D
Acq: 20 Jan 2023 17:51




Instrument :
BNA_M
ClientSampleId :
PB150235BL


Tgt Ion:240 Resp: 232588
Ion Ratio Lower Upper
240 100
120 6.6 6.2 9.2
236 25.6 20.2 30.2



#78
Pyrene
Concen: 0.019 ng
RT: 19.774 min Scan# 2820
Delta R.T. 0.012 min
Lab File: BM038504.D
Acq: 20 Jan 2023 17:51

Tgt Ion:202 Resp: 303
Ion Ratio Lower Upper
202 100
200 0.0 17.0 25.4#
203 0.0 14.2 21.2#









#79
 Terphenyl-d14
 Concen: 111.960 ng
 RT: 19.956 min Scan# 2851
 Delta R.T. -0.006 min
 Lab File: BM038504.D
 Acq: 20 Jan 2023 17:51




Instrument :
 BNA_M
 ClientSampleId :
 PB150235BL


Tgt Ion:244 Resp: 1335744
 Ion Ratio Lower Upper
 244 100
 212 7.6 6.0 9.0
 122 8.0 6.7 10.1



#80
 Butylbenzylphthalate
 Concen: 0.077 ng
 RT: 20.644 min Scan# 2968
 Delta R.T. -0.006 min
 Lab File: BM038504.D
 Acq: 20 Jan 2023 17:51

Tgt Ion:149 Resp: 430
 Ion Ratio Lower Upper
 149 100
 91 0.0 57.5 86.3#
 206 0.0 18.2 27.2#





#81

Benzo(a)anthracene

Concen: 0.047 ng

RT: 21.521 min Scan# 3117

Delta R.T. 0.018 min

Lab File: BM038504.D

Acq: 20 Jan 2023 17:51

Instrument :

BNA_M

ClientSampleId :

PB150235BL





Tgt Ion:228 Resp: 768

Ion Ratio Lower Upper

228 100

226 0.0 22.2 33.2#

229 0.0 16.0 24.0#



#82

3,3'-Dichlorobenzidine

Concen: 0.031 ng

RT: 21.450 min Scan# 3105

Delta R.T. 0.012 min

Lab File: BM038504.D

Acq: 20 Jan 2023 17:51




Tgt Ion:252 Resp: 163


Ion Ratio Lower Upper

252 100

254 0.0 51.9 77.9#

126 0.0 7.7 11.5#









#83
Chrysene
Concen: 0.009 ng
RT: 21.556 min Scan# 3123
Delta R.T. -0.000 min
Lab File: BM038504.D
Acq: 20 Jan 2023 17:51




Instrument :
BNA_M
ClientSampleId :
PB150235BL


Tgt Ion:228 Resp: 149
Ion Ratio Lower Upper
228 100
226 0.0 24.4 36.6#
229 0.0 15.7 23.5#



#84
Bis(2-ethylhexyl)phthalate
Concen: 0.022 ng
RT: 21.415 min Scan# 3099
Delta R.T. -0.000 min
Lab File: BM038504.D
Acq: 20 Jan 2023 17:51

Tgt Ion:149 Resp: 181
Ion Ratio Lower Upper
149 100
167 0.0 21.3 31.9#
279 0.0 3.0 4.4#





#85

Di-n-octyl phthalate

Concen: 0.013 ng

RT: 22.350 min Scan# 31

Delta R.T. 0.006 min

Lab File: BM038504.D

Acq: 20 Jan 2023 17:51

Instrument :

BNA_M

ClientSampleId :

PB150235BL





Tgt Ion:149 Resp: 183

Ion Ratio Lower Upper

149 100

167 0.0 1.3 1.9#

43 0.0 6.7 10.1#



#86

Perylene-d12

Concen: 20.000 ng

RT: 23.932 min Scan# 3527

Delta R.T. -0.000 min

Lab File: BM038504.D

Acq: 20 Jan 2023 17:51




Tgt Ion:264 Resp: 286896


Ion Ratio Lower Upper

264 100

260 22.3 18.5 27.7

265 22.3 17.8 26.6





#87

Indeno(1,2,3-cd)pyrene

Concen: 0.006 ng

RT: 26.468 min Scan# 3956

Delta R.T. 0.012 min

Lab File: BM038504.D

Acq: 20 Jan 2023 17:51

Instrument :

BNA_M

ClientSampleId :

PB150235BL





Tgt Ion:276 Resp: 133

Ion Ratio Lower Upper

276 100

138 0.0 15.8 23.8#

277 0.0 20.1 30.1#



#88

Benzo(b)fluoranthene

Concen: 0.026 ng

RT: 23.191 min Scan# 3401

Delta R.T. -0.006 min

Lab File: BM038504.D

Acq: 20 Jan 2023 17:51




Tgt Ion:252 Resp: 451


Ion Ratio Lower Upper

252 100

253 101.2 17.4 26.0#

125 84.9 6.4 9.6#









#89
Benzo(k)fluoranthene
Concen: 0.041 ng
RT: 23.274 min Scan# 3410
Delta R.T. 0.029 min
Lab File: BM038504.D
Acq: 20 Jan 2023 17:51




Instrument :
BNA_M
ClientSampleId :
PB150235BL


Tgt Ion:252 Resp: 728
Ion Ratio Lower Upper
252 100
253 77.1 17.8 26.8#
125 46.8 6.8 10.2#



#90
Benzo(a)pyrene
Concen: 0.045 ng
RT: 23.803 min Scan# 3505
Delta R.T. -0.024 min
Lab File: BM038504.D
Acq: 20 Jan 2023 17:51

Tgt Ion:252 Resp: 779
Ion Ratio Lower Upper
252 100
253 67.3 17.4 26.0#
125 86.4 7.3 10.9#





#92

Benzo(g,h,i)perylene

Concen: 0.003 ng

RT: 27.232 min Scan# 40

Delta R.T. -0.000 min

Lab File: BM038504.D


Acq: 20 Jan 2023 17:51

Instrument :

BNA_M

ClientSampleId :

PB150235BL



Tgt Ion:276 Resp: 56

Ion Ratio Lower Upper

276 100

277 0.0 18.8 28.2#

138 0.0 13.8 20.8#

