



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012323\
Data File : BM038505.D
Acq On : 23 Jan 2023 10:00
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM012023.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Sat Jan 21 01:21:49 2023






AutoFind: Scans 2477, 2478, 2479; Background Corrected with Scan 2471

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.0	49205	PASS
68	69	0.00	2	1.4	843	PASS
69	198	0.00	100	42.2	59363	PASS
70	69	0.00	2	0.6	348	PASS
127	198	10	80	44.8	62896	PASS
197	198	0.00	2	0.2	340	PASS
198	198	100	100	100.0	140517	PASS
199	198	5	9	6.9	9765	PASS
275	198	10	60	25.9	36365	PASS
365	198	1	100	4.1	5784	PASS
441	198	0.01	100	13.8	19372	PASS
442	442	50	100	100.0	118200	PASS
443	442	15	24	20.5	24187	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012323\
Data File : BM038505.D
Acq On : 23 Jan 2023 10:00
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Jan 24 06:42:04 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM012023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jan 21 01:21:49 2023
Response via : Initial Calibration



TIC: BM038505.D\data.ms

(70) Pentachlorophenol (C)

16.992min (-0.006) 0.00 ng




response 191649

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.60	65.45
264.00	61.70	60.19
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012323\
 Data File : BM038505.D
 Acq On : 23 Jan 2023 10:00
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP

Quant Time: Jan 24 06:42:04 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM012023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 21 01:21:49 2023
 Response via : Initial Calibration



TIC: BM038505.D\data.ms

(77) Benzidine

19.586min (+ 0.000) 0.00 ng

response 1185781

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.60	14.04
183.00	11.60	10.97
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
1/23/2023	BNA_M	<u>BM038505.D</u>
Compound Name	Response	Retention Time
DDT	640519	20.809
DDD	5779	20.415
DDE	177	19.851
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
5956	646475	0.92

Instrument :
BNA_M
ClientSampleId :
DFTPP