

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012422\
 Data File : BM033837.D
 Acq On : 24 Jan 2022 14:30
 Operator : CG/JU
 Sample : SSTD3.248
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD3.2048

Quant Time: Jan 24 15:12:32 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM012422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 24 14:03:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.943	152	2123	0.400	ng/ul	0.00
4) Naphthalene-d8	10.756	136	7482	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.580	164	3966	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.316	188	7950	0.400	ng/ul	0.00
17) Chrysene-d12	21.480	240	6576	0.400	ng/ul	# 0.00
23) Perylene-d12	23.883	264	6399	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.296	96	6769	2.993	ng/ul	0.00
6) 2-Methylnaphthalene-d10	0.000	152	0d	0.000	ng/ul	
18) Fluoranthene-d10	0.000	212	0d	0.000	ng/ul	
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.330	88	7103	2.879	ng/ul	# 76
14) Pentachlorophenol	16.969	266	11312	3.606	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012422\
Data File : BM033837.D
Acq On : 24 Jan 2022 14:30
Operator : CG/JU
Sample : SST3.248
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD3.2048

Quant Time: Jan 24 15:12:32 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM012422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 24 14:03:45 2022
Response via : Initial Calibration

