

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012422\
 Data File : BM033853.D
 Acq On : 25 Jan 2022 00:57
 Operator : CG/JU
 Sample : N1199-18
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0Q13

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 01/25/2022
 Supervised By :Yogesh Patel 01/31/2022

Quant Time: Jan 25 02:22:15 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM012422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 24 15:20:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.943	152	2019	0.400	ng/ul	0.00
4) Naphthalene-d8	10.752	136	7541	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.576	164	6465	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.320	188	10573	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.477	240	7712	0.400	ng/ul	# 0.00
23) Perylene-d12	23.875	264	7993	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.296	96	9399	4.453	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.339	152	4317	0.415	ng/ul	0.00
18) Fluoranthene-d10	19.334	212	11360	0.478	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.330	88	134	0.061	ng/ul#	48
5) Naphthalene	10.806	128	35709	1.723	ng/ul#	94
8) 1-Methylnaphthalene	12.632	142	111257	8.424	ng/ul	99
10) Acenaphthylene	14.299	152	11206	0.388	ng/ul#	60
11) Acenaphthene	14.640	153	28869	1.248	ng/ul	95
12) Fluorene	15.622	166	108095	4.096	ng/ul	96
14) Pentachlorophenol	17.000	266	9487369	1952.934	ng/ul	98
15) Phenanthrene	17.361	178	116679	3.464	ng/ul	95
16) Anthracene	17.448	178	27520	0.873	ng/ul	95
19) Fluoranthene	19.365	202	5428	0.154	ng/ul#	83
20) Pyrene	19.727	202	7671m	0.213	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012422\
Data File : BM033853.D
Acq On : 25 Jan 2022 00:57
Operator : CG/JU
Sample : N1199-18
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0Q13

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 01/25/2022
Supervised By :Yogesh Patel 01/31/2022

Quant Time: Jan 25 02:22:15 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM012422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 24 15:20:03 2022
Response via : Initial Calibration

