

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012422\
 Data File : BM033855.D
 Acq On : 25 Jan 2022 02:08
 Operator : CG/JU
 Sample : N1199-22
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0Q22

Quant Time: Jan 25 03:03:57 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM012422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 24 15:20:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.943	152	2106	0.400	ng/ul	0.00
4) Naphthalene-d8	10.751	136	7699	0.400	ng/ul #	0.00
9) Acenaphthene-d10	14.580	164	4689	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.313	188	9306	0.400	ng/ul	0.00
17) Chrysene-d12	21.477	240	7377	0.400	ng/ul #	0.00
23) Perylene-d12	23.875	264	7961	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.296	96	9525	4.326	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.344	152	3787	0.357	ng/ul	0.00
18) Fluoranthene-d10	19.334	212	9244	0.407	ng/ul	0.00
Target Compounds						
					Qvalue	
5) Naphthalene	10.805	128	708	0.033	ng/ul#	82
8) 1-Methylnaphthalene	12.631	142	1803	0.134	ng/ul	99
12) Fluorene	15.622	166	456	0.024	ng/ul#	93
14) Pentachlorophenol	16.969	266	5299	1.239	ng/ul	97
15) Phenanthrene	17.354	178	1299	0.044	ng/ul#	87
16) Anthracene	17.448	178	593	0.021	ng/ul#	65

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012422\
Data File : BM033855.D
Acq On : 25 Jan 2022 02:08
Operator : CG/JU
Sample : N1199-22
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0Q22

Quant Time: Jan 25 03:03:57 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM012422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 24 15:20:03 2022
Response via : Initial Calibration

