

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM012519\  
 Data File : BM018659.D  
 Acq On : 25 Jan 2019 19:10  
 Operator : JU/SJ  
 Sample : K1234-07  
 Misc :  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 WWB-3

Quant Time: Jan 28 02:48:19 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM010919.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Jan 18 22:01:30 2019  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min)      |
|-----------------------------|-------|------|----------|--------|-------|---------------|
| 1) 1,4-Dichlorobenzene-d4   | 7.73  | 152  | 244038   | 20.00  | ng    | 0.00          |
| 21) Naphthalene-d8          | 10.52 | 136  | 955693   | 20.00  | ng    | 0.00          |
| 39) Acenaphthene-d10        | 14.38 | 164  | 527088   | 20.00  | ng    | 0.00          |
| 64) Phenanthrene-d10        | 17.13 | 188  | 1191615  | 20.00  | ng    | 0.00          |
| 76) Chrysene-d12            | 21.32 | 240  | 1290960  | 20.00  | ng    | -0.01         |
| 87) Perylene-d12            | 23.59 | 264  | 1307086  | 20.00  | ng    | -0.01         |
| System Monitoring Compounds |       |      |          |        |       |               |
| 5) 2-Fluorophenol           | 5.32  | 112  | 1946578  | 147.29 | ng    | 0.00          |
| 7) Phenol-d6                | 6.91  | 99   | 2520922  | 150.18 | ng    | 0.00          |
| 23) Nitrobenzene-d5         | 8.90  | 82   | 1541345  | 93.50  | ng    | 0.00          |
| 42) 2,4,6-Tribromophenol    | 15.88 | 330  | 1061501  | 158.17 | ng    | 0.00          |
| 45) 2-Fluorobiphenyl        | 13.00 | 172  | 3871810  | 95.85  | ng    | 0.00          |
| 79) Terphenyl-d14           | 19.76 | 244  | 5906139  | 96.44  | ng    | 0.00          |
| Target Compounds            |       |      |          |        |       |               |
| 50) Dimethylphthalate       | 13.84 | 163  | 282038   | 6.515  | ng    | Qvalue<br>100 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM012519\  
Data File : BM018659.D  
Acq On : 25 Jan 2019 19:10  
Operator : JU/SJ  
Sample : K1234-07  
Misc :  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
WWB-3

Quant Time: Jan 28 02:48:19 2019  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM010919.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Jan 18 22:01:30 2019  
Response via : Initial Calibration

