

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM012519\
 Data File : BM018661.D
 Acq On : 25 Jan 2019 20:23
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 28 02:50:13 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM010919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 18 22:01:30 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	70	0.00
2	1,4-Dioxane	0.441	0.399	9.5	64	0.00
3	Pyridine	1.101	1.096	0.5	66	0.00
4	n-Nitrosodimethylamine	0.455	0.444	2.4	65	0.00
5 S	2-Fluorophenol	1.083	1.109	-2.4	70	0.00
6	Aniline	1.535	1.672	-8.9	73	-0.01
7 S	Phenol-d6	1.376	1.436	-4.4	71	0.00
8	2-Chlorophenol	1.265	1.325	-4.7	72	0.00
9	Benzaldehyde	0.770	0.750	2.6	68	0.00
10 C	Phenol	1.398	1.441	-3.1	71	0.00
11	bis(2-Chloroethyl)ether	1.098	1.123	-2.3	70	0.00
12	1,3-Dichlorobenzene	1.507	1.531	-1.6	71	0.00
13 C	1,4-Dichlorobenzene	1.527	1.546	-1.2	71	0.00
14	1,2-Dichlorobenzene	1.448	1.488	-2.8	72	0.00
15	Benzyl Alcohol	0.994	1.070	-7.6	73	0.00
16	2,2'-oxybis(1-Chloropropane	1.404	1.433	-2.1	72	0.00
17	2-Methylphenol	0.949	1.008	-6.2	72	0.00
18	Hexachloroethane	0.544	0.545	-0.2	71	0.00
19 P	n-Nitroso-di-n-propylamine	0.895	0.898	-0.3	69	0.00
20	3+4-Methylphenols	1.310	1.388	-6.0	71	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	69	0.00
22	Acetophenone	0.495	0.492	0.6	68	0.00
23 S	Nitrobenzene-d5	0.345	0.350	-1.4	69	0.00
24	Nitrobenzene	0.339	0.338	0.3	68	0.00
25	Isophorone	0.522	0.545	-4.4	70	-0.01
26 C	2-Nitrophenol	0.164	0.193	-17.7	76	0.00
27	2,4-Dimethylphenol	0.246	0.258	-4.9	70	0.00
28	bis(2-Chloroethoxy)methane	0.361	0.367	-1.7	69	-0.01
29 C	2,4-Dichlorophenol	0.293	0.320	-9.2	72	0.00
30	1,2,4-Trichlorobenzene	0.364	0.374	-2.7	72	0.00
31	Naphthalene	0.963	0.978	-1.6	70	0.00
32	Benzoic acid	0.150	0.231	-54.0#	103	-0.02
33	4-Chloroaniline	0.379	0.432	-14.0	74	0.00
34 C	Hexachlorobutadiene	0.230	0.242	-5.2	73	0.00
35	Caprolactam	0.100	0.107	-7.0	70	-0.02
36 C	4-Chloro-3-methylphenol	0.307	0.327	-6.5	69	0.00
37	2-Methylnaphthalene	0.681	0.687	-0.9	69	0.00
38	1-Methylnaphthalene	0.658	0.666	-1.2	69	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	67	0.00
40	1,2,4,5-Tetrachlorobenzene	0.699	0.730	-4.4	70	0.00
41 P	Hexachlorocyclopentadiene	0.384	0.232	39.6#	39#	-0.01
42 S	2,4,6-Tribromophenol	0.255	0.288	-12.9	72	0.00
43 C	2,4,6-Trichlorophenol	0.425	0.477	-12.2	72	0.00
44	2,4,5-Trichlorophenol	0.447	0.506	-13.2	72	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM012519\
 Data File : BM018661.D
 Acq On : 25 Jan 2019 20:23
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 28 02:50:13 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM010919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 18 22:01:30 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.533	1.557	-1.6	68	0.00
46	1,1'-Biphenyl	1.746	1.772	-1.5	68	0.00
47	2-Chloronaphthalene	1.354	1.390	-2.7	69	0.00
48	2-Nitroaniline	0.333	0.366	-9.9	70	0.00
49	Acenaphthylene	1.973	2.036	-3.2	68	0.00
50	Dimethylphthalate	1.643	1.644	-0.1	66	0.00
51	2,6-Dinitrotoluene	0.348	0.368	-5.7	68	-0.01
52 C	Acenaphthene	1.207	1.229	-1.8	69	0.00
53	3-Nitroaniline	0.351	0.394	-12.3	71	0.00
54 P	2,4-Dinitrophenol	0.158	0.132	16.5	55	0.00
55	Dibenzofuran	1.995	2.014	-1.0	68	0.00
56 P	4-Nitrophenol	0.303	0.327	-7.9	70	0.00
57	2,4-Dinitrotoluene	0.492	0.508	-3.3	68	0.00
58	Fluorene	1.486	1.522	-2.4	68	0.00
59	2,3,4,6-Tetrachlorophenol	0.396	0.435	-9.8	71	0.00
60	Diethylphthalate	1.658	1.688	-1.8	67	0.00
61	4-Chlorophenyl-phenylether	0.857	0.868	-1.3	68	0.00
62	4-Nitroaniline	0.383	0.389	-1.6	66	0.00
63	Azobenzene	1.352	1.403	-3.8	68	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	67	0.00
65	4,6-Dinitro-2-methylphenol	0.118	0.103	12.7	58	0.00
66 c	n-Nitrosodiphenylamine	0.620	0.638	-2.9	67	0.00
67	4-Bromophenyl-phenylether	0.224	0.232	-3.6	69	0.00
68	Hexachlorobenzene	0.251	0.257	-2.4	69	0.00
69	Atrazine	0.224	0.220	1.8	63	0.00
70 C	Pentachlorophenol	0.164	0.185	-12.8	74	0.00
71	Phenanthrene	1.064	1.075	-1.0	68	0.00
72	Anthracene	1.023	1.075	-5.1	69	0.00
73	Carbazole	1.051	1.099	-4.6	68	0.00
74	Di-n-butylphthalate	1.151	1.293	-12.3	71	0.00
75 C	Fluoranthene	1.226	1.275	-4.0	69	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	65	0.00
77	Benzidine	0.494	0.583	-18.0	67	0.00
78	Pyrene	1.185	1.260	-6.3	68	0.00
79 S	Terphenyl-d14	0.949	1.016	-7.1	68	0.00
80	Butylbenzylphthalate	0.506	0.577	-14.0	74	0.00
81	Benzo(a)anthracene	1.178	1.219	-3.5	66	0.00
82	3,3'-Dichlorobenzidine	0.446	0.496	-11.2	69	0.00
83	Chrysene	1.138	1.161	-2.0	66	0.00
84	Bis(2-ethylhexyl)phthalate	0.730	0.833	-14.1	73	0.00
85 c	Di-n-octyl phthalate	1.229	1.407	-14.5	73	0.00
86	Indeno(1,2,3-cd)pyrene	1.312	1.345	-2.5	66	0.00
87 I	Perylene-d12	1.000	1.000	0.0	65	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM012519\
Data File : BM018661.D
Acq On : 25 Jan 2019 20:23
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Jan 28 02:50:13 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM010919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 18 22:01:30 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.135	1.155	-1.8	67	0.00
89	Benzo(k)fluoranthene	1.144	1.171	-2.4	65	0.00
90 C	Benzo(a)pyrene	1.087	1.121	-3.1	66	0.00
91	Dibenzo(a,h)anthracene	1.101	1.130	-2.6	66	0.00
92	Benzo(q,h,i)perylene	1.072	1.101	-2.7	66	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0