

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012822\
 Data File : BM033920.D
 Acq On : 29 Jan 2022 00:51
 Operator : CG/JU
 Sample : PB142334BS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS334

Quant Time: Jan 29 01:22:57 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM010622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jan 28 23:26:29 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 01/31/2022
 Supervised By :mohammad ahmed 02/01/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.931	152	97051	20.000	ng/u1	0.00
20) Naphthalene-d8	10.742	136	405388	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.565	164	263908	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.306	188	553703	20.000	ng/u1	0.00
79) Chrysene-d12	21.465	240	563855	20.000	ng/u1	0.00
88) Perylene-d12	23.865	264	559160	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.290	96	12844	4.972	ng/uL	0.00
4) Pyridine-d5	3.713	84	171243	23.872	ng/u1	0.00
7) Phenol-d5	7.089	99	226155	25.945	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.254	67	139874	26.350	ng/u1	0.00
11) 2-Chlorophenol-d4	7.460	132	185864	27.889	ng/u1	0.00
15) 4-Methylphenol-d8	8.642	113	181964	25.218	ng/u1	0.00
21) Nitrobenzene-d5	9.089	128	88634	27.686	ng/u1	0.00
24) 2-Nitrophenol-d4	9.819	143	97035	28.599	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.360	165	183185	28.162	ng/u1	0.00
31) 4-Chloroaniline-d4	10.872	131	230789	23.869	ng/u1	0.00
46) Dimethylphthalate-d6	13.977	166	595436	28.637	ng/u1	0.00
49) Acenaphthylene-d8	14.260	160	690719	28.158	ng/u1	0.00
54) 4-Nitrophenol-d4	14.754	143	104367	26.153	ng/u1	0.00
60) Fluorene-d10	15.554	176	492912	28.574	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.671	200	93764	28.468	ng/u1	0.00
73) Anthracene-d10	17.406	188	795342	29.602	ng/u1	0.00
81) Pyrene-d10	19.689	212	941506	30.920	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.706	264	903149	31.043	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.325	88	27251	10.106	ng/uL#	88
5) Pyridine	3.737	79	181369	24.655	ng/u1#	84
6) Benzaldehyde	7.060	77	127888	25.425	ng/u1#	83
8) Phenol	7.113	94	233600	26.086	ng/u1#	87
10) Bis(2-Chloroethyl)ether	7.348	93	184345	26.430	ng/u1	92
12) 2-Chlorophenol	7.489	128	189331	27.046	ng/u1	91
13) 2-Methylphenol	8.372	108	175632	25.507	ng/u1	96
14) 2,2'-oxybis(1-Chloropr...	8.460	45	254865	25.973	ng/u1#	95
16) Acetophenone	8.754	105	294870	25.974	ng/u1#	87
17) N-Nitroso-di-n-propyla...	8.742	70	149964	25.509	ng/u1#	88
18) 4-Methylphenol	8.701	108	192190	25.536	ng/u1	93
19) Hexachloroethane	9.019	117	83870	28.252	ng/u1	88
22) Nitrobenzene	9.136	77	220807	27.200	ng/u1#	89
23) Isophorone	9.666	82	415873	26.805	ng/u1	96
25) 2-Nitrophenol	9.854	139	104637	28.584	ng/u1#	82
26) 2,4-Dimethylphenol	9.913	107	233106	27.530	ng/u1	96
27) Bis(2-Chloroethoxy)met...	10.148	93	251196	27.012	ng/u1	99
29) 2,4-Dichlorophenol	10.389	162	181778	27.776	ng/u1	97
30) Naphthalene	10.795	128	614186	27.353	ng/u1	99
32) 4-Chloroaniline	10.901	127	223339	22.852	ng/u1	99
33) Hexachlorobutadiene	11.089	225	132934	29.474	ng/u1	95
34) Caprolactam	11.666	113	60266m	26.927	ng/u1	
35) 4-Chloro-3-methylphenol	12.024	107	218326	27.331	ng/u1	87
36) 2-Methylnaphthalene	12.401	142	424835	27.472	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012822\
 Data File : BM033920.D
 Acq On : 29 Jan 2022 00:51
 Operator : CG/JU
 Sample : PB142334BS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS334


Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 01/31/2022
 Supervised By :mohammad ahmed 02/01/2022

Quant Time: Jan 29 01:22:57 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM010622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jan 28 23:26:29 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.619	142	431895	27.112	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.771	216	225422	28.585	ng/ul	98
40) Hexachlorocyclopentadiene	12.754	237	142623	27.542	ng/ul	100
41) 2,4,6-Trichlorophenol	13.007	196	147154	29.037	ng/ul	96
42) 2,4,5-Trichlorophenol	13.077	196	160589	28.642	ng/ul	92
43) 1,1'-Biphenyl	13.407	154	579290	27.990	ng/ul	99
44) 2-Chloronaphthalene	13.448	162	436728	28.083	ng/ul	98
45) 2-Nitroaniline	13.648	65	145997m	28.856	ng/ul	
47) Dimethylphthalate	14.024	163	594633	28.605	ng/ul	100
48) 2,6-Dinitrotoluene	14.142	165	117661	29.529	ng/ul	95
50) Acenaphthylene	14.289	152	729677	28.008	ng/ul	100
51) 3-Nitroaniline	14.465	138	112629	27.505	ng/ul	87
52) Acenaphthene	14.630	153	482325	28.125	ng/ul	96
53) 2,4-Dinitrophenol	14.671	184	59517	24.818	ng/ul	88
55) 4-Nitrophenol	14.771	109	100310	25.946	ng/ul	84
56) Dibenzofuran	14.965	168	697325	28.351	ng/ul	99
57) 2,4-Dinitrotoluene	14.924	165	174682	29.280	ng/ul#	81
58) 2,3,4,6-Tetrachlorophenol	15.189	232	141205	29.714	ng/ul#	99
59) Diethylphthalate	15.383	149	641595	29.213	ng/ul	99
61) Fluorene	15.612	166	571491	28.608	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.607	204	288645	28.790	ng/ul	98
63) 4-Nitroaniline	15.624	138	127549	32.332	ng/ul	85
66) 4,6-Dinitro-2-methylph...	15.683	198	96525	29.162	ng/ul#	98
67) N-Nitrosodiphenylamine	15.818	169	498885	29.676	ng/ul	99
68) 4-Bromophenyl-phenylether	16.495	248	175023	31.811	ng/ul	98
69) Hexachlorobenzene	16.618	284	202872	30.278	ng/ul	96
70) Atrazine	16.765	200	187500	28.023	ng/ul	99
71) Pentachlorophenol	16.954	266	128892	29.407	ng/ul	98
72) Phenanthrene	17.348	178	931975	29.510	ng/ul	99
74) Anthracene	17.436	178	942834	29.397	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.371	216	227702	27.336	ng/uL	98
76) Pentachlorobenzene	14.889	250	233663	29.865	ng/uL	97
77) Carbazole	17.706	167	851184	29.585	ng/ul	99
78) Di-n-butylphthalate	18.271	149	1122448	30.980	ng/ul	98
80) Fluoranthene	19.353	202	1115553	30.820	ng/ul	98
82) Pyrene	19.718	202	1147222	30.658	ng/ul	99
83) Butylbenzylphthalate	20.606	149	513936	31.748	ng/ul	89
84) 3,3'-Dichlorobenzidine	21.383	252	328430	26.180	ng/ul	96
85) Benzo(a)anthracene	21.453	228	1094287	29.643	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.383	149	783529	30.861	ng/ul	99
87) Chrysene	21.506	228	1079616	30.123	ng/ul	99
89) Di-n-octyl phthalate	22.300	149	1342186	29.270	ng/ul	100
90) Benzo(b)fluoranthene	23.130	252	1125628	29.639	ng/ul	99
91) Benzo(k)fluoranthene	23.183	252	1093681	30.940	ng/ul	100
93) Benzo(a)pyrene	23.759	252	1103484	31.069	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.365	276	1254124	35.087	ng/ul	96
95) Dibenzo(a,h)anthracene	26.382	278	1083637	35.182	ng/ul	98
96) Benzo(g,h,i)perylene	27.135	276	1048135	35.935	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 10.106 ng/uL
RT: 3.325 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Instrument :
BNA_M
ClientSampleId :
SLCS334




Tgt Ion: 88 Resp: 2725
Ion Ratio Lower Upper
88 100
43 31.9 36.2 54.4
58 78.0 68.5 102.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022




#5
Pyridine
Concen: 24.655 ng/uL
RT: 3.737 min Scan# 85
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Tgt Ion: 79 Resp: 181369
Ion Ratio Lower Upper
79 100
52 77.5 78.1 117.1
51 40.1 36.0 54.0





#6
Benzaldehyde
Concen: 25.425 ng/ul
RT: 7.060 min Scan# 610
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Instrument :
BNA_M
ClientSampleId :
SLCS334




Tgt Ion: 77 Resp: 12788
Ion Ratio Lower Upper
77 100
105 95.6 65.6 98.4
106 91.6 60.6 90.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022




#8
Phenol
Concen: 26.086 ng/ul
RT: 7.113 min Scan# 659
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Tgt Ion: 94 Resp: 233600
Ion Ratio Lower Upper
94 100
65 31.6 31.7 47.5#
66 47.1 45.4 68.0





#10
Bis(2-Chloroethyl)ether
Concen: 26.430 ng/ul
RT: 7.348 min Scan# 699
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Instrument :
BNA_M
ClientSampleId :
SLCS334




Tgt Ion: 93 Resp: 184349
Ion Ratio Lower Upper
93 100
63 79.3 71.7 107.5
95 32.9 26.3 39.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022




#12
2-Chlorophenol
Concen: 27.046 ng/ul
RT: 7.489 min Scan# 723
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51



Tgt Ion:128 Resp: 189331
Ion Ratio Lower Upper
128 100
64 51.1 49.4 74.2
130 32.9 26.8 40.2









#13
2-Methylphenol
Concen: 25.507 ng/ul
RT: 8.372 min Scan# 81
Delta R.T. -0.006 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51

Instrument :
BNA_M
ClientSampleId :
SLCS334

Tgt Ion:108 Resp: 17563
Ion Ratio Lower Upper
108 100
107 91.8 70.2 105.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022



#14
2,2'-oxybis(1-Chloropropane)
Concen: 25.973 ng/ul
RT: 8.460 min Scan# 888
Delta R.T. -0.012 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51


Tgt Ion: 45 Resp: 254865
Ion Ratio Lower Upper
45 100
77 15.1 9.9 14.9#
79 11.1 8.3 12.5





#16
Acetophenone
Concen: 25.974 ng/ul
RT: 8.754 min Scan# 91
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Instrument :
BNA_M
ClientSampleId :
SLCS334



Tgt Ion:105 Resp: 294870
Ion Ratio Lower Upper
105 100
77 79.7 73.1 109.7
51 30.2 30.5 45.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022



#17
N-Nitroso-di-n-propylamine
Concen: 25.509 ng/ul
RT: 8.742 min Scan# 936
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51


Tgt Ion: 70 Resp: 149964
Ion Ratio Lower Upper
70 100
42 59.2 56.9 85.3
101 10.0 6.9 10.3
130 20.8 13.8 20.8





#18
4-Methylphenol
Concen: 25.536 ng/ul
RT: 8.701 min Scan# 91
Delta R.T. -0.006 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51



Instrument :
BNA_M
ClientSampleId :
SLCS334



Tgt Ion:108 Resp: 192190
Ion Ratio Lower Upper
108 100
107 116.3 87.4 131.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022



#19
Hexachloroethane
Concen: 28.252 ng/ul
RT: 9.019 min Scan# 983
Delta R.T. -0.006 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51


Tgt Ion:117 Resp: 83870
Ion Ratio Lower Upper
117 100
201 97.2 69.6 104.4
199 59.8 40.6 60.8





#22
Nitrobenzene
Concen: 27.200 ng/u1
RT: 9.136 min Scan# 1003
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Instrument :
BNA_M
ClientSampleId :
SLCS334




Tgt Ion: 77 Resp: 220807
Ion Ratio Lower Upper
77 100
123 45.2 29.0 43.6
65 14.5 11.4 17.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022




#23
Isophorone
Concen: 26.805 ng/u1
RT: 9.666 min Scan# 1093
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Tgt Ion: 82 Resp: 415873
Ion Ratio Lower Upper
82 100
95 7.4 6.5 9.7
138 17.0 11.8 17.8





#25
2-Nitrophenol
Concen: 28.584 ng/ul
RT: 9.854 min Scan# 1125
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Instrument :
BNA_M
ClientSampleId :
SLCS334



Tgt Ion:139 Resp: 104637
Ion Ratio Lower Upper
139 100
65 55.9 59.0 88.6
109 37.5 37.2 55.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022



#26
2,4-Dimethylphenol
Concen: 27.530 ng/ul
RT: 9.913 min Scan# 1135
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51


Tgt Ion:107 Resp: 233106
Ion Ratio Lower Upper
107 100
121 47.8 34.8 52.2
122 77.7 60.0 90.0





#27
Bis(2-Chloroethoxy)methane
Concen: 27.012 ng/ul
RT: 10.148 min Scan# 1175
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Instrument :
BNA_M
ClientSampleId :
SLCS334




Tgt Ion: 93 Resp: 251190
Ion Ratio Lower Upper
93 100
95 31.5 25.3 37.9
123 11.5 7.9 11.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022




#29
2,4-Dichlorophenol
Concen: 27.776 ng/ul
RT: 10.389 min Scan# 1216
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Tgt Ion:162 Resp: 181778
Ion Ratio Lower Upper
162 100
164 66.4 50.0 75.0
98 38.7 30.6 45.8





#30
Naphthalene
Concen: 27.353 ng/ul
RT: 10.795 min Scan# 1285
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Instrument :
BNA_M
ClientSampleId :
SLCS334




Tgt Ion:128 Resp: 614180
Ion Ratio Lower Upper
128 100
129 10.9 9.1 13.7
127 13.0 10.2 15.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022




#32
4-Chloroaniline
Concen: 22.852 ng/ul
RT: 10.901 min Scan# 1303
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Tgt Ion:127 Resp: 223339
Ion Ratio Lower Upper
127 100
129 32.0 25.4 38.0





#33
Hexachlorobutadiene
Concen: 29.474 ng/ul
RT: 11.089 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Instrument :
BNA_M
ClientSampleId :
SLCS334




Tgt Ion:225 Resp: 132934
Ion Ratio Lower Upper
225 100
223 60.8 52.2 78.2
227 63.0 52.8 79.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022




#34
Caprolactam
Concen: 26.927 ng/ul m
RT: 11.666 min Scan# 1433
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Tgt Ion:113 Resp: 60266
Ion Ratio Lower Upper
113 100
55 172.4 157.9 236.9
56 136.1 131.8 197.6





#35
4-Chloro-3-methylphenol
Concen: 27.331 ng/ul
RT: 12.024 min Scan# 1494
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Instrument :
BNA_M
ClientSampleId :
SLCS334




Tgt Ion:107 Resp: 218320
Ion Ratio Lower Upper
107 100
144 23.2 17.1 25.7
142 74.0 49.4 74.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022




#36
2-Methylnaphthalene
Concen: 27.472 ng/ul
RT: 12.401 min Scan# 1558
Delta R.T. -0.006 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Tgt Ion:142 Resp: 424835
Ion Ratio Lower Upper
142 100
141 87.8 71.1 106.7





#37
1-Methylnaphthalene
Concen: 27.112 ng/ul
RT: 12.619 min Scan# 1595
Delta R.T. -0.006 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Instrument :
BNA_M
ClientSampleId :
SLCS334




Tgt Ion:142 Resp: 43189
Ion Ratio Lower Upper
142 100
141 92.7 75.5 113.3
116 4.4 4.0 6.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 28.585 ng/ul
RT: 12.771 min Scan# 1621
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Tgt Ion:216 Resp: 225422
Ion Ratio Lower Upper
216 100
214 78.3 61.3 91.9
179 22.1 18.0 27.0
108 18.5 16.6 25.0





#40
Hexachlorocyclopentadiene
Concen: 27.542 ng/ul
RT: 12.754 min Scan# 1618
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Instrument :
BNA_M
ClientSampleId :
SLCS334




Tgt Ion:237 Resp: 142623
Ion Ratio Lower Upper
237 100
235 63.2 50.6 76.0
272 11.5 9.6 14.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022




#41
2,4,6-Trichlorophenol
Concen: 29.037 ng/ul
RT: 13.007 min Scan# 1661
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51



Tgt Ion:196 Resp: 147154
Ion Ratio Lower Upper
196 100
198 93.9 78.9 118.3
200 30.5 23.7 35.5






#42
2,4,5-Trichlorophenol
Concen: 28.642 ng/u1
RT: 13.077 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51

Instrument :

BNA_M

ClientSampleId :

SLCS334






Tgt Ion:196 Resp: 160589
Ion Ratio Lower Upper
196 100
198 94.0 82.2 123.2
200 30.7 27.9 41.9


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022




#43
1,1'-Biphenyl
Concen: 27.990 ng/u1
RT: 13.407 min Scan# 1729
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Tgt Ion:154 Resp: 579290
Ion Ratio Lower Upper
154 100
153 40.6 33.1 49.7
76 12.8 10.7 16.1





#44
2-Chloronaphthalene
Concen: 28.083 ng/ul
RT: 13.448 min Scan# 1736
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Instrument :
BNA_M
ClientSampleId :
SLCS334




Tgt Ion:162 Resp: 436728
Ion Ratio Lower Upper
162 100
164 32.9 26.2 39.4
127 44.5 33.7 50.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022




#45
2-Nitroaniline
Concen: 28.856 ng/ul m
RT: 13.648 min Scan# 1770
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Tgt Ion: 65 Resp: 145997
Ion Ratio Lower Upper
65 100
92 66.7 47.8 71.6
138 100.0 61.0 91.6#





#47
Dimethylphthalate
Concen: 28.605 ng/ul
RT: 14.024 min Scan# 1834
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Instrument :
BNA_M
ClientSampleId :
SLCS334




Tgt Ion:163 Resp: 59463
Ion Ratio Lower Upper
163 100
194 4.0 3.3 4.9
164 10.0 8.2 12.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022




#48
2,6-Dinitrotoluene
Concen: 29.529 ng/ul
RT: 14.142 min Scan# 1854
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Tgt Ion:165 Resp: 117661
Ion Ratio Lower Upper
165 100
89 65.4 55.7 83.5
121 22.6 20.0 30.0





#50
Acenaphthylene
Concen: 28.008 ng/u1
RT: 14.289 min Scan# 1880
Delta R.T. -0.006 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Instrument :
BNA_M
ClientSampleId :
SLCS334



Tgt Ion:152 Resp: 72967
Ion Ratio Lower Upper
152 100
151 19.8 15.8 23.8
153 12.7 10.1 15.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022



#51
3-Nitroaniline
Concen: 27.505 ng/u1
RT: 14.465 min Scan# 1909
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51


Tgt Ion:138 Resp: 112629
Ion Ratio Lower Upper
138 100
108 12.6 10.8 16.2
92 121.1 110.6 166.0





#52
Acenaphthene
Concen: 28.125 ng/ul
RT: 14.630 min Scan# 1938
Delta R.T. -0.006 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Instrument :
BNA_M
ClientSampleId :
SLCS334




Tgt Ion:153 Resp: 482329
Ion Ratio Lower Upper
153 100
152 47.0 39.4 59.2
154 86.8 72.7 109.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022




#53
2,4-Dinitrophenol
Concen: 24.818 ng/ul
RT: 14.671 min Scan# 1944
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Tgt Ion:184 Resp: 59517
Ion Ratio Lower Upper
184 100
63 75.1 73.8 110.8
154 75.2 63.5 95.3





#55
4-Nitrophenol
Concen: 25.946 ng/ul
RT: 14.771 min Scan# 1961
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Instrument :
BNA_M
ClientSampleId :
SLCS334




Tgt Ion:109 Resp: 100310
Ion Ratio Lower Upper
109 100
139 113.7 77.6 116.4
65 121.5 112.2 168.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022




#56
Dibenzofuran
Concen: 28.351 ng/ul
RT: 14.965 min Scan# 1994
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Tgt Ion:168 Resp: 697325
Ion Ratio Lower Upper
168 100
139 42.0 34.0 51.0





#57
2,4-Dinitrotoluene
Concen: 29.280 ng/ul
RT: 14.924 min Scan# 1987
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Instrument :
BNA_M
ClientSampleId :
SLCS334




Tgt Ion:165 Resp: 17468
Ion Ratio Lower Upper
165 100
63 48.9 51.8 77.8
182 2.7 2.6 4.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022




#58
2,3,4,6-Tetrachlorophenol
Concen: 29.714 ng/ul
RT: 15.189 min Scan# 2032
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Tgt Ion:232 Resp: 141205
Ion Ratio Lower Upper
232 100
131 52.4 41.4 62.2
130 3.2 2.0 3.0#
166 34.6 27.4 41.0





#59
Diethylphthalate
Concen: 29.213 ng/ul
RT: 15.383 min Scan# 2065
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Instrument :
BNA_M
ClientSampleId :
SLCS334




Tgt Ion:149 Resp: 641598
Ion Ratio Lower Upper
149 100
177 20.8 17.1 25.7
150 12.7 9.9 14.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022




#61
Fluorene
Concen: 28.608 ng/ul
RT: 15.612 min Scan# 2104
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Tgt Ion:166 Resp: 571491
Ion Ratio Lower Upper
166 100
165 98.9 77.9 116.9
167 13.2 10.3 15.5





#62
4-Chlorophenyl-phenylether
Concen: 28.790 ng/ul
RT: 15.607 min Scan# 2103
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Instrument :
BNA_M
ClientSampleId :
SLCS334




Tgt Ion:204 Resp: 288649
Ion Ratio Lower Upper
204 100
206 33.0 27.8 41.8
141 70.1 54.9 82.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022




#63
4-Nitroaniline
Concen: 32.332 ng/ul
RT: 15.624 min Scan# 2106
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Tgt Ion:138 Resp: 127549
Ion Ratio Lower Upper
138 100
92 54.2 48.0 72.0
108 92.4 89.5 134.3





#66
4,6-Dinitro-2-methylphenol
Concen: 29.162 ng/ul
RT: 15.683 min Scan# 2116
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Instrument :
BNA_M
ClientSampleId :
SLCS334




Tgt Ion:198 Resp: 9652
Ion Ratio Lower Upper
198 100
182 3.8 4.2 6.2
77 32.1 26.3 39.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022




#67
N-Nitrosodiphenylamine
Concen: 29.676 ng/ul
RT: 15.818 min Scan# 2139
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Tgt Ion:169 Resp: 498885
Ion Ratio Lower Upper
169 100
168 66.8 52.9 79.3
167 35.9 29.9 44.9





#68
4-Bromophenyl-phenylether
Concen: 31.811 ng/ul
RT: 16.495 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Instrument :
BNA_M
ClientSampleId :
SLCS334




Tgt Ion:248 Resp: 17502
Ion Ratio Lower Upper
248 100
250 97.2 79.0 118.4
141 90.0 69.7 104.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022




#69
Hexachlorobenzene
Concen: 30.278 ng/ul
RT: 16.618 min Scan# 2275
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Tgt Ion:284 Resp: 202872
Ion Ratio Lower Upper
284 100
142 40.0 34.3 51.5
249 31.1 26.1 39.1





#70
Atrazine
Concen: 28.023 ng/ul
RT: 16.765 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Instrument :
BNA_M
ClientSampleId :
SLCS334




Tgt Ion:200 Resp: 187500
Ion Ratio Lower Upper
200 100
173 26.8 20.6 30.8
215 46.7 36.9 55.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022




#71
Pentachlorophenol
Concen: 29.407 ng/ul
RT: 16.954 min Scan# 2332
Delta R.T. -0.006 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Tgt Ion:266 Resp: 128892
Ion Ratio Lower Upper
266 100
264 60.8 49.1 73.7
268 62.3 47.8 71.8





#72
Phenanthrene
Concen: 29.510 ng/u1
RT: 17.348 min Scan# 2399
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Instrument :
BNA_M
ClientSampleId :
SLCS334




Tgt Ion:178 Resp: 93197
Ion Ratio Lower Upper
178 100
179 15.3 11.7 17.5
176 19.2 15.1 22.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022




#74
Anthracene
Concen: 29.397 ng/u1
RT: 17.436 min Scan# 2414
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Tgt Ion:178 Resp: 942834
Ion Ratio Lower Upper
178 100
179 15.6 12.3 18.5
176 18.4 14.5 21.7





#75
1,2,3,4-Tetrachlorobenzene
Concen: 27.336 ng/uL
RT: 13.371 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Instrument :
BNA_M
ClientSampleId :
SLCS334




Tgt Ion:216 Resp: 22770
Ion Ratio Lower Upper
216 100
214 79.6 65.4 98.0
218 48.0 38.8 58.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022




#76
Pentachlorobenzene
Concen: 29.865 ng/uL
RT: 14.889 min Scan# 1981
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Tgt Ion:250 Resp: 233663
Ion Ratio Lower Upper
250 100
248 62.6 51.6 77.4
252 62.3 52.0 78.0





#77
 Carbazole
 Concen: 29.585 ng/u1
 RT: 17.706 min Scan# 2460
 Delta R.T. -0.000 min
 Lab File: BM033920.D
 Acq: 29 Jan 2022 00:51




Instrument :
 BNA_M
 ClientSampleId :
 SLCS334




Tgt Ion:167 Resp: 851184
 Ion Ratio Lower Upper
 167 100
 166 21.2 16.9 25.3
 139 13.9 10.5 15.7

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 01/31/2022
 Supervised By :mohammad ahmed 02/01/2022




#78
 Di-n-butylphthalate
 Concen: 30.980 ng/u1
 RT: 18.271 min Scan# 2556
 Delta R.T. -0.000 min
 Lab File: BM033920.D
 Acq: 29 Jan 2022 00:51




Tgt Ion:149 Resp: 1122448
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.0 10.6
 104 5.6 5.1 7.7





#80
Fluoranthene
Concen: 30.820 ng/u1
RT: 19.353 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Instrument :
BNA_M
ClientSampleId :
SLCS334




Tgt Ion:202 Resp: 111555
Ion Ratio Lower Upper
202 100
101 11.2 8.2 12.4
100 8.9 6.9 10.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022




#82
Pyrene
Concen: 30.658 ng/u1
RT: 19.718 min Scan# 2802
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51



Tgt Ion:202 Resp: 1147222
Ion Ratio Lower Upper
202 100
101 12.1 9.8 14.6
100 10.2 8.6 13.0






#83
Butylbenzylphthalate
Concen: 31.748 ng/ul
RT: 20.606 min Scan# 2953
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51

Instrument :

BNA_M

ClientSampleId :

SLCS334





Tgt Ion:149 Resp: 513930
Ion Ratio Lower Upper
149 100
91 74.8 69.3 103.9
206 19.5 16.1 24.1


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022




#84
3,3'-Dichlorobenzidine
Concen: 26.180 ng/ul
RT: 21.383 min Scan# 3085
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Tgt Ion:252 Resp: 328430
Ion Ratio Lower Upper
252 100
254 65.1 49.4 74.2
126 10.7 8.2 12.4





#85
Benzo(a)anthracene
Concen: 29.643 ng/ul
RT: 21.453 min Scan# 3097
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Instrument :
BNA_M
ClientSampleId :
SLCS334




Tgt Ion:228 Resp: 109428
Ion Ratio Lower Upper
228 100
229 19.5 15.0 22.4
226 27.1 21.4 32.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022




#86
Bis(2-ethylhexyl)phthalate
Concen: 30.861 ng/ul
RT: 21.383 min Scan# 3085
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Tgt Ion:149 Resp: 783529
Ion Ratio Lower Upper
149 100
167 26.8 21.8 32.8
279 3.9 3.2 4.8





#87
Chrysene
Concen: 30.123 ng/ul
RT: 21.506 min Scan# 3106
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Instrument :
BNA_M
ClientSampleId :
SLCS334



Tgt Ion:228 Resp: 1079610
Ion Ratio Lower Upper
228 100
226 29.8 24.5 36.7
229 19.4 15.9 23.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022



#89
Di-n-octyl phthalate
Concen: 29.270 ng/ul
RT: 22.300 min Scan# 3241
Delta R.T. -0.006 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51


Tgt Ion:149 Resp: 1342186





#90
Benzo(b)fluoranthene
Concen: 29.639 ng/u1
RT: 23.130 min Scan# 3382
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Instrument :
BNA_M
ClientSampleId :
SLCS334



Tgt Ion:252 Resp: 1125628
Ion Ratio Lower Upper
252 100
253 22.2 17.7 26.5
125 10.2 7.8 11.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022



#91
Benzo(k)fluoranthene
Concen: 30.940 ng/u1
RT: 23.183 min Scan# 3391
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51


Tgt Ion:252 Resp: 1093681
Ion Ratio Lower Upper
252 100
253 22.0 17.6 26.4
125 9.5 7.4 11.2





#93
Benzo(a)pyrene
Concen: 31.069 ng/u1
RT: 23.759 min Scan# 3489
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Instrument :
BNA_M
ClientSampleId :
SLCS334




Tgt Ion:252 Resp: 1103484
Ion Ratio Lower Upper
252 100
253 21.4 17.4 26.0
125 10.2 7.8 11.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022




#94
Indeno(1,2,3-cd)pyrene
Concen: 35.087 ng/u1
RT: 26.365 min Scan# 3932
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Tgt Ion:276 Resp: 1254124
Ion Ratio Lower Upper
276 100
138 23.0 16.5 24.7
277 24.5 20.6 31.0





#95
Dibenzo(a,h)anthracene
Concen: 35.182 ng/ul
RT: 26.382 min Scan# 3935
Delta R.T. -0.000 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51




Instrument :
BNA_M
ClientSampleId :
SLCS334




Tgt Ion:278 Resp: 108363
Ion Ratio Lower Upper
278 100
139 15.1 11.0 16.4
279 22.9 18.1 27.1

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 01/31/2022
Supervised By :mohammad ahmed 02/01/2022



#96
Benzo(g,h,i)perylene
Concen: 35.935 ng/ul
RT: 27.135 min Scan# 4063
Delta R.T. 0.006 min
Lab File: BM033920.D
Acq: 29 Jan 2022 00:51



Tgt Ion:276 Resp: 1048135
Ion Ratio Lower Upper
276 100
138 20.5 15.6 23.4
277 23.6 18.7 28.1

