

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM013023\
 Data File : BM038611.D
 Acq On : 30 Jan 2023 18:39
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV032

Quant Time: Jan 31 01:05:48 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM013023.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jan 31 01:02:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.957	152	34182	20.000	ng/ul	0.00
20) Naphthalene-d8	10.763	136	112582	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.592	164	84473	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.333	188	182835	20.000	ng/ul	0.00
79) Chrysene-d12	21.509	240	171995	20.000	ng/ul	0.00
88) Perylene-d12	23.915	264	234336	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.334	96	7723	8.056	ng/uL	0.00
4) Pyridine-d5	3.763	84	46038	19.178	ng/ul	0.00
7) Phenol-d5	7.122	99	42609	18.674	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.287	67	25527	19.167	ng/ul	0.00
11) 2-Chlorophenol-d4	7.487	132	37357	19.743	ng/ul	0.00
15) 4-Methylphenol-d8	8.669	113	33711	18.517	ng/ul	0.00
21) Nitrobenzene-d5	9.128	128	16613	19.467	ng/ul	0.00
24) 2-Nitrophenol-d4	9.851	143	23824	21.140	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.386	165	51540	20.664	ng/ul	0.00
31) 4-Chloroaniline-d4	10.910	131	44629	18.412	ng/ul	0.00
46) Dimethylphthalate-d6	14.004	166	135329	19.394	ng/ul	0.00
49) Acenaphthylene-d8	14.286	160	149920	20.190	ng/ul	0.00
54) 4-Nitrophenol-d4	14.863	143	105	0.109	ng/ul	0.06
60) Fluorene-d10	15.580	176	117950	18.901	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.710	200	26858	17.161	ng/ul	0.00
73) Anthracene-d10	17.433	188	178966	19.446	ng/ul	0.00
81) Pyrene-d10	19.721	212	195040	22.008	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.756	264	244636	20.339	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.369	88	7785	7.537	ng/uL	97
5) Pyridine	3.781	79	46330	18.793	ng/ul	99
6) Benzaldehyde	7.098	77	20687	19.664	ng/ul	91
8) Phenol	7.145	94	42808	18.739	ng/ul	97
10) Bis(2-Chloroethyl)ether	7.381	93	30911	18.938	ng/ul	89
12) 2-Chlorophenol	7.516	128	37664	19.836	ng/ul	96
13) 2-Methylphenol	8.404	108	33170	19.111	ng/ul	93
14) 2,2'-oxybis(1-Chloropr...	8.492	45	33004	19.572	ng/ul	94
16) Acetophenone	8.787	105	59911	18.628	ng/ul	94
17) N-Nitroso-di-n-propyla...	8.769	70	30522	20.650	ng/ul#	98
18) 4-Methylphenol	8.734	108	34560	18.759	ng/ul	99
19) Hexachloroethane	9.034	117	17963	18.756	ng/ul#	83
22) Nitrobenzene	9.169	77	50918	19.824	ng/ul	98
23) Isophorone	9.692	82	80344	20.672	ng/ul	99
25) 2-Nitrophenol	9.881	139	23370	20.921	ng/ul	98
26) 2,4-Dimethylphenol	9.939	107	48557	20.012	ng/ul	98
27) Bis(2-Chloroethoxy)met...	10.175	93	40890	20.633	ng/ul	97
29) 2,4-Dichlorophenol	10.416	162	48804	20.588	ng/ul	97
30) Naphthalene	10.816	128	112822	20.198	ng/ul	98
32) 4-Chloroaniline	10.933	127	45038	18.517	ng/ul	99
33) Hexachlorobutadiene	11.086	225	46863	19.022	ng/ul	95
34) Caprolactam	11.722	113	6617	14.188	ng/ul	90
35) 4-Chloro-3-methylphenol	12.051	107	38941	20.796	ng/ul#	91
36) 2-Methylnaphthalene	12.422	142	87020	20.945	ng/ul	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM013023\
 Data File : BM038611.D
 Acq On : 30 Jan 2023 18:39
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV032

Quant Time: Jan 31 01:05:48 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM013023.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jan 31 01:02:54 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.639	142	89427	20.851	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.786	216	74339	17.878	ng/ul	95
40) Hexachlorocyclopentadiene	12.757	237	53790	18.128	ng/ul	97
41) 2,4,6-Trichlorophenol	13.027	196	47809	18.884	ng/ul	94
42) 2,4,5-Trichlorophenol	13.104	196	50460	18.664	ng/ul	93
43) 1,1'-Biphenyl	13.427	154	132782	20.443	ng/ul	100
44) 2-Chloronaphthalene	13.469	162	108215	20.055	ng/ul	96
45) 2-Nitroaniline	13.686	65	27232	18.868	ng/ul#	79
47) Dimethylphthalate	14.051	163	135129	19.680	ng/ul	100
48) 2,6-Dinitrotoluene	14.180	165	25109	19.745	ng/ul	98
50) Acenaphthylene	14.316	152	150412	20.256	ng/ul	98
51) 3-Nitroaniline	14.510	138	16659	18.200	ng/ul	92
52) Acenaphthene	14.657	153	108623	19.653	ng/ul	97
53) 2,4-Dinitrophenol	14.721	184	16306	15.951	ng/ul	91
55) 4-Nitrophenol	14.816	109	26115	16.126	ng/ul#	90
56) Dibenzofuran	14.986	168	162049	19.713	ng/ul	94
57) 2,4-Dinitrotoluene	14.963	165	36817	19.643	ng/ul	89
58) 2,3,4,6-Tetrachlorophenol	15.216	232	41136	17.378	ng/ul#	97
59) Diethylphthalate	15.404	149	123879	19.245	ng/ul	99
61) Fluorene	15.633	166	132007	19.067	ng/ul	95
62) 4-Chlorophenyl-phenyle...	15.627	204	84213	17.287	ng/ul	96
63) 4-Nitroaniline	15.668	138	13744	18.212	ng/ul	82
66) 4,6-Dinitro-2-methylph...	15.721	198	26223	17.754	ng/ul	95
67) N-Nitrosodiphenylamine	15.845	169	110191	21.376	ng/ul	96
68) 4-Bromophenyl-phenylether	16.521	248	50795	19.717	ng/ul	98
69) Hexachlorobenzene	16.633	284	58208	19.355	ng/ul	97
70) Atrazine	16.792	200	47702	18.443	ng/ul	96
71) Pentachlorophenol	16.980	266	32610	16.952	ng/ul	94
72) Phenanthrene	17.374	178	191170	20.008	ng/ul	99
74) Anthracene	17.468	178	199500	19.939	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.392	216	69928	19.899	ng/ul	96
76) Pentachlorobenzene	14.904	250	71660	19.597	ng/ul	98
77) Carbazole	17.739	167	154489	17.788	ng/ul	97
78) Di-n-butylphthalate	18.292	149	172956	20.584	ng/ul	98
80) Fluoranthene	19.386	202	231065	22.494	ng/ul	99
82) Pyrene	19.751	202	240044	22.019	ng/ul	99
83) Butylbenzylphthalate	20.639	149	67019	21.314	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.427	252	83174	19.685	ng/ul#	97
85) Benzo(a)anthracene	21.492	228	241541	19.825	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.403	149	93036	20.794	ng/ul	96
87) Chrysene	21.545	228	225591	20.004	ng/ul	99
89) Di-n-octyl phthalate	22.327	149	169810	17.110	ng/ul	100
90) Benzo(b)fluoranthene	23.180	252	295245	20.379	ng/ul	99
91) Benzo(k)fluoranthene	23.227	252	279707	19.369	ng/ul	98
93) Benzo(a)pyrene	23.809	252	265330	19.776	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.421	276	415853	19.609	ng/ul	100
95) Dibenzo(a,h)anthracene	26.427	278	353703	19.651	ng/ul	96
96) Benzo(g,h,i)perylene	27.191	276	344282	20.001	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM013023\
Data File : BM038611.D
Acq On : 30 Jan 2023 18:39
Operator : CG/JU
Sample : SSTDICV020
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SICV032

Quant Time: Jan 31 01:05:48 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM013023.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jan 31 01:02:54 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.537 ng/uL
RT: 3.369 min Scan# 31
Delta R.T. 0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39




Instrument :
BNA_M
ClientSampleId :
SICV032


Tgt Ion: 88 Resp: 7785
Ion Ratio Lower Upper
88 100
43 38.5 33.8 50.8
58 64.0 52.1 78.1



#5
Pyridine
Concen: 18.793 ng/uL
RT: 3.781 min Scan# 101
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39

Tgt Ion: 79 Resp: 46330
Ion Ratio Lower Upper
79 100
52 82.6 67.4 101.0
51 51.2 40.7 61.1









#6
Benzaldehyde
Concen: 19.664 ng/ul
RT: 7.098 min Scan# 665
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39




Instrument :
BNA_M
ClientSampleId :
SICV032


Tgt Ion: 77 Resp: 20687
Ion Ratio Lower Upper
77 100
105 88.7 68.3 102.5
106 85.4 57.7 86.5



#8
Phenol
Concen: 18.739 ng/ul
RT: 7.145 min Scan# 673
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39

Tgt Ion: 94 Resp: 42808
Ion Ratio Lower Upper
94 100
65 44.4 38.2 57.2
66 71.4 57.9 86.9









#10
Bis(2-Chloroethyl)ether
Concen: 18.938 ng/ul
RT: 7.381 min Scan# 713
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39




Instrument :
BNA_M
ClientSampleId :
SICV032


Tgt Ion: 93 Resp: 30911
Ion Ratio Lower Upper
93 100
63 78.7 73.9 110.9
95 33.7 28.0 42.0



#12
2-Chlorophenol
Concen: 19.836 ng/ul
RT: 7.516 min Scan# 736
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39

Tgt Ion:128 Resp: 37664
Ion Ratio Lower Upper
128 100
64 46.4 40.0 60.0
130 31.9 24.4 36.6





#13

2-Methylphenol

Concen: 19.111 ng/ul

RT: 8.404 min Scan# 887

Delta R.T. -0.000 min

Lab File: BM038611.D

Acq: 30 Jan 2023 18:39

Instrument :

BNA_M

ClientSampleId :





SICV032

Tgt Ion:108 Resp: 33170

Ion Ratio Lower Upper

108 100

107 89.3 66.6 100.0



#14

2,2'-oxybis(1-Chloropropane)

Concen: 19.572 ng/ul

RT: 8.492 min Scan# 902

Delta R.T. 0.012 min

Lab File: BM038611.D

Acq: 30 Jan 2023 18:39




Tgt Ion: 45 Resp: 33004


Ion Ratio Lower Upper

45 100

77 18.8 16.9 25.3

79 17.4 11.8 17.6









#16
Acetophenone
Concen: 18.628 ng/ul
RT: 8.787 min Scan# 911
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39




Instrument :
BNA_M
ClientSampleId :
SICV032


Tgt Ion:	105	Resp:	59911
Ion Ratio	Lower	Upper	
105	100		
77	92.6	78.6	117.8
51	42.8	37.2	55.8



#17
N-Nitroso-di-n-propylamine
Concen: 20.650 ng/ul
RT: 8.769 min Scan# 949
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39

Tgt Ion:	70	Resp:	30522
Ion Ratio	Lower	Upper	
70	100		
42	77.9	63.6	95.4
101	10.0	6.3	9.5#
130	22.5	17.6	26.4





#18

4-Methylphenol

Concen: 18.759 ng/ul

RT: 8.734 min Scan# 943

Delta R.T. -0.000 min

Lab File: BM038611.D

Acq: 30 Jan 2023 18:39

Instrument :

BNA_M

ClientSampleId :





SICV032

Tgt Ion:108 Resp: 34560

Ion Ratio Lower Upper

108 100

107 120.5 95.1 142.7



#19

Hexachloroethane

Concen: 18.756 ng/ul

RT: 9.034 min Scan# 994

Delta R.T. -0.000 min

Lab File: BM038611.D

Acq: 30 Jan 2023 18:39




Tgt Ion:117 Resp: 17963


Ion Ratio Lower Upper

117 100

201 174.6 124.4 186.6

199 115.4 76.3 114.5#









#22
Nitrobenzene
Concen: 19.824 ng/ul
RT: 9.169 min Scan# 1017
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39




Instrument :
BNA_M
ClientSampleId :
SICV032


Tgt Ion: 77 Resp: 50918
Ion Ratio Lower Upper
77 100
123 38.1 29.5 44.3
65 16.4 13.6 20.4



#23
Isophorone
Concen: 20.672 ng/ul
RT: 9.692 min Scan# 1106
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39

Tgt Ion: 82 Resp: 80344
Ion Ratio Lower Upper
82 100
95 11.3 9.3 13.9
138 17.6 13.4 20.0









#25
2-Nitrophenol
Concen: 20.921 ng/ul
RT: 9.881 min Scan# 1138
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39




Instrument :
BNA_M
ClientSampleId :
SICV032


Tgt Ion:139 Resp: 23370
Ion Ratio Lower Upper
139 100
65 71.1 58.6 87.8
109 59.5 46.8 70.2



#26
2,4-Dimethylphenol
Concen: 20.012 ng/ul
RT: 9.939 min Scan# 1148
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39

Tgt Ion:107 Resp: 48557
Ion Ratio Lower Upper
107 100
121 42.6 32.6 48.8
122 70.0 54.8 82.2









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.633 ng/ul
 RT: 10.175 min Scan# 1188
 Delta R.T. -0.006 min
 Lab File: BM038611.D
 Acq: 30 Jan 2023 18:39




Instrument :
 BNA_M
 ClientSampleId :
 SICV032


Tgt Ion: 93 Resp: 40890
 Ion Ratio Lower Upper
 93 100
 95 33.9 25.3 37.9
 123 11.0 8.6 13.0



#29
 2,4-Dichlorophenol
 Concen: 20.588 ng/ul
 RT: 10.416 min Scan# 1229
 Delta R.T. -0.000 min
 Lab File: BM038611.D
 Acq: 30 Jan 2023 18:39

Tgt Ion: 162 Resp: 48804
 Ion Ratio Lower Upper
 162 100
 164 71.5 56.3 84.5
 98 27.5 24.6 37.0









#30
Naphthalene
Concen: 20.198 ng/ul
RT: 10.816 min Scan# 1297
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39




Instrument :
BNA_M
ClientSampleId :
SICV032


Tgt Ion:128 Resp: 112822
Ion Ratio Lower Upper
128 100
129 10.4 8.0 12.0
127 14.0 10.5 15.7



#32
4-Chloroaniline
Concen: 18.517 ng/ul
RT: 10.933 min Scan# 1317
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39

Tgt Ion:127 Resp: 45038
Ion Ratio Lower Upper
127 100
129 34.3 26.9 40.3









#33
Hexachlorobutadiene
Concen: 19.022 ng/ul
RT: 11.086 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39




Instrument :
BNA_M
ClientSampleId :
SICV032


Tgt Ion:225 Resp: 46863
Ion Ratio Lower Upper
225 100
223 65.9 48.0 72.0
227 65.5 53.6 80.4



#34
Caprolactam
Concen: 14.188 ng/ul
RT: 11.722 min Scan# 1451
Delta R.T. 0.006 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39

Tgt Ion:113 Resp: 6617
Ion Ratio Lower Upper
113 100
55 146.1 107.6 161.4
56 131.0 96.5 144.7









#35
4-Chloro-3-methylphenol
Concen: 20.796 ng/ul
RT: 12.051 min Scan# 1507
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39




Instrument :
BNA_M
ClientSampleId :
SICV032


Tgt Ion:107 Resp: 38941
Ion Ratio Lower Upper
107 100
144 28.7 18.1 27.1#
142 82.5 60.2 90.4



#36
2-Methylnaphthalene
Concen: 20.945 ng/ul
RT: 12.422 min Scan# 1570
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39

Tgt Ion:142 Resp: 87020
Ion Ratio Lower Upper
142 100
141 90.3 73.7 110.5





#37

1-Methylnaphthalene

Concen: 20.851 ng/ul

RT: 12.639 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM038611.D

Acq: 30 Jan 2023 18:39

Instrument :

BNA_M

ClientSampleId :

SICV032





Tgt Ion:142 Resp: 89427

Ion Ratio Lower Upper

142 100

141 91.5 75.2 112.8

116 5.3 5.0 7.4



#39

1,2,4,5-Tetrachlorobenzene

Concen: 17.878 ng/ul

RT: 12.786 min Scan# 1632

Delta R.T. -0.000 min

Lab File: BM038611.D

Acq: 30 Jan 2023 18:39

Tgt Ion:216 Resp: 74339




Ion Ratio Lower Upper


216 100

214 82.4 61.1 91.7

179 20.8 17.0 25.4

108 13.8 11.4 17.2





#40

Hexachlorocyclopentadiene

Concen: 18.128 ng/ul

RT: 12.757 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM038611.D

Acq: 30 Jan 2023 18:39

Instrument :

BNA_M

ClientSampleId :

SICV032





Tgt Ion:237 Resp: 53790

Ion Ratio Lower Upper

237 100

235 63.9 49.2 73.8

272 12.9 11.0 16.6



#41

2,4,6-Trichlorophenol

Concen: 18.884 ng/ul

RT: 13.027 min Scan# 1673

Delta R.T. -0.006 min

Lab File: BM038611.D

Acq: 30 Jan 2023 18:39




Tgt Ion:196 Resp: 47809


Ion Ratio Lower Upper

196 100

198 97.6 83.6 125.4

200 32.2 27.5 41.3









#42
2,4,5-Trichlorophenol
Concen: 18.664 ng/ul
RT: 13.104 min Scan# 1686
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39




Instrument :
BNA_M
ClientSampleId :
SICV032


Tgt Ion:196 Resp: 50460
Ion Ratio Lower Upper
196 100
198 99.1 73.2 109.8
200 29.7 25.8 38.8



#43
1,1'-Biphenyl
Concen: 20.443 ng/ul
RT: 13.427 min Scan# 1741
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39

Tgt Ion:154 Resp: 132782
Ion Ratio Lower Upper
154 100
153 41.3 33.2 49.8
76 8.6 6.7 10.1









#44
2-Chloronaphthalene
Concen: 20.055 ng/ul
RT: 13.469 min Scan# 1748
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39




Instrument :
BNA_M
ClientSampleId :
SICV032


Tgt Ion:162 Resp: 108215
Ion Ratio Lower Upper
162 100
164 34.8 26.6 40.0
127 36.0 31.6 47.4



#45
2-Nitroaniline
Concen: 18.868 ng/ul
RT: 13.686 min Scan# 1785
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39

Tgt Ion: 65 Resp: 27232
Ion Ratio Lower Upper
65 100
92 69.0 41.5 62.3#
138 95.3 63.3 94.9#





#47

Dimethylphthalate

Concen: 19.680 ng/ul

RT: 14.051 min Scan# 1847

Delta R.T. -0.000 min

Lab File: BM038611.D

Acq: 30 Jan 2023 18:39

Instrument :

BNA_M

ClientSampleId :

SICV032





Tgt Ion:163 Resp: 135129

Ion Ratio Lower Upper

163 100

194 4.6 3.5 5.3

164 10.0 8.0 12.0



#48

2,6-Dinitrotoluene

Concen: 19.745 ng/ul

RT: 14.180 min Scan# 1869

Delta R.T. -0.000 min

Lab File: BM038611.D

Acq: 30 Jan 2023 18:39




Tgt Ion:165 Resp: 25109


Ion Ratio Lower Upper

165 100

89 82.4 66.7 100.1

121 20.0 18.2 27.4









#50
Acenaphthylene
Concen: 20.256 ng/ul
RT: 14.316 min Scan# 1892
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39




Instrument :
BNA_M
ClientSampleId :
SICV032


Tgt Ion:152 Resp: 150412
Ion Ratio Lower Upper
152 100
151 21.1 17.7 26.5
153 12.8 10.6 16.0



#51
3-Nitroaniline
Concen: 18.200 ng/ul
RT: 14.510 min Scan# 1925
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39

Tgt Ion:138 Resp: 16659
Ion Ratio Lower Upper
138 100
108 10.5 8.2 12.4
92 108.5 94.2 141.4









#52
Acenaphthene
Concen: 19.653 ng/ul
RT: 14.657 min Scan# 1950
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39




Instrument :
BNA_M
ClientSampleId :
SICV032


Tgt Ion:153 Resp: 108623
Ion Ratio Lower Upper
153 100
152 45.8 40.6 61.0
154 86.0 68.7 103.1



#53
2,4-Dinitrophenol
Concen: 15.951 ng/ul
RT: 14.721 min Scan# 1961
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39

Tgt Ion:184 Resp: 16306
Ion Ratio Lower Upper
184 100
63 74.8 59.7 89.5
154 84.0 79.6 119.4









#55
4-Nitrophenol
Concen: 16.126 ng/ul
RT: 14.816 min Scan# 1977
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39




Instrument :
BNA_M
ClientSampleId :
SICV032


Tgt Ion:109 Resp: 26115
Ion Ratio Lower Upper
109 100
139 63.8 41.3 61.9#
65 83.7 63.4 95.2



#56
Dibenzofuran
Concen: 19.713 ng/ul
RT: 14.986 min Scan# 2006
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39

Tgt Ion:168 Resp: 162049
Ion Ratio Lower Upper
168 100
139 40.0 35.4 53.0









#57
2,4-Dinitrotoluene
Concen: 19.643 ng/ul
RT: 14.963 min Scan# 2002
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39




Instrument :
BNA_M
ClientSampleId :
SICV032


Tgt Ion:165 Resp: 36817
Ion Ratio Lower Upper
165 100
63 57.5 53.1 79.7
182 3.3 2.8 4.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 17.378 ng/ul
RT: 15.216 min Scan# 2045
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39

Tgt Ion:232 Resp: 41136
Ion Ratio Lower Upper
232 100
131 34.6 30.4 45.6
130 2.1 2.4 3.6#
166 28.0 22.2 33.2









#59
Diethylphthalate
Concen: 19.245 ng/ul
RT: 15.404 min Scan# 2077
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39




Instrument :
BNA_M
ClientSampleId :
SICV032


Tgt Ion:149 Resp: 123879
Ion Ratio Lower Upper
149 100
177 23.2 18.0 27.0
150 12.4 10.4 15.6



#61
Fluorene
Concen: 19.067 ng/ul
RT: 15.633 min Scan# 2116
Delta R.T. -0.006 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39

Tgt Ion:166 Resp: 132007
Ion Ratio Lower Upper
166 100
165 97.9 83.3 124.9
167 13.2 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 17.287 ng/ul
RT: 15.627 min Scan# 2115
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39




Instrument :
BNA_M
ClientSampleId :
SICV032


Tgt Ion:204 Resp: 84213
Ion Ratio Lower Upper
204 100
206 33.0 27.9 41.9
141 49.0 41.4 62.0



#63
4-Nitroaniline
Concen: 18.212 ng/ul
RT: 15.668 min Scan# 2122
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39

Tgt Ion:138 Resp: 13744
Ion Ratio Lower Upper
138 100
92 53.7 51.9 77.9
108 165.6 156.2 234.4









#66
 4,6-Dinitro-2-methylphenol
 Concen: 17.754 ng/ul
 RT: 15.721 min Scan# 2132
 Delta R.T. -0.006 min
 Lab File: BM038611.D
 Acq: 30 Jan 2023 18:39




Instrument :
 BNA_M
 ClientSampleId :
 SICV032


Tgt Ion:198 Resp: 26223
 Ion Ratio Lower Upper
 198 100
 182 3.6 2.7 4.1
 77 26.7 23.9 35.9



#67
 N-Nitrosodiphenylamine
 Concen: 21.376 ng/ul
 RT: 15.845 min Scan# 2152
 Delta R.T. -0.000 min
 Lab File: BM038611.D
 Acq: 30 Jan 2023 18:39

Tgt Ion:169 Resp: 110191
 Ion Ratio Lower Upper
 169 100
 168 70.0 59.1 88.7
 167 37.8 31.6 47.4









#68
4-Bromophenyl-phenylether
Concen: 19.717 ng/ul
RT: 16.521 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39




Instrument :
BNA_M
ClientSampleId :
SICV032


Tgt Ion:248 Resp: 50795
Ion Ratio Lower Upper
248 100
250 94.3 77.8 116.6
141 64.8 52.3 78.5



#69
Hexachlorobenzene
Concen: 19.355 ng/ul
RT: 16.633 min Scan# 2286
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39

Tgt Ion:284 Resp: 58208
Ion Ratio Lower Upper
284 100
142 32.7 28.0 42.0
249 31.2 23.9 35.9





#70

Atrazine

Concen: 18.443 ng/ul

RT: 16.792 min Scan# 21

Delta R.T. -0.000 min

Lab File: BM038611.D

Acq: 30 Jan 2023 18:39

Instrument :

BNA_M

ClientSampleId :

SICV032





Tgt Ion:200 Resp: 47702

Ion Ratio Lower Upper

200 100

173 21.8 15.2 22.8

215 47.4 36.6 54.8



#71

Pentachlorophenol

Concen: 16.952 ng/ul

RT: 16.980 min Scan# 2345

Delta R.T. -0.000 min

Lab File: BM038611.D

Acq: 30 Jan 2023 18:39




Tgt Ion:266 Resp: 32610


Ion Ratio Lower Upper

266 100

264 63.1 47.2 70.8

268 62.7 45.8 68.6





#72

Phenanthrene

Concen: 20.008 ng/ul

RT: 17.374 min Scan# 2412

Delta R.T. -0.000 min

Lab File: BM038611.D

Acq: 30 Jan 2023 18:39

Instrument :

BNA_M

ClientSampleId :

SICV032





Tgt Ion:178 Resp: 191170

Ion Ratio Lower Upper

178 100

179 15.6 11.7 17.5

176 20.2 16.2 24.2



#74

Anthracene

Concen: 19.939 ng/ul

RT: 17.468 min Scan# 2428

Delta R.T. -0.000 min

Lab File: BM038611.D

Acq: 30 Jan 2023 18:39




Tgt Ion:178 Resp: 199500


Ion Ratio Lower Upper

178 100

179 14.8 12.3 18.5

176 19.4 16.3 24.5









#75
1,2,3,4-Tetrachlorobenzene
Concen: 19.899 ng/uL
RT: 13.392 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39




Instrument :
BNA_M
ClientSampleId :
SICV032


Tgt Ion:216 Resp: 69928
Ion Ratio Lower Upper
216 100
214 74.6 62.5 93.7
218 43.9 37.4 56.0



#76
Pentachlorobenzene
Concen: 19.597 ng/uL
RT: 14.904 min Scan# 1992
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39

Tgt Ion:250 Resp: 71660
Ion Ratio Lower Upper
250 100
248 62.7 50.6 75.8
252 61.8 51.3 76.9









#77
 Carbazole
 Concen: 17.788 ng/ul
 RT: 17.739 min Scan# 2474
 Delta R.T. -0.000 min
 Lab File: BM038611.D
 Acq: 30 Jan 2023 18:39

Instrument :
 BNA_M
 ClientSampleId :
 SICV032

Tgt Ion:167 Resp: 154489




Ion	Ratio	Lower	Upper
167	100		
166	22.7	19.5	29.3
139	12.7	11.3	16.9




#78
 Di-n-butylphthalate
 Concen: 20.584 ng/ul
 RT: 18.292 min Scan# 2568
 Delta R.T. -0.000 min
 Lab File: BM038611.D
 Acq: 30 Jan 2023 18:39

Tgt Ion:149 Resp: 172956

Ion	Ratio	Lower	Upper
149	100		
150	8.6	7.2	10.8
104	7.0	6.2	9.4









#80
Fluoranthene
Concen: 22.494 ng/ul
RT: 19.386 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39




Instrument :
BNA_M
ClientSampleId :
SICV032


Tgt Ion:202 Resp: 231065
Ion Ratio Lower Upper
202 100
101 3.0 2.8 4.2
100 2.9 2.4 3.6



#82
Pyrene
Concen: 22.019 ng/ul
RT: 19.751 min Scan# 2816
Delta R.T. -0.000 min
Lab File: BM038611.D
Acq: 30 Jan 2023 18:39

Tgt Ion:202 Resp: 240044
Ion Ratio Lower Upper
202 100
101 3.8 3.3 4.9
100 3.0 2.8 4.2





#83

Butylbenzylphthalate

Concen: 21.314 ng/ul

RT: 20.639 min Scan# 2967

Delta R.T. -0.000 min

Lab File: BM038611.D

Acq: 30 Jan 2023 18:39

Instrument :

BNA_M

ClientSampleId :

SICV032





Tgt Ion:149 Resp: 67019

Ion Ratio Lower Upper

149 100

91 68.8 59.0 88.4

206 24.7 19.0 28.6



#84

3,3'-Dichlorobenzidine

Concen: 19.685 ng/ul

RT: 21.427 min Scan# 3101

Delta R.T. -0.000 min

Lab File: BM038611.D

Acq: 30 Jan 2023 18:39




Tgt Ion:252 Resp: 83174


Ion Ratio Lower Upper

252 100

254 63.3 52.7 79.1

126 4.6 5.0 7.4#





#85

Benzo(a)anthracene

Concen: 19.825 ng/ul

RT: 21.492 min Scan# 3112

Delta R.T. -0.000 min

Lab File: BM038611.D

Acq: 30 Jan 2023 18:39

Instrument :

BNA_M

ClientSampleId :

SICV032





Tgt Ion:228 Resp: 241541

Ion Ratio Lower Upper

228 100

229 19.8 15.8 23.6

226 28.0 22.2 33.4



#86

Bis(2-ethylhexyl)phthalate

Concen: 20.794 ng/ul

RT: 21.403 min Scan# 3097

Delta R.T. -0.000 min

Lab File: BM038611.D

Acq: 30 Jan 2023 18:39




Tgt Ion:149 Resp: 93036


Ion Ratio Lower Upper

149 100

167 28.8 21.0 31.4

279 6.6 5.1 7.7





#87

Chrysene

Concen: 20.004 ng/ul

RT: 21.545 min Scan# 31

Delta R.T. -0.000 min

Lab File: BM038611.D

Acq: 30 Jan 2023 18:39

Instrument :

BNA_M

ClientSampleId :

SICV032





Tgt Ion:228 Resp: 225591

Ion Ratio Lower Upper

228 100

226 30.6 24.5 36.7

229 20.5 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 17.110 ng/ul




RT: 22.327 min Scan# 3254


Delta R.T. -0.006 min

Lab File: BM038611.D

Acq: 30 Jan 2023 18:39

Tgt Ion:149 Resp: 169810





#90

Benzo(b)fluoranthene

Concen: 20.379 ng/ul

RT: 23.180 min Scan# 3399

Delta R.T. -0.000 min

Lab File: BM038611.D

Acq: 30 Jan 2023 18:39

Instrument :

BNA_M

ClientSampleId :

SICV032





Tgt Ion:252 Resp: 295245

Ion Ratio Lower Upper

252 100

253 22.8 17.8 26.6

125 3.2 2.6 3.8



#91

Benzo(k)fluoranthene

Concen: 19.369 ng/ul

RT: 23.227 min Scan# 3407

Delta R.T. -0.000 min

Lab File: BM038611.D

Acq: 30 Jan 2023 18:39




Tgt Ion:252 Resp: 279707


Ion Ratio Lower Upper

252 100

253 21.8 18.2 27.2

125 2.9 2.6 4.0





#93

Benzo(a)pyrene

Concen: 19.776 ng/ul

RT: 23.809 min Scan# 3

Delta R.T. -0.000 min

Lab File: BM038611.D

Acq: 30 Jan 2023 18:39

Instrument :

BNA_M

ClientSampleId :

SICV032





Tgt Ion:252 Resp: 265330

Ion Ratio Lower Upper

252 100

253 22.0 17.8 26.8

125 2.8 2.8 4.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.609 ng/ul

RT: 26.421 min Scan# 3950

Delta R.T. -0.000 min

Lab File: BM038611.D

Acq: 30 Jan 2023 18:39




Tgt Ion:276 Resp: 415853


Ion Ratio Lower Upper

276 100

138 7.7 6.1 9.1

277 25.5 20.5 30.7





#95

Dibenzo(a,h)anthracene

Concen: 19.651 ng/ul

RT: 26.427 min Scan# 3951

Delta R.T. -0.006 min

Lab File: BM038611.D

Acq: 30 Jan 2023 18:39

Instrument :

BNA_M

ClientSampleId :

SICV032





Tgt Ion:278 Resp: 353703

Ion Ratio Lower Upper

278 100

139 4.5 4.3 6.5

279 21.5 18.9 28.3



#96

Benzo(g,h,i)perylene

Concen: 20.001 ng/ul

RT: 27.191 min Scan# 4081

Delta R.T. -0.006 min

Lab File: BM038611.D

Acq: 30 Jan 2023 18:39

Tgt Ion:276 Resp: 344282

Ion Ratio Lower Upper

276 100

138 7.1 6.0 9.0

277 24.6 19.2 28.8

