

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM013123\
 Data File : BM038633.D
 Acq On : 31 Jan 2023 23:29
 Operator : CG/JU
 Sample : 01277-03
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 EX9H3

Quant Time: Feb 01 02:01:59 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM013023.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jan 31 01:02:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.951	152	68628	20.000	ng/u1	0.00
20) Naphthalene-d8	10.757	136	238720	20.000	ng/u1	#-0.01
38) Acenaphthene-d10	14.586	164	129506	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.327	188	240977	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.503	240	206642	20.000	ng/u1	# 0.00
88) Perylene-d12	23.909	264	324656	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.334	96	11172	5.804	ng/uL	0.00
4) Pyridine-d5	3.769	84	36123	7.495	ng/u1	0.00
7) Phenol-d5	7.116	99	60083	13.116	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.281	67	125698	47.010	ng/u1	0.00
11) 2-Chlorophenol-d4	7.481	132	135404	35.642	ng/u1	0.00
15) 4-Methylphenol-d8	8.663	113	98909	27.061	ng/u1	-0.01
21) Nitrobenzene-d5	9.122	128	73010	40.348	ng/u1	0.00
24) 2-Nitrophenol-d4	9.845	143	88341	36.968	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.381	165	127099	24.032	ng/u1	0.00
31) 4-Chloroaniline-d4	10.916	131	77112	15.004	ng/u1	0.00
46) Dimethylphthalate-d6	13.998	166	339933	31.776	ng/u1	0.00
49) Acenaphthylene-d8	14.280	160	450613	39.582	ng/u1	0.00
54) 4-Nitrophenol-d4	14.804	143	15208	10.276	ng/u1	0.00
60) Fluorene-d10	15.574	176	301975	31.564	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.704	200	47212	22.888	ng/u1	0.00
73) Anthracene-d10	17.427	188	455556	37.557	ng/u1	0.00
81) Pyrene-d10	19.715	212	523517	49.167	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.750	264	665791	39.954	ng/u1	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM013123\
Data File : BM038633.D
Acq On : 31 Jan 2023 23:29
Operator : CG/JU
Sample : 01277-03
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
EX9H3

Quant Time: Feb 01 02:01:59 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM013023.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jan 31 01:02:54 2023
Response via : Initial Calibration

