

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM013123\
 Data File : BM038637.D
 Acq On : 01 Feb 2023 01:56
 Operator : CG/JU
 Sample : 01277-07
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 EX9H1

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/01/2023
 Supervised By :mohammad ahmed 02/01/2023

Quant Time: Feb 01 02:30:47 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM013023.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jan 31 01:02:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.951	152	43011	20.000	ng/u1	0.00
20) Naphthalene-d8	10.763	136	117216	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.586	164	108099	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.327	188	307022	20.000	ng/u1	0.00
79) Chrysene-d12	21.503	240	324799	20.000	ng/u1	0.00
88) Perylene-d12	23.909	264	413543	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.334	96	4758m	3.944	ng/uL	0.00
4) Pyridine-d5	3.769	84	23413	7.751	ng/u1	0.00
7) Phenol-d5	7.116	99	16209	5.646	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.281	67	38833	23.173	ng/u1	0.00
11) 2-Chlorophenol-d4	7.481	132	47101	19.783	ng/u1	0.00
15) 4-Methylphenol-d8	8.669	113	29424	12.845	ng/u1	0.00
21) Nitrobenzene-d5	9.122	128	25376	28.560	ng/u1	0.00
24) 2-Nitrophenol-d4	9.845	143	35413	30.181	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.380	165	76508	29.462	ng/u1	0.00
31) 4-Chloroaniline-d4	10.910	131	33775	13.384	ng/u1	0.00
46) Dimethylphthalate-d6	13.998	166	244094	27.335	ng/u1	0.00
49) Acenaphthylene-d8	14.280	160	269865	28.399	ng/u1	0.00
54) 4-Nitrophenol-d4	14.810	143	7419m	6.006	ng/u1	0.01
60) Fluorene-d10	15.574	176	252732	31.649	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.704	200	54078	20.577	ng/u1	0.00
73) Anthracene-d10	17.427	188	448486m	29.020	ng/u1	0.00
81) Pyrene-d10	19.715	212	548976	32.802	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.756	264	628365	29.603	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM013123\
Data File : BM038637.D
Acq On : 01 Feb 2023 01:56
Operator : CG/JU
Sample : 01277-07
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
EX9H1

Quant Time: Feb 01 02:30:47 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM013023.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jan 31 01:02:54 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/01/2023
Supervised By :mohammad ahmed 02/01/2023

