

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM020119\
 Data File : BM018706.D
 Acq On : 01 Feb 2019 13:55
 Operator : JU/SJ
 Sample : K1297-01
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 A0C17-SS1

Quant Time: Feb 01 17:18:27 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM010919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 18 22:01:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.72	152	238749	20.00	ng	-0.02
21) Naphthalene-d8	10.51	136	911534	20.00	ng	-0.02
39) Acenaphthene-d10	14.37	164	467579	20.00	ng	-0.02
64) Phenanthrene-d10	17.12	188	1034532	20.00	ng	-0.02
76) Chrysene-d12	21.31	240	1282319	20.00	ng	-0.02
87) Perylene-d12	23.57	264	1338064	20.00	ng	-0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.30	112	1421001	109.90	ng	-0.02
7) Phenol-d6	6.89	99	1788367	108.90	ng	-0.02
23) Nitrobenzene-d5	8.89	82	1069261	68.00	ng	-0.02
42) 2,4,6-Tribromophenol	15.86	330	617032	103.64	ng	-0.02
45) 2-Fluorobiphenyl	12.98	172	1659957	46.32	ng	-0.02
79) Terphenyl-d14	19.75	244	2620672	43.08	ng	-0.02
Target Compounds						
10) Phenol	6.92	94	117813	7.060	ng	96
50) Dimethylphthalate	13.83	163	258945	6.743	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM020119\
Data File : BM018706.D
Acq On : 01 Feb 2019 13:55
Operator : JU/SJ
Sample : K1297-01
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
A0C17-SS1

Quant Time: Feb 01 17:18:27 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM010919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 18 22:01:30 2019
Response via : Initial Calibration

