


Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM020119\
 Data File : BM018704.D
 Acq On : 01 Feb 2019 10:53
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM010919.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jan 18 22:01:30 2019



AutoFind: Scans 2458, 2459, 2460; Background Corrected with Scan 2450

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.1	120609	PASS
68	69	0.00	2	2.0	2546	PASS
69	198	0.00	100	33.3	129008	PASS
70	69	0.00	2	0.3	432	PASS
127	198	10	80	45.9	177813	PASS
197	198	0.00	2	0.4	1575	PASS
198	198	100	100	100.0	387584	PASS
199	198	5	9	6.7	25874	PASS
275	198	10	60	29.3	113429	PASS
365	198	1	100	4.2	16164	PASS
441	198	0.01	100	14.6	56434	PASS
442	198	50	100	94.3	365546	PASS
443	442	15	24	19.1	69810	PASS