

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM020223\
 Data File : BM038659.D
 Acq On : 02 Feb 2023 12:20
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 02/06/2023
 Supervised By : Jagrut Upadhyay 02/06/2023

Quant Time: Feb 03 05:27:38 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM020123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 02 03:31:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.934	152	59481	20.000	ng	0.00
21) Naphthalene-d8	10.739	136	200909	20.000	ng	-0.01
39) Acenaphthene-d10	14.568	164	148684	20.000	ng	0.00
64) Phenanthrene-d10	17.315	188	318133	20.000	ng	0.00
76) Chrysene-d12	21.492	240	280638	20.000	ng	# 0.00
86) Perylene-d12	23.886	264	333736	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.498	112	241528	79.543	ng	0.00
7) Phenol-d6	7.116	99	280465	80.986	ng	0.00
23) Nitrobenzene-d5	9.104	82	281699	79.641	ng	-0.01
42) 2,4,6-Tribromophenol	16.063	330	187283	92.671	ng	0.00
45) 2-Fluorobiphenyl	13.198	172	985707	84.401	ng	0.00
79) Terphenyl-d14	19.927	244	1422828	89.468	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.340	88	48222	36.024	ng	94
3) Pyridine	3.757	79	122866	38.599	ng	96
4) n-Nitrosodimethylamine	3.669	42	53142	36.981	ng	94
6) Aniline	7.263	93	171233	39.638	ng	95
8) 2-Chlorophenol	7.498	128	134764	40.674	ng	97
9) Benzaldehyde	7.075	77	54986	36.024	ng	87
10) Phenol	7.145	94	139737	39.940	ng	98
11) bis(2-Chloroethyl)ether	7.357	93	111455	39.337	ng	97
12) 1,3-Dichlorobenzene	7.816	146	170146	40.192	ng	96
13) 1,4-Dichlorobenzene	7.969	146	172891	40.467	ng	98
14) 1,2-Dichlorobenzene	8.281	146	168109	41.073	ng	99
15) Benzyl Alcohol	8.187	79	106936	41.952	ng	99
16) 2,2'-oxybis(1-Chloropr...	8.457	45	95857m	35.718	ng	
17) 2-Methylphenol	8.398	107	109104	40.656	ng	99
18) Hexachloroethane	9.004	117	57299	41.528	ng	97
19) n-Nitroso-di-n-propyla...	8.745	70	89354	42.805	ng	95
20) 3+4-Methylphenols	8.728	107	150605	42.092	ng	99
22) Acetophenone	8.769	105	196330	40.977	ng	# 98
24) Nitrobenzene	9.151	77	141810	39.500	ng	94
25) Isophorone	9.675	82	252164	40.696	ng	98
26) 2-Nitrophenol	9.857	139	87462	43.780	ng	95
27) 2,4-Dimethylphenol	9.928	122	121903	42.293	ng	98
28) bis(2-Chloroethoxy)met...	10.157	93	150280	39.686	ng	99
29) 2,4-Dichlorophenol	10.404	162	171230	41.814	ng	99
30) 1,2,4-Trichlorobenzene	10.604	180	179908	41.301	ng	99
31) Naphthalene	10.792	128	422891	41.018	ng	99
32) Benzoic acid	10.081	122	76895	32.956	ng	99
33) 4-Chloroaniline	10.916	127	186454	41.266	ng	98
34) Hexachlorobutadiene	11.063	225	98518	45.148	ng	96
35) Caprolactam	11.716	113	37793	42.853	ng	94
36) 4-Chloro-3-methylphenol	12.051	107	127894	42.237	ng	95
37) 2-Methylnaphthalene	12.398	142	351760	42.221	ng	99
38) 1-Methylnaphthalene	12.616	142	326479	41.587	ng	96
40) 1,2,4,5-Tetrachloroben...	12.763	216	167781	43.359	ng	98
41) Hexachlorocyclopentadiene	12.733	237	93134	38.787	ng	98
43) 2,4,6-Trichlorophenol	13.016	196	132575	43.611	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM020223\
 Data File : BM038659.D
 Acq On : 02 Feb 2023 12:20
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 02/06/2023
 Supervised By : Jagrut Upadhyay 02/06/2023

Quant Time: Feb 03 05:27:38 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM020123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 02 03:31:45 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	13.098	196	151630	42.386	ng	96
46) 1,1'-Biphenyl	13.404	154	488818	41.503	ng	99
47) 2-Chloronaphthalene	13.451	162	387053	41.763	ng	99
48) 2-Nitroaniline	13.669	65	78728	38.109	ng	98
49) Acenaphthylene	14.292	152	611933	41.920	ng	99
50) Dimethylphthalate	14.033	163	513770	42.343	ng	100
51) 2,6-Dinitrotoluene	14.163	165	106805	43.310	ng	98
52) Acenaphthene	14.633	154	363368	41.689	ng	96
53) 3-Nitroaniline	14.498	138	100477	39.237	ng	94
54) 2,4-Dinitrophenol	14.704	184	54960	34.239	ng	91
55) Dibenzofuran	14.968	168	625899	41.700	ng	100
56) 4-Nitrophenol	14.833	139	75365	38.105	ng	94
57) 2,4-Dinitrotoluene	14.945	165	147673	40.834	ng	97
58) Fluorene	15.615	166	507363	42.721	ng	100
59) 2,3,4,6-Tetrachlorophenol	15.204	232	109392	43.936	ng	96
60) Diethylphthalate	15.386	149	494198	42.994	ng	99
61) 4-Chlorophenyl-phenyle...	15.610	204	225275	44.713	ng	96
62) 4-Nitroaniline	15.657	138	102631m	39.093	ng	
63) Azobenzene	15.904	77	351757	39.775	ng	96
65) 4,6-Dinitro-2-methylph...	15.710	198	81390	40.640	ng	92
66) n-Nitrosodiphenylamine	15.827	169	428170	41.219	ng	99
67) 4-Bromophenyl-phenylether	16.504	248	136879	44.560	ng	99
68) Hexachlorobenzene	16.615	284	164228	43.746	ng	96
69) Atrazine	16.780	200	135739	45.086	ng	100
70) Pentachlorophenol	16.968	266	103921	37.974	ng	97
71) Phenanthrene	17.357	178	780714	41.330	ng	99
72) Anthracene	17.445	178	805706	41.693	ng	100
73) Carbazole	17.727	167	739301	41.021	ng	99
74) Di-n-butylphthalate	18.274	149	915074	44.941	ng	100
75) Fluoranthene	19.368	202	865384	43.633	ng	99
77) Benzidine	19.556	184	255813	31.486	ng	98
78) Pyrene	19.733	202	902321	41.762	ng	99
80) Butylbenzylphthalate	20.621	149	435520	40.794	ng	95
81) Benzo(a)anthracene	21.474	228	840901	42.360	ng	99
82) 3,3'-Dichlorobenzidine	21.409	252	297890	43.618	ng	96
83) Chrysene	21.527	228	804678	42.071	ng	100
84) Bis(2-ethylhexyl)phtha...	21.386	149	642898	41.291	ng	98
85) Di-n-octyl phthalate	22.309	149	1108956	43.604	ng	99
87) Indeno(1,2,3-cd)pyrene	26.385	276	1087722	41.378	ng	97
88) Benzo(b)fluoranthene	23.156	252	885923m	44.320	ng	
89) Benzo(k)fluoranthene	23.203	252	868627	42.792	ng	99
90) Benzo(a)pyrene	23.780	252	843144	42.652	ng	# 98
91) Dibenzo(a,h)anthracene	26.397	278	925360	41.841	ng	98
92) Benzo(g,h,i)perylene	27.156	276	878586	39.300	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Manual Integrations APPROVED


Quant Time: Feb 03 05:27:38 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM020123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 02 03:31:45 2023
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.934 min Scan# 80
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:152 Resp: 5948
Ion Ratio Lower Upper
152 100
150 146.6 119.5 179.3
115 44.0 37.6 56.4




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023



#2
1,4-Dioxane
Concen: 36.024 ng
RT: 3.340 min Scan# 26
Delta R.T. -0.012 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20


Tgt Ion: 88 Resp: 48222
Ion Ratio Lower Upper
88 100
58 53.3 47.0 70.4
43 26.0 22.6 33.8





#3
Pyridine
Concen: 38.599 ng
RT: 3.757 min Scan# 91
Delta R.T. -0.012 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion: 79 Resp: 122866
Ion Ratio Lower Upper
79 100
52 51.4 43.7 65.5
51 32.1 26.6 40.0

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023




#4
n-Nitrosodimethylamine
Concen: 36.981 ng
RT: 3.669 min Scan# 82
Delta R.T. -0.012 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20



Tgt Ion: 42 Resp: 53142
Ion Ratio Lower Upper
42 100
74 130.3 98.4 147.6
44 5.1 3.5 5.3









#5
2-Fluorophenol
Concen: 79.543 ng
RT: 5.498 min Scan# 392
Delta R.T. 0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion: 112 Resp: 241528
Ion Ratio Lower Upper
112 100
64 43.1 37.9 56.9
63 27.3 23.3 34.9




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023



#6
Aniline
Concen: 39.638 ng
RT: 7.263 min Scan# 693
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20

Tgt Ion: 93 Resp: 171233
Ion Ratio Lower Upper
93 100
66 34.6 29.8 44.8
65 18.1 16.9 25.3









#7
Phenol-d6
Concen: 80.986 ng
RT: 7.116 min Scan# 60
Delta R.T. 0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion: 99 Resp: 280465
Ion Ratio Lower Upper
99 100
42 16.1 14.6 22.0
71 39.1 31.0 46.4




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023



#8
2-Chlorophenol
Concen: 40.674 ng
RT: 7.498 min Scan# 733
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20

Tgt Ion:128 Resp: 134764
Ion Ratio Lower Upper
128 100
130 32.7 13.2 53.2
64 34.6 17.3 57.3









#9
Benzaldehyde
Concen: 36.024 ng
RT: 7.075 min Scan# 60
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion: 77 Resp: 54986
Ion Ratio Lower Upper
77 100
105 114.7 80.8 120.8
106 105.5 73.8 113.8




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023



#10
Phenol
Concen: 39.940 ng
RT: 7.145 min Scan# 673
Delta R.T. 0.012 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20


Tgt Ion: 94 Resp: 139737
Ion Ratio Lower Upper
94 100
65 30.3 12.3 52.3
66 43.3 24.2 64.2





#11
bis(2-Chloroethyl)ether
Concen: 39.337 ng
RT: 7.357 min Scan# 709
Delta R.T. -0.012 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 93 Resp: 111455
Ion Ratio Lower Upper
93 100
63 65.6 48.5 88.5
95 33.1 13.8 53.8




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023



#12
1,3-Dichlorobenzene
Concen: 40.192 ng
RT: 7.816 min Scan# 787
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20


Tgt Ion:146 Resp: 170146
Ion Ratio Lower Upper
146 100
148 68.4 51.5 77.3
75 26.9 21.1 31.7





#13
1,4-Dichlorobenzene
Concen: 40.467 ng
RT: 7.969 min Scan# 814
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:146 Resp: 172891
Ion Ratio Lower Upper
146 100
148 64.8 53.1 79.7
111 31.6 25.9 38.9

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023




#14
1,2-Dichlorobenzene
Concen: 41.073 ng
RT: 8.281 min Scan# 866
Delta R.T. -0.012 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20




Tgt Ion:146 Resp: 168109
Ion Ratio Lower Upper
146 100
148 65.8 52.9 79.3
111 34.0 26.2 39.2





#15
Benzyl Alcohol
Concen: 41.952 ng
RT: 8.187 min Scan# 851
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 79 Resp: 106936
Ion Ratio Lower Upper
79 100
108 77.0 61.9 92.9
77 68.4 55.8 83.8




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023



#16
2,2'-oxybis(1-Chloropropane)
Concen: 35.718 ng m
RT: 8.457 min Scan# 896
Delta R.T. -0.012 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20


Tgt Ion: 45 Resp: 95857
Ion Ratio Lower Upper
45 100
77 21.3 1.1 41.1
79 16.5 0.0 34.6





#17
2-Methylphenol
Concen: 40.656 ng
RT: 8.398 min Scan# 885
Delta R.T. 0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion	Ratio	Lower	Upper
107	100		
108	110.6	89.6	134.4
77	51.4	40.7	61.1
79	48.2	39.5	59.3




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023



#18
Hexachloroethane
Concen: 41.528 ng
RT: 9.004 min Scan# 989
Delta R.T. -0.012 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20


Tgt Ion	Ratio	Lower	Upper
117	100		
119	98.4	78.6	117.8
201	119.2	90.2	135.2





#19
n-Nitroso-di-n-propylamine
Concen: 42.805 ng
RT: 8.745 min Scan# 945
Delta R.T. -0.012 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion: 70 Resp: 89354
Ion Ratio Lower Upper
70 100
42 50.1 43.4 65.2
101 10.1 8.2 12.2
130 26.1 18.8 28.2

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023




#20
3+4-Methylphenols
Concen: 42.092 ng
RT: 8.728 min Scan# 942
Delta R.T. 0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20




Tgt Ion:107 Resp: 150605
Ion Ratio Lower Upper
107 100
108 88.7 67.2 107.2
77 36.5 17.5 57.5
79 31.2 10.8 50.8





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.739 min Scan# 1284
Delta R.T. -0.012 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040






Tgt Ion:136 Resp: 200909

Ion	Ratio	Lower	Upper
136	100		
137	10.5	8.8	13.2
54	4.9	4.2	6.4
68	3.1	2.9	4.3

Manual Integrations
APPROVED




Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023




#22
Acetophenone
Concen: 40.977 ng
RT: 8.769 min Scan# 949
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20

Tgt Ion:105 Resp: 196330

Ion	Ratio	Lower	Upper
105	100		
71	0.4	0.9	1.3
51	22.9	19.3	28.9
120	22.1	18.0	27.0





#23
Nitrobenzene-d5
Concen: 79.641 ng
RT: 9.104 min Scan# 1006
Delta R.T. -0.012 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20

Instrument :

BNA_M

ClientSampleId :





SSTDCCC040

Tgt Ion: 82 Resp: 281699
Ion Ratio Lower Upper
82 100
128 48.0 36.0 54.0
54 44.0 36.8 55.2

Manual Integrations




APPROVED


Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023



#24
Nitrobenzene
Concen: 39.500 ng
RT: 9.151 min Scan# 1014
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20

Tgt Ion: 77 Resp: 141810
Ion Ratio Lower Upper
77 100
123 50.5 36.5 54.7
65 13.7 10.7 16.1









#25
Isophorone
Concen: 40.696 ng
RT: 9.675 min Scan# 1104
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion: 82 Resp: 252164
Ion Ratio Lower Upper
82 100
95 8.4 6.4 9.6
138 22.4 17.0 25.6




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023



#26
2-Nitrophenol
Concen: 43.780 ng
RT: 9.857 min Scan# 1134
Delta R.T. -0.012 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20

Tgt Ion:139 Resp: 87462
Ion Ratio Lower Upper
139 100
109 36.1 24.5 36.7
65 42.0 32.6 49.0





#27

2,4-Dimethylphenol

Concen: 42.293 ng

RT: 9.928 min Scan# 1146

Delta R.T. -0.000 min

Lab File: BM038659.D


Acq: 02 Feb 2023 12:20

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040



Tgt Ion:122 Resp: 121903

Ion Ratio Lower Upper

122 100

107 116.2 94.9 142.3




121 55.6 45.5 68.3

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 02/06/2023

Supervised By :Jagrut Upadhyay 02/06/2023



#28

bis(2-Chloroethoxy)methane

Concen: 39.686 ng

RT: 10.157 min Scan# 1185

Delta R.T. -0.006 min

Lab File: BM038659.D

Acq: 02 Feb 2023 12:20




Tgt Ion: 93 Resp: 150280


Ion Ratio Lower Upper

93 100

95 32.7 26.6 40.0

123 12.9 9.6 14.4





#29

2,4-Dichlorophenol

Concen: 41.814 ng

RT: 10.404 min Scan# 1227

Delta R.T. 0.006 min

Lab File: BM038659.D


Acq: 02 Feb 2023 12:20

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040






Tgt Ion:	162	Resp:	171230
Ion Ratio	Lower	Upper	
162	100		
164	64.2	44.3	84.3
98	23.9	3.1	43.1

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023



#30

1,2,4-Trichlorobenzene

Concen: 41.301 ng




RT: 10.604 min Scan# 1261


Delta R.T. -0.006 min

Lab File: BM038659.D

Acq: 02 Feb 2023 12:20


Tgt Ion:	180	Resp:	179908
Ion Ratio	Lower	Upper	
180	100		
182	95.4	76.2	114.2
145	27.9	20.8	31.2





#31
Naphthalene
Concen: 41.018 ng
RT: 10.792 min Scan# 1171
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:128 Resp: 422891
Ion Ratio Lower Upper
128 100
129 11.3 9.1 13.7
127 13.4 9.9 14.9




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023



#32
Benzoic acid
Concen: 32.956 ng
RT: 10.081 min Scan# 1172
Delta R.T. 0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20


Tgt Ion:122 Resp: 76895
Ion Ratio Lower Upper
122 100
105 118.0 99.2 139.2
77 91.1 72.3 112.3





#33
4-Chloroaniline
Concen: 41.266 ng
RT: 10.916 min Scan# 1314
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:127 Resp: 186454
Ion Ratio Lower Upper
127 100
129 33.3 27.1 40.7
65 26.5 22.2 33.4
92 16.3 13.6 20.4

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023




#34
Hexachlorobutadiene
Concen: 45.148 ng
RT: 11.063 min Scan# 1339
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20




Tgt Ion:225 Resp: 98518
Ion Ratio Lower Upper
225 100
223 61.1 51.0 76.4
227 65.5 49.2 73.8





#35
 Caprolactam
 Concen: 42.853 ng
 RT: 11.716 min Scan# 1450
 Delta R.T. -0.011 min
 Lab File: BM038659.D
 Acq: 02 Feb 2023 12:20




Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040




Tgt Ion:113 Resp: 37793
 Ion Ratio Lower Upper
 113 100
 55 120.2 111.2 151.2
 56 100.3 82.0 122.0

Manual Integrations
 APPROVED



Reviewed By :Christian Giraldo 02/06/2023
 Supervised By :Jagrut Upadhyay 02/06/2023




#36
 4-Chloro-3-methylphenol
 Concen: 42.237 ng
 RT: 12.051 min Scan# 1507
 Delta R.T. 0.012 min
 Lab File: BM038659.D
 Acq: 02 Feb 2023 12:20



Tgt Ion:107 Resp: 127894
 Ion Ratio Lower Upper
 107 100
 144 33.8 26.5 39.7
 142 102.4 77.0 115.6









#37
2-Methylnaphthalene
Concen: 42.221 ng
RT: 12.398 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:142 Resp: 351760
Ion Ratio Lower Upper
142 100
141 88.4 71.8 107.8
115 36.4 28.7 43.1




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023



#38
1-Methylnaphthalene
Concen: 41.587 ng
RT: 12.616 min Scan# 1603
Delta R.T. -0.012 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20

Tgt Ion:142 Resp: 326479
Ion Ratio Lower Upper
142 100
141 90.9 75.5 113.3
116 3.9 3.1 4.7





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.568 min Scan# 1935

Delta R.T. -0.006 min

Lab File: BM038659.D


Acq: 02 Feb 2023 12:20

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040



Tgt Ion:164 Resp: 148684

Ion Ratio Lower Upper

164 100

162 99.0 81.4 122.0




160 46.4 36.5 54.7

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 02/06/2023

Supervised By :Jagrut Upadhyay 02/06/2023



#40

1,2,4,5-Tetrachlorobenzene

Concen: 43.359 ng

RT: 12.763 min Scan# 1628

Delta R.T. -0.006 min

Lab File: BM038659.D

Acq: 02 Feb 2023 12:20

Tgt Ion:216 Resp: 167781




Ion Ratio Lower Upper


216 100

214 78.1 63.4 95.0

179 25.8 21.0 31.4


108 19.4 16.8 25.2





#41
Hexachlorocyclopentadiene
Concen: 38.787 ng
RT: 12.733 min Scan# 1624
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion: 237 Resp: 93134
Ion Ratio Lower Upper
237 100
235 61.8 43.8 83.8
272 13.5 0.0 33.8

Manual Integrations
APPROVED



Reviewed By : Christian Giraldo 02/06/2023
Supervised By : Jagrut Upadhyay 02/06/2023




#42
2,4,6-Tribromophenol
Concen: 92.671 ng
RT: 16.063 min Scan# 2189
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20




Tgt Ion: 330 Resp: 187283
Ion Ratio Lower Upper
330 100
332 97.0 77.8 116.6
141 36.5 31.0 46.6





#43
2,4,6-Trichlorophenol
Concen: 43.611 ng
RT: 13.016 min Scan# 1671
Delta R.T. -0.000 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:196 Resp: 132575
Ion Ratio Lower Upper
196 100
198 95.0 73.7 110.5
200 30.7 24.3 36.5

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023




#44
2,4,5-Trichlorophenol
Concen: 42.386 ng
RT: 13.098 min Scan# 1685
Delta R.T. 0.012 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20



Tgt Ion:196 Resp: 151630
Ion Ratio Lower Upper
196 100
198 98.3 73.1 109.7
97 44.4 35.5 53.3
132 23.3 18.2 27.4





#45

2-Fluorobiphenyl

Concen: 84.401 ng

RT: 13.198 min Scan# 1702

Delta R.T. -0.006 min

Lab File: BM038659.D

Acq: 02 Feb 2023 12:20

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Tgt Ion:172 Resp: 98570





Ion	Ratio	Lower	Upper
172	100		
171	35.8	28.7	43.1
170	24.4	20.0	30.0

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 02/06/2023

Supervised By :Jagrut Upadhyay 02/06/2023



#46

1,1'-Biphenyl

Concen: 41.503 ng

RT: 13.404 min Scan# 1737




Delta R.T. -0.012 min


Lab File: BM038659.D

Acq: 02 Feb 2023 12:20

Tgt Ion:154 Resp: 488818


Ion	Ratio	Lower	Upper
154	100		
153	39.8	19.9	59.9
76	7.1	0.0	27.9





#47
2-Chloronaphthalene
Concen: 41.763 ng
RT: 13.451 min Scan# 1745
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:162 Resp: 387053
Ion Ratio Lower Upper
162 100
127 30.0 23.6 35.4
164 33.5 26.7 40.1

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023




#48
2-Nitroaniline
Concen: 38.109 ng
RT: 13.669 min Scan# 1782
Delta R.T. -0.005 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20




Tgt Ion: 65 Resp: 78728
Ion Ratio Lower Upper
65 100
92 72.8 55.9 83.9
138 140.2 110.6 166.0





#49
Acenaphthylene
Concen: 41.920 ng
RT: 14.292 min Scan# 1888
Delta R.T. -0.012 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:152 Resp: 611933
Ion Ratio Lower Upper
152 100
151 20.3 16.1 24.1
153 13.4 10.0 15.0

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023




#50
Dimethylphthalate
Concen: 42.343 ng
RT: 14.033 min Scan# 1844
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20



Tgt Ion:163 Resp: 513770
Ion Ratio Lower Upper
163 100
194 3.4 2.6 4.0
164 10.4 8.2 12.4









#51
2,6-Dinitrotoluene
Concen: 43.310 ng
RT: 14.163 min Scan# 1867
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:165 Resp: 106805
Ion Ratio Lower Upper
165 100
63 43.5 35.8 53.8
89 47.9 39.3 58.9




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023



#52
Acenaphthene
Concen: 41.689 ng
RT: 14.633 min Scan# 1946
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20

Tgt Ion:154 Resp: 363368
Ion Ratio Lower Upper
154 100
153 112.3 93.7 140.5
152 54.3 44.6 67.0






#53
3-Nitroaniline
Concen: 39.237 ng
RT: 14.498 min Scan# 1923
Delta R.T. -0.000 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040






Tgt Ion:138 Resp: 100472
Ion Ratio Lower Upper
138 100
108 7.4 5.8 8.8
92 86.1 73.9 110.9


Manual Integrations

APPROVED



Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023




#54
2,4-Dinitrophenol
Concen: 34.239 ng
RT: 14.704 min Scan# 1958
Delta R.T. -0.000 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20




Tgt Ion:184 Resp: 54960
Ion Ratio Lower Upper
184 100
63 54.6 45.9 68.9
154 79.9 54.9 82.3





#55
Dibenzofuran
Concen: 41.700 ng
RT: 14.968 min Scan# 20
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:168 Resp: 625899
Ion Ratio Lower Upper
168 100
139 34.6 27.8 41.6
169 13.1 10.9 16.3




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023



#56
4-Nitrophenol
Concen: 38.105 ng
RT: 14.833 min Scan# 1980
Delta R.T. 0.029 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20


Tgt Ion:139 Resp: 75365
Ion Ratio Lower Upper
139 100
109 89.5 63.8 103.8
65 76.7 62.5 102.5





#57
2,4-Dinitrotoluene
Concen: 40.834 ng
RT: 14.945 min Scan# 1999
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:165 Resp: 147673
Ion Ratio Lower Upper
165 100
63 37.1 29.4 44.0
89 60.9 51.8 77.6
182 2.1 1.4 2.2

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023




#58
Fluorene
Concen: 42.721 ng
RT: 15.615 min Scan# 2113
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20




Tgt Ion:166 Resp: 507363
Ion Ratio Lower Upper
166 100
165 100.0 80.2 120.4
167 13.4 10.7 16.1





#59
2,3,4,6-Tetrachlorophenol
Concen: 43.936 ng
RT: 15.204 min Scan# 2043
Delta R.T. -0.000 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:232 Resp: 109392
Ion Ratio Lower Upper
232 100
131 48.3 41.2 61.8
130 2.8 2.4 3.6
166 38.2 31.8 47.6

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023




#60
Diethylphthalate
Concen: 42.994 ng
RT: 15.386 min Scan# 2074
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20




Tgt Ion:149 Resp: 494198
Ion Ratio Lower Upper
149 100
177 23.1 18.1 27.1
150 12.0 10.0 15.0





#61
4-Chlorophenyl-phenylether
Concen: 44.713 ng
RT: 15.610 min Scan# 2113
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:204 Resp: 225275
Ion Ratio Lower Upper
204 100
206 31.6 25.3 37.9
141 59.7 51.3 76.9




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023



#62
4-Nitroaniline
Concen: 39.093 ng m
RT: 15.657 min Scan# 2120
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20


Tgt Ion:138 Resp: 102631
Ion Ratio Lower Upper
138 100
92 43.0 23.4 63.4
108 100.5 75.1 115.1





#63
Azobenzene
Concen: 39.775 ng
RT: 15.904 min Scan# 2162
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion: 77 Resp: 35175
Ion Ratio Lower Upper
77 100
182 35.2 12.0 52.0
105 19.8 0.0 37.5
51 28.7 8.9 48.9

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.315 min Scan# 2402
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20



Tgt Ion:188 Resp: 318133
Ion Ratio Lower Upper
188 100
94 4.9 4.2 6.4
80 5.5 4.6 7.0









#65
4,6-Dinitro-2-methylphenol
Concen: 40.640 ng
RT: 15.710 min Scan# 2110
Delta R.T. -0.005 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:198 Resp: 81390
Ion Ratio Lower Upper
198 100
51 28.0 13.0 53.0
105 33.3 17.9 57.9




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023



#66
n-Nitrosodiphenylamine
Concen: 41.219 ng
RT: 15.827 min Scan# 2149
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20

Tgt Ion:169 Resp: 428170
Ion Ratio Lower Upper
169 100
168 71.2 57.6 86.4
167 39.4 32.3 48.5





#67
4-Bromophenyl-phenylether
Concen: 44.560 ng
RT: 16.504 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20

Instrument :

BNA_M

ClientSampleId :





SSTDCCC040

Tgt Ion:248 Resp: 136879
Ion Ratio Lower Upper
248 100
250 99.8 79.4 119.2
141 75.8 61.0 91.6

Manual Integrations




APPROVED


Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023



#68
Hexachlorobenzene
Concen: 43.746 ng
RT: 16.615 min Scan# 2283
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20


Tgt Ion:284 Resp: 164228
Ion Ratio Lower Upper
284 100
142 34.3 29.8 44.6
249 29.3 22.2 33.4





#69
 Atrazine
 Concen: 45.086 ng
 RT: 16.780 min Scan# 2311
 Delta R.T. -0.006 min
 Lab File: BM038659.D
 Acq: 02 Feb 2023 12:20




Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040




Tgt Ion:200 Resp: 135739
 Ion Ratio Lower Upper
 200 100
 173 27.6 7.9 47.9
 215 41.0 20.8 60.8

Manual Integrations
 APPROVED



Reviewed By :Christian Giraldo 02/06/2023
 Supervised By :Jagrut Upadhyay 02/06/2023




#70
 Pentachlorophenol
 Concen: 37.974 ng
 RT: 16.968 min Scan# 2343
 Delta R.T. 0.000 min
 Lab File: BM038659.D
 Acq: 02 Feb 2023 12:20




Tgt Ion:266 Resp: 103921
 Ion Ratio Lower Upper
 266 100
 268 62.5 51.8 77.6
 264 61.5 50.6 76.0





#71
Phenanthrene
Concen: 41.330 ng
RT: 17.357 min Scan# 2410
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:178 Resp: 780714
Ion Ratio Lower Upper
178 100
176 20.4 16.7 25.1
179 15.1 12.5 18.7

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023




#72
Anthracene
Concen: 41.693 ng
RT: 17.445 min Scan# 2424
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20




Tgt Ion:178 Resp: 805706
Ion Ratio Lower Upper
178 100
176 19.4 15.6 23.4
179 15.4 12.4 18.6





#73
 Carbazole
 Concen: 41.021 ng
 RT: 17.727 min Scan# 2472
 Delta R.T. -0.000 min
 Lab File: BM038659.D
 Acq: 02 Feb 2023 12:20




Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040




Tgt Ion:167 Resp: 739301
 Ion Ratio Lower Upper
 167 100
 166 21.9 17.8 26.6
 139 11.6 9.4 14.2

Manual Integrations
 APPROVED



Reviewed By :Christian Giraldo 02/06/2023
 Supervised By :Jagrut Upadhyay 02/06/2023




#74
 Di-n-butylphthalate
 Concen: 44.941 ng
 RT: 18.274 min Scan# 2565
 Delta R.T. -0.006 min
 Lab File: BM038659.D
 Acq: 02 Feb 2023 12:20




Tgt Ion:149 Resp: 915074
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.4 11.0
 104 4.7 3.8 5.6





#75
Fluoranthene
Concen: 43.633 ng
RT: 19.368 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:202 Resp: 865384
Ion Ratio Lower Upper
202 100
101 6.4 0.0 27.4
203 17.6 0.0 37.3




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.492 min Scan# 3112
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20

Tgt Ion:240 Resp: 280638
Ion Ratio Lower Upper
240 100
120 6.9 7.0 10.4#
236 26.4 20.9 31.3





#77

Benzidine

Concen: 31.486 ng

RT: 19.556 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM038659.D

Acq: 02 Feb 2023 12:20

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040





Tgt Ion:184 Resp: 255813

Ion	Ratio	Lower	Upper
184	100		
185	14.1	10.8	16.2
183	11.7	8.8	13.2

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023



#78

Pyrene

Concen: 41.762 ng

RT: 19.733 min Scan# 2813




Delta R.T. -0.006 min


Lab File: BM038659.D

Acq: 02 Feb 2023 12:20

Tgt Ion:202 Resp: 902321

Ion	Ratio	Lower	Upper
202	100		
200	22.4	17.8	26.6
203	17.2	13.5	20.3






#79
 Terphenyl-d14
 Concen: 89.468 ng
 RT: 19.927 min Scan# 2847
 Delta R.T. -0.006 min
 Lab File: BM038659.D
 Acq: 02 Feb 2023 12:20

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040






Tgt Ion:244 Resp: 1422828
 Ion Ratio Lower Upper
 244 100
 212 7.8 6.6 9.8
 122 8.5 7.8 11.8


Manual Integrations

APPROVED



Reviewed By :Christian Giraldo 02/06/2023
 Supervised By :Jagrut Upadhyay 02/06/2023




#80
 Butylbenzylphthalate
 Concen: 40.794 ng
 RT: 20.621 min Scan# 2964
 Delta R.T. -0.006 min
 Lab File: BM038659.D
 Acq: 02 Feb 2023 12:20




Tgt Ion:149 Resp: 435520
 Ion Ratio Lower Upper
 149 100
 91 53.5 45.6 68.4
 206 17.3 12.5 18.7





#81
Benzo(a)anthracene
Concen: 42.360 ng
RT: 21.474 min Scan# 3110
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:228 Resp: 840901
Ion Ratio Lower Upper
228 100
226 28.2 22.2 33.4
229 19.1 15.5 23.3

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023




#82
3,3'-Dichlorobenzidine
Concen: 43.618 ng
RT: 21.409 min Scan# 3098
Delta R.T. -0.006 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20



Tgt Ion:252 Resp: 297890
Ion Ratio Lower Upper
252 100
254 62.8 52.8 79.2
126 9.1 8.2 12.2





#83

Chrysene

Concen: 42.071 ng

RT: 21.527 min Scan# 3119

Delta R.T. -0.006 min

Lab File: BM038659.D

Acq: 02 Feb 2023 12:20

Instrument :

BNA_M

ClientSampleId :





SSTDCCC040

Tgt Ion:228 Resp: 804678
 Ion Ratio Lower Upper
 228 100
 226 30.8 24.7 37.1
 229 19.6 15.5 23.3

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 02/06/2023
 Supervised By :Jagrut Upadhyay 02/06/2023



#84

Bis(2-ethylhexyl)phthalate

Concen: 41.291 ng




RT: 21.386 min Scan# 3094


Delta R.T. -0.006 min

Lab File: BM038659.D

Acq: 02 Feb 2023 12:20

Tgt Ion:149 Resp: 642898
 Ion Ratio Lower Upper
 149 100
 167 27.2 20.8 31.2
 279 3.6 2.5 3.7





#85

Di-n-octyl phthalate

Concen: 43.604 ng

RT: 22.309 min Scan# 31

Delta R.T. -0.006 min

Lab File: BM038659.D

Acq: 02 Feb 2023 12:20

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Tgt Ion:149 Resp: 1108956

Ion Ratio Lower Upper

149 100

167 1.6 1.3 1.9





43 5.1 4.6 6.8

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 02/06/2023

Supervised By :Jagrut Upadhyay 02/06/2023



#86

Perylene-d12

Concen: 20.000 ng

RT: 23.886 min Scan# 3519

Delta R.T. -0.006 min

Lab File: BM038659.D

Acq: 02 Feb 2023 12:20




Tgt Ion:264 Resp: 333736


Ion Ratio Lower Upper

264 100

260 25.0 19.6 29.4

265 23.3 17.9 26.9





#87

Indeno(1,2,3-cd)pyrene

Concen: 41.378 ng

RT: 26.385 min Scan# 3944

Delta R.T. -0.018 min

Lab File: BM038659.D

Acq: 02 Feb 2023 12:20

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Tgt Ion:276 Resp: 108772

Ion Ratio Lower Upper

276 100

138 18.6 17.2 25.8





277 25.1 20.2 30.4

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 02/06/2023

Supervised By :Jagrut Upadhyay 02/06/2023



#88

Benzo(b)fluoranthene

Concen: 44.320 ng m

RT: 23.156 min Scan# 3395

Delta R.T. -0.006 min

Lab File: BM038659.D

Acq: 02 Feb 2023 12:20




Tgt Ion:252 Resp: 885923


Ion Ratio Lower Upper

252 100

253 21.5 17.4 26.2

125 6.8 7.0 10.6#









#89
Benzo(k)fluoranthene
Concen: 42.792 ng
RT: 23.203 min Scan# 3405
Delta R.T. -0.012 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:252 Resp: 86862
Ion Ratio Lower Upper
252 100
253 21.6 17.4 26.0
125 7.1 6.6 9.8




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023



#90
Benzo(a)pyrene
Concen: 42.652 ng
RT: 23.780 min Scan# 3501
Delta R.T. -0.012 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20

Tgt Ion:252 Resp: 843144
Ion Ratio Lower Upper
252 100
253 22.2 17.7 26.5
125 7.6 7.8 11.8#









#91
Dibenzo(a,h)anthracene
Concen: 41.841 ng
RT: 26.397 min Scan# 3949
Delta R.T. -0.018 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:278 Resp: 925360
Ion Ratio Lower Upper
278 100
139 11.3 10.5 15.7
279 23.6 18.7 28.1

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 02/06/2023
Supervised By :Jagrut Upadhyay 02/06/2023



#92
Benzo(g,h,i)perylene
Concen: 39.300 ng
RT: 27.156 min Scan# 4075
Delta R.T. -0.012 min
Lab File: BM038659.D
Acq: 02 Feb 2023 12:20

Tgt Ion:276 Resp: 878586
Ion Ratio Lower Upper
276 100
277 23.6 19.1 28.7
138 16.2 15.3 22.9

