

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM021919\
 Data File : BM018867.D
 Acq On : 19 Feb 2019 21:26
 Operator : JU/SJ
 Sample : K1518-08
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 CABLE-BRIDGE-FOUNDATION-D-4

Quant Time: Feb 20 03:13:48 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM021119.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 11 16:02:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.70	152	305238	20.00	ng	-0.01
21) Naphthalene-d8	10.49	136	1219400	20.00	ng	-0.01
39) Acenaphthene-d10	14.35	164	682367	20.00	ng	-0.01
64) Phenanthrene-d10	17.10	188	1638066	20.00	ng	-0.01
76) Chrysene-d12	21.30	240	1614094	20.00	ng	-0.01
87) Perylene-d12	23.55	264	1372985	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.30	112	2420595	129.93	ng	0.00
7) Phenol-d6	6.88	99	2951112	112.45	ng	0.00
23) Nitrobenzene-d5	8.87	82	2424545	91.94	ng	-0.02
42) 2,4,6-Tribromophenol	15.85	330	1384059	140.71	ng	-0.01
45) 2-Fluorobiphenyl	12.97	172	4676768	90.98	ng	-0.01
79) Terphenyl-d14	19.74	244	8452364	108.03	ng	0.00


Target Compounds Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM021919\
Data File : BM018867.D
Acq On : 19 Feb 2019 21:26
Operator : JU/SJ
Sample : K1518-08
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
CABLE-BRIDGE-FOUNDATION-D-4

Quant Time: Feb 20 03:13:48 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM021119.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 11 16:02:12 2019
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.70 min Scan# 785
Delta R.T. -0.01 min
Lab File: BM018867.D
Acq: 19 Feb 2019 21:26

Instrument :
BNA_M
ClientSampleId :
CABLE-BRIDGE-FOUNDATION-D-4

Tot Ion:152 Resp: 305238
Ion Ratio Lower Upper
152 100
150 159.0 127.3 190.9
115 64.8 50.8 76.2




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 129.934 ng
RT: 5.30 min Scan# 376
Delta R.T. -0.01 min
Lab File: BM018867.D
Acq: 19 Feb 2019 21:26

Tgt Ion:112 Resp: 2420595
Ion Ratio Lower Upper
112 100
64 65.7 50.6 75.8
63 34.1 25.7 38.5



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BM0
Ion 63.00 (62.70 to 63.70): BM0





#7

Phenol-d6

Concen: 112.448 ng

RT: 6.88 min Scan# 645

Delta R.T. -0.01 min

Lab File: BM018867.D

Acq: 19 Feb 2019 21:26

Instrument :


BNA_M

ClientSampleId :

CABLE-BRIDGE-FOUNDATION-D-4

Tot Ion: 99 Resp: 2951112



Ion	Ratio	Lower	Upper
99	100		
42	14.3	11.0	16.6
71	39.6	31.3	46.9



Abundance Ion 99.00 (98.70 to 99.70): BM018867.D (-629) (-)

Ion 42.00 (41.70 to 42.70): BM018867.D (-629) (-)

Ion 71.00 (70.70 to 71.70): BM018867.D (-629) (-)



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 10.49 min Scan# 1259


Delta R.T. -0.01 min

Lab File: BM018867.D

Acq: 19 Feb 2019 21:26

Tgt Ion: 136 Resp: 1219400

Ion	Ratio	Lower	Upper
136	100		
137	10.8	8.8	13.2
54	6.7	5.0	7.6
68	7.2	5.5	8.3




Abundance Ion 136.00 (135.70 to 136.70): BM018867.D (-1244) (-)

Ion 137.00 (136.70 to 137.70): BM018867.D (-1244) (-)

Ion 54.00 (53.70 to 54.70): BM018867.D (-1244) (-)

Ion 68.00 (67.70 to 68.70): BM018867.D (-1244) (-)









#23
 Nitrobenzene-d5
 Concen: 91.937 ng
 RT: 8.87 min Scan# 983
 Delta R.T. -0.02 min
 Lab File: BM018867.D
 Acq: 19 Feb 2019 21:26




Instrument :
 BNA_M
 ClientSampleId :
 CABLE-BRIDGE-FOUNDATION-D-4


Tot Ion: 82 Resp: 2424545
 Ion Ratio Lower Upper
 82 100
 128 35.8 29.9 44.9
 54 38.2 29.1 43.7



#39
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 14.35 min Scan# 1915
 Delta R.T. -0.01 min
 Lab File: BM018867.D
 Acq: 19 Feb 2019 21:26

Tgt Ion:164 Resp: 682367
 Ion Ratio Lower Upper
 164 100
 162 103.3 83.4 125.0
 160 45.2 36.6 55.0





#42

2,4,6-Tribromophenol

Concen: 140.712 ng

RT: 15.85 min Scan# 2170

Delta R.T. -0.01 min

Lab File: BM018867.D

Acq: 19 Feb 2019 21:26

Instrument :

BNA_M

ClientSampleId :

CABLE-BRIDGE-FOUNDATION-D-4


Tot Ion:330 Resp: 1384059

Ion Ratio Lower Upper

330 100

332 96.5 76.9 115.3

141 37.6 28.7 43.1






Abundance

Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E



#45

2-Fluorobiphenyl

Concen: 90.982 ng

RT: 12.97 min Scan# 1680

Delta R.T. -0.01 min

Lab File: BM018867.D

Acq: 19 Feb 2019 21:26


Tgt Ion:172 Resp: 4676768

Ion Ratio Lower Upper

172 100

171 36.1 29.0 43.6

170 23.9 18.6 28.0






Abundance

Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.10 min Scan# 2383
Delta R.T. -0.01 min
Lab File: BM018867.D
Acq: 19 Feb 2019 21:26

Instrument :
BNA_M
ClientSampleId :
CABLE-BRIDGE-FOUNDATION-D-4

Tot Ion:188 Resp: 1638066
Ion Ratio Lower Upper
188 100
94 10.0 8.2 12.2
80 11.9 9.7 14.5




Abundance Ion 188.00 (187.70 to 188.70): E
Ion 94.00 (93.70 to 94.70): BM
Ion 80.00 (79.70 to 80.70): BM






#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.30 min Scan# 3096
Delta R.T. -0.01 min
Lab File: BM018867.D
Acq: 19 Feb 2019 21:26

Tgt Ion:240 Resp: 1614094
Ion Ratio Lower Upper
240 100
120 11.2 8.9 13.3
236 26.0 20.9 31.3



Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E









#79
 Terphenyl-d14
 Concen: 108.026 ng
 RT: 19.74 min Scan# 2831
 Delta R.T. -0.01 min
 Lab File: BM018867.D
 Acq: 19 Feb 2019 21:26

Instrument :
 BNA_M
 ClientSampleId :
 CABLE-BRIDGE-FOUNDATION-D-4

Tot Ion:244 Resp: 8452364
 Ion Ratio Lower Upper
 244 100
 212 7.1 5.6 8.4
 122 11.2 9.3 13.9



#87
 Perylene-d12
 Concen: 20.000 ng
 RT: 23.55 min Scan# 3479
 Delta R.T. -0.02 min
 Lab File: BM018867.D
 Acq: 19 Feb 2019 21:26

Tgt Ion:264 Resp: 1372985
 Ion Ratio Lower Upper
 264 100
 260 24.5 18.9 28.3
 265 22.0 17.3 25.9

