

Data File : BM025012.D
Acq On : 22 Feb 2020 06:21
Operator : CG/JU
Sample : L1507-10
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :

Quant Time: Feb 24 04:13:24 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM020620MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Feb 24 03:25:45 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.33	152	182521	20.00	ng/ul	0.00
18) Naphthalene-d8	10.09	136	727611	20.00	ng/ul	0.00
36) Acenaphthene-d10	13.99	164	487566	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.75	188	1055445	20.00	ng/ul	0.00
78) Chrysene-d12	20.96	240	1207231	20.00	ng/ul	0.00
86) Perylene-d12	23.04	264	1357634	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	2.95	96	5517	1.45	ng/uL	0.00
5) Phenol-d5	6.54	99	55255	4.56	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.69	67	43065	6.29	ng/ul	0.00
9) 2-Chlorophenol-d4	6.88	132	54371	5.01	ng/ul	0.00
13) 4-Methylphenol-d8	8.06	113	27641	2.70	ng/ul	0.00
19) Nitrobenzene-d5	8.48	128	26159m	4.97	ng/ul	0.00
22) 2-Nitrophenol-d4	9.20	143	25430	4.09	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.74	165	47060m	3.75	ng/ul	0.01
29) 4-Chloroaniline-d4	10.26	131	45774m	3.90	ng/ul	0.02
44) Dimethylphthalate-d6	13.42	166	192833	5.24	ng/ul	0.00
47) Acenaphthylene-d8	13.68	160	221050	5.14	ng/ul	0.00
52) 4-Nitrophenol-d4	14.29	143	4312m	0.72	ng/ul	0.06
58) Fluorene-d10	15.00	176	174802	5.36	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.15	200	13060	1.90	ng/ul	0.01
71) Anthracene-d10	16.85	188	262472	5.45	ng/ul	0.00
79) Pyrene-d10	19.16	212	355731	5.89	ng/ul	0.00
90) Benzo(a)pyrene-d12	22.91	264	409877	6.01	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BM025012.D
Acq On : 22 Feb 2020 06:21
Operator : CG/JU
Sample : L1507-10
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :

Quant Time: Feb 24 04:13:24 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM020620MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Feb 24 03:25:45 2020
Response via : Initial Calibration

