


Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM022120\
 Data File : BM024988.D
 Acq On : 21 Feb 2020 15:46
 Operator : CG/JU
 Sample : DFTPP24
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP24

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM020620MA.M
 Title : SVOA CALIBRATION
 Last Update : Mon Feb 24 05:59:54 2020



AutoFind: Scans 2441, 2442, 2443; Background Corrected with Scan 2435

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	17.5	60368	PASS
68	69	0.00	2	1.8	2212	PASS
69	198	0.00	100	35.5	122501	PASS
70	69	0.00	2	0.5	635	PASS
127	198	10	80	40.0	138061	PASS
197	198	0.00	2	0.6	2124	PASS
198	198	100	100	100.0	345109	PASS
199	198	5	9	6.6	22779	PASS
275	198	10	60	32.1	110930	PASS
365	198	1	100	4.6	15833	PASS
441	198	0.01	100	13.8	47797	PASS
442	198	50	100	94.3	325290	PASS
443	442	15	24	19.1	62261	PASS

Scan 2435 (17.210 min): BM024988.D
DFTPP24

m/z	Abundance
43.90	749.0
73.00	529.0
207.00	1548.0
207.90	493.0
281.10	750.0

Instrument :
BNA_M
ClientSampleId :
DFTPP24

m/z	Abundance
38.00	939.0
39.00	5571.0
40.10	492.0
43.90	1055.0
49.10	616.0
50.00	20392.0
51.00	72352.0
52.00	4509.0
54.90	548.0
55.90	2866.0
57.00	5221.0
57.90	300.0
61.00	1379.0
61.90	1812.0
63.00	5193.0
64.00	579.0
65.10	2321.0
68.00	2885.0
69.00	148480.0
69.90	908.0
73.00	1608.0
74.00	16068.0
75.00	24056.0
76.10	8260.0
77.00	154816.0
78.00	10882.0
79.00	11999.0
80.00	8407.0
81.00	11941.0
81.90	2235.0
83.00	2806.0
83.80	395.0
85.00	1897.0
86.00	4006.0
87.10	1785.0
88.00	516.0
91.00	2577.0
91.90	3262.0
93.00	20560.0
94.00	1251.0
95.10	428.0
96.00	902.0
96.80	514.0
98.00	16888.0
99.00	14003.0
100.10	886.0
101.00	6515.0
101.80	358.0
102.90	2121.0
104.00	4487.0
105.00	4173.0
106.10	1292.0
107.00	48712.0
108.00	6980.0
109.10	1871.0
110.00	84000.0
111.00	11888.0
111.90	1651.0
112.90	753.0
115.00	317.0
116.00	2706.0
117.00	44656.0
118.00	3473.0
118.90	529.0
120.10	549.0
122.00	3768.0
123.00	4613.0
123.90	2543.0
125.00	1810.0
127.00	159232.0
128.00	13370.0
129.00	63152.0
130.00	6629.0
130.90	754.0
131.90	599.0
133.00	639.0
134.00	2217.0
135.00	4928.0
136.00	1982.0
137.00	3345.0
137.80	522.0
138.90	343.0
140.00	1064.0
141.00	9925.0
141.90	3219.0
143.00	2107.0
144.00	687.0
144.90	621.0
146.00	1949.0
147.00	4582.0
148.00	11779.0
149.00	2698.0

Instrument :
BNA_M
ClientSampleId :
DFTPP24

150.00	518.0
151.10	1008.0
151.80	664.0
153.00	3343.0
153.90	2479.0
155.00	4829.0
156.00	7296.0
157.00	1412.0
157.90	1460.0
158.90	1027.0
159.90	2157.0
161.00	4382.0
162.00	1157.0
163.00	466.0
163.80	435.0
165.00	3303.0
165.90	2376.0
167.00	16832.0
168.00	11538.0
168.90	1990.0
170.00	851.0
170.80	786.0
172.00	1893.0
172.90	1710.0
174.00	3877.0
175.00	7663.0
175.90	1833.0
177.00	3297.0
177.80	1300.0
179.00	14674.0
180.00	10982.0
181.00	4926.0
182.00	858.0
183.00	626.0
184.00	1145.0
185.00	6467.0
186.00	53104.0
187.00	16289.0
188.00	2207.0
188.90	3356.0
190.00	585.0
191.00	1550.0
192.00	4785.0
193.00	4995.0
193.90	1136.0
195.00	691.0
196.00	11167.0
197.00	2353.0
197.90	372352.0
198.90	24144.0
200.00	1844.0
201.60	1825.0
203.00	2481.0
204.10	13928.0
205.00	23056.0
206.00	103456.0
207.00	15280.0
208.00	3711.0
208.90	1411.0
210.90	3696.0
212.90	399.0
215.00	1255.0
215.90	2300.0
216.90	28672.0
218.00	3455.0
221.00	14340.0
222.00	2922.0
223.00	6082.0
224.00	58920.0
225.00	15021.0
225.90	2150.0
227.00	25176.0
228.00	3175.0
229.00	4799.0
229.90	515.0
230.90	1840.0
231.90	535.0
232.80	488.0
233.90	1536.0
234.90	2339.0
235.90	1002.0
237.00	2906.0
237.90	449.0
238.90	618.0
240.00	961.0
240.90	2005.0
242.00	2505.0
243.00	3945.0
244.00	45840.0
245.00	6073.0
246.00	9747.0
246.90	1822.0
247.90	548.0
248.90	1805.0
250.00	433.0
250.80	478.0

Instrument :
BNA_M
ClientSampleId :
DFTPP24

251.90	429.0
253.00	1035.0
254.00	1782.0
255.00	244800.0
256.00	36048.0
256.90	3132.0
258.00	14213.0
258.90	1982.0
259.90	357.0
260.90	333.0
263.80	338.0
265.00	6196.0
266.00	882.0
269.80	313.0
271.00	901.0
272.00	693.0
273.00	6600.0
274.00	18544.0
275.00	112496.0
276.00	13653.0
277.00	8865.0
277.90	1872.0
279.10	444.0
281.00	595.0
281.90	482.0
282.90	1037.0
284.00	708.0
285.10	1384.0
289.00	505.0
291.90	309.0
292.90	2341.0
293.90	729.0
295.00	714.0
296.00	30256.0
297.00	5443.0
297.90	491.0
301.00	566.0
301.90	525.0
303.00	3782.0
304.10	1066.0
309.00	312.0
314.00	1565.0
314.90	3326.0
315.90	1696.0
317.00	451.0
321.00	1310.0
321.90	830.0
323.00	9920.0
324.00	1910.0
326.90	2198.0
328.00	1063.0
332.00	828.0
333.00	1000.0
334.00	5943.0
335.10	1585.0
336.10	341.0
341.00	1318.0
341.90	627.0
346.00	2081.0
347.00	649.0
352.00	2481.0
353.00	2353.0
354.00	3474.0
354.90	535.0
355.20	630.0
364.90	14421.0
365.90	2156.0
370.90	821.0
372.00	6031.0
373.00	1245.0
383.00	1302.0
389.80	540.0
391.00	307.0
392.10	396.0
401.90	1835.0
403.10	3043.0
403.90	1288.0
420.90	1971.0
421.90	2130.0
423.00	15805.0
424.10	6019.0
424.90	907.0
441.00	39416.0
442.00	266816.0
443.00	53872.0
444.00	4645.0
445.10	425.0

Instrument :
BNA_M
ClientSampleId :
DFTPP24

m/z	Abundance
38.00	917.0
39.00	4854.0
40.00	354.0
44.00	894.0
48.90	813.0
50.00	20712.0
51.00	71720.0
52.00	3558.0
53.10	326.0
55.00	502.0
56.00	2419.0
57.00	6103.0
61.00	1166.0
62.00	1452.0
63.00	4771.0
63.80	632.0
65.00	1985.0
67.00	312.0
68.00	2233.0
69.00	142208.0
70.00	566.0
73.10	1750.0
74.00	14764.0
75.00	22720.0
76.00	7274.0
77.00	145536.0
78.10	9942.0
79.00	12347.0
80.00	9444.0
81.00	11435.0
82.00	2872.0
83.00	2097.0
83.90	380.0
85.10	1761.0
86.00	3411.0
86.90	1606.0
88.00	842.0
91.00	2556.0
92.00	3472.0
93.00	19664.0
93.90	1108.0
95.00	450.0
96.00	974.0
97.00	706.0
98.00	15017.0
99.00	11625.0
100.00	964.0
101.00	6405.0
101.90	364.0
103.00	2361.0
104.00	4432.0
104.90	3812.0
106.00	1712.0
107.00	50632.0
108.00	7414.0
108.90	1080.0
110.00	82920.0
111.00	12607.0
112.10	1926.0
112.90	387.0
116.00	2980.0
117.00	45096.0
118.00	3527.0
118.90	510.0
119.80	404.0
122.00	3811.0
123.00	5372.0
124.00	2409.0
125.00	2215.0
126.10	934.0
127.00	166016.0
128.00	12402.0
129.00	66232.0
129.90	5647.0
131.00	1501.0
132.10	518.0
132.90	324.0
134.00	2083.0
135.00	5563.0
136.00	3027.0
137.00	3017.0
138.00	666.0
139.90	689.0
141.00	9171.0
142.00	3068.0
142.90	2093.0
143.90	496.0
145.90	1887.0
146.90	4957.0
148.00	12337.0
149.00	2335.0
149.90	779.0

Instrument :
BNA_M
ClientSampleId :
DFTPP24

151.00	1147.0
151.80	901.0
153.00	3614.0
153.90	2102.0
155.00	4751.0
156.00	8224.0
157.00	1080.0
157.90	1974.0
158.90	1768.0
160.00	3191.0
161.00	5337.0
162.00	1223.0
164.00	517.0
164.90	3795.0
166.00	2936.0
167.00	19984.0
168.00	11563.0
168.90	1846.0
170.10	796.0
170.90	819.0
171.80	2035.0
173.00	2101.0
173.90	3835.0
175.00	8261.0
175.90	2177.0
176.90	3116.0
178.00	1510.0
179.00	16840.0
180.00	10447.0
181.00	5105.0
181.90	633.0
182.90	555.0
184.00	1614.0
185.00	6816.0
186.00	57352.0
187.00	16400.0
188.00	1826.0
188.90	4164.0
189.90	678.0
191.10	1725.0
192.00	5141.0
192.90	5529.0
194.00	1209.0
194.90	570.0
195.90	12660.0
197.10	2753.0
197.90	417280.0
198.90	29008.0
200.00	2168.0
201.60	1763.0
203.00	3000.0
204.00	16123.0
205.00	26120.0
206.00	116992.0
207.00	16584.0
208.00	4346.0
208.90	1636.0
209.90	1999.0
211.00	4245.0
211.90	1021.0
212.90	381.0
215.00	1305.0
216.00	2667.0
216.90	30960.0
217.90	3843.0
221.00	17368.0
221.70	5057.0
223.00	7506.0
224.00	69144.0
225.00	16856.0
226.10	1872.0
227.00	29376.0
228.00	4524.0
228.90	6139.0
230.10	705.0
230.90	2087.0
231.90	390.0
232.80	608.0
233.90	1991.0
234.90	2268.0
235.90	1308.0
237.00	2774.0
238.00	484.0
238.90	1052.0
239.90	923.0
240.90	2061.0
242.00	3648.0
243.00	3876.0
244.00	51184.0
245.00	6640.0
246.00	11510.0
246.90	1860.0
247.90	351.0
248.90	1899.0
249.90	424.0
250.80	404.0

Instrument :
BNA_M
ClientSampleId :
DFTPP24

251.80	567.0
252.90	927.0
254.10	2233.0
255.00	286720.0
256.00	42136.0
257.00	3835.0
258.00	16424.0
258.90	2799.0
259.90	311.0
260.90	465.0
264.00	609.0
265.00	7046.0
265.80	1032.0
270.00	454.0
271.00	712.0
271.90	940.0
273.00	8234.0
274.00	21592.0
275.00	134144.0
276.00	18424.0
277.00	11994.0
278.00	2020.0
279.10	534.0
280.90	651.0
281.90	506.0
282.90	1429.0
283.90	748.0
285.00	1928.0
285.80	344.0
289.00	529.0
289.90	385.0
291.10	420.0
292.00	663.0
292.90	2991.0
294.00	689.0
295.10	1003.0
296.00	40240.0
296.90	4980.0
300.90	526.0
301.90	957.0
303.00	4410.0
304.00	1382.0
307.90	624.0
309.00	318.0
310.00	558.0
313.00	485.0
314.10	1855.0
314.90	4546.0
315.90	3138.0
316.90	438.0
321.00	1270.0
321.90	451.0
323.00	12446.0
324.10	2242.0
326.00	341.0
326.90	2644.0
328.00	1663.0
329.10	301.0
331.90	717.0
332.90	1334.0
334.00	9497.0
335.00	2130.0
335.90	387.0
340.90	1685.0
342.10	516.0
345.90	2974.0
346.90	725.0
352.00	4452.0
353.00	2830.0
354.00	4653.0
355.00	989.0
359.10	410.0
365.00	18896.0
365.90	3144.0
369.90	529.0
371.00	934.0
372.10	7310.0
373.10	1286.0
383.00	2142.0
384.00	641.0
390.10	992.0
391.10	488.0
400.80	340.0
401.90	2365.0
403.00	3355.0
404.00	1731.0
405.00	355.0
421.00	3219.0
422.00	3506.0
423.10	21496.0
424.00	6400.0
425.00	601.0
441.00	56200.0
442.00	395968.0
443.00	75128.0
444.00	6773.0

Instrument :
BNA_M
ClientSampleId :
DFTPP24

444.90

601.0

Instrument :
BNA_M
ClientSampleId :
DFTPP24

m/z	Abundance
38.10	392.0
39.00	2691.0
39.90	448.0
43.90	876.0
49.00	427.0
50.00	10040.0
51.00	37032.0
52.10	1580.0
56.00	1507.0
57.00	3410.0
60.90	758.0
61.90	991.0
63.00	2485.0
65.00	1165.0
68.10	1519.0
69.00	76816.0
70.00	432.0
73.00	1188.0
74.00	8655.0
75.00	11791.0
76.00	3621.0
77.00	77656.0
78.00	4852.0
79.00	5979.0
80.00	3937.0
81.00	5006.0
82.00	1588.0
83.00	1234.0
83.90	404.0
84.90	1311.0
86.00	1754.0
87.00	781.0
87.80	400.0
90.90	1375.0
92.00	1494.0
93.00	10370.0
94.00	553.0
96.10	745.0
97.10	327.0
98.00	8519.0
99.00	6543.0
99.90	506.0
101.00	3870.0
102.80	1120.0
104.00	2412.0
105.00	1961.0
106.10	812.0
107.00	25496.0
107.90	4418.0
109.10	865.0
110.00	45080.0
111.00	6746.0
111.90	1616.0
113.00	369.0
116.00	1568.0
117.00	23864.0
118.00	1957.0
121.90	2065.0
123.00	3194.0
124.00	1220.0
124.90	1334.0
127.00	88936.0
128.00	7134.0
129.00	36344.0
130.00	3311.0
131.10	619.0
132.10	305.0
132.80	466.0
134.00	972.0
135.00	3048.0
135.90	1324.0
136.90	1912.0
140.10	389.0
141.00	5547.0
141.90	1561.0
143.00	923.0
144.00	334.0
144.80	475.0
146.00	852.0
146.90	3083.0
148.00	7281.0
149.00	1682.0
149.90	685.0
151.00	679.0
151.80	722.0
153.00	1863.0
153.90	1053.0
154.90	2424.0
156.00	4434.0
157.10	986.0
157.80	1032.0
159.10	674.0

Instrument :
BNA_M
ClientSampleId :
DFTPP24

159.90	1518.0
161.00	3233.0
162.00	596.0
165.00	2529.0
165.90	1697.0
167.00	9245.0
168.00	7163.0
168.90	1234.0
169.90	598.0
170.90	600.0
171.90	797.0
172.90	1161.0
174.00	2235.0
175.00	4539.0
175.90	1184.0
176.90	1944.0
177.90	991.0
179.00	9798.0
179.90	6507.0
180.90	2513.0
182.00	543.0
183.90	1147.0
185.00	4544.0
186.00	34408.0
187.00	8946.0
187.90	1060.0
189.00	2777.0
190.00	615.0
190.80	1148.0
192.00	3178.0
193.00	3610.0
193.80	684.0
195.00	530.0
196.00	6943.0
196.90	1266.0
197.90	245696.0
198.90	15187.0
200.00	1618.0
201.40	1043.0
202.90	1948.0
204.00	8722.0
205.00	15403.0
206.00	71728.0
207.00	11079.0
208.00	2388.0
209.00	1058.0
210.30	1205.0
211.00	2836.0
214.90	976.0
216.10	1465.0
216.90	20096.0
218.00	2535.0
219.00	315.0
219.80	331.0
221.00	9930.0
221.90	2202.0
223.00	4515.0
224.00	40528.0
225.00	9738.0
225.90	1323.0
227.00	16632.0
227.90	2601.0
228.90	3851.0
230.00	575.0
231.00	1863.0
234.00	1016.0
235.00	1031.0
235.90	748.0
237.00	2300.0
239.10	989.0
240.00	512.0
241.00	964.0
242.00	2473.0
243.00	2451.0
244.00	32448.0
245.00	4329.0
246.00	6532.0
247.00	1457.0
248.00	418.0
249.00	1146.0
252.00	481.0
252.90	748.0
254.10	1201.0
255.00	184000.0
256.00	25888.0
257.00	2597.0
258.00	11384.0
258.90	1568.0
260.00	344.0
265.00	5163.0
266.00	622.0
271.10	428.0
272.10	560.0
273.00	5820.0
274.00	15152.0
275.00	86152.0

Instrument :
BNA_M
ClientSampleId :
DFTPP24

276.00	11115.0
277.00	7529.0
278.00	1378.0
278.90	349.0
281.10	583.0
281.90	408.0
282.80	830.0
283.90	744.0
285.00	1380.0
286.00	451.0
293.00	1848.0
293.80	311.0
295.20	716.0
296.00	28168.0
296.90	3677.0
297.90	376.0
301.20	316.0
301.90	574.0
303.00	3095.0
304.00	714.0
308.70	303.0
314.00	1411.0
315.00	2817.0
316.00	973.0
320.90	701.0
321.90	511.0
323.00	8807.0
324.00	1369.0
326.90	1932.0
328.00	1096.0
329.00	345.0
332.00	659.0
332.90	714.0
334.00	6095.0
335.00	1695.0
341.10	1140.0
345.90	2049.0
346.90	580.0
352.00	2695.0
352.90	1537.0
354.00	2853.0
355.10	796.0
358.80	340.0
365.00	14182.0
366.00	1999.0
370.00	480.0
370.90	863.0
372.00	4676.0
373.00	1024.0
383.00	1315.0
383.90	398.0
390.00	617.0
390.90	598.0
392.20	401.0
401.10	516.0
402.00	2409.0
403.00	2903.0
404.00	1227.0
405.10	454.0
420.90	2771.0
422.10	2618.0
423.00	18224.0
424.00	5032.0
425.00	590.0
441.00	47776.0
442.00	313088.0
443.10	57784.0
444.00	5305.0

Instrument :
BNA_M
ClientSampleId :
DFTPP24