

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM022520\
 Data File : BM025043.D
 Acq On : 25 Feb 2020 14:37
 Operator : CG/JU
 Sample : L1486-09
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DBCT9

Manual Integrations
 APPROVED

mohammad
 2/26/2020 6:48:31 AM

Quant Time: Feb 25 15:13:40 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM022420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 25 11:03:43 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.70	152	2743	0.40	ng/ul	0.00
2) Naphthalene-d8	10.47	136	11031	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.32	164	7331	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.07	188	16684m	0.40	ng/ul	0.00
16) Chrysene-d12	21.25	240	17537	0.40	ng/ul	0.00
20) Perylene-d12	23.45	264	19880	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.06	152	5034	0.24	ng/ul	0.00
14) Fluoranthene-d10	19.09	212	14609	0.26	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	10.52	128	33222	0.989	ng/ul#	83
5) 2-Methylnaphthalene	12.14	142	584	0.024	ng/ul	92
7) Acenaphthylene	14.04	152	2734	0.065	ng/ul#	79
8) Acenaphthene	14.39	153	514480	18.116	ng/ul	97
9) Fluorene	15.38	166	510083	14.314	ng/ul	95
12) Phenanthrene	17.11	178	34400	0.586	ng/ul	96
13) Anthracene	17.20	178	2229	0.039	ng/ul#	84
15) Fluoranthene	19.13	202	1890	0.022	ng/ul	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM022520\
Data File : BM025043.D
Acq On : 25 Feb 2020 14:37
Operator : CG/JU
Sample : L1486-09
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DBCT9

Manual Integrations
APPROVED

mohammad
2/26/2020 6:48:31 AM

Quant Time: Feb 25 15:13:40 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM022420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 25 11:03:43 2020
Response via : Initial Calibration

