

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM022520\
 Data File : BM025050.D
 Acq On : 25 Feb 2020 18:51
 Operator : CG/JU
 Sample : L1486-16
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DBCW7

Manual Integrations
 APPROVED

mohammad
 2/26/2020 6:49:00 AM

Quant Time: Feb 25 23:59:28 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM022420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 25 11:03:43 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.70	152	2423	0.40	ng/ul	0.00
2) Naphthalene-d8	10.47	136	10164	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.32	164	7085	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.07	188	15847m	0.40	ng/ul	0.00
16) Chrysene-d12	21.25	240	16389	0.40	ng/ul	0.00
20) Perylene-d12	23.45	264	18508	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.06	152	6188	0.32	ng/ul	0.00
14) Fluoranthene-d10	19.09	212	17493	0.32	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	10.52	128	3283	0.106	ng/ul#	90
5) 2-Methylnaphthalene	12.13	142	471	0.021	ng/ul	92
8) Acenaphthene	14.39	153	1088	0.040	ng/ul	98
9) Fluorene	15.37	166	731	0.021	ng/ul#	95
12) Phenanthrene	17.11	178	1849m	0.033	ng/ul	
13) Anthracene	17.20	178	3586	0.066	ng/ul	98
15) Fluoranthene	19.13	202	2561	0.032	ng/ul	92
17) Pyrene	19.49	202	2064	0.028	ng/ul#	90
19) Chrysene	21.29	228	1441	0.020	ng/ul#	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM022520\
Data File : BM025050.D
Acq On : 25 Feb 2020 18:51
Operator : CG/JU
Sample : L1486-16
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DBCW7

Manual Integrations
APPROVED

mohammad
2/26/2020 6:49:00 AM

Quant Time: Feb 25 23:59:28 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM022420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 25 11:03:43 2020
Response via : Initial Calibration

