

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM022520\
 Data File : BM025056.D
 Acq On : 25 Feb 2020 22:28
 Operator : CG/JU
 Sample : L1486-12DL 5X
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DBCW3DL

Quant Time: Feb 26 00:09:31 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM022420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 25 11:03:43 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.70	152	2808	0.40	ng/ul	0.00
2) Naphthalene-d8	10.47	136	10990	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.32	164	6948	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.07	188	15698	0.40	ng/ul	0.00
16) Chrysene-d12	21.25	240	16178	0.40	ng/ul	0.00
20) Perylene-d12	23.45	264	18601	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.06	152	1112	0.05	ng/ul	0.00
14) Fluoranthene-d10	19.09	212	3166	0.06	ng/ul	0.00
Target Compounds				Ovalue		
8) Acenaphthene	14.39	153	3365	0.125	ng/ul	99
9) Fluorene	15.37	166	10346	0.306	ng/ul	98
12) Phenanthrene	17.11	178	5692	0.103	ng/ul	98
13) Anthracene	17.20	178	4448	0.083	ng/ul	99
15) Fluoranthene	19.12	202	28342	0.353	ng/ul	98
17) Pyrene	19.48	202	16564	0.227	ng/ul	99
18) Benzo(a)anthracene	21.23	228	2126	0.030	ng/ul	94
19) Chrysene	21.28	228	2381	0.034	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM022520\
Data File : BM025056.D
Acq On : 25 Feb 2020 22:28
Operator : CG/JU
Sample : L1486-12DL 5X
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DBCW3DL

Quant Time: Feb 26 00:09:31 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM022420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 25 11:03:43 2020
Response via : Initial Calibration

