

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM022520\
 Data File : BM025041.D
 Acq On : 25 Feb 2020 13:25
 Operator : CG/JU
 Sample : PB126774BL
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK74

Manual Integrations
 APPROVED

mohammad
 2/26/2020 6:48:23 AM

Quant Time: Feb 25 14:03:38 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM022420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 25 11:03:43 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.70	152	1929	0.40	ng/ul	0.00
2) Naphthalene-d8	10.47	136	8002	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.33	164	5528	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.07	188	12558m	0.40	ng/ul	0.00
16) Chrysene-d12	21.25	240	13253	0.40	ng/ul	0.00
20) Perylene-d12	23.45	264	14897	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.06	152	5907	0.39	ng/ul	0.00
14) Fluoranthene-d10	19.09	212	16575	0.38	ng/ul	0.00

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM022520\
Data File : BM025041.D
Acq On : 25 Feb 2020 13:25
Operator : CG/JU
Sample : PB126774BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK74

Manual Integrations
APPROVED

mohammad
2/26/2020 6:48:23 AM

Quant Time: Feb 25 14:03:38 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM022420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 25 11:03:43 2020
Response via : Initial Calibration

