

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM022520\
Data File : BM025052.D
Acq On : 25 Feb 2020 20:04
Operator : CG/JU
Sample : L1486-18
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DBCX0

Manual Integrations
APPROVED

mohammad
2/26/2020 6:49:09 AM

Quant Time: Feb 26 00:03:21 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM022420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 25 11:03:43 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.70	152	3178	0.40	ng/ul	0.00
2) Naphthalene-d8	10.47	136	12651	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.33	164	8186	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.07	188	18166m	0.40	ng/ul	0.00
16) Chrysene-d12	21.25	240	19228	0.40	ng/ul	0.00
20) Perylene-d12	23.45	264	22751	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.06	152	6250	0.26	ng/ul	0.00
14) Fluoranthene-d10	19.09	212	17300	0.28	ng/ul	0.00

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM022520\
Data File : BM025052.D
Acq On : 25 Feb 2020 20:04
Operator : CG/JU
Sample : L1486-18
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DBCX0

Manual Integrations
APPROVED

mohammad
2/26/2020 6:49:09 AM

Quant Time: Feb 26 00:03:21 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM022420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 25 11:03:43 2020
Response via : Initial Calibration

