

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM022720\
 Data File : BM025113.D
 Acq On : 27 Feb 2020 21:03
 Operator : CG/JU
 Sample : L1582-04
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DBDJ6

Quant Time: Feb 28 04:56:11 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM022420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 25 11:03:43 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.69	152	3030	0.40	ng/ul	0.00
2) Naphthalene-d8	10.48	136	10858	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.31	164	24607	0.40	ng/ul	-0.02
10) Phenanthrene-d10	17.06	188	19219	0.40	ng/ul	0.00
16) Chrysene-d12	21.24	240	19814	0.40	ng/ul	0.00
20) Perylene-d12	23.44	264	21857	0.40	ng/ul	-0.01
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.06	152	7663	0.38	ng/ul	0.00
14) Fluoranthene-d10	19.09	212	18786	0.28	ng/ul	0.00
Target Compounds						
					Ovalue	
3) Naphthalene	10.52	128	44166405	1336.244	ng/ul#	90
5) 2-Methylnaphthalene	12.15	142	10800133	452.832	ng/ul	98
7) Acenaphthylene	14.03	152	141492	0.997	ng/ul#	90
8) Acenaphthene	14.39	153	9311191	97.681	ng/ul	95
9) Fluorene	15.38	166	5903301	49.353	ng/ul	94
12) Phenanthrene	17.11	178	9738098	144.055	ng/ul	97
13) Anthracene	17.19	178	293769	4.484	ng/ul	96
15) Fluoranthene	19.12	202	770078	7.842	ng/ul	95
17) Pyrene	19.48	202	397804	4.456	ng/ul	96
18) Benzo(a)anthracene	21.23	228	13522	0.158	ng/ul	99
19) Chrysene	21.28	228	15760	0.185	ng/ul	99
21) Benzo(b)fluoranthene	22.79	252	4111	0.048	ng/ul	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM022720\
Data File : BM025113.D
Acq On : 27 Feb 2020 21:03
Operator : CG/JU
Sample : L1582-04
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DBDJ6

Quant Time: Feb 28 04:56:11 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM022420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 25 11:03:43 2020
Response via : Initial Calibration

