

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM022820\
 Data File : BM025156.D
 Acq On : 29 Feb 2020 05:32
 Operator : CG/JU
 Sample : L1506-07DL 5X
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DBCY5DL

Manual Integrations
 APPROVED

Sohil
 3/3/2020 8:39:29 AM

Quant Time: Mar 02 01:00:03 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM022420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 02 00:36:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.69	152	3097	0.40	ng/ul	0.00
2) Naphthalene-d8	10.46	136	12047	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.31	164	7336	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.05	188	17179	0.40	ng/ul	0.00
16) Chrysene-d12	21.24	240	17775	0.40	ng/ul	0.00
20) Perylene-d12	23.43	264	18865	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.05	152	362	0.02	ng/ul	0.00
14) Fluoranthene-d10	19.08	212	1008	0.02	ng/ul	0.00
Target Compounds				Ovalue		
12) Phenanthrene	17.09	178	26275	0.435	ng/ul	96
13) Anthracene	17.18	178	13115	0.224	ng/ul	96
15) Fluoranthene	19.11	202	87011	0.991	ng/ul	94
17) Pyrene	19.47	202	73087	0.913	ng/ul	96
18) Benzo(a)anthracene	21.22	228	32918	0.429	ng/ul	99
19) Chrysene	21.27	228	30279	0.396	ng/ul	98
21) Benzo(b)fluoranthene	22.78	252	26275m	0.355	ng/ul	
22) Benzo(k)fluoranthene	22.82	252	9602m	0.131	ng/ul	
23) Benzo(a)pyrene	23.34	252	14598	0.216	ng/ul#	94
24) Indeno(1,2,3-cd)pyrene	25.62	276	6127	0.072	ng/ul#	100
25) Dibenzo(a,h)anthracene	25.63	278	1833	0.026	ng/ul#	59
26) Benzo(g,h,i)perylene	26.29	276	3898	0.055	ng/ul	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM022820\
Data File : BM025156.D
Acq On : 29 Feb 2020 05:32
Operator : CG/JU
Sample : L1506-07DL 5X
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DBCY5DL

Manual Integrations
APPROVED

Sohil
3/3/2020 8:39:29 AM

Quant Time: Mar 02 01:00:03 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM022420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 02 00:36:45 2020
Response via : Initial Calibration

