

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022823\
 Data File : BM038812.D
 Acq On : 28 Feb 2023 15:32
 Operator : CG/JU
 Sample : SSTD00533
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD005040

Quant Time: Mar 01 01:33:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 01 01:29:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.010	152	255800	20.000	ng/ul	0.00
20) Naphthalene-d8	10.828	136	1101404	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.645	164	687527	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.392	188	1533309	20.000	ng/ul	0.00
79) Chrysene-d12	21.568	240	1527564	20.000	ng/ul	0.00
88) Perylene-d12	24.027	264	1768679	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.381	96	13487	2.810	ng/uL	0.00
4) Pyridine-d5	3.810	84	71340	5.705	ng/ul	0.00
7) Phenol-d5	7.157	99	97407	5.854	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.340	67	70181	7.673	ng/ul	0.00
11) 2-Chlorophenol-d4	7.540	132	76978	5.038	ng/ul	0.00
15) 4-Methylphenol-d8	8.716	113	75481	5.058	ng/ul	0.00
21) Nitrobenzene-d5	9.175	128	30418	3.717	ng/ul	0.00
24) 2-Nitrophenol-d4	9.904	143	25401	2.290	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.439	165	70622	3.169	ng/ul	0.00
31) 4-Chloroaniline-d4	10.957	131	117234	4.549	ng/ul	0.00
46) Dimethylphthalate-d6	14.057	166	243407	4.461	ng/ul	0.00
49) Acenaphthylene-d8	14.345	160	270879	4.456	ng/ul	0.00
54) 4-Nitrophenol-d4	14.821	143	34947	3.934	ng/ul	0.00
60) Fluorene-d10	15.633	176	214721	4.637	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.745	200	18159	1.940	ng/ul	0.00
73) Anthracene-d10	17.486	188	331735	4.584	ng/ul	0.00
81) Pyrene-d10	19.774	212	374353	4.047	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.868	264	368092	4.019	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.422	88	13870	2.826	ng/uL	91
5) Pyridine	3.828	79	86454	6.878	ng/ul#	89
6) Benzaldehyde	7.145	77	62598	9.430	ng/ul	97
8) Phenol	7.187	94	109233	6.571	ng/ul	99
10) Bis(2-Chloroethyl)ether	7.434	93	92511	6.935	ng/ul	98
12) 2-Chlorophenol	7.569	128	80126	5.397	ng/ul	99
13) 2-Methylphenol	8.451	108	76930	5.536	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.475	45	154	0.013	ng/ul#	1
16) Acetophenone	8.839	105	141695	5.877	ng/ul	98
17) N-Nitroso-di-n-propyla...	8.822	70	60999	5.727	ng/ul	97
18) 4-Methylphenol	8.775	108	85657	5.582	ng/ul	98
19) Hexachloroethane	9.104	117	32918	5.110	ng/ul	100
22) Nitrobenzene	9.222	77	90065	5.188	ng/ul	96
23) Isophorone	9.745	82	182313	5.397	ng/ul	100
25) 2-Nitrophenol	9.934	139	31767	2.900	ng/ul	98
26) 2,4-Dimethylphenol	9.992	107	93337	5.049	ng/ul	98
27) Bis(2-Chloroethoxy)met...	10.234	93	120374	5.928	ng/ul	99
29) 2,4-Dichlorophenol	10.469	162	73604	3.397	ng/ul	100
30) Naphthalene	10.881	128	292096	5.359	ng/ul	99
32) 4-Chloroaniline	10.986	127	121698	4.820	ng/ul	100
33) Hexachlorobutadiene	11.169	225	50857	4.515	ng/ul	96
34) Caprolactam	11.733	113	16408	3.242	ng/ul	90
35) 4-Chloro-3-methylphenol	12.098	107	80673	4.779	ng/ul	99
36) 2-Methylnaphthalene	12.480	142	184914	4.329	ng/ul	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022823\
 Data File : BM038812.D
 Acq On : 28 Feb 2023 15:32
 Operator : CG/JU
 Sample : SST00533
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST005040


Quant Time: Mar 01 01:33:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 01 01:29:16 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.704	142	187460	4.253	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.845	216	96376	5.965	ng/ul	99
40) Hexachlorocyclopentadiene	12.833	237	41266	4.480	ng/ul#	84
41) 2,4,6-Trichlorophenol	13.080	196	54850	4.216	ng/ul	98
42) 2,4,5-Trichlorophenol	13.145	196	57065	4.024	ng/ul	98
43) 1,1'-Biphenyl	13.486	154	253728	5.007	ng/ul	98
44) 2-Chloronaphthalene	13.527	162	198657	4.828	ng/ul	99
45) 2-Nitroaniline	13.727	65	32018	4.083	ng/ul	93
47) Dimethylphthalate	14.104	163	248880	4.605	ng/ul	100
48) 2,6-Dinitrotoluene	14.222	165	27835	2.596	ng/ul	96
50) Acenaphthylene	14.369	152	313876	4.977	ng/ul	100
51) 3-Nitroaniline	14.545	138	33549	3.789	ng/ul#	99
52) Acenaphthene	14.710	153	223414	5.006	ng/ul	99
53) 2,4-Dinitrophenol	14.745	184	10571	1.602	ng/ul	96
55) 4-Nitrophenol	14.833	109	33487	4.330	ng/ul	94
56) Dibenzofuran	15.045	168	309353	4.861	ng/ul	98
57) 2,4-Dinitrotoluene	14.998	165	44438	2.869	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.263	232	51202	4.772	ng/ul	96
59) Diethylphthalate	15.468	149	233240	4.321	ng/ul	99
61) Fluorene	15.692	166	254703	4.796	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.686	204	125999	5.487	ng/ul	97
63) 4-Nitroaniline	15.698	138	36716	5.374	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.757	198	20295	2.237	ng/ul#	92
67) N-Nitrosodiphenylamine	15.898	169	204322	4.501	ng/ul	98
68) 4-Bromophenyl-phenylether	16.580	248	70571	4.913	ng/ul	97
69) Hexachlorobenzene	16.692	284	86533	4.727	ng/ul	98
70) Atrazine	16.845	200	57180	3.472	ng/ul	98
71) Pentachlorophenol	17.033	266	38892	3.449	ng/ul	95
72) Phenanthrene	17.433	178	409362	4.763	ng/ul	98
74) Anthracene	17.521	178	402253	4.600	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.451	216	101949	6.140	ng/uL	100
76) Pentachlorobenzene	14.963	250	103560	5.500	ng/uL	97
77) Carbazole	17.792	167	358461	4.434	ng/ul	99
78) Di-n-butylphthalate	18.368	149	316525	3.017	ng/ul	98
80) Fluoranthene	19.445	202	435224	3.734	ng/ul	100
82) Pyrene	19.803	202	490731	4.022	ng/ul	100
83) Butylbenzylphthalate	20.709	149	79739	1.318	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.486	252	98071	2.693	ng/ul	96
85) Benzo(a)anthracene	21.550	228	473738	4.468	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.492	149	143880	1.602	ng/ul	98
87) Chrysene	21.603	228	476436	4.864	ng/ul	99
89) Di-n-octyl phthalate	22.445	149	229164	1.261	ng/ul	100
90) Benzo(b)fluoranthene	23.274	252	465579	4.027	ng/ul	99
91) Benzo(k)fluoranthene	23.321	252	493356	4.317	ng/ul	98
93) Benzo(a)pyrene	23.915	252	430437	4.448	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.585	276	557434	4.543	ng/ul	98
95) Dibenzo(a,h)anthracene	26.609	278	468258	4.561	ng/ul	99
96) Benzo(g,h,i)perylene	27.368	276	461279	4.777	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
BNA_M
ClientSampleId :
SSTD005040

Instrument :
BNA_M
ClientSampleId :
SSTD005040









#2
1,4-Dioxane
Concen: 2.826 ng/uL
RT: 3.422 min Scan# 51
Delta R.T. 0.006 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion: 88 Resp: 13870
Ion Ratio Lower Upper
88 100
43 26.0 21.5 32.3
58 61.1 56.4 84.6



#5
Pyridine
Concen: 6.878 ng/uL
RT: 3.828 min Scan# 126
Delta R.T. 0.006 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32

Tgt Ion: 79 Resp: 86454
Ion Ratio Lower Upper
79 100
52 77.4 57.8 86.6
51 47.1 27.8 41.6#









#6
Benzaldehyde
Concen: 9.430 ng/ul
RT: 7.145 min Scan# 690
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion: 77 Resp: 62598
Ion Ratio Lower Upper
77 100
105 94.4 77.9 116.9
106 94.4 78.3 117.5



#8
Phenol
Concen: 6.571 ng/ul
RT: 7.187 min Scan# 697
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32

Tgt Ion: 94 Resp: 109233
Ion Ratio Lower Upper
94 100
65 27.7 22.1 33.1
66 37.0 28.6 43.0









#10
 Bis(2-Chloroethyl)ether
 Concen: 6.935 ng/ul
 RT: 7.434 min Scan# 739
 Delta R.T. -0.000 min
 Lab File: BM038812.D
 Acq: 28 Feb 2023 15:32




Instrument :
 BNA_M
 ClientSampleId :
 SSTD005040


Tgt Ion: 93	Resp:	92511
Ion	Ratio	Lower Upper
93	100	
63	75.1	60.3 90.5
95	29.5	25.8 38.6



#12
 2-Chlorophenol
 Concen: 5.397 ng/ul
 RT: 7.569 min Scan# 762
 Delta R.T. -0.000 min
 Lab File: BM038812.D
 Acq: 28 Feb 2023 15:32

Tgt Ion: 128	Resp:	80126
Ion	Ratio	Lower Upper
128	100	
64	48.2	38.9 58.3
130	31.9	25.1 37.7





#13

2-Methylphenol

Concen: 5.536 ng/ul

RT: 8.451 min Scan# 912

Delta R.T. -0.000 min

Lab File: BM038812.D

Acq: 28 Feb 2023 15:32

Instrument :

BNA_M

ClientSampleId :





SSTD005040

Tgt Ion:108 Resp: 76930

Ion Ratio Lower Upper

108 100

107 84.8 69.0 103.6



#14

2,2'-oxybis(1-Chloropropane)

Concen: 0.013 ng/ul

RT: 8.475 min Scan# 916

Delta R.T. -0.077 min

Lab File: BM038812.D

Acq: 28 Feb 2023 15:32




Tgt Ion: 45 Resp: 154


Ion Ratio Lower Upper

45 100

77 518.2 12.4 18.6#

79 554.5 9.4 14.0#









#16
Acetophenone
Concen: 5.877 ng/ul
RT: 8.839 min Scan# 91
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:105 Resp: 141695
Ion Ratio Lower Upper
105 100
77 77.1 62.6 93.8
51 28.4 20.8 31.2



#17
N-Nitroso-di-n-propylamine
Concen: 5.727 ng/ul
RT: 8.822 min Scan# 975
Delta R.T. -0.006 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32

Tgt Ion: 70 Resp: 60999
Ion Ratio Lower Upper
70 100
42 56.8 43.5 65.3
101 8.8 7.2 10.8
130 18.4 16.0 24.0





#18

4-Methylphenol

Concen: 5.582 ng/ul

RT: 8.775 min Scan# 968

Delta R.T. -0.006 min

Lab File: BM038812.D

Acq: 28 Feb 2023 15:32

Instrument :

BNA_M

ClientSampleId :





SSTD005040

Tgt Ion:108 Resp: 85657

Ion Ratio Lower Upper

108 100

107 112.9 89.0 133.6



#19

Hexachloroethane

Concen: 5.110 ng/ul

RT: 9.104 min Scan# 1023

Delta R.T. -0.000 min

Lab File: BM038812.D

Acq: 28 Feb 2023 15:32




Tgt Ion:117 Resp: 32918


Ion Ratio Lower Upper

117 100

201 95.5 76.6 115.0

199 58.4 46.6 70.0









#22
Nitrobenzene
Concen: 5.188 ng/ul
RT: 9.222 min Scan# 1043
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion: 77 Resp: 90065
Ion Ratio Lower Upper
77 100
123 41.4 35.3 52.9
65 12.5 10.6 15.8



#23
Isophorone
Concen: 5.397 ng/ul
RT: 9.745 min Scan# 1132
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32

Tgt Ion: 82 Resp: 182313
Ion Ratio Lower Upper
82 100
95 6.5 5.2 7.8
138 16.1 13.1 19.7









#25
2-Nitrophenol
Concen: 2.900 ng/ul
RT: 9.934 min Scan# 1164
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:139 Resp: 31767
Ion Ratio Lower Upper
139 100
65 55.3 42.6 63.8
109 28.5 23.1 34.7



#26
2,4-Dimethylphenol
Concen: 5.049 ng/ul
RT: 9.992 min Scan# 1174
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32

Tgt Ion:107 Resp: 93337
Ion Ratio Lower Upper
107 100
121 49.3 39.8 59.8
122 83.4 69.0 103.4









#27
 Bis(2-Chloroethoxy)methane
 Concen: 5.928 ng/ul
 RT: 10.234 min Scan# 1215
 Delta R.T. -0.000 min
 Lab File: BM038812.D
 Acq: 28 Feb 2023 15:32




Instrument :
 BNA_M
 ClientSampleId :
 SSTD005040


Tgt Ion: 93 Resp: 120374
 Ion Ratio Lower Upper
 93 100
 95 31.9 25.3 37.9
 123 11.0 9.3 13.9



#29
 2,4-Dichlorophenol
 Concen: 3.397 ng/ul
 RT: 10.469 min Scan# 1255
 Delta R.T. -0.000 min
 Lab File: BM038812.D
 Acq: 28 Feb 2023 15:32

Tgt Ion: 162 Resp: 73604
 Ion Ratio Lower Upper
 162 100
 164 65.6 52.6 78.8
 98 35.9 29.0 43.6









#30
Naphthalene
Concen: 5.359 ng/ul
RT: 10.881 min Scan# 1325
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:128 Resp: 292096
Ion Ratio Lower Upper
128 100
129 10.3 8.6 12.8
127 12.5 10.2 15.4



#32
4-Chloroaniline
Concen: 4.820 ng/ul
RT: 10.986 min Scan# 1343
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32

Tgt Ion:127 Resp: 121698
Ion Ratio Lower Upper
127 100
129 32.8 26.3 39.5





#33

Hexachlorobutadiene

Concen: 4.515 ng/ul

RT: 11.169 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM038812.D

Acq: 28 Feb 2023 15:32

Instrument :

BNA_M

ClientSampleId :

SSTD005040





Tgt Ion:225 Resp: 50857

Ion Ratio Lower Upper

225 100

223 59.6 50.9 76.3

227 64.7 50.4 75.6



#34

Caprolactam

Concen: 3.242 ng/ul

RT: 11.733 min Scan# 1470

Delta R.T. -0.006 min

Lab File: BM038812.D

Acq: 28 Feb 2023 15:32




Tgt Ion:113 Resp: 16408


Ion Ratio Lower Upper

113 100

55 158.1 137.0 205.6

56 137.5 99.9 149.9









#35
 4-Chloro-3-methylphenol
 Concen: 4.779 ng/ul
 RT: 12.098 min Scan# 1532
 Delta R.T. -0.000 min
 Lab File: BM038812.D
 Acq: 28 Feb 2023 15:32




Instrument :
 BNA_M
 ClientSampleId :
 SSTD005040


Tgt Ion:107 Resp: 80673
 Ion Ratio Lower Upper
 107 100
 144 25.8 21.1 31.7
 142 81.3 64.7 97.1



#36
 2-Methylnaphthalene
 Concen: 4.329 ng/ul
 RT: 12.480 min Scan# 1597
 Delta R.T. -0.006 min
 Lab File: BM038812.D
 Acq: 28 Feb 2023 15:32

Tgt Ion:142 Resp: 184914
 Ion Ratio Lower Upper
 142 100
 141 86.5 71.0 106.6









#37
1-Methylnaphthalene
Concen: 4.253 ng/ul
RT: 12.704 min Scan# 1635
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:142 Resp: 187460
Ion Ratio Lower Upper
142 100
141 90.7 73.4 110.0
116 3.9 3.2 4.8



#39
1,2,4,5-Tetrachlorobenzene
Concen: 5.965 ng/ul
RT: 12.845 min Scan# 1659
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32

Tgt Ion:216 Resp: 96376
Ion Ratio Lower Upper
216 100
214 78.3 62.2 93.4
179 20.7 16.4 24.6
108 18.3 15.3 22.9





#40

Hexachlorocyclopentadiene

Concen: 4.480 ng/ul

RT: 12.833 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM038812.D

Acq: 28 Feb 2023 15:32

Instrument :

BNA_M

ClientSampleId :

SSTD005040





Tgt Ion:237 Resp: 41266

Ion Ratio Lower Upper

237 100

235 74.2 48.1 72.1#

272 14.3 10.0 15.0



#41

2,4,6-Trichlorophenol

Concen: 4.216 ng/ul

RT: 13.080 min Scan# 1699

Delta R.T. -0.000 min

Lab File: BM038812.D

Acq: 28 Feb 2023 15:32




Tgt Ion:196 Resp: 54850


Ion Ratio Lower Upper

196 100

198 94.3 74.8 112.2

200 27.6 24.3 36.5





#42

2,4,5-Trichlorophenol

Concen: 4.024 ng/ul

RT: 13.145 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM038812.D

Acq: 28 Feb 2023 15:32

Instrument :

BNA_M

ClientSampleId :

SSTD005040





Tgt Ion:196 Resp: 57065

Ion Ratio Lower Upper

196 100

198 98.3 77.5 116.3

200 29.1 24.8 37.2



#43

1,1'-Biphenyl

Concen: 5.007 ng/ul

RT: 13.486 min Scan# 1768

Delta R.T. -0.000 min

Lab File: BM038812.D

Acq: 28 Feb 2023 15:32




Tgt Ion:154 Resp: 253728


Ion Ratio Lower Upper

154 100

153 40.7 31.3 46.9

76 13.7 11.0 16.6





#44

2-Chloronaphthalene

Concen: 4.828 ng/ul

RT: 13.527 min Scan# 11

Delta R.T. -0.000 min

Lab File: BM038812.D

Acq: 28 Feb 2023 15:32

Instrument :

BNA_M

ClientSampleId :

SSTD005040





Tgt Ion:162 Resp: 198657

Ion Ratio Lower Upper

162 100

164 32.1 26.6 40.0

127 39.0 31.5 47.3



#45

2-Nitroaniline

Concen: 4.083 ng/ul

RT: 13.727 min Scan# 1809

Delta R.T. -0.000 min

Lab File: BM038812.D

Acq: 28 Feb 2023 15:32




Tgt Ion: 65 Resp: 32018


Ion Ratio Lower Upper

65 100

92 63.6 54.2 81.2

138 97.2 84.3 126.5





#47

Dimethylphthalate

Concen: 4.605 ng/ul

RT: 14.104 min Scan# 1873

Delta R.T. -0.000 min

Lab File: BM038812.D

Acq: 28 Feb 2023 15:32

Instrument :

BNA_M

ClientSampleId :

SSTD005040





Tgt Ion:163 Resp: 248880

Ion Ratio Lower Upper

163 100

194 3.9 3.0 4.6

164 10.3 8.4 12.6



#48

2,6-Dinitrotoluene

Concen: 2.596 ng/ul

RT: 14.222 min Scan# 1893

Delta R.T. -0.000 min

Lab File: BM038812.D

Acq: 28 Feb 2023 15:32




Tgt Ion:165 Resp: 27835


Ion Ratio Lower Upper

165 100

89 69.0 51.4 77.2

121 24.3 19.4 29.2









#50
Acenaphthylene
Concen: 4.977 ng/ul
RT: 14.369 min Scan# 1918
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:152 Resp: 313876
Ion Ratio Lower Upper
152 100
151 19.6 15.9 23.9
153 13.0 10.4 15.6



#51
3-Nitroaniline
Concen: 3.789 ng/ul
RT: 14.545 min Scan# 1948
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32

Tgt Ion:138 Resp: 33549
Ion Ratio Lower Upper
138 100
108 15.8 10.3 15.5#
92 122.9 97.7 146.5









#52
Acenaphthene
Concen: 5.006 ng/ul
RT: 14.710 min Scan# 1976
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:153 Resp: 223414
Ion Ratio Lower Upper
153 100
152 49.2 39.1 58.7
154 88.5 70.4 105.6



#53
2,4-Dinitrophenol
Concen: 1.602 ng/ul
RT: 14.745 min Scan# 1982
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32

Tgt Ion:184 Resp: 10571
Ion Ratio Lower Upper
184 100
63 71.8 61.9 92.9
154 62.3 50.7 76.1









#55
4-Nitrophenol
Concen: 4.330 ng/ul
RT: 14.833 min Scan# 1997
Delta R.T. -0.006 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:109 Resp: 33487
Ion Ratio Lower Upper
109 100
139 123.8 108.8 163.2
65 137.6 107.8 161.6



#56
Dibenzofuran
Concen: 4.861 ng/ul
RT: 15.045 min Scan# 2033
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32

Tgt Ion:168 Resp: 309353
Ion Ratio Lower Upper
168 100
139 37.6 30.9 46.3









#57
2,4-Dinitrotoluene
Concen: 2.869 ng/ul
RT: 14.998 min Scan# 2025
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:165 Resp: 44438
Ion Ratio Lower Upper
165 100
63 51.5 38.9 58.3
182 2.6 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 4.772 ng/ul
RT: 15.263 min Scan# 2070
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32

Tgt Ion:232 Resp: 51202
Ion Ratio Lower Upper
232 100
131 48.4 40.6 60.8
130 2.6 2.0 3.0
166 35.2 25.9 38.9









#59
Diethylphthalate
Concen: 4.321 ng/ul
RT: 15.468 min Scan# 2105
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:149 Resp: 233240
Ion Ratio Lower Upper
149 100
177 22.2 17.2 25.8
150 12.4 10.1 15.1



#61
Fluorene
Concen: 4.796 ng/ul
RT: 15.692 min Scan# 2143
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32

Tgt Ion:166 Resp: 254703
Ion Ratio Lower Upper
166 100
165 96.9 78.2 117.4
167 13.7 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 5.487 ng/ul
RT: 15.686 min Scan# 2142
Delta R.T. -0.006 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:204 Resp: 125999
Ion Ratio Lower Upper
204 100
206 31.2 26.3 39.5
141 59.1 45.5 68.3



#63
4-Nitroaniline
Concen: 5.374 ng/ul
RT: 15.698 min Scan# 2144
Delta R.T. -0.006 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32

Tgt Ion:138 Resp: 36716
Ion Ratio Lower Upper
138 100
92 59.0 41.8 62.8
108 82.2 65.0 97.6









#66
4,6-Dinitro-2-methylphenol
Concen: 2.237 ng/ul
RT: 15.757 min Scan# 2154
Delta R.T. -0.006 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:198 Resp: 20295
Ion Ratio Lower Upper
198 100
182 9.2 3.5 5.3#
77 29.7 21.0 31.6



#67
N-Nitrosodiphenylamine
Concen: 4.501 ng/ul
RT: 15.898 min Scan# 2178
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32

Tgt Ion:169 Resp: 204322
Ion Ratio Lower Upper
169 100
168 68.0 53.5 80.3
167 34.8 29.1 43.7









#68
4-Bromophenyl-phenylether
Concen: 4.913 ng/ul
RT: 16.580 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:248 Resp: 70571
Ion Ratio Lower Upper
248 100
250 97.4 77.6 116.4
141 74.1 63.7 95.5



#69
Hexachlorobenzene
Concen: 4.727 ng/ul
RT: 16.692 min Scan# 2313
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32

Tgt Ion:284 Resp: 86533
Ion Ratio Lower Upper
284 100
142 37.5 29.4 44.2
249 31.9 24.0 36.0









#70
Atrazine
Concen: 3.472 ng/ul
RT: 16.845 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:200 Resp: 57180
Ion Ratio Lower Upper
200 100
173 26.2 21.3 31.9
215 49.0 38.0 57.0



#71
Pentachlorophenol
Concen: 3.449 ng/ul
RT: 17.033 min Scan# 2371
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32

Tgt Ion:266 Resp: 38892
Ion Ratio Lower Upper
266 100
264 70.5 51.0 76.6
268 63.6 51.7 77.5









#72
Phenanthrene
Concen: 4.763 ng/ul
RT: 17.433 min Scan# 2439
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:178 Resp: 409362
Ion Ratio Lower Upper
178 100
179 14.9 12.3 18.5
176 18.3 15.7 23.5



#74
Anthracene
Concen: 4.600 ng/ul
RT: 17.521 min Scan# 2454
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32

Tgt Ion:178 Resp: 402253
Ion Ratio Lower Upper
178 100
179 15.2 12.5 18.7
176 18.2 14.9 22.3









#75
1,2,3,4-Tetrachlorobenzene
Concen: 6.140 ng/uL
RT: 13.451 min Scan# 1762
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:216 Resp: 101949
Ion Ratio Lower Upper
216 100
214 77.6 62.4 93.6
218 48.2 38.8 58.2



#76
Pentachlorobenzene
Concen: 5.500 ng/uL
RT: 14.963 min Scan# 2019
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32

Tgt Ion:250 Resp: 103560
Ion Ratio Lower Upper
250 100
248 63.8 49.6 74.4
252 61.6 51.2 76.8









#77
 Carbazole
 Concen: 4.434 ng/ul
 RT: 17.792 min Scan# 2500
 Delta R.T. -0.000 min
 Lab File: BM038812.D
 Acq: 28 Feb 2023 15:32




Instrument :
 BNA_M
 ClientSampleId :
 SSTD005040


Tgt Ion:167 Resp: 358461
 Ion Ratio Lower Upper
 167 100
 166 21.1 17.2 25.8
 139 12.9 10.6 16.0



#78
 Di-n-butylphthalate
 Concen: 3.017 ng/ul
 RT: 18.368 min Scan# 2598
 Delta R.T. -0.000 min
 Lab File: BM038812.D
 Acq: 28 Feb 2023 15:32

Tgt Ion:149 Resp: 316525
 Ion Ratio Lower Upper
 149 100
 150 8.4 7.3 10.9
 104 6.1 4.2 6.4









#80
Fluoranthene
Concen: 3.734 ng/ul
RT: 19.445 min Scan# 2781
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:202 Resp: 435224
Ion Ratio Lower Upper
202 100
101 11.2 9.0 13.4
100 8.7 6.9 10.3



#82
Pyrene
Concen: 4.022 ng/ul
RT: 19.803 min Scan# 2842
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32

Tgt Ion:202 Resp: 490731
Ion Ratio Lower Upper
202 100
101 14.0 11.2 16.8
100 11.1 9.1 13.7









#83
Butylbenzylphthalate
Concen: 1.318 ng/ul
RT: 20.709 min Scan# 2996
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:149 Resp: 79739
Ion Ratio Lower Upper
149 100
91 78.3 60.5 90.7
206 20.2 15.8 23.8



#84
3,3'-Dichlorobenzidine
Concen: 2.693 ng/ul
RT: 21.486 min Scan# 3128
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32

Tgt Ion:252 Resp: 98071
Ion Ratio Lower Upper
252 100
254 67.5 51.8 77.6
126 12.6 9.0 13.4









#85
Benzo(a)anthracene
Concen: 4.468 ng/ul
RT: 21.550 min Scan# 3139
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:228 Resp: 473738
Ion Ratio Lower Upper
228 100
229 19.6 15.8 23.6
226 27.3 21.4 32.2



#86
Bis(2-ethylhexyl)phthalate
Concen: 1.602 ng/ul
RT: 21.492 min Scan# 3129
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32

Tgt Ion:149 Resp: 143880
Ion Ratio Lower Upper
149 100
167 27.7 21.4 32.2
279 3.4 3.1 4.7









#87
Chrysene
Concen: 4.864 ng/ul
RT: 21.603 min Scan# 3149
Delta R.T. -0.006 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:228 Resp: 476436
Ion Ratio Lower Upper
228 100
226 29.7 23.6 35.4
229 19.3 15.8 23.6



#89
Di-n-octyl phthalate
Concen: 1.261 ng/ul
RT: 22.445 min Scan# 3291
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32

Tgt Ion:149 Resp: 229164









#90
Benzo(b)fluoranthene
Concen: 4.027 ng/ul
RT: 23.274 min Scan# 3432
Delta R.T. -0.000 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32




Instrument :
BNA_M
ClientSampleId :
SSTD005040


Tgt Ion:252 Resp: 465579
Ion Ratio Lower Upper
252 100
253 21.8 17.2 25.8
125 10.6 8.6 13.0



#91
Benzo(k)fluoranthene
Concen: 4.317 ng/ul
RT: 23.321 min Scan# 3440
Delta R.T. -0.006 min
Lab File: BM038812.D
Acq: 28 Feb 2023 15:32

Tgt Ion:252 Resp: 493356
Ion Ratio Lower Upper
252 100
253 20.8 17.7 26.5
125 10.1 8.4 12.6





#93

Benzo(a)pyrene

Concen: 4.448 ng/ul

RT: 23.915 min Scan# 3

Delta R.T. -0.000 min

Lab File: BM038812.D

Acq: 28 Feb 2023 15:32

Instrument :

BNA_M

ClientSampleId :

SSTD005040





Tgt Ion:252 Resp: 430437

Ion Ratio Lower Upper

252 100

253 21.8 17.8 26.8

125 11.4 9.4 14.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 4.543 ng/ul

RT: 26.585 min Scan# 3995

Delta R.T. -0.006 min

Lab File: BM038812.D

Acq: 28 Feb 2023 15:32




Tgt Ion:276 Resp: 557434


Ion Ratio Lower Upper

276 100

138 24.2 20.2 30.4

277 24.1 19.8 29.8





#95

Dibenzo(a,h)anthracene

Concen: 4.561 ng/ul

RT: 26.609 min Scan# 3999

Delta R.T. -0.006 min

Lab File: BM038812.D

Acq: 28 Feb 2023 15:32

Instrument :

BNA_M

ClientSampleId :

SSTD005040





Tgt Ion:278 Resp: 468258

Ion Ratio Lower Upper

278 100

139 16.3 13.4 20.2

279 23.1 19.1 28.7



#96

Benzo(g,h,i)perylene

Concen: 4.777 ng/ul

RT: 27.368 min Scan# 4128

Delta R.T. -0.006 min

Lab File: BM038812.D

Acq: 28 Feb 2023 15:32

Tgt Ion:276 Resp: 461279

Ion Ratio Lower Upper

276 100

138 23.1 19.4 29.0

277 23.4 18.7 28.1

