

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022823\
 Data File : BM038813.D
 Acq On : 28 Feb 2023 16:09
 Operator : CG/JU
 Sample : SSTD01034
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD010041

Quant Time: Mar 01 01:33:37 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 01 01:29:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.016	152	180080	20.000	ng/u1	0.00
20) Naphthalene-d8	10.828	136	786376	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.651	164	504197	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.392	188	1120142	20.000	ng/u1	0.00
79) Chrysene-d12	21.568	240	1169306	20.000	ng/u1	0.00
88) Perylene-d12	24.027	264	1344911	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.381	96	18526	5.482	ng/uL	0.00
4) Pyridine-d5	3.805	84	107985	12.266	ng/u1	0.00
7) Phenol-d5	7.158	99	146566	12.512	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.340	67	102876	15.977	ng/u1	0.00
11) 2-Chlorophenol-d4	7.540	132	114378	10.633	ng/u1	0.00
15) 4-Methylphenol-d8	8.716	113	112576	10.717	ng/u1	0.00
21) Nitrobenzene-d5	9.175	128	45941	7.862	ng/u1	0.00
24) 2-Nitrophenol-d4	9.904	143	41426	5.230	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.440	165	108908	6.844	ng/u1	0.00
31) 4-Chloroaniline-d4	10.963	131	180042	9.785	ng/u1	0.00
46) Dimethylphthalate-d6	14.057	166	371266	9.278	ng/u1	0.00
49) Acenaphthylene-d8	14.339	160	421681	9.459	ng/u1	0.00
54) 4-Nitrophenol-d4	14.822	143	59964	9.204	ng/u1	0.00
60) Fluorene-d10	15.639	176	326575	9.617	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.745	200	33079	4.838	ng/u1	0.00
73) Anthracene-d10	17.486	188	513788	9.719	ng/u1	0.00
81) Pyrene-d10	19.774	212	590101	8.333	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.862	264	612413	8.793	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.417	88	20708	5.993	ng/uL	99
5) Pyridine	3.828	79	125261	14.156	ng/u1	96
6) Benzaldehyde	7.146	77	91668	19.615	ng/u1	98
8) Phenol	7.187	94	158728	13.563	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.434	93	135621	14.442	ng/u1	98
12) 2-Chlorophenol	7.569	128	118156	11.305	ng/u1	98
13) 2-Methylphenol	8.452	108	116627	11.921	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.552	45	172164	20.899	ng/u1	99
16) Acetophenone	8.840	105	212737	12.533	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.822	70	95334	12.715	ng/u1	96
18) 4-Methylphenol	8.775	108	127597	11.811	ng/u1	98
19) Hexachloroethane	9.104	117	47901	10.562	ng/u1	95
22) Nitrobenzene	9.222	77	137218	11.070	ng/u1	98
23) Isophorone	9.746	82	284514	11.797	ng/u1	100
25) 2-Nitrophenol	9.934	139	49959	6.387	ng/u1	99
26) 2,4-Dimethylphenol	9.993	107	138685	10.507	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.234	93	181753	12.537	ng/u1	99
29) 2,4-Dichlorophenol	10.469	162	112456	7.270	ng/u1	97
30) Naphthalene	10.881	128	438207	11.261	ng/u1	99
32) 4-Chloroaniline	10.987	127	184383	10.227	ng/u1	100
33) Hexachlorobutadiene	11.169	225	73722	9.168	ng/u1	97
34) Caprolactam	11.734	113	29651	8.206	ng/u1	99
35) 4-Chloro-3-methylphenol	12.098	107	126977	10.534	ng/u1	96
36) 2-Methylnaphthalene	12.487	142	278407	9.129	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022823\
 Data File : BM038813.D
 Acq On : 28 Feb 2023 16:09
 Operator : CG/JU
 Sample : SSTD01034
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SSTD010041


Quant Time: Mar 01 01:33:37 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 01 01:29:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.698	142	282218	8.968	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.845	216	144512	12.196	ng/ul	99
40) Hexachlorocyclopentadiene	12.834	237	62445	9.244	ng/ul	95
41) 2,4,6-Trichlorophenol	13.081	196	85612	8.972	ng/ul	99
42) 2,4,5-Trichlorophenol	13.145	196	95062	9.141	ng/ul	100
43) 1,1'-Biphenyl	13.487	154	384887	10.356	ng/ul	100
44) 2-Chloronaphthalene	13.528	162	300382	9.954	ng/ul	99
45) 2-Nitroaniline	13.728	65	51892	9.023	ng/ul	95
47) Dimethylphthalate	14.104	163	375993	9.487	ng/ul	98
48) 2,6-Dinitrotoluene	14.216	165	46479	5.911	ng/ul	90
50) Acenaphthylene	14.369	152	476586	10.305	ng/ul	99
51) 3-Nitroaniline	14.545	138	57284	8.822	ng/ul#	98
52) Acenaphthene	14.710	153	333056	10.177	ng/ul	99
53) 2,4-Dinitrophenol	14.745	184	18696	3.864	ng/ul	94
55) 4-Nitrophenol	14.834	109	53244	9.388	ng/ul	94
56) Dibenzofuran	15.045	168	465018	9.963	ng/ul	98
57) 2,4-Dinitrotoluene	14.998	165	77583	6.829	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.263	232	82165	10.443	ng/ul#	98
59) Diethylphthalate	15.469	149	364780	9.216	ng/ul	99
61) Fluorene	15.692	166	388075	9.965	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.692	204	187075	11.109	ng/ul	98
63) 4-Nitroaniline	15.698	138	66310	13.236	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.757	198	37418	5.645	ng/ul#	98
67) N-Nitrosodiphenylamine	15.898	169	316227	9.536	ng/ul	99
68) 4-Bromophenyl-phenylether	16.580	248	107228	10.219	ng/ul	99
69) Hexachlorobenzene	16.692	284	133436	9.979	ng/ul	100
70) Atrazine	16.851	200	93327	7.758	ng/ul	97
71) Pentachlorophenol	17.033	266	65529	7.955	ng/ul	95
72) Phenanthrene	17.433	178	622393	9.913	ng/ul	99
74) Anthracene	17.522	178	623738	9.764	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.451	216	150132	12.378	ng/uL	98
76) Pentachlorobenzene	14.963	250	154852	11.258	ng/uL	99
77) Carbazole	17.792	167	562351	9.523	ng/ul	100
78) Di-n-butylphthalate	18.369	149	529222	6.906	ng/ul	100
80) Fluoranthene	19.445	202	687248	7.704	ng/ul	99
82) Pyrene	19.804	202	765759	8.200	ng/ul	100
83) Butylbenzylphthalate	20.710	149	146687	3.167	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.486	252	175199	6.285	ng/ul	98
85) Benzo(a)anthracene	21.551	228	761240	9.379	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.492	149	268439	3.904	ng/ul	100
87) Chrysene	21.604	228	748740	9.985	ng/ul	99
89) Di-n-octyl phthalate	22.445	149	437282	3.166	ng/ul	100
90) Benzo(b)fluoranthene	23.274	252	768144	8.737	ng/ul	99
91) Benzo(k)fluoranthene	23.321	252	793304	9.128	ng/ul	99
93) Benzo(a)pyrene	23.915	252	713386	9.695	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.592	276	942162	10.097	ng/ul	100
95) Dibenzo(a,h)anthracene	26.609	278	795636	10.192	ng/ul	99
96) Benzo(g,h,i)perylene	27.374	276	776875	10.581	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Mar 01 01:33:37 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022823.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Mar 01 01:29:16 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 5.993 ng/uL
RT: 3.417 min Scan# 50
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09




Instrument :
BNA_M
ClientSampleId :
SSTD010041


Tgt Ion: 88 Resp: 20708
Ion Ratio Lower Upper
88 100
43 27.7 21.5 32.3
58 70.1 56.4 84.6



#5
Pyridine
Concen: 14.156 ng/uL
RT: 3.828 min Scan# 126
Delta R.T. 0.006 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09

Tgt Ion: 79 Resp: 125261
Ion Ratio Lower Upper
79 100
52 73.3 57.8 86.6
51 39.7 27.8 41.6









#6
Benzaldehyde
Concen: 19.615 ng/ul
RT: 7.146 min Scan# 690
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09




Instrument :
BNA_M
ClientSampleId :
SSTD010041


Tgt Ion: 77 Resp: 91668
Ion Ratio Lower Upper
77 100
105 99.9 77.9 116.9
106 98.5 78.3 117.5



#8
Phenol
Concen: 13.563 ng/ul
RT: 7.187 min Scan# 697
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09

Tgt Ion: 94 Resp: 158728
Ion Ratio Lower Upper
94 100
65 28.2 22.1 33.1
66 34.6 28.6 43.0









#10
Bis(2-Chloroethyl)ether
Concen: 14.442 ng/ul
RT: 7.434 min Scan# 739
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09




Instrument :
BNA_M
ClientSampleId :
SSTD010041


Tgt Ion: 93 Resp: 135621
Ion Ratio Lower Upper
93 100
63 72.9 60.3 90.5
95 31.5 25.8 38.6



#12
2-Chlorophenol
Concen: 11.305 ng/ul
RT: 7.569 min Scan# 762
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09

Tgt Ion:128 Resp: 118156
Ion Ratio Lower Upper
128 100
64 47.0 38.9 58.3
130 31.1 25.1 37.7





#13

2-Methylphenol

Concen: 11.921 ng/ul

RT: 8.452 min Scan# 912

Delta R.T. -0.000 min

Lab File: BM038813.D

Acq: 28 Feb 2023 16:09

Instrument :

BNA_M

ClientSampleId :





SSTD010041

Tgt Ion:108 Resp: 116627

Ion Ratio Lower Upper

108 100

107 88.8 69.0 103.6



#14

2,2'-oxybis(1-Chloropropane)

Concen: 20.899 ng/ul

RT: 8.552 min Scan# 929

Delta R.T. -0.000 min

Lab File: BM038813.D

Acq: 28 Feb 2023 16:09




Tgt Ion: 45 Resp: 172164


Ion Ratio Lower Upper

45 100

77 16.0 12.4 18.6

79 12.3 9.4 14.0









#16
Acetophenone
Concen: 12.533 ng/ul
RT: 8.840 min Scan# 91
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09




Instrument :
BNA_M
ClientSampleId :
SSTD010041


Tgt Ion:105 Resp: 212737
Ion Ratio Lower Upper
105 100
77 77.0 62.6 93.8
51 27.2 20.8 31.2



#17
N-Nitroso-di-n-propylamine
Concen: 12.715 ng/ul
RT: 8.822 min Scan# 975
Delta R.T. -0.006 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09

Tgt Ion: 70 Resp: 95334
Ion Ratio Lower Upper
70 100
42 57.6 43.5 65.3
101 9.5 7.2 10.8
130 18.5 16.0 24.0





#18

4-Methylphenol

Concen: 11.811 ng/ul

RT: 8.775 min Scan# 968

Delta R.T. -0.006 min

Lab File: BM038813.D

Acq: 28 Feb 2023 16:09

Instrument :

BNA_M

ClientSampleId :





SSTD010041

Tgt Ion:108 Resp: 127597

Ion Ratio Lower Upper

108 100

107 113.1 89.0 133.6



#19

Hexachloroethane

Concen: 10.562 ng/ul

RT: 9.104 min Scan# 1023

Delta R.T. -0.000 min

Lab File: BM038813.D

Acq: 28 Feb 2023 16:09




Tgt Ion:117 Resp: 47901


Ion Ratio Lower Upper

117 100

201 100.3 76.6 115.0

199 61.8 46.6 70.0









#22
Nitrobenzene
Concen: 11.070 ng/ul
RT: 9.222 min Scan# 1043
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09




Instrument :
BNA_M
ClientSampleId :
SSTD010041


Tgt Ion: 77 Resp: 137218
Ion Ratio Lower Upper
77 100
123 43.1 35.3 52.9
65 12.0 10.6 15.8



#23
Isophorone
Concen: 11.797 ng/ul
RT: 9.746 min Scan# 1132
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09

Tgt Ion: 82 Resp: 284514
Ion Ratio Lower Upper
82 100
95 6.5 5.2 7.8
138 16.7 13.1 19.7





#25

2-Nitrophenol

Concen: 6.387 ng/ul

RT: 9.934 min Scan# 1164

Delta R.T. -0.000 min

Lab File: BM038813.D

Acq: 28 Feb 2023 16:09

Instrument :

BNA_M

ClientSampleId :

SSTD010041





Tgt Ion:139 Resp: 49959

Ion Ratio Lower Upper

139 100

65 53.3 42.6 63.8

109 29.7 23.1 34.7



#26

2,4-Dimethylphenol

Concen: 10.507 ng/ul

RT: 9.993 min Scan# 1174

Delta R.T. -0.000 min

Lab File: BM038813.D

Acq: 28 Feb 2023 16:09




Tgt Ion:107 Resp: 138685


Ion Ratio Lower Upper

107 100

121 50.8 39.8 59.8

122 88.0 69.0 103.4









#27
 Bis(2-Chloroethoxy)methane
 Concen: 12.537 ng/ul
 RT: 10.234 min Scan# 1215
 Delta R.T. -0.000 min
 Lab File: BM038813.D
 Acq: 28 Feb 2023 16:09




Instrument :
 BNA_M
 ClientSampleId :
 SSTD010041


Tgt Ion: 93 Resp: 181753
 Ion Ratio Lower Upper
 93 100
 95 32.2 25.3 37.9
 123 11.3 9.3 13.9



#29
 2,4-Dichlorophenol
 Concen: 7.270 ng/ul
 RT: 10.469 min Scan# 1255
 Delta R.T. -0.000 min
 Lab File: BM038813.D
 Acq: 28 Feb 2023 16:09

Tgt Ion: 162 Resp: 112456
 Ion Ratio Lower Upper
 162 100
 164 68.6 52.6 78.8
 98 35.5 29.0 43.6









#30
Naphthalene
Concen: 11.261 ng/ul
RT: 10.881 min Scan# 1325
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09




Instrument :
BNA_M
ClientSampleId :
SSTD010041


Tgt Ion:128 Resp: 438207
Ion Ratio Lower Upper
128 100
129 10.5 8.6 12.8
127 12.5 10.2 15.4



#32
4-Chloroaniline
Concen: 10.227 ng/ul
RT: 10.987 min Scan# 1343
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09

Tgt Ion:127 Resp: 184383
Ion Ratio Lower Upper
127 100
129 32.6 26.3 39.5









#33
Hexachlorobutadiene
Concen: 9.168 ng/ul
RT: 11.169 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09




Instrument :
BNA_M
ClientSampleId :
SSTD010041


Tgt Ion:225 Resp: 73722
Ion Ratio Lower Upper
225 100
223 61.1 50.9 76.3
227 60.8 50.4 75.6



#34
Caprolactam
Concen: 8.206 ng/ul
RT: 11.734 min Scan# 1470
Delta R.T. -0.006 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09

Tgt Ion:113 Resp: 29651
Ion Ratio Lower Upper
113 100
55 172.9 137.0 205.6
56 126.3 99.9 149.9









#35
 4-Chloro-3-methylphenol
 Concen: 10.534 ng/ul
 RT: 12.098 min Scan# 1532
 Delta R.T. -0.000 min
 Lab File: BM038813.D
 Acq: 28 Feb 2023 16:09




Instrument :
 BNA_M
 ClientSampleId :
 SSTD010041


Tgt Ion:107 Resp: 126977
 Ion Ratio Lower Upper
 107 100
 144 25.4 21.1 31.7
 142 77.3 64.7 97.1



#36
 2-Methylnaphthalene
 Concen: 9.129 ng/ul
 RT: 12.487 min Scan# 1598
 Delta R.T. -0.000 min
 Lab File: BM038813.D
 Acq: 28 Feb 2023 16:09

Tgt Ion:142 Resp: 278407
 Ion Ratio Lower Upper
 142 100
 141 89.7 71.0 106.6









#37
1-Methylnaphthalene
Concen: 8.968 ng/ul
RT: 12.698 min Scan# 1634
Delta R.T. -0.006 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09




Instrument :
BNA_M
ClientSampleId :
SSTD010041


Tgt Ion:142 Resp: 282218
Ion Ratio Lower Upper
142 100
141 91.5 73.4 110.0
116 3.8 3.2 4.8



#39
1,2,4,5-Tetrachlorobenzene
Concen: 12.196 ng/ul
RT: 12.845 min Scan# 1659
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09

Tgt Ion:216 Resp: 144512
Ion Ratio Lower Upper
216 100
214 77.5 62.2 93.4
179 21.4 16.4 24.6
108 20.7 15.3 22.9









#40
Hexachlorocyclopentadiene
Concen: 9.244 ng/ul
RT: 12.834 min Scan# 1657
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09




Instrument :
BNA_M
ClientSampleId :
SSTD010041


Tgt Ion:237 Resp: 62445
Ion Ratio Lower Upper
237 100
235 64.2 48.1 72.1
272 13.7 10.0 15.0



#41
2,4,6-Trichlorophenol
Concen: 8.972 ng/ul
RT: 13.081 min Scan# 1699
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09

Tgt Ion:196 Resp: 85612
Ion Ratio Lower Upper
196 100
198 92.0 74.8 112.2
200 30.0 24.3 36.5









#42
2,4,5-Trichlorophenol
Concen: 9.141 ng/ul
RT: 13.145 min Scan# 1710
Delta R.T. -0.006 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09




Instrument :
BNA_M
ClientSampleId :
SSTD010041


Tgt Ion:196 Resp: 95062
Ion Ratio Lower Upper
196 100
198 96.5 77.5 116.3
200 31.1 24.8 37.2



#43
1,1'-Biphenyl
Concen: 10.356 ng/ul
RT: 13.487 min Scan# 1768
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09

Tgt Ion:154 Resp: 384887
Ion Ratio Lower Upper
154 100
153 38.9 31.3 46.9
76 13.7 11.0 16.6









#44
2-Chloronaphthalene
Concen: 9.954 ng/ul
RT: 13.528 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09




Instrument :
BNA_M
ClientSampleId :
SSTD010041


Tgt Ion:162 Resp: 300382
Ion Ratio Lower Upper
162 100
164 32.0 26.6 40.0
127 39.0 31.5 47.3



#45
2-Nitroaniline
Concen: 9.023 ng/ul
RT: 13.728 min Scan# 1809
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09

Tgt Ion: 65 Resp: 51892
Ion Ratio Lower Upper
65 100
92 71.9 54.2 81.2
138 100.6 84.3 126.5









#47
Dimethylphthalate
Concen: 9.487 ng/ul
RT: 14.104 min Scan# 1873
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09




Instrument :
BNA_M
ClientSampleId :
SSTD010041


Tgt Ion:163 Resp: 375993
Ion Ratio Lower Upper
163 100
194 4.0 3.0 4.6
164 9.8 8.4 12.6



#48
2,6-Dinitrotoluene
Concen: 5.911 ng/ul
RT: 14.216 min Scan# 1892
Delta R.T. -0.006 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09

Tgt Ion:165 Resp: 46479
Ion Ratio Lower Upper
165 100
89 73.3 51.4 77.2
121 28.1 19.4 29.2









#50
Acenaphthylene
Concen: 10.305 ng/ul
RT: 14.369 min Scan# 1918
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09




Instrument :
BNA_M
ClientSampleId :
SSTD010041


Tgt Ion:152 Resp: 476586
Ion Ratio Lower Upper
152 100
151 19.5 15.9 23.9
153 12.9 10.4 15.6



#51
3-Nitroaniline
Concen: 8.822 ng/ul
RT: 14.545 min Scan# 1948
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09

Tgt Ion:138 Resp: 57284
Ion Ratio Lower Upper
138 100
108 16.4 10.3 15.5#
92 120.8 97.7 146.5









#52
Acenaphthene
Concen: 10.177 ng/ul
RT: 14.710 min Scan# 1976
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09




Instrument :
BNA_M
ClientSampleId :
SSTD010041


Tgt Ion:153 Resp: 333056
Ion Ratio Lower Upper
153 100
152 48.8 39.1 58.7
154 89.0 70.4 105.6



#53
2,4-Dinitrophenol
Concen: 3.864 ng/ul
RT: 14.745 min Scan# 1982
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09

Tgt Ion:184 Resp: 18696
Ion Ratio Lower Upper
184 100
63 71.0 61.9 92.9
154 66.7 50.7 76.1









#55
4-Nitrophenol
Concen: 9.388 ng/ul
RT: 14.834 min Scan# 1997
Delta R.T. -0.006 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09




Instrument :
BNA_M
ClientSampleId :
SSTD010041


Tgt Ion:109 Resp: 53244
Ion Ratio Lower Upper
109 100
139 123.8 108.8 163.2
65 133.0 107.8 161.6



#56
Dibenzofuran
Concen: 9.963 ng/ul
RT: 15.045 min Scan# 2033
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09

Tgt Ion:168 Resp: 465018
Ion Ratio Lower Upper
168 100
139 37.6 30.9 46.3









#57
2,4-Dinitrotoluene
Concen: 6.829 ng/ul
RT: 14.998 min Scan# 2025
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09




Instrument :
BNA_M
ClientSampleId :
SSTD010041


Tgt Ion:165 Resp: 77583
Ion Ratio Lower Upper
165 100
63 49.3 38.9 58.3
182 2.5 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 10.443 ng/ul
RT: 15.263 min Scan# 2070
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09

Tgt Ion:232 Resp: 82165
Ion Ratio Lower Upper
232 100
131 49.7 40.6 60.8
130 3.1 2.0 3.0#
166 30.9 25.9 38.9









#59
Diethylphthalate
Concen: 9.216 ng/ul
RT: 15.469 min Scan# 2105
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09




Instrument :
BNA_M
ClientSampleId :
SSTD010041


Tgt Ion:149 Resp: 364780
Ion Ratio Lower Upper
149 100
177 22.4 17.2 25.8
150 12.3 10.1 15.1



#61
Fluorene
Concen: 9.965 ng/ul
RT: 15.692 min Scan# 2143
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09

Tgt Ion:166 Resp: 388075
Ion Ratio Lower Upper
166 100
165 97.8 78.2 117.4
167 13.3 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 11.109 ng/ul
RT: 15.692 min Scan# 2143
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09




Instrument :
BNA_M
ClientSampleId :
SSTD010041


Tgt Ion:204 Resp: 187075
Ion Ratio Lower Upper
204 100
206 33.6 26.3 39.5
141 58.5 45.5 68.3



#63
4-Nitroaniline
Concen: 13.236 ng/ul
RT: 15.698 min Scan# 2144
Delta R.T. -0.006 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09

Tgt Ion:138 Resp: 66310
Ion Ratio Lower Upper
138 100
92 53.9 41.8 62.8
108 86.3 65.0 97.6









#66
4,6-Dinitro-2-methylphenol
Concen: 5.645 ng/ul
RT: 15.757 min Scan# 2155
Delta R.T. -0.006 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09




Instrument :
BNA_M
ClientSampleId :
SSTD010041


Tgt Ion:198 Resp: 37418
Ion Ratio Lower Upper
198 100
182 7.6 3.5 5.3#
77 26.9 21.0 31.6



#67
N-Nitrosodiphenylamine
Concen: 9.536 ng/ul
RT: 15.898 min Scan# 2178
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09

Tgt Ion:169 Resp: 316227
Ion Ratio Lower Upper
169 100
168 66.8 53.5 80.3
167 37.2 29.1 43.7









#68
4-Bromophenyl-phenylether
Concen: 10.219 ng/ul
RT: 16.580 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09




Instrument :
BNA_M
ClientSampleId :
SSTD010041


Tgt Ion:248 Resp: 107228
Ion Ratio Lower Upper
248 100
250 96.6 77.6 116.4
141 78.1 63.7 95.5



#69
Hexachlorobenzene
Concen: 9.979 ng/ul
RT: 16.692 min Scan# 2313
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09

Tgt Ion:284 Resp: 133436
Ion Ratio Lower Upper
284 100
142 37.1 29.4 44.2
249 30.1 24.0 36.0





#70

Atrazine

Concen: 7.758 ng/ul

RT: 16.851 min Scan# 21

Delta R.T. -0.000 min

Lab File: BM038813.D

Acq: 28 Feb 2023 16:09

Instrument :

BNA_M

ClientSampleId :

SSTD010041





Tgt Ion:200 Resp: 93327

Ion Ratio Lower Upper

200 100

173 27.7 21.3 31.9

215 50.3 38.0 57.0



#71

Pentachlorophenol

Concen: 7.955 ng/ul

RT: 17.033 min Scan# 2371

Delta R.T. -0.000 min

Lab File: BM038813.D

Acq: 28 Feb 2023 16:09




Tgt Ion:266 Resp: 65529


Ion Ratio Lower Upper

266 100

264 63.8 51.0 76.6

268 72.9 51.7 77.5









#72
Phenanthrene
Concen: 9.913 ng/ul
RT: 17.433 min Scan# 2439
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09




Instrument :
BNA_M
ClientSampleId :
SSTD010041


Tgt Ion:178 Resp: 622393
Ion Ratio Lower Upper
178 100
179 14.8 12.3 18.5
176 19.1 15.7 23.5



#74
Anthracene
Concen: 9.764 ng/ul
RT: 17.522 min Scan# 2454
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09

Tgt Ion:178 Resp: 623738
Ion Ratio Lower Upper
178 100
179 15.1 12.5 18.7
176 18.6 14.9 22.3





#75

1,2,3,4-Tetrachlorobenzene

Concen: 12.378 ng/uL

RT: 13.451 min Scan# 11

Delta R.T. -0.000 min

Lab File: BM038813.D

Acq: 28 Feb 2023 16:09

Instrument :

BNA_M

ClientSampleId :

SSTD010041





Tgt Ion:216 Resp: 150132

Ion Ratio Lower Upper

216 100

214 77.6 62.4 93.6

218 51.7 38.8 58.2



#76

Pentachlorobenzene

Concen: 11.258 ng/uL

RT: 14.963 min Scan# 2019

Delta R.T. -0.000 min

Lab File: BM038813.D

Acq: 28 Feb 2023 16:09




Tgt Ion:250 Resp: 154852


Ion Ratio Lower Upper

250 100

248 63.4 49.6 74.4

252 63.7 51.2 76.8









#77
 Carbazole
 Concen: 9.523 ng/ul
 RT: 17.792 min Scan# 2500
 Delta R.T. -0.000 min
 Lab File: BM038813.D
 Acq: 28 Feb 2023 16:09




Instrument :
 BNA_M
 ClientSampleId :
 SSTD010041


Tgt Ion:167 Resp: 562351
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.2 25.8
 139 13.4 10.6 16.0



#78
 Di-n-butylphthalate
 Concen: 6.906 ng/ul
 RT: 18.369 min Scan# 2598
 Delta R.T. -0.000 min
 Lab File: BM038813.D
 Acq: 28 Feb 2023 16:09

Tgt Ion:149 Resp: 529222
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.3 10.9
 104 5.6 4.2 6.4









#80
Fluoranthene
Concen: 7.704 ng/u1
RT: 19.445 min Scan# 2781
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09




Instrument :
BNA_M
ClientSampleId :
SSTD010041


Tgt Ion:202 Resp: 687248
Ion Ratio Lower Upper
202 100
101 11.4 9.0 13.4
100 8.8 6.9 10.3



#82
Pyrene
Concen: 8.200 ng/u1
RT: 19.804 min Scan# 2842
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09

Tgt Ion:202 Resp: 765759
Ion Ratio Lower Upper
202 100
101 13.8 11.2 16.8
100 11.6 9.1 13.7





#83

Butylbenzylphthalate

Concen: 3.167 ng/ul

RT: 20.710 min Scan# 2996

Delta R.T. -0.000 min

Lab File: BM038813.D

Acq: 28 Feb 2023 16:09

Instrument :

BNA_M

ClientSampleId :

SSTD010041





Tgt Ion:149 Resp: 146687

Ion Ratio Lower Upper

149 100

91 76.3 60.5 90.7

206 18.8 15.8 23.8



#84

3,3'-Dichlorobenzidine

Concen: 6.285 ng/ul

RT: 21.486 min Scan# 3128

Delta R.T. -0.000 min

Lab File: BM038813.D

Acq: 28 Feb 2023 16:09




Tgt Ion:252 Resp: 175199


Ion Ratio Lower Upper

252 100

254 66.4 51.8 77.6

126 11.6 9.0 13.4





#85

Benzo(a)anthracene

Concen: 9.379 ng/ul

RT: 21.551 min Scan# 3139

Delta R.T. -0.000 min

Lab File: BM038813.D

Acq: 28 Feb 2023 16:09

Instrument :

BNA_M

ClientSampleId :

SSTD010041





Tgt Ion:228 Resp: 761240

Ion Ratio Lower Upper

228 100

229 19.3 15.8 23.6

226 26.6 21.4 32.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 3.904 ng/ul

RT: 21.492 min Scan# 3129

Delta R.T. -0.000 min

Lab File: BM038813.D

Acq: 28 Feb 2023 16:09




Tgt Ion:149 Resp: 268439


Ion Ratio Lower Upper

149 100

167 26.8 21.4 32.2

279 3.8 3.1 4.7









#87
Chrysene
Concen: 9.985 ng/ul
RT: 21.604 min Scan# 3148
Delta R.T. -0.006 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09




Instrument :
BNA_M
ClientSampleId :
SSTD010041


Tgt Ion:228 Resp: 748740
Ion Ratio Lower Upper
228 100
226 30.2 23.6 35.4
229 19.8 15.8 23.6



#89
Di-n-octyl phthalate
Concen: 3.166 ng/ul
RT: 22.445 min Scan# 3291
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09

Tgt Ion:149 Resp: 437282









#90
Benzo(b)fluoranthene
Concen: 8.737 ng/ul
RT: 23.274 min Scan# 3432
Delta R.T. -0.000 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09




Instrument :
BNA_M
ClientSampleId :
SSTD010041


Tgt Ion:252 Resp: 768144
Ion Ratio Lower Upper
252 100
253 22.1 17.2 25.8
125 10.5 8.6 13.0



#91
Benzo(k)fluoranthene
Concen: 9.128 ng/ul
RT: 23.321 min Scan# 3440
Delta R.T. -0.006 min
Lab File: BM038813.D
Acq: 28 Feb 2023 16:09

Tgt Ion:252 Resp: 793304
Ion Ratio Lower Upper
252 100
253 21.9 17.7 26.5
125 11.0 8.4 12.6





#93

Benzo(a)pyrene

Concen: 9.695 ng/ul

RT: 23.915 min Scan# 3

Delta R.T. -0.000 min

Lab File: BM038813.D

Acq: 28 Feb 2023 16:09

Instrument :

BNA_M

ClientSampleId :

SSTD010041





Tgt Ion:252 Resp: 713386

Ion Ratio Lower Upper

252 100

253 22.6 17.8 26.8

125 12.5 9.4 14.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 10.097 ng/ul

RT: 26.592 min Scan# 3996

Delta R.T. -0.000 min

Lab File: BM038813.D

Acq: 28 Feb 2023 16:09




Tgt Ion:276 Resp: 942162


Ion Ratio Lower Upper

276 100

138 24.9 20.2 30.4

277 24.9 19.8 29.8





#95

Dibenzo(a,h)anthracene

Concen: 10.192 ng/ul

RT: 26.609 min Scan# 3999

Delta R.T. -0.006 min

Lab File: BM038813.D

Acq: 28 Feb 2023 16:09

Instrument :

BNA_M

ClientSampleId :

SSTD010041





Tgt Ion:278 Resp: 795636

Ion Ratio Lower Upper

278 100

139 16.2 13.4 20.2

279 23.4 19.1 28.7



#96

Benzo(g,h,i)perylene

Concen: 10.581 ng/ul

RT: 27.374 min Scan# 4129

Delta R.T. -0.000 min

Lab File: BM038813.D

Acq: 28 Feb 2023 16:09

Tgt Ion:276 Resp: 776875

Ion Ratio Lower Upper

276 100

138 23.0 19.4 29.0

277 24.3 18.7 28.1

