

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022823\
 Data File : BM038818.D
 Acq On : 28 Feb 2023 19:17
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV046

Quant Time: Mar 01 02:16:32 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 01 02:09:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.016	152	260738	20.000	ng/u1	0.00
20) Naphthalene-d8	10.828	136	1182809	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.651	164	770216	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.392	188	1693716	20.000	ng/u1	0.00
79) Chrysene-d12	21.568	240	1714615	20.000	ng/u1	0.00
88) Perylene-d12	24.027	264	1967256	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.381	96	47829	6.850	ng/uL	0.00
4) Pyridine-d5	3.805	84	315175	16.417	ng/u1	0.00
7) Phenol-d5	7.157	99	412154	16.861	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.340	67	292626	18.735	ng/u1	0.00
11) 2-Chlorophenol-d4	7.540	132	322987	17.732	ng/u1	0.00
15) 4-Methylphenol-d8	8.716	113	326862	16.973	ng/u1	0.00
21) Nitrobenzene-d5	9.181	128	157911	19.017	ng/u1	0.00
24) 2-Nitrophenol-d4	9.904	143	150738	19.373	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.439	165	320496	17.726	ng/u1	0.00
31) 4-Chloroaniline-d4	10.963	131	500801	16.618	ng/u1	0.00
46) Dimethylphthalate-d6	14.063	166	1034065	17.334	ng/u1	0.00
49) Acenaphthylene-d8	14.345	160	1223436	17.547	ng/u1	0.00
54) 4-Nitrophenol-d4	14.821	143	209617	17.553	ng/u1	0.00
60) Fluorene-d10	15.639	176	902356	17.151	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.745	200	145471	16.521	ng/u1	0.00
73) Anthracene-d10	17.486	188	1438561	17.225	ng/u1	0.00
81) Pyrene-d10	19.774	212	1661105	17.614	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.868	264	1728980	17.440	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.416	88	48097	6.409	ng/uL	97
5) Pyridine	3.822	79	336347	16.274	ng/u1	98
6) Benzaldehyde	7.145	77	298494	25.213	ng/u1	99
8) Phenol	7.187	94	448325	17.135	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.434	93	358241	17.199	ng/u1	99
12) 2-Chlorophenol	7.569	128	341845	18.221	ng/u1	99
13) 2-Methylphenol	8.451	108	345264	17.654	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.551	45	454528	17.382	ng/u1	100
16) Acetophenone	8.839	105	592472	17.702	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.828	70	287948	18.732	ng/u1	99
18) 4-Methylphenol	8.781	108	376411	17.334	ng/u1	98
19) Hexachloroethane	9.104	117	132703	17.656	ng/u1	96
22) Nitrobenzene	9.222	77	406362	17.888	ng/u1	98
23) Isophorone	9.745	82	819440	17.734	ng/u1	99
25) 2-Nitrophenol	9.934	139	183858	19.926	ng/u1	97
26) 2,4-Dimethylphenol	9.992	107	405244	17.877	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.239	93	496484	17.509	ng/u1	100
29) 2,4-Dichlorophenol	10.469	162	333943	18.220	ng/u1	99
30) Naphthalene	10.881	128	1171233	17.353	ng/u1	100
32) 4-Chloroaniline	10.986	127	530061	17.372	ng/u1	99
33) Hexachlorobutadiene	11.169	225	196192	17.204	ng/u1	99
34) Caprolactam	11.739	113	120470	19.168	ng/u1	98
35) 4-Chloro-3-methylphenol	12.098	107	374371	18.068	ng/u1	99
36) 2-Methylnaphthalene	12.480	142	812155	18.435	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM022823\
 Data File : BM038818.D
 Acq On : 28 Feb 2023 19:17
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV046


Quant Time: Mar 01 02:16:32 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 01 02:09:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.704	142	761579	17.157	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.845	216	410056	17.798	ng/ul	100
40) Hexachlorocyclopentadiene	12.833	237	214011	17.217	ng/ul	99
41) 2,4,6-Trichlorophenol	13.080	196	263764	18.151	ng/ul	97
42) 2,4,5-Trichlorophenol	13.151	196	289406	18.095	ng/ul	99
43) 1,1'-Biphenyl	13.486	154	1075981	17.711	ng/ul	100
44) 2-Chloronaphthalene	13.527	162	823720	17.191	ng/ul	99
45) 2-Nitroaniline	13.727	65	218008	20.161	ng/ul	96
47) Dimethylphthalate	14.110	163	1052314	17.406	ng/ul	99
48) 2,6-Dinitrotoluene	14.222	165	212445	22.207	ng/ul	96
50) Acenaphthylene	14.369	152	1366597	17.673	ng/ul	100
51) 3-Nitroaniline	14.545	138	249897	20.813	ng/ul	97
52) Acenaphthene	14.710	153	913702	17.166	ng/ul	98
53) 2,4-Dinitrophenol	14.745	184	92366	16.984	ng/ul	98
55) 4-Nitrophenol	14.839	109	169533	17.773	ng/ul	98
56) Dibenzofuran	15.045	168	1309289	17.718	ng/ul	98
57) 2,4-Dinitrotoluene	14.998	165	289032	19.544	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.263	232	270555	19.430	ng/ul	98
59) Diethylphthalate	15.468	149	1054127	17.621	ng/ul	100
61) Fluorene	15.692	166	1080299	17.436	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.692	204	519350	17.363	ng/ul	99
63) 4-Nitroaniline	15.704	138	268059	21.585	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.763	198	158120	17.352	ng/ul	97
67) N-Nitrosodiphenylamine	15.898	169	920314	17.965	ng/ul	100
68) 4-Bromophenyl-phenylether	16.580	248	302487	17.352	ng/ul	99
69) Hexachlorobenzene	16.692	284	364129	17.245	ng/ul	99
70) Atrazine	16.851	200	303048	17.487	ng/ul	99
71) Pentachlorophenol	17.033	266	218554	17.023	ng/ul	100
72) Phenanthrene	17.433	178	1683338	17.202	ng/ul	99
74) Anthracene	17.521	178	1744388	17.507	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.451	216	400909	16.796	ng/ul	99
76) Pentachlorobenzene	14.963	250	411342	16.772	ng/ul	99
77) Carbazole	17.792	167	1642799	17.755	ng/ul	99
78) Di-n-butylphthalate	18.368	149	1729786	18.375	ng/ul	100
80) Fluoranthene	19.445	202	1881541	17.348	ng/ul	99
82) Pyrene	19.804	202	2091713	17.587	ng/ul	99
83) Butylbenzylphthalate	20.709	149	670180	21.207	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.486	252	754558	22.459	ng/ul	100
85) Benzo(a)anthracene	21.550	228	2106142	17.778	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.492	149	1098456	20.908	ng/ul	98
87) Chrysene	21.609	228	2011333	17.361	ng/ul	99
89) Di-n-octyl phthalate	22.445	149	1793645	18.133	ng/ul	100
90) Benzo(b)fluoranthene	23.274	252	2127466	17.417	ng/ul	99
91) Benzo(k)fluoranthene	23.327	252	2224971	17.733	ng/ul	100
93) Benzo(a)pyrene	23.915	252	1977419	17.476	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.591	276	2618901	17.509	ng/ul	99
95) Dibenzo(a,h)anthracene	26.615	278	2228187	17.658	ng/ul	99
96) Benzo(g,h,i)perylene	27.379	276	2143935	17.506	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Mar 01 02:16:32 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM022823.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Mar 01 02:09:21 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 6.409 ng/uL
RT: 3.416 min Scan# 50
Delta R.T. -0.001 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17




Instrument :
BNA_M
ClientSampleId :
SICV046


Tgt Ion: 88 Resp: 48097
Ion Ratio Lower Upper
88 100
43 27.9 21.5 32.3
58 73.5 56.4 84.6



#5
Pyridine
Concen: 16.274 ng/uL
RT: 3.822 min Scan# 125
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17

Tgt Ion: 79 Resp: 336347
Ion Ratio Lower Upper
79 100
52 70.6 57.8 86.6
51 33.6 27.8 41.6









#6
Benzaldehyde
Concen: 25.213 ng/ul
RT: 7.145 min Scan# 690
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17




Instrument :
BNA_M
ClientSampleId :
SICV046


Tgt Ion: 77 Resp: 298494
Ion Ratio Lower Upper
77 100
105 98.9 77.9 116.9
106 98.5 78.3 117.5



#8
Phenol
Concen: 17.135 ng/ul
RT: 7.187 min Scan# 697
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17

Tgt Ion: 94 Resp: 448325
Ion Ratio Lower Upper
94 100
65 27.3 22.1 33.1
66 34.4 28.6 43.0









#10
Bis(2-Chloroethyl)ether
Concen: 17.199 ng/ul
RT: 7.434 min Scan# 739
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17




Instrument :
BNA_M
ClientSampleId :
SICV046


Tgt Ion: 93 Resp: 358241
Ion Ratio Lower Upper
93 100
63 74.6 60.3 90.5
95 32.7 25.8 38.6



#12
2-Chlorophenol
Concen: 18.221 ng/ul
RT: 7.569 min Scan# 762
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17

Tgt Ion:128 Resp: 341845
Ion Ratio Lower Upper
128 100
64 47.0 38.9 58.3
130 31.4 25.1 37.7





#13

2-Methylphenol

Concen: 17.654 ng/ul

RT: 8.451 min Scan# 912

Delta R.T. -0.000 min

Lab File: BM038818.D

Acq: 28 Feb 2023 19:17

Instrument :

BNA_M

ClientSampleId :





SICV046

Tgt Ion:108 Resp: 345264

Ion Ratio Lower Upper

108 100

107 86.6 69.0 103.6



#14

2,2'-oxybis(1-Chloropropane)

Concen: 17.382 ng/ul

RT: 8.551 min Scan# 929

Delta R.T. -0.000 min

Lab File: BM038818.D

Acq: 28 Feb 2023 19:17




Tgt Ion: 45 Resp: 454528


Ion Ratio Lower Upper

45 100

77 15.3 12.4 18.6

79 11.8 9.4 14.0









#16
Acetophenone
Concen: 17.702 ng/ul
RT: 8.839 min Scan# 91
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17




Instrument :
BNA_M
ClientSampleId :
SICV046


Tgt Ion:105 Resp: 592472
Ion Ratio Lower Upper
105 100
77 75.9 62.6 93.8
51 25.2 20.8 31.2



#17
N-Nitroso-di-n-propylamine
Concen: 18.732 ng/ul
RT: 8.828 min Scan# 976
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17

Tgt Ion: 70 Resp: 287948
Ion Ratio Lower Upper
70 100
42 54.1 43.5 65.3
101 9.7 7.2 10.8
130 21.1 16.0 24.0





#18

4-Methylphenol

Concen: 17.334 ng/ul

RT: 8.781 min Scan# 968

Delta R.T. -0.000 min

Lab File: BM038818.D

Acq: 28 Feb 2023 19:17

Instrument :

BNA_M

ClientSampleId :





SICV046

Tgt Ion:108 Resp: 376411

Ion Ratio Lower Upper

108 100

107 109.5 89.0 133.6



#19

Hexachloroethane

Concen: 17.656 ng/ul

RT: 9.104 min Scan# 1023

Delta R.T. -0.000 min

Lab File: BM038818.D

Acq: 28 Feb 2023 19:17




Tgt Ion:117 Resp: 132703


Ion Ratio Lower Upper

117 100

201 98.0 76.6 115.0

199 62.5 46.6 70.0









#22
Nitrobenzene
Concen: 17.888 ng/ul
RT: 9.222 min Scan# 1043
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17




Instrument :
BNA_M
ClientSampleId :
SICV046


Tgt Ion: 77 Resp: 406362
Ion Ratio Lower Upper
77 100
123 45.6 35.3 52.9
65 13.0 10.6 15.8



#23
Isophorone
Concen: 17.734 ng/ul
RT: 9.745 min Scan# 1132
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17

Tgt Ion: 82 Resp: 819440
Ion Ratio Lower Upper
82 100
95 6.7 5.2 7.8
138 17.0 13.1 19.7









#25
2-Nitrophenol
Concen: 19.926 ng/ul
RT: 9.934 min Scan# 1164
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17




Instrument :
BNA_M
ClientSampleId :
SICV046


Tgt Ion:139 Resp: 183858
Ion Ratio Lower Upper
139 100
65 50.6 42.6 63.8
109 29.2 23.1 34.7



#26
2,4-Dimethylphenol
Concen: 17.877 ng/ul
RT: 9.992 min Scan# 1174
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17

Tgt Ion:107 Resp: 405244
Ion Ratio Lower Upper
107 100
121 48.7 39.8 59.8
122 85.7 69.0 103.4









#27
 Bis(2-Chloroethoxy)methane
 Concen: 17.509 ng/ul
 RT: 10.239 min Scan# 1216
 Delta R.T. 0.006 min
 Lab File: BM038818.D
 Acq: 28 Feb 2023 19:17




Instrument :
 BNA_M
 ClientSampleId :
 SICV046


Tgt Ion: 93 Resp: 496484
 Ion Ratio Lower Upper
 93 100
 95 31.8 25.3 37.9
 123 11.6 9.3 13.9



#29
 2,4-Dichlorophenol
 Concen: 18.220 ng/ul
 RT: 10.469 min Scan# 1255
 Delta R.T. -0.000 min
 Lab File: BM038818.D
 Acq: 28 Feb 2023 19:17

Tgt Ion:162 Resp: 333943
 Ion Ratio Lower Upper
 162 100
 164 65.0 52.6 78.8
 98 36.6 29.0 43.6









#30
Naphthalene
Concen: 17.353 ng/ul
RT: 10.881 min Scan# 1325
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17




Instrument :
BNA_M
ClientSampleId :
SICV046


Tgt Ion:128 Resp: 1171233
Ion Ratio Lower Upper
128 100
129 10.7 8.6 12.8
127 12.9 10.2 15.4



#32
4-Chloroaniline
Concen: 17.372 ng/ul
RT: 10.986 min Scan# 1343
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17

Tgt Ion:127 Resp: 530061
Ion Ratio Lower Upper
127 100
129 32.3 26.3 39.5





#33

Hexachlorobutadiene

Concen: 17.204 ng/ul

RT: 11.169 min Scan# 11

Delta R.T. -0.000 min

Lab File: BM038818.D

Acq: 28 Feb 2023 19:17

Instrument :

BNA_M

ClientSampleId :

SICV046





Tgt Ion:225 Resp: 196192

Ion Ratio Lower Upper

225 100

223 63.7 50.9 76.3

227 64.4 50.4 75.6



#34

Caprolactam

Concen: 19.168 ng/ul

RT: 11.739 min Scan# 1471

Delta R.T. -0.000 min

Lab File: BM038818.D

Acq: 28 Feb 2023 19:17




Tgt Ion:113 Resp: 120470


Ion Ratio Lower Upper

113 100

55 168.6 137.0 205.6

56 123.8 99.9 149.9









#35
4-Chloro-3-methylphenol
Concen: 18.068 ng/ul
RT: 12.098 min Scan# 1532
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17




Instrument :
BNA_M
ClientSampleId :
SICV046


Tgt Ion:107 Resp: 374371
Ion Ratio Lower Upper
107 100
144 26.6 21.1 31.7
142 81.7 64.7 97.1



#36
2-Methylnaphthalene
Concen: 18.435 ng/ul
RT: 12.480 min Scan# 1597
Delta R.T. -0.006 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17

Tgt Ion:142 Resp: 812155
Ion Ratio Lower Upper
142 100
141 87.9 71.0 106.6





#37

1-Methylnaphthalene

Concen: 17.157 ng/ul

RT: 12.704 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM038818.D

Acq: 28 Feb 2023 19:17

Instrument :

BNA_M

ClientSampleId :

SICV046





Tgt Ion:142 Resp: 761579

Ion Ratio Lower Upper

142 100

141 91.9 73.4 110.0

116 3.8 3.2 4.8



#39

1,2,4,5-Tetrachlorobenzene

Concen: 17.798 ng/ul

RT: 12.845 min Scan# 1659

Delta R.T. -0.000 min

Lab File: BM038818.D

Acq: 28 Feb 2023 19:17

Tgt Ion:216 Resp: 410056




Ion Ratio Lower Upper


216 100

214 77.9 62.2 93.4

179 20.4 16.4 24.6

108 19.4 15.3 22.9





#40

Hexachlorocyclopentadiene

Concen: 17.217 ng/ul

RT: 12.833 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM038818.D

Acq: 28 Feb 2023 19:17

Instrument :

BNA_M

ClientSampleId :

SICV046





Tgt Ion:237 Resp: 214011

Ion Ratio Lower Upper

237 100

235 61.1 48.1 72.1

272 12.8 10.0 15.0



#41

2,4,6-Trichlorophenol

Concen: 18.151 ng/ul

RT: 13.080 min Scan# 1699

Delta R.T. -0.000 min

Lab File: BM038818.D

Acq: 28 Feb 2023 19:17




Tgt Ion:196 Resp: 263764


Ion Ratio Lower Upper

196 100

198 96.4 74.8 112.2

200 31.1 24.3 36.5









#42
2,4,5-Trichlorophenol
Concen: 18.095 ng/ul
RT: 13.151 min Scan# 1711
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17




Instrument :
BNA_M
ClientSampleId :
SICV046


Tgt Ion:196 Resp: 289406
Ion Ratio Lower Upper
196 100
198 98.4 77.5 116.3
200 31.2 24.8 37.2



#43
1,1'-Biphenyl
Concen: 17.711 ng/ul
RT: 13.486 min Scan# 1768
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17

Tgt Ion:154 Resp: 1075981
Ion Ratio Lower Upper
154 100
153 39.2 31.3 46.9
76 13.6 11.0 16.6





#44

2-Chloronaphthalene

Concen: 17.191 ng/ul

RT: 13.527 min Scan# 11

Delta R.T. -0.000 min

Lab File: BM038818.D

Acq: 28 Feb 2023 19:17

Instrument :

BNA_M

ClientSampleId :

SICV046





Tgt Ion:162 Resp: 823720

Ion Ratio Lower Upper

162 100

164 32.6 26.6 40.0

127 39.0 31.5 47.3



#45

2-Nitroaniline

Concen: 20.161 ng/ul

RT: 13.727 min Scan# 1809

Delta R.T. -0.000 min

Lab File: BM038818.D

Acq: 28 Feb 2023 19:17




Tgt Ion: 65 Resp: 218008


Ion Ratio Lower Upper

65 100

92 71.7 54.2 81.2

138 109.6 84.3 126.5





#47

Dimethylphthalate

Concen: 17.406 ng/ul

RT: 14.110 min Scan# 1874

Delta R.T. 0.006 min

Lab File: BM038818.D

Acq: 28 Feb 2023 19:17

Instrument :

BNA_M

ClientSampleId :

SICV046





Tgt Ion:163 Resp: 1052314

Ion Ratio Lower Upper

163 100

194 4.2 3.0 4.6

164 9.9 8.4 12.6



#48

2,6-Dinitrotoluene

Concen: 22.207 ng/ul

RT: 14.222 min Scan# 1893

Delta R.T. -0.000 min

Lab File: BM038818.D

Acq: 28 Feb 2023 19:17




Tgt Ion:165 Resp: 212445


Ion Ratio Lower Upper

165 100

89 61.5 51.4 77.2

121 21.7 19.4 29.2









#50
Acenaphthylene
Concen: 17.673 ng/ul
RT: 14.369 min Scan# 1918
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17




Instrument :
BNA_M
ClientSampleId :
SICV046


Tgt Ion:152 Resp: 1366597
Ion Ratio Lower Upper
152 100
151 20.0 15.9 23.9
153 13.2 10.4 15.6



#51
3-Nitroaniline
Concen: 20.813 ng/ul
RT: 14.545 min Scan# 1948
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17

Tgt Ion:138 Resp: 249897
Ion Ratio Lower Upper
138 100
108 11.9 10.3 15.5
92 118.8 97.7 146.5









#52
Acenaphthene
Concen: 17.166 ng/ul
RT: 14.710 min Scan# 1976
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17




Instrument :
BNA_M
ClientSampleId :
SICV046


Tgt Ion:153 Resp: 913702
Ion Ratio Lower Upper
153 100
152 47.9 39.1 58.7
154 89.7 70.4 105.6



#53
2,4-Dinitrophenol
Concen: 16.984 ng/ul
RT: 14.745 min Scan# 1982
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17

Tgt Ion:184 Resp: 92366
Ion Ratio Lower Upper
184 100
63 75.0 61.9 92.9
154 62.4 50.7 76.1









#55
4-Nitrophenol
Concen: 17.773 ng/ul
RT: 14.839 min Scan# 1998
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17




Instrument :
BNA_M
ClientSampleId :
SICV046


Tgt Ion:109 Resp: 169533
Ion Ratio Lower Upper
109 100
139 140.4 108.8 163.2
65 133.7 107.8 161.6



#56
Dibenzofuran
Concen: 17.718 ng/ul
RT: 15.045 min Scan# 2033
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17

Tgt Ion:168 Resp: 1309289
Ion Ratio Lower Upper
168 100
139 37.4 30.9 46.3









#57
2,4-Dinitrotoluene
Concen: 19.544 ng/ul
RT: 14.998 min Scan# 2025
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17




Instrument :
BNA_M
ClientSampleId :
SICV046


Tgt Ion:165 Resp: 289032
Ion Ratio Lower Upper
165 100
63 47.1 38.9 58.3
182 2.5 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.430 ng/ul
RT: 15.263 min Scan# 2070
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17

Tgt Ion:232 Resp: 270555
Ion Ratio Lower Upper
232 100
131 48.7 40.6 60.8
130 2.5 2.0 3.0
166 31.3 25.9 38.9









#59
Diethylphthalate
Concen: 17.621 ng/ul
RT: 15.468 min Scan# 2105
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17




Instrument :
BNA_M
ClientSampleId :
SICV046


Tgt Ion:149 Resp: 1054127
Ion Ratio Lower Upper
149 100
177 21.5 17.2 25.8
150 12.2 10.1 15.1



#61
Fluorene
Concen: 17.436 ng/ul
RT: 15.692 min Scan# 2143
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17

Tgt Ion:166 Resp: 1080299
Ion Ratio Lower Upper
166 100
165 96.9 78.2 117.4
167 13.6 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 17.363 ng/ul
RT: 15.692 min Scan# 2143
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17




Instrument :
BNA_M
ClientSampleId :
SICV046


Tgt Ion:204 Resp: 519350
Ion Ratio Lower Upper
204 100
206 32.2 26.3 39.5
141 56.2 45.5 68.3



#63
4-Nitroaniline
Concen: 21.585 ng/ul
RT: 15.704 min Scan# 2145
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17

Tgt Ion:138 Resp: 268059
Ion Ratio Lower Upper
138 100
92 49.4 41.8 62.8
108 78.0 65.0 97.6









#66
4,6-Dinitro-2-methylphenol
Concen: 17.352 ng/ul
RT: 15.763 min Scan# 2155
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17




Instrument :
BNA_M
ClientSampleId :
SICV046


Tgt Ion:198 Resp: 158120
Ion Ratio Lower Upper
198 100
182 3.9 3.5 5.3
77 24.5 21.0 31.6



#67
N-Nitrosodiphenylamine
Concen: 17.965 ng/ul
RT: 15.898 min Scan# 2178
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17

Tgt Ion:169 Resp: 920314
Ion Ratio Lower Upper
169 100
168 67.2 53.5 80.3
167 36.6 29.1 43.7









#68
4-Bromophenyl-phenylether
Concen: 17.352 ng/ul
RT: 16.580 min Scan# 2294
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17




Instrument :
BNA_M
ClientSampleId :
SICV046


Tgt Ion:248 Resp: 302487
Ion Ratio Lower Upper
248 100
250 97.8 77.6 116.4
141 78.3 63.7 95.5



#69
Hexachlorobenzene
Concen: 17.245 ng/ul
RT: 16.692 min Scan# 2313
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17

Tgt Ion:284 Resp: 364129
Ion Ratio Lower Upper
284 100
142 37.6 29.4 44.2
249 30.4 24.0 36.0





#70

Atrazine

Concen: 17.487 ng/ul

RT: 16.851 min Scan# 21

Delta R.T. -0.000 min

Lab File: BM038818.D

Acq: 28 Feb 2023 19:17

Instrument :

BNA_M

ClientSampleId :

SICV046





Tgt Ion:200 Resp: 303048

Ion Ratio Lower Upper

200 100

173 26.7 21.3 31.9

215 48.6 38.0 57.0



#71

Pentachlorophenol

Concen: 17.023 ng/ul

RT: 17.033 min Scan# 2371

Delta R.T. -0.000 min

Lab File: BM038818.D

Acq: 28 Feb 2023 19:17




Tgt Ion:266 Resp: 218554


Ion Ratio Lower Upper

266 100

264 63.6 51.0 76.6

268 64.8 51.7 77.5





#72

Phenanthrene

Concen: 17.202 ng/ul

RT: 17.433 min Scan# 2439

Delta R.T. -0.000 min

Lab File: BM038818.D

Acq: 28 Feb 2023 19:17

Instrument :

BNA_M

ClientSampleId :

SICV046





Tgt Ion:178 Resp: 1683338

Ion Ratio Lower Upper

178 100

179 15.5 12.3 18.5

176 19.2 15.7 23.5



#74

Anthracene

Concen: 17.507 ng/ul

RT: 17.521 min Scan# 2454

Delta R.T. -0.000 min

Lab File: BM038818.D

Acq: 28 Feb 2023 19:17




Tgt Ion:178 Resp: 1744388


Ion Ratio Lower Upper

178 100

179 15.7 12.5 18.7

176 19.0 14.9 22.3





#75

1,2,3,4-Tetrachlorobenzene

Concen: 16.796 ng/uL

RT: 13.451 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM038818.D

Acq: 28 Feb 2023 19:17

Instrument :

BNA_M

ClientSampleId :

SICV046





Tgt Ion:216 Resp: 400909

Ion Ratio Lower Upper

216 100

214 78.9 62.4 93.6

218 48.0 38.8 58.2



#76

Pentachlorobenzene

Concen: 16.772 ng/uL

RT: 14.963 min Scan# 2019

Delta R.T. -0.000 min

Lab File: BM038818.D

Acq: 28 Feb 2023 19:17




Tgt Ion:250 Resp: 411342


Ion Ratio Lower Upper

250 100

248 63.2 49.6 74.4

252 63.1 51.2 76.8









#77
 Carbazole
 Concen: 17.755 ng/ul
 RT: 17.792 min Scan# 2500
 Delta R.T. -0.000 min
 Lab File: BM038818.D
 Acq: 28 Feb 2023 19:17




Instrument :
 BNA_M
 ClientSampleId :
 SICV046


Tgt Ion:167 Resp: 1642799
 Ion Ratio Lower Upper
 167 100
 166 21.3 17.2 25.8
 139 12.8 10.6 16.0



#78
 Di-n-butylphthalate
 Concen: 18.375 ng/ul
 RT: 18.368 min Scan# 2598
 Delta R.T. -0.000 min
 Lab File: BM038818.D
 Acq: 28 Feb 2023 19:17

Tgt Ion:149 Resp: 1729786
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.3 10.9
 104 5.2 4.2 6.4









#80
Fluoranthene
Concen: 17.348 ng/ul
RT: 19.445 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17




Instrument :
BNA_M
ClientSampleId :
SICV046


Tgt Ion:202 Resp: 1881541
Ion Ratio Lower Upper
202 100
101 11.6 9.0 13.4
100 8.8 6.9 10.3



#82
Pyrene
Concen: 17.587 ng/ul
RT: 19.804 min Scan# 2842
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17

Tgt Ion:202 Resp: 2091713
Ion Ratio Lower Upper
202 100
101 14.8 11.2 16.8
100 11.6 9.1 13.7





#83

Butylbenzylphthalate

Concen: 21.207 ng/ul

RT: 20.709 min Scan# 2996

Delta R.T. -0.000 min

Lab File: BM038818.D

Acq: 28 Feb 2023 19:17

Instrument :

BNA_M

ClientSampleId :

SICV046





Tgt Ion:149 Resp: 670180

Ion Ratio Lower Upper

149 100

91 73.2 60.5 90.7

206 20.0 15.8 23.8



#84

3,3'-Dichlorobenzidine

Concen: 22.459 ng/ul

RT: 21.486 min Scan# 3128

Delta R.T. -0.000 min

Lab File: BM038818.D

Acq: 28 Feb 2023 19:17




Tgt Ion:252 Resp: 754558


Ion Ratio Lower Upper

252 100

254 64.7 51.8 77.6

126 11.1 9.0 13.4





#85

Benzo(a)anthracene

Concen: 17.778 ng/ul

RT: 21.550 min Scan# 3139

Delta R.T. -0.000 min

Lab File: BM038818.D

Acq: 28 Feb 2023 19:17

Instrument :

BNA_M

ClientSampleId :

SICV046





Tgt Ion:228 Resp: 2106142

Ion Ratio Lower Upper

228 100

229 19.9 15.8 23.6

226 27.0 21.4 32.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 20.908 ng/ul

RT: 21.492 min Scan# 3129

Delta R.T. -0.000 min

Lab File: BM038818.D

Acq: 28 Feb 2023 19:17




Tgt Ion:149 Resp: 1098456


Ion Ratio Lower Upper

149 100

167 27.9 21.4 32.2

279 3.9 3.1 4.7









#87
Chrysene
Concen: 17.361 ng/ul
RT: 21.609 min Scan# 3149
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17




Instrument :
BNA_M
ClientSampleId :
SICV046


Tgt Ion:228 Resp: 2011333
Ion Ratio Lower Upper
228 100
226 29.9 23.6 35.4
229 19.4 15.8 23.6



#89
Di-n-octyl phthalate
Concen: 18.133 ng/ul
RT: 22.445 min Scan# 3291
Delta R.T. -0.000 min
Lab File: BM038818.D
Acq: 28 Feb 2023 19:17

Tgt Ion:149 Resp: 1793645





#90

Benzo(b)fluoranthene

Concen: 17.417 ng/ul

RT: 23.274 min Scan# 3432

Delta R.T. -0.000 min

Lab File: BM038818.D

Acq: 28 Feb 2023 19:17

Instrument :

BNA_M

ClientSampleId :

SICV046





Tgt Ion:252 Resp: 2127466

Ion Ratio Lower Upper

252 100

253 21.8 17.2 25.8

125 11.1 8.6 13.0



#91

Benzo(k)fluoranthene

Concen: 17.733 ng/ul

RT: 23.327 min Scan# 3441

Delta R.T. -0.000 min

Lab File: BM038818.D

Acq: 28 Feb 2023 19:17




Tgt Ion:252 Resp: 2224971


Ion Ratio Lower Upper

252 100

253 22.0 17.7 26.5

125 10.7 8.4 12.6





#93

Benzo(a)pyrene

Concen: 17.476 ng/ul

RT: 23.915 min Scan# 3

Delta R.T. -0.000 min

Lab File: BM038818.D

Acq: 28 Feb 2023 19:17

Instrument :

BNA_M

ClientSampleId :

SICV046





Tgt Ion:252 Resp: 1977419

Ion Ratio Lower Upper

252 100

253 21.6 17.8 26.8

125 11.8 9.4 14.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 17.509 ng/ul

RT: 26.591 min Scan# 3996

Delta R.T. -0.000 min

Lab File: BM038818.D

Acq: 28 Feb 2023 19:17




Tgt Ion:276 Resp: 2618901


Ion Ratio Lower Upper

276 100

138 24.6 20.2 30.4

277 24.6 19.8 29.8





#95

Dibenzo(a,h)anthracene

Concen: 17.658 ng/ul

RT: 26.615 min Scan# 4000

Delta R.T. -0.000 min

Lab File: BM038818.D

Acq: 28 Feb 2023 19:17

Instrument :

BNA_M

ClientSampleId :

SICV046





Tgt Ion:278 Resp: 2228187

Ion Ratio Lower Upper

278 100

139 16.9 13.4 20.2

279 23.2 19.1 28.7



#96

Benzo(g,h,i)perylene

Concen: 17.506 ng/ul

RT: 27.379 min Scan# 4130

Delta R.T. 0.006 min

Lab File: BM038818.D

Acq: 28 Feb 2023 19:17

Tgt Ion:276 Resp: 2143935

Ion Ratio Lower Upper

276 100

138 22.6 19.4 29.0

277 23.5 18.7 28.1

