

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM030519\
 Data File : BM019045.D
 Acq On : 05 Mar 2019 20:18
 Operator : JU/SJ
 Sample : K1705-20
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PST-028-B105-022719

Quant Time: Mar 06 06:31:49 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM021119.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 05 03:36:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.67	152	252010	20.00	ng	0.00
21) Naphthalene-d8	10.46	136	911896	20.00	ng	0.00
39) Acenaphthene-d10	14.33	164	459873	20.00	ng	0.00
64) Phenanthrene-d10	17.08	188	933463	20.00	ng	0.00
76) Chrysene-d12	21.28	240	892544	20.00	ng	0.00
87) Perylene-d12	23.52	264	990443	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.27	112	2042514	132.80	ng	0.00
7) Phenol-d6	6.86	99	2698147	124.52	ng	0.00
23) Nitrobenzene-d5	8.85	82	1692068	85.80	ng	0.00
42) 2,4,6-Tribromophenol	15.83	330	797691	120.33	ng	0.00
45) 2-Fluorobiphenyl	12.95	172	2680533	77.38	ng	0.00
79) Terphenyl-d14	19.72	244	3058722	70.69	ng	-0.01
Target Compounds						
50) Dimethylphthalate	13.79	163	192197	5.015	ng	100
75) Fluoranthene	19.14	202	119425	2.061	ng	99
78) Pyrene	19.51	202	116594	2.249	ng	99
84) Bis(2-ethylhexyl)phthalate	21.19	149	84627	2.447	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM030519\
Data File : BM019045.D
Acq On : 05 Mar 2019 20:18
Operator : JU/SJ
Sample : K1705-20
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PST-028-B105-022719

Quant Time: Mar 06 06:31:49 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM021119.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 05 03:36:13 2019
Response via : Initial Calibration

