

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM030620\
 Data File : BM025371.D
 Acq On : 08 Mar 2020 05:27
 Operator : CG/JU
 Sample : L1616-07
 Misc :
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Mar 09 01:14:28 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM030320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 04 08:02:03 2020
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.67	152	4541	0.40	ng/ul	0.00
2) Naphthalene-d8	10.43	136	20287	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.29	164	14615	0.40	ng/ul	-0.01
10) Phenanthrene-d10	0.00	188	0	0.00	ng/ul	-17.05
16) Chrysene-d12	21.22	240	26518	0.40	ng/ul	0.00
20) Perylene-d12	23.41	264	23932	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.03	152	10740	0.28	ng/ul	-0.01
14) Fluoranthene-d10	19.07	212	28478	0.00	ng/ul	0.00
Target Compounds						
17) Pyrene	19.43	202	3875	0.039	ng/ul#	Ovalue 1


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM030620\
Data File : BM025371.D
Acq On : 08 Mar 2020 05:27
Operator : CG/JU
Sample : L1616-07
Misc :
ALS Vial : 57 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :

Quant Time: Mar 09 01:14:28 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM030320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 04 08:02:03 2020
Response via : Initial Calibration





#17

Pvrene

Concen: 0.039 ng/ul

RT: 19.43 min Scan# 5251

Delta R.T. -0.03 min

Lab File: BM025371.D

Acq: 08 Mar 2020 05:27

Instrument :

BNA_M

ClientSampleId :



Tot Ion:202 Resp: 3875

Ion Ratio Lower Upper

202 100

101 136.6 8.7 13.1#

100 928.4 7.3 10.9#



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): B

