

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM030620\
 Data File : BM025391.D
 Acq On : 08 Mar 2020 18:34
 Operator : CG/JU
 Sample : L1617-10
 Misc :
 ALS Vial : 77 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Mar 09 01:15:52 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM030320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 04 08:02:03 2020
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.67	152	4290	0.40	ng/ul	-0.01
2) Naphthalene-d8	10.43	136	19117	0.40	ng/ul	-0.01
6) Acenaphthene-d10	14.29	164	14186	0.40	ng/ul	-0.01
10) Phenanthrene-d10	17.03	188	34024	0.40	ng/ul	-0.01
16) Chrysene-d12	21.22	240	30438	0.40	ng/ul	0.00
20) Perylene-d12	23.41	264	27558	0.40	ng/ul	-0.01
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.03	152	5676	0.16	ng/ul	-0.01
14) Fluoranthene-d10	19.07	212	18649	0.17	ng/ul	0.00
Target Compounds						
17) Pyrene	19.43	202	3280	0.028	ng/ul#	Ovalue 1


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM030620\
Data File : BM025391.D
Acq On : 08 Mar 2020 18:34
Operator : CG/JU
Sample : L1617-10
Misc :
ALS Vial : 77 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :

Quant Time: Mar 09 01:15:52 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM030320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 04 08:02:03 2020
Response via : Initial Calibration





#17

Pvrene

Concen: 0.028 ng/ul

RT: 19.43 min Scan# 5251

Delta R.T. -0.04 min

Lab File: BM025391.D

Acq: 08 Mar 2020 18:34

Instrument :

BNA_M

ClientSampleId :


Tot Ion:202 Resp: 3280

Ion Ratio Lower Upper

202 100

101 135.8 8.7 13.1#

100 910.5 7.3 10.9#



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): B

