

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM030620\
 Data File : BM025384.D
 Acq On : 08 Mar 2020 13:47
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 70 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.418

Quant Time: Mar 09 00:54:01 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM030320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 04 08:02:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.67	152	5538	0.40	ng/ul	0.00
2) Naphthalene-d8	10.43	136	24686	0.40	ng/ul	-0.01
6) Acenaphthene-d10	14.29	164	17292	0.40	ng/ul	-0.01
10) Phenanthrene-d10	17.04	188	38015	0.40	ng/ul	-0.01
16) Chrysene-d12	21.22	240	27918	0.40	ng/ul	0.00
20) Perylene-d12	23.41	264	25702	0.40	ng/ul	-0.01
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.03	152	15519	0.34	ng/ul	-0.01
14) Fluoranthene-d10	19.06	212	36994	0.30	ng/ul	-0.01
Target Compounds				Ovalue		
3) Naphthalene	10.48	128	24704	0.333	ng/ul#	82
5) 2-Methylnaphthalene	12.10	142	17942	0.332	ng/ul	100
7) Acenaphthylene	14.01	152	26128	0.323	ng/ul	94
8) Acenaphthene	14.35	153	20525	0.327	ng/ul	98
9) Fluorene	15.34	166	25769	0.327	ng/ul	96
11) Pentachlorophenol	16.69	266	4077	0.410	ng/ul	97
12) Phenanthrene	17.08	178	38620	0.315	ng/ul	96
13) Anthracene	17.17	178	39441	0.363	ng/ul	96
15) Fluoranthene	19.09	202	45105	0.294	ng/ul	98
17) Pyrene	19.46	202	44696	0.422	ng/ul	97
18) Benzo(a)anthracene	21.20	228	37750	0.368	ng/ul	99
19) Chrysene	21.26	228	36972	0.328	ng/ul	99
21) Benzo(b)fluoranthene	22.76	252	33519	0.326	ng/ul	95
22) Benzo(k)fluoranthene	22.80	252	32770	0.304	ng/ul	99
23) Benzo(a)pyrene	23.31	252	29743	0.345	ng/ul#	95
24) Indeno(1,2,3-cd)pyrene	25.58	276	35728	0.332	ng/ul#	100
25) Dibenzo(a,h)anthracene	25.59	278	29655	0.333	ng/ul#	96
26) Benzo(g,h,i)perylene	26.24	276	29855	0.327	ng/ul#	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM030620\
Data File : BM025384.D
Acq On : 08 Mar 2020 13:47
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 70 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.418

Quant Time: Mar 09 00:54:01 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM030320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 04 08:02:03 2020
Response via : Initial Calibration

