

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030822\
 Data File : BM034143.D
 Acq On : 08 Mar 2022 09:19
 Operator : CG/JU
 Sample : SSTD00539
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD005039

Quant Time: Mar 08 11:31:07 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM030822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 08 11:29:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.989	152	143580	20.000	ng/ul	0.00
20) Naphthalene-d8	10.807	136	567794	20.000	ng/ul	# 0.00
38) Acenaphthene-d10	14.630	164	396077	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.365	188	876571	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.542	240	932379	20.000	ng/ul	# 0.00
88) Perylene-d12	23.983	264	904322	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.366	96	5647	1.813	ng/uL	0.00
4) Pyridine-d5	3.796	84	35253	4.308	ng/ul	0.00
7) Phenol-d5	7.137	99	43985	4.289	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.313	67	27180	4.774	ng/ul	0.00
11) 2-Chlorophenol-d4	7.519	132	39723	4.493	ng/ul	0.00
15) 4-Methylphenol-d8	8.689	113	36055	4.131	ng/ul	0.00
21) Nitrobenzene-d5	9.148	128	19091	4.451	ng/ul	0.00
24) 2-Nitrophenol-d4	9.878	143	18701	3.836	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.419	165	42216	4.279	ng/ul	0.00
31) 4-Chloroaniline-d4	10.931	131	55352	4.449	ng/ul	0.00
46) Dimethylphthalate-d6	14.030	166	133280	4.447	ng/ul	0.00
49) Acenaphthylene-d8	14.319	160	159264	4.364	ng/ul	0.00
54) 4-Nitrophenol-d4	14.801	143	17955	3.498	ng/ul	0.00
60) Fluorene-d10	15.613	176	121512	4.691	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.724	200	17885	3.125	ng/ul	0.00
73) Anthracene-d10	17.465	188	194177	4.581	ng/ul	0.00
81) Pyrene-d10	19.753	212	230766	4.591	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.824	264	206532	4.481	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.407	88	6131	2.006	ng/uL#	78
5) Pyridine	3.813	79	36999	4.431	ng/ul#	70
6) Benzaldehyde	7.119	77	31677	5.440	ng/ul#	80
8) Phenol	7.166	94	46625	4.657	ng/ul#	85
10) Bis(2-Chloroethyl)ether	7.401	93	38284	4.855	ng/ul#	84
12) 2-Chlorophenol	7.548	128	41433	4.722	ng/ul#	79
13) 2-Methylphenol	8.431	108	34534	4.389	ng/ul	97
14) 2,2'-oxybis(1-Chloropr...	8.525	45	34498	4.009	ng/ul#	80
16) Acetophenone	8.813	105	62675	4.617	ng/ul#	81
17) N-Nitroso-di-n-propyla...	8.795	70	28123	4.542	ng/ul#	81
18) 4-Methylphenol	8.754	108	39190	4.593	ng/ul	99
19) Hexachloroethane	9.084	117	17303	4.709	ng/ul#	53
22) Nitrobenzene	9.195	77	44270	4.812	ng/ul#	85
23) Isophorone	9.725	82	73719	4.346	ng/ul#	92
25) 2-Nitrophenol	9.907	139	20508	4.075	ng/ul#	74
26) 2,4-Dimethylphenol	9.972	107	46048	4.568	ng/ul	89
27) Bis(2-Chloroethoxy)met...	10.207	93	50061	4.868	ng/ul	94
29) 2,4-Dichlorophenol	10.448	162	41965	4.429	ng/ul	91
30) Naphthalene	10.854	128	141806	4.984	ng/ul	99
32) 4-Chloroaniline	10.960	127	56452	4.590	ng/ul	97
33) Hexachlorobutadiene	11.154	225	35279	4.848	ng/ul	94
34) Caprolactam	11.707	113	9309	3.618	ng/ul#	41
35) 4-Chloro-3-methylphenol	12.077	107	38449	4.334	ng/ul#	71
36) 2-Methylnaphthalene	12.460	142	96632	4.652	ng/ul	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030822\
 Data File : BM034143.D
 Acq On : 08 Mar 2022 09:19
 Operator : CG/JU
 Sample : SST00539
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST005039

Quant Time: Mar 08 11:31:07 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM030822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 08 11:29:45 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.677	142	98890	4.697	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.830	216	63410	4.819	ng/ul#	96
40) Hexachlorocyclopentadiene	12.819	237	35912	4.002	ng/ul	96
41) 2,4,6-Trichlorophenol	13.060	196	33558	4.175	ng/ul	99
42) 2,4,5-Trichlorophenol	13.130	196	37563	4.249	ng/ul	94
43) 1,1'-Biphenyl	13.466	154	139968	4.794	ng/ul#	94
44) 2-Chloronaphthalene	13.507	162	109187	4.814	ng/ul	94
45) 2-Nitroaniline	13.701	65	20136	3.943	ng/ul#	49
47) Dimethylphthalate	14.077	163	134446	4.634	ng/ul	99
48) 2,6-Dinitrotoluene	14.195	165	22890	4.032	ng/ul#	85
50) Acenaphthylene	14.348	152	163558	4.563	ng/ul	99
51) 3-Nitroaniline	14.519	138	21922	3.970	ng/ul#	78
52) Acenaphthene	14.689	153	113011	4.751	ng/ul	97
53) 2,4-Dinitrophenol	14.724	184	9845	2.614	ng/ul#	63
55) 4-Nitrophenol	14.813	109	16425	3.515	ng/ul#	75
56) Dibenzofuran	15.024	168	169868	4.862	ng/ul	92
57) 2,4-Dinitrotoluene	14.977	165	33473	4.057	ng/ul#	65
58) 2,3,4,6-Tetrachlorophenol	15.248	232	32511	4.250	ng/ul#	84
59) Diethylphthalate	15.436	149	128316	4.430	ng/ul	98
61) Fluorene	15.671	166	136990	4.702	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.665	204	76067	4.882	ng/ul#	86
63) 4-Nitroaniline	15.671	138	21639	3.830	ng/ul#	74
66) 4,6-Dinitro-2-methylph...	15.736	198	18347	3.280	ng/ul#	87
67) N-Nitrosodiphenylamine	15.871	169	115178	4.525	ng/ul	98
68) 4-Bromophenyl-phenylether	16.560	248	41862	4.377	ng/ul#	88
69) Hexachlorobenzene	16.677	284	55755	4.807	ng/ul#	89
70) Atrazine	16.818	200	40740	3.900	ng/ul	97
71) Pentachlorophenol	17.012	266	26214	3.471	ng/ul	96
72) Phenanthrene	17.407	178	224944	4.731	ng/ul	98
74) Anthracene	17.501	178	222290	4.617	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.430	216	68141	4.925	ng/ul	94
76) Pentachlorobenzene	14.948	250	65004	4.749	ng/ul	97
77) Carbazole	17.765	167	182762	4.355	ng/ul	99
78) Di-n-butylphthalate	18.336	149	189496	3.873	ng/ul	98
80) Fluoranthene	19.418	202	263094	4.530	ng/ul#	91
82) Pyrene	19.777	202	282529	4.710	ng/ul#	91
83) Butylbenzylphthalate	20.677	149	79196	3.744	ng/ul#	76
84) 3,3'-Dichlorobenzidine	21.453	252	69846	3.534	ng/ul#	97
85) Benzo(a)anthracene	21.524	228	275597	4.835	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.453	149	114968	3.525	ng/ul#	97
87) Chrysene	21.577	228	275270	4.919	ng/ul	98
89) Di-n-octyl phthalate	22.394	149	186421	3.350	ng/ul	100
90) Benzo(b)fluoranthene	23.236	252	271294	4.731	ng/ul	98
91) Benzo(k)fluoranthene	23.236	252	271294	5.025	ng/ul	98
93) Benzo(a)pyrene	23.871	252	257220	4.685	ng/ul#	96
94) Indeno(1,2,3-cd)pyrene	26.535	276	274592	4.494	ng/ul	97
95) Dibenzo(a,h)anthracene	26.547	278	236494	4.471	ng/ul#	96
96) Benzo(g,h,i)perylene	27.312	276	231496	4.486	ng/ul#	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030822\
 Data File : BM034143.D
 Acq On : 08 Mar 2022 09:19
 Operator : CG/JU
 Sample : SSTD00539
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD005039

Quant Time: Mar 08 11:31:07 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM030822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 08 11:29:45 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 2.006 ng/uL
RT: 3.407 min Scan# 29
Delta R.T. -0.000 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19




Instrument :
BNA_M
ClientSampleId :
SSTD005039


Tgt Ion: 88 Resp: 6131
Ion Ratio Lower Upper
88 100
43 29.6 36.2 54.4#
58 66.5 68.5 102.7#



#5
Pyridine
Concen: 4.431 ng/uL
RT: 3.813 min Scan# 98
Delta R.T. -0.000 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19

Tgt Ion: 79 Resp: 36999
Ion Ratio Lower Upper
79 100
52 60.9 78.1 117.1#
51 34.8 36.0 54.0#









#6
Benzaldehyde
Concen: 5.440 ng/ul
RT: 7.119 min Scan# 660
Delta R.T. -0.000 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19




Instrument :
BNA_M
ClientSampleId :
SSTD005039


Tgt Ion: 77 Resp: 31677
Ion Ratio Lower Upper
77 100
105 97.0 65.6 98.4
106 95.3 60.6 90.8#



#8
Phenol
Concen: 4.657 ng/ul
RT: 7.166 min Scan# 668
Delta R.T. -0.000 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19

Tgt Ion: 94 Resp: 46625
Ion Ratio Lower Upper
94 100
65 32.1 31.7 47.5
66 44.5 45.4 68.0#





#10

Bis(2-Chloroethyl)ether

Concen: 4.855 ng/ul

RT: 7.401 min Scan# 708

Delta R.T. -0.006 min

Lab File: BM034143.D

Acq: 08 Mar 2022 09:19

Instrument :

BNA_M

ClientSampleId :

SSTD005039





Tgt Ion: 93 Resp: 38284

Ion Ratio Lower Upper

93 100

63 70.2 71.7 107.5#

95 29.9 26.3 39.5



#12

2-Chlorophenol

Concen: 4.722 ng/ul

RT: 7.548 min Scan# 733

Delta R.T. -0.006 min

Lab File: BM034143.D

Acq: 08 Mar 2022 09:19




Tgt Ion:128 Resp: 41433


Ion Ratio Lower Upper

128 100

64 38.3 49.4 74.2#

130 35.0 26.8 40.2









#13
2-Methylphenol
Concen: 4.389 ng/ul
RT: 8.431 min Scan# 883
Delta R.T. -0.000 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19




Instrument :
BNA_M
ClientSampleId :
SSTD005039


Tgt Ion:108 Resp: 34534
Ion Ratio Lower Upper
108 100
107 90.8 70.2 105.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 4.009 ng/ul
RT: 8.525 min Scan# 899
Delta R.T. -0.000 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19

Tgt Ion: 45 Resp: 34498
Ion Ratio Lower Upper
45 100
77 21.1 9.9 14.9#
79 17.0 8.3 12.5#









#16
Acetophenone
Concen: 4.617 ng/ul
RT: 8.813 min Scan# 948
Delta R.T. -0.000 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19




Instrument :
BNA_M
ClientSampleId :
SSTD005039


Tgt Ion:105 Resp: 62675
Ion Ratio Lower Upper
105 100
77 74.3 73.1 109.7
51 24.6 30.5 45.7#



#17
N-Nitroso-di-n-propylamine
Concen: 4.542 ng/ul
RT: 8.795 min Scan# 945
Delta R.T. -0.006 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19

Tgt Ion: 70 Resp: 28123
Ion Ratio Lower Upper
70 100
42 53.0 56.9 85.3#
101 10.0 6.9 10.3
130 23.3 13.8 20.8#









#18
4-Methylphenol
Concen: 4.593 ng/ul
RT: 8.754 min Scan# 918
Delta R.T. -0.006 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19




Instrument :
BNA_M
ClientSampleId :
SSTD005039


Tgt Ion:108 Resp: 39190
Ion Ratio Lower Upper
108 100
107 108.2 87.4 131.2



#19
Hexachloroethane
Concen: 4.709 ng/ul
RT: 9.084 min Scan# 994
Delta R.T. -0.000 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19

Tgt Ion:117 Resp: 17303
Ion Ratio Lower Upper
117 100
201 131.7 69.6 104.4#
199 81.4 40.6 60.8#









#22
Nitrobenzene
Concen: 4.812 ng/ul
RT: 9.195 min Scan# 1013
Delta R.T. -0.000 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19




Instrument :
BNA_M
ClientSampleId :
SSTD005039


Tgt Ion: 77 Resp: 44270
Ion Ratio Lower Upper
77 100
123 47.8 29.0 43.6#
65 12.7 11.4 17.2



#23
Isophorone
Concen: 4.346 ng/ul
RT: 9.725 min Scan# 1103
Delta R.T. -0.000 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19

Tgt Ion: 82 Resp: 73719
Ion Ratio Lower Upper
82 100
95 7.6 6.5 9.7
138 19.8 11.8 17.8#









#25
2-Nitrophenol
Concen: 4.075 ng/ul
RT: 9.907 min Scan# 1134
Delta R.T. -0.006 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19




Instrument :
BNA_M
ClientSampleId :
SSTD005039


Tgt Ion:139 Resp: 20508
Ion Ratio Lower Upper
139 100
65 45.5 59.0 88.6#
109 36.8 37.2 55.8#



#26
2,4-Dimethylphenol
Concen: 4.568 ng/ul
RT: 9.972 min Scan# 1145
Delta R.T. -0.000 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19

Tgt Ion:107 Resp: 46048
Ion Ratio Lower Upper
107 100
121 50.6 34.8 52.2
122 84.1 60.0 90.0





#27

Bis(2-Chloroethoxy)methane

Concen: 4.868 ng/ul

RT: 10.207 min Scan# 1185

Delta R.T. -0.006 min

Lab File: BM034143.D

Acq: 08 Mar 2022 09:19

Instrument :

BNA_M

ClientSampleId :

SSTD005039





Tgt Ion: 93 Resp: 50061

Ion Ratio Lower Upper

93 100

95 35.0 25.3 37.9

123 11.5 7.9 11.9



#29

2,4-Dichlorophenol

Concen: 4.429 ng/ul

RT: 10.448 min Scan# 1226

Delta R.T. -0.000 min

Lab File: BM034143.D

Acq: 08 Mar 2022 09:19




Tgt Ion:162 Resp: 41965


Ion Ratio Lower Upper

162 100

164 67.6 50.0 75.0

98 30.8 30.6 45.8









#30
Naphthalene
Concen: 4.984 ng/ul
RT: 10.854 min Scan# 1295
Delta R.T. -0.006 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19




Instrument :
BNA_M
ClientSampleId :
SSTD005039


Tgt Ion:128 Resp: 141806
Ion Ratio Lower Upper
128 100
129 10.6 9.1 13.7
127 12.7 10.2 15.4



#32
4-Chloroaniline
Concen: 4.590 ng/ul
RT: 10.960 min Scan# 1313
Delta R.T. -0.000 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19

Tgt Ion:127 Resp: 56452
Ion Ratio Lower Upper
127 100
129 30.0 25.4 38.0









#33
Hexachlorobutadiene
Concen: 4.848 ng/ul
RT: 11.154 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19




Instrument :
BNA_M
ClientSampleId :
SSTD005039


Tgt Ion:225 Resp: 35279
Ion Ratio Lower Upper
225 100
223 66.6 52.2 78.2
227 57.4 52.8 79.2



#34
Caprolactam
Concen: 3.618 ng/ul
RT: 11.707 min Scan# 1440
Delta R.T. -0.012 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19

Tgt Ion:113 Resp: 9309
Ion Ratio Lower Upper
113 100
55 103.2 157.9 236.9#
56 92.2 131.8 197.6#









#35
4-Chloro-3-methylphenol
Concen: 4.334 ng/ul
RT: 12.077 min Scan# 1503
Delta R.T. -0.000 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19




Instrument :
BNA_M
ClientSampleId :
SSTD005039


Tgt Ion:107 Resp: 38449
Ion Ratio Lower Upper
107 100
144 27.4 17.1 25.7#
142 88.0 49.4 74.2#



#36
2-Methylnaphthalene
Concen: 4.652 ng/ul
RT: 12.460 min Scan# 1568
Delta R.T. -0.006 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19

Tgt Ion:142 Resp: 96632
Ion Ratio Lower Upper
142 100
141 86.2 71.1 106.7









#37
1-Methylnaphthalene
Concen: 4.697 ng/ul
RT: 12.677 min Scan# 1605
Delta R.T. -0.006 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19




Instrument :
BNA_M
ClientSampleId :
SSTD005039


Tgt Ion:142 Resp: 98890
Ion Ratio Lower Upper
142 100
141 91.1 75.5 113.3
116 4.3 4.0 6.0



#39
1,2,4,5-Tetrachlorobenzene
Concen: 4.819 ng/ul
RT: 12.830 min Scan# 1631
Delta R.T. -0.000 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19

Tgt Ion:216 Resp: 63410
Ion Ratio Lower Upper
216 100
214 78.0 61.3 91.9
179 20.7 18.0 27.0
108 14.1 16.6 25.0#





#40

Hexachlorocyclopentadiene

Concen: 4.002 ng/ul

RT: 12.819 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM034143.D

Acq: 08 Mar 2022 09:19

Instrument :

BNA_M

ClientSampleId :

SSTD005039





Tgt Ion:237 Resp: 35912

Ion Ratio Lower Upper

237 100

235 66.2 50.6 76.0

272 13.8 9.6 14.4



#41

2,4,6-Trichlorophenol

Concen: 4.175 ng/ul

RT: 13.060 min Scan# 1670

Delta R.T. -0.006 min

Lab File: BM034143.D

Acq: 08 Mar 2022 09:19




Tgt Ion:196 Resp: 33558


Ion Ratio Lower Upper

196 100

198 99.4 78.9 118.3

200 28.9 23.7 35.5









#42
2,4,5-Trichlorophenol
Concen: 4.249 ng/ul
RT: 13.130 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19




Instrument :
BNA_M
ClientSampleId :
SSTD005039


Tgt Ion:196 Resp: 37563
Ion Ratio Lower Upper
196 100
198 96.1 82.2 123.2
200 31.9 27.9 41.9



#43
1,1'-Biphenyl
Concen: 4.794 ng/ul
RT: 13.466 min Scan# 1739
Delta R.T. -0.000 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19

Tgt Ion:154 Resp: 139968
Ion Ratio Lower Upper
154 100
153 38.8 33.1 49.7
76 8.9 10.7 16.1#









#44
2-Chloronaphthalene
Concen: 4.814 ng/ul
RT: 13.507 min Scan# 1746
Delta R.T. -0.000 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19




Instrument :
BNA_M
ClientSampleId :
SSTD005039


Tgt Ion:162 Resp: 109187
Ion Ratio Lower Upper
162 100
164 34.4 26.2 39.4
127 36.6 33.7 50.5



#45
2-Nitroaniline
Concen: 3.943 ng/ul
RT: 13.701 min Scan# 1779
Delta R.T. -0.000 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19

Tgt Ion: 65 Resp: 20136
Ion Ratio Lower Upper
65 100
92 81.6 47.8 71.6#
138 135.1 61.0 91.6#









#47
Dimethylphthalate
Concen: 4.634 ng/ul
RT: 14.077 min Scan# 1843
Delta R.T. -0.006 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19




Instrument :
BNA_M
ClientSampleId :
SSTD005039


Tgt Ion:163 Resp: 134446
Ion Ratio Lower Upper
163 100
194 4.4 3.3 4.9
164 9.9 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 4.032 ng/ul
RT: 14.195 min Scan# 1863
Delta R.T. -0.000 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19

Tgt Ion:165 Resp: 22890
Ion Ratio Lower Upper
165 100
89 55.2 55.7 83.5#
121 20.6 20.0 30.0









#50
Acenaphthylene
Concen: 4.563 ng/ul
RT: 14.348 min Scan# 1889
Delta R.T. -0.006 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19




Instrument :
BNA_M
ClientSampleId :
SSTD005039


Tgt Ion:152 Resp: 163558
Ion Ratio Lower Upper
152 100
151 19.7 15.8 23.8
153 13.2 10.1 15.1



#51
3-Nitroaniline
Concen: 3.970 ng/ul
RT: 14.519 min Scan# 1918
Delta R.T. -0.000 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19

Tgt Ion:138 Resp: 21922
Ion Ratio Lower Upper
138 100
108 11.7 10.8 16.2
92 109.5 110.6 166.0#









#52
Acenaphthene
Concen: 4.751 ng/ul
RT: 14.689 min Scan# 1947
Delta R.T. -0.006 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19




Instrument :
BNA_M
ClientSampleId :
SSTD005039


Tgt Ion:153 Resp: 113011
Ion Ratio Lower Upper
153 100
152 47.6 39.4 59.2
154 88.2 72.7 109.1



#53
2,4-Dinitrophenol
Concen: 2.614 ng/ul
RT: 14.724 min Scan# 1953
Delta R.T. -0.000 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19

Tgt Ion:184 Resp: 9845
Ion Ratio Lower Upper
184 100
63 48.8 73.8 110.8#
154 54.6 63.5 95.3#









#55
4-Nitrophenol
Concen: 3.515 ng/ul
RT: 14.813 min Scan# 1969
Delta R.T. -0.006 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19




Instrument :
BNA_M
ClientSampleId :
SSTD005039


Tgt Ion:109 Resp: 16425
Ion Ratio Lower Upper
109 100
139 122.0 77.6 116.4#
65 109.9 112.2 168.4#



#56
Dibenzofuran
Concen: 4.862 ng/ul
RT: 15.024 min Scan# 2004
Delta R.T. -0.000 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19

Tgt Ion:168 Resp: 169868
Ion Ratio Lower Upper
168 100
139 37.7 34.0 51.0









#57
2,4-Dinitrotoluene
Concen: 4.057 ng/ul
RT: 14.977 min Scan# 1996
Delta R.T. -0.000 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19




Instrument :
BNA_M
ClientSampleId :
SSTD005039


Tgt Ion:165 Resp: 33473
Ion Ratio Lower Upper
165 100
63 35.9 51.8 77.8#
182 2.7 2.6 4.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 4.250 ng/ul
RT: 15.248 min Scan# 2042
Delta R.T. -0.000 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19

Tgt Ion:232 Resp: 32511
Ion Ratio Lower Upper
232 100
131 38.4 41.4 62.2#
130 2.3 2.0 3.0
166 27.3 27.4 41.0#









#59
Diethylphthalate
Concen: 4.430 ng/ul
RT: 15.436 min Scan# 2074
Delta R.T. -0.006 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19




Instrument :
BNA_M
ClientSampleId :
SSTD005039


Tgt Ion:149 Resp: 128316
Ion Ratio Lower Upper
149 100
177 22.8 17.1 25.7
150 12.1 9.9 14.9



#61
Fluorene
Concen: 4.702 ng/ul
RT: 15.671 min Scan# 2114
Delta R.T. -0.000 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19

Tgt Ion:166 Resp: 136990
Ion Ratio Lower Upper
166 100
165 99.7 77.9 116.9
167 12.9 10.3 15.5









#62
4-Chlorophenyl-phenylether
Concen: 4.882 ng/ul
RT: 15.665 min Scan# 2113
Delta R.T. -0.000 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19




Instrument :
BNA_M
ClientSampleId :
SSTD005039


Tgt Ion:204 Resp: 76067
Ion Ratio Lower Upper
204 100
206 31.1 27.8 41.8
141 54.1 54.9 82.3#



#63
4-Nitroaniline
Concen: 3.830 ng/ul
RT: 15.671 min Scan# 2114
Delta R.T. -0.006 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19

Tgt Ion:138 Resp: 21639
Ion Ratio Lower Upper
138 100
92 45.6 48.0 72.0#
108 79.7 89.5 134.3#









#66
4,6-Dinitro-2-methylphenol
Concen: 3.280 ng/ul
RT: 15.736 min Scan# 2125
Delta R.T. -0.006 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19




Instrument :
BNA_M
ClientSampleId :
SSTD005039


Tgt Ion:198 Resp: 18347
Ion Ratio Lower Upper
198 100
182 4.6 4.2 6.2
77 24.4 26.3 39.5#



#67
N-Nitrosodiphenylamine
Concen: 4.525 ng/ul
RT: 15.871 min Scan# 2148
Delta R.T. -0.006 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19

Tgt Ion:169 Resp: 115178
Ion Ratio Lower Upper
169 100
168 66.8 52.9 79.3
167 35.6 29.9 44.9









#68
4-Bromophenyl-phenylether
Concen: 4.377 ng/ul
RT: 16.560 min Scan# 218
Delta R.T. -0.000 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19




Instrument :
BNA_M
ClientSampleId :
SSTD005039


Tgt Ion:248 Resp: 41862
Ion Ratio Lower Upper
248 100
250 95.2 79.0 118.4
141 66.2 69.7 104.5#



#69
Hexachlorobenzene
Concen: 4.807 ng/ul
RT: 16.677 min Scan# 2285
Delta R.T. -0.006 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19

Tgt Ion:284 Resp: 55755
Ion Ratio Lower Upper
284 100
142 32.5 34.3 51.5#
249 30.6 26.1 39.1









#70
 Atrazine
 Concen: 3.900 ng/ul
 RT: 16.818 min Scan# 21
 Delta R.T. -0.006 min
 Lab File: BM034143.D
 Acq: 08 Mar 2022 09:19




Instrument :
 BNA_M
 ClientSampleId :
 SSTD005039


Tgt Ion:200 Resp: 40740
 Ion Ratio Lower Upper
 200 100
 173 23.9 20.6 30.8
 215 48.0 36.9 55.3



#71
 Pentachlorophenol
 Concen: 3.471 ng/ul
 RT: 17.012 min Scan# 2342
 Delta R.T. -0.006 min
 Lab File: BM034143.D
 Acq: 08 Mar 2022 09:19

Tgt Ion:266 Resp: 26214
 Ion Ratio Lower Upper
 266 100
 264 61.0 49.1 73.7
 268 65.3 47.8 71.8









#72
Phenanthrene
Concen: 4.731 ng/ul
RT: 17.407 min Scan# 2410
Delta R.T. -0.006 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19




Instrument :
BNA_M
ClientSampleId :
SSTD005039


Tgt Ion:178 Resp: 224944
Ion Ratio Lower Upper
178 100
179 15.2 11.7 17.5
176 17.9 15.1 22.7



#74
Anthracene
Concen: 4.617 ng/ul
RT: 17.501 min Scan# 2425
Delta R.T. -0.000 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19

Tgt Ion:178 Resp: 222290
Ion Ratio Lower Upper
178 100
179 14.9 12.3 18.5
176 17.9 14.5 21.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 4.925 ng/uL
RT: 13.430 min Scan# 1733
Delta R.T. -0.006 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19




Instrument :
BNA_M
ClientSampleId :
SSTD005039


Tgt Ion:216 Resp: 68141
Ion Ratio Lower Upper
216 100
214 74.1 65.4 98.0
218 46.4 38.8 58.2



#76
Pentachlorobenzene
Concen: 4.749 ng/uL
RT: 14.948 min Scan# 1991
Delta R.T. -0.000 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19

Tgt Ion:250 Resp: 65004
Ion Ratio Lower Upper
250 100
248 61.9 51.6 77.4
252 62.7 52.0 78.0









#77
 Carbazole
 Concen: 4.355 ng/ul
 RT: 17.765 min Scan# 2470
 Delta R.T. -0.000 min
 Lab File: BM034143.D
 Acq: 08 Mar 2022 09:19




Instrument :
 BNA_M
 ClientSampleId :
 SSTD005039


Tgt Ion:167 Resp: 182762
 Ion Ratio Lower Upper
 167 100
 166 21.0 16.9 25.3
 139 12.7 10.5 15.7



#78
 Di-n-butylphthalate
 Concen: 3.873 ng/ul
 RT: 18.336 min Scan# 2567
 Delta R.T. -0.000 min
 Lab File: BM034143.D
 Acq: 08 Mar 2022 09:19

Tgt Ion:149 Resp: 189496
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.0 10.6
 104 5.6 5.1 7.7









#80
Fluoranthene
Concen: 4.530 ng/ul
RT: 19.418 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19




Instrument :
BNA_M
ClientSampleId :
SSTD005039


Tgt Ion:202 Resp: 263094
Ion Ratio Lower Upper
202 100
101 6.9 8.2 12.4#
100 5.6 6.9 10.3#



#82
Pyrene
Concen: 4.710 ng/ul
RT: 19.777 min Scan# 2812
Delta R.T. -0.006 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19

Tgt Ion:202 Resp: 282529
Ion Ratio Lower Upper
202 100
101 8.5 9.8 14.6#
100 7.8 8.6 13.0#









#83
Butylbenzylphthalate
Concen: 3.744 ng/ul
RT: 20.677 min Scan# 2965
Delta R.T. -0.000 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19




Instrument :
BNA_M
ClientSampleId :
SSTD005039


Tgt Ion:149 Resp: 79196
Ion Ratio Lower Upper
149 100
91 61.9 69.3 103.9#
206 24.7 16.1 24.1#



#84
3,3'-Dichlorobenzidine
Concen: 3.534 ng/ul
RT: 21.453 min Scan# 3097
Delta R.T. -0.000 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19

Tgt Ion:252 Resp: 69846
Ion Ratio Lower Upper
252 100
254 64.0 49.4 74.2
126 8.1 8.2 12.4#





#85

Benzo(a)anthracene

Concen: 4.835 ng/ul

RT: 21.524 min Scan# 3109

Delta R.T. -0.000 min

Lab File: BM034143.D

Acq: 08 Mar 2022 09:19

Instrument :

BNA_M

ClientSampleId :

SSTD005039





Tgt Ion:228 Resp: 275597

Ion Ratio Lower Upper

228 100

229 20.2 15.0 22.4

226 25.6 21.4 32.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 3.525 ng/ul

RT: 21.453 min Scan# 3097

Delta R.T. -0.000 min

Lab File: BM034143.D

Acq: 08 Mar 2022 09:19




Tgt Ion:149 Resp: 114968


Ion Ratio Lower Upper

149 100

167 28.7 21.8 32.8

279 5.4 3.2 4.8#





#87

Chrysene

Concen: 4.919 ng/ul

RT: 21.577 min Scan# 3118

Delta R.T. -0.000 min

Lab File: BM034143.D

Acq: 08 Mar 2022 09:19

Instrument :

BNA_M

ClientSampleId :

SSTD005039





Tgt Ion:228 Resp: 275270

Ion Ratio Lower Upper

228 100

226 29.4 24.5 36.7

229 19.3 15.9 23.9



#89

Di-n-octyl phthalate

Concen: 3.350 ng/ul




RT: 22.394 min Scan# 3257


Delta R.T. -0.000 min

Lab File: BM034143.D

Acq: 08 Mar 2022 09:19

Tgt Ion:149 Resp: 186421









#90
Benzo(b)fluoranthene
Concen: 4.731 ng/ul
RT: 23.236 min Scan# 3400
Delta R.T. -0.006 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19




Instrument :
BNA_M
ClientSampleId :
SSTD005039


Tgt Ion:252 Resp: 271294
Ion Ratio Lower Upper
252 100
253 22.6 17.7 26.5
125 8.0 7.8 11.6



#91
Benzo(k)fluoranthene
Concen: 5.025 ng/ul
RT: 23.236 min Scan# 3400
Delta R.T. -0.053 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19

Tgt Ion:252 Resp: 271294
Ion Ratio Lower Upper
252 100
253 22.6 17.6 26.4
125 8.0 7.4 11.2





#93

Benzo(a)pyrene

Concen: 4.685 ng/ul

RT: 23.871 min Scan# 3

Delta R.T. -0.006 min

Lab File: BM034143.D

Acq: 08 Mar 2022 09:19

Instrument :

BNA_M

ClientSampleId :

SSTD005039





Tgt Ion:252 Resp: 257220

Ion Ratio Lower Upper

252 100

253 22.9 17.4 26.0

125 7.3 7.8 11.6#



#94

Indeno(1,2,3-cd)pyrene

Concen: 4.494 ng/ul

RT: 26.535 min Scan# 3961

Delta R.T. -0.006 min

Lab File: BM034143.D

Acq: 08 Mar 2022 09:19




Tgt Ion:276 Resp: 274592


Ion Ratio Lower Upper

276 100

138 18.7 16.5 24.7

277 24.8 20.6 31.0









#95
Dibenzo(a,h)anthracene
Concen: 4.471 ng/ul
RT: 26.547 min Scan# 3963
Delta R.T. -0.012 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19

Instrument :
BNA_M
ClientSampleId :
SSTD005039

Tgt Ion:278 Resp: 236494
Ion Ratio Lower Upper
278 100
139 9.9 11.0 16.4#
279 22.6 18.1 27.1



#96
Benzo(g,h,i)perylene
Concen: 4.486 ng/ul
RT: 27.312 min Scan# 4093
Delta R.T. -0.012 min
Lab File: BM034143.D
Acq: 08 Mar 2022 09:19

Tgt Ion:276 Resp: 231496
Ion Ratio Lower Upper
276 100
138 15.4 15.6 23.4#
277 24.1 18.7 28.1

