

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030822\
 Data File : BM034145.D
 Acq On : 08 Mar 2022 10:31
 Operator : CG/JU
 Sample : SST02041
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST020041

Quant Time: Mar 08 11:32:16 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM030822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 08 11:29:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.995	152	140002	20.000	ng/ul	0.00
20) Naphthalene-d8	10.807	136	586283	20.000	ng/ul	# 0.00
38) Acenaphthene-d10	14.630	164	409167	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.371	188	910042	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.542	240	1040532	20.000	ng/ul	0.00
88) Perylene-d12	23.989	264	1037318	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.372	96	27611	9.093	ng/uL	0.00
4) Pyridine-d5	3.790	84	189292	23.724	ng/ul	0.00
7) Phenol-d5	7.143	99	233860	23.388	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.313	67	129889	23.399	ng/ul	0.00
11) 2-Chlorophenol-d4	7.519	132	204508	23.725	ng/ul	0.00
15) 4-Methylphenol-d8	8.695	113	200996	23.617	ng/ul	0.00
21) Nitrobenzene-d5	9.154	128	98839	22.319	ng/ul	0.00
24) 2-Nitrophenol-d4	9.884	143	109075	21.668	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.419	165	225105	22.097	ng/ul	0.00
31) 4-Chloroaniline-d4	10.936	131	311646	24.258	ng/ul	0.00
46) Dimethylphthalate-d6	14.036	166	678872	21.929	ng/ul	0.00
49) Acenaphthylene-d8	14.324	160	852776	22.617	ng/ul	0.00
54) 4-Nitrophenol-d4	14.807	143	125139	23.599	ng/ul	0.00
60) Fluorene-d10	15.618	176	601726	22.488	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.724	200	117234	19.732	ng/ul	0.00
73) Anthracene-d10	17.465	188	984738	22.377	ng/ul	0.00
81) Pyrene-d10	19.754	212	1210216	21.575	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.830	264	1218274	23.045	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.408	88	28636	9.608	ng/uL#	73
5) Pyridine	3.813	79	191143	23.474	ng/ul#	68
6) Benzaldehyde	7.260	77	220	0.039	ng/ul	86
8) Phenol	7.166	94	236331	24.210	ng/ul#	81
10) Bis(2-Chloroethyl)ether	7.407	93	188662	24.536	ng/ul#	84
12) 2-Chlorophenol	7.554	128	208331	24.352	ng/ul#	79
13) 2-Methylphenol	8.431	108	185932	24.236	ng/ul	96
14) 2,2'-oxybis(1-Chloropr...	8.525	45	167268	19.935	ng/ul#	80
16) Acetophenone	8.813	105	316782	23.935	ng/ul#	81
17) N-Nitroso-di-n-propyla...	8.801	70	151363	25.073	ng/ul#	79
18) 4-Methylphenol	8.760	108	205451	24.693	ng/ul	97
19) Hexachloroethane	9.084	117	85859	23.966	ng/ul#	70
22) Nitrobenzene	9.195	77	227320	23.930	ng/ul#	88
23) Isophorone	9.725	82	411844	23.516	ng/ul#	91
25) 2-Nitrophenol	9.913	139	119004	22.898	ng/ul#	73
26) 2,4-Dimethylphenol	9.972	107	239456	23.005	ng/ul	91
27) Bis(2-Chloroethoxy)met...	10.213	93	257196	24.221	ng/ul#	96
29) 2,4-Dichlorophenol	10.448	162	223587	22.853	ng/ul	95
30) Naphthalene	10.860	128	688354	23.428	ng/ul	98
32) 4-Chloroaniline	10.960	127	315987	24.881	ng/ul	96
33) Hexachlorobutadiene	11.154	225	165944	22.083	ng/ul	99
34) Caprolactam	11.719	113	38965	14.665	ng/ul#	49
35) 4-Chloro-3-methylphenol	12.078	107	222653	24.304	ng/ul#	74
36) 2-Methylnaphthalene	12.466	142	496316	23.141	ng/ul	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030822\
 Data File : BM034145.D
 Acq On : 08 Mar 2022 10:31
 Operator : CG/JU
 Sample : SSTD02041
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020041

Quant Time: Mar 08 11:32:16 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM030822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 08 11:29:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.683	142	514355	23.658	ng/ul#	96
39) 1,2,4,5-Tetrachloroben...	12.830	216	310624	22.851	ng/ul#	96
40) Hexachlorocyclopentadiene	12.819	237	203641	21.968	ng/ul	99
41) 2,4,6-Trichlorophenol	13.066	196	193309	23.280	ng/ul	99
42) 2,4,5-Trichlorophenol	13.136	196	215536	23.601	ng/ul	95
43) 1,1'-Biphenyl	13.466	154	698144	23.148	ng/ul#	95
44) 2-Chloronaphthalene	13.507	162	545134	23.264	ng/ul	95
45) 2-Nitroaniline	13.701	65	130701	24.777	ng/ul#	65
47) Dimethylphthalate	14.083	163	688761	22.982	ng/ul	99
48) 2,6-Dinitrotoluene	14.195	165	138217	23.566	ng/ul#	85
50) Acenaphthylene	14.354	152	867742	23.433	ng/ul	99
51) 3-Nitroaniline	14.519	138	95698	16.775	ng/ul#	78
52) Acenaphthene	14.695	153	563626	22.935	ng/ul	97
53) 2,4-Dinitrophenol	14.724	184	75582	19.425	ng/ul#	70
55) 4-Nitrophenol	14.819	109	109599	22.706	ng/ul#	75
56) Dibenzofuran	15.024	168	833696	23.100	ng/ul	94
57) 2,4-Dinitrotoluene	14.977	165	204665	24.015	ng/ul#	68
58) 2,3,4,6-Tetrachlorophenol	15.248	232	186907	23.651	ng/ul#	88
59) Diethylphthalate	15.442	149	686482	22.944	ng/ul	99
61) Fluorene	15.671	166	684693	22.748	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.666	204	366094	22.742	ng/ul	88
63) 4-Nitroaniline	15.677	138	84213	14.428	ng/ul#	76
66) 4,6-Dinitro-2-methylph...	15.742	198	118199	20.357	ng/ul#	82
67) N-Nitrosodiphenylamine	15.877	169	594744	22.507	ng/ul	99
68) 4-Bromophenyl-phenylether	16.560	248	217256	21.882	ng/ul#	88
69) Hexachlorobenzene	16.683	284	285602	23.718	ng/ul#	84
70) Atrazine	16.824	200	234575	21.631	ng/ul	96
71) Pentachlorophenol	17.018	266	169908	21.668	ng/ul	98
72) Phenanthrene	17.413	178	1125476	22.800	ng/ul	99
74) Anthracene	17.501	178	1139924	22.804	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.436	216	337064	23.464	ng/uL	97
76) Pentachlorobenzene	14.948	250	315072	22.173	ng/uL	100
77) Carbazole	17.765	167	1009501	23.172	ng/ul	99
78) Di-n-butylphthalate	18.336	149	1134067	22.328	ng/ul	98
80) Fluoranthene	19.424	202	1414820	21.830	ng/ul#	92
82) Pyrene	19.783	202	1464324	21.872	ng/ul#	91
83) Butylbenzylphthalate	20.677	149	525280	22.251	ng/ul#	77
84) 3,3'-Dichlorobenzidine	21.453	252	405777	18.396	ng/ul	96
85) Benzo(a)anthracene	21.524	228	1505434	23.668	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.453	149	781800	21.480	ng/ul	100
87) Chrysene	21.577	228	1478501	23.673	ng/ul	99
89) Di-n-octyl phthalate	22.395	149	1301296	20.384	ng/ul	100
90) Benzo(b)fluoranthene	23.242	252	1579792	24.016	ng/ul#	97
91) Benzo(k)fluoranthene	23.289	252	1463883	23.640	ng/ul#	98
93) Benzo(a)pyrene	24.036	252	9196	0.146	ng/ul	96
94) Indeno(1,2,3-cd)pyrene	26.541	276	1535840	21.913	ng/ul	96
95) Dibenzo(a,h)anthracene	26.559	278	1314701	21.668	ng/ul	96
96) Benzo(g,h,i)perylene	27.324	276	1276636	21.567	ng/ul	96





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 9.608 ng/uL
RT: 3.408 min Scan# 29
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31




Instrument :
BNA_M
ClientSampleId :
SSTD020041


Tgt Ion: 88 Resp: 28636
Ion Ratio Lower Upper
88 100
43 26.5 36.2 54.4#
58 61.8 68.5 102.7#



#5
Pyridine
Concen: 23.474 ng/uL
RT: 3.813 min Scan# 98
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31

Tgt Ion: 79 Resp: 191143
Ion Ratio Lower Upper
79 100
52 59.6 78.1 117.1#
51 32.7 36.0 54.0#









#6
Benzaldehyde
Concen: 0.039 ng/ul
RT: 7.260 min Scan# 61
Delta R.T. 0.141 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31




Instrument :
BNA_M
ClientSampleId :
SSTD020041


Tgt Ion: 77	Resp: 220
Ion Ratio	Lower Upper
77 100	
105 69.0	65.6 98.4
106 64.3	60.6 90.8



#8
Phenol
Concen: 24.210 ng/ul
RT: 7.166 min Scan# 668
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31

Tgt Ion: 94	Resp: 236331
Ion Ratio	Lower Upper
94 100	
65 29.0	31.7 47.5#
66 41.6	45.4 68.0#









#10
Bis(2-Chloroethyl)ether
Concen: 24.536 ng/ul
RT: 7.407 min Scan# 709
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31




Instrument :
BNA_M
ClientSampleId :
SSTD020041


Tgt Ion: 93 Resp: 188662
Ion Ratio Lower Upper
93 100
63 69.5 71.7 107.5#
95 32.8 26.3 39.5



#12
2-Chlorophenol
Concen: 24.352 ng/ul
RT: 7.554 min Scan# 734
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31

Tgt Ion:128 Resp: 208331
Ion Ratio Lower Upper
128 100
64 38.2 49.4 74.2#
130 31.8 26.8 40.2









#13
2-Methylphenol
Concen: 24.236 ng/ul
RT: 8.431 min Scan# 883
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31




Instrument :
BNA_M
ClientSampleId :
SSTD020041


Tgt Ion:108 Resp: 185932
Ion Ratio Lower Upper
108 100
107 91.8 70.2 105.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 19.935 ng/ul
RT: 8.525 min Scan# 899
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31

Tgt Ion: 45 Resp: 167268
Ion Ratio Lower Upper
45 100
77 22.2 9.9 14.9#
79 15.4 8.3 12.5#









#16
Acetophenone
Concen: 23.935 ng/ul
RT: 8.813 min Scan# 948
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31




Instrument :
BNA_M
ClientSampleId :
SSTD020041


Tgt Ion:105 Resp: 316782
Ion Ratio Lower Upper
105 100
77 75.5 73.1 109.7
51 23.1 30.5 45.7#



#17
N-Nitroso-di-n-propylamine
Concen: 25.073 ng/ul
RT: 8.801 min Scan# 946
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31

Tgt Ion: 70 Resp: 151363
Ion Ratio Lower Upper
70 100
42 49.4 56.9 85.3#
101 10.2 6.9 10.3
130 21.4 13.8 20.8#









#18
4-Methylphenol
Concen: 24.693 ng/ul
RT: 8.760 min Scan# 939
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31




Instrument :
BNA_M
ClientSampleId :
SSTD020041


Tgt Ion:108 Resp: 205451
Ion Ratio Lower Upper
108 100
107 112.1 87.4 131.2



#19
Hexachloroethane
Concen: 23.966 ng/ul
RT: 9.084 min Scan# 994
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31

Tgt Ion:117 Resp: 85859
Ion Ratio Lower Upper
117 100
201 113.5 69.6 104.4#
199 73.4 40.6 60.8#









#22
Nitrobenzene
Concen: 23.930 ng/ul
RT: 9.195 min Scan# 1013
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31




Instrument :
BNA_M
ClientSampleId :
SSTD020041


Tgt Ion: 77 Resp: 227320
Ion Ratio Lower Upper
77 100
123 45.1 29.0 43.6#
65 13.1 11.4 17.2



#23
Isophorone
Concen: 23.516 ng/ul
RT: 9.725 min Scan# 1103
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31

Tgt Ion: 82 Resp: 411844
Ion Ratio Lower Upper
82 100
95 8.7 6.5 9.7
138 20.1 11.8 17.8#









#25
2-Nitrophenol
Concen: 22.898 ng/ul
RT: 9.913 min Scan# 1135
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31




Instrument :
BNA_M
ClientSampleId :
SSTD020041


Tgt Ion:139 Resp: 119004
Ion Ratio Lower Upper
139 100
65 47.8 59.0 88.6#
109 32.3 37.2 55.8#



#26
2,4-Dimethylphenol
Concen: 23.005 ng/ul
RT: 9.972 min Scan# 1145
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31

Tgt Ion:107 Resp: 239456
Ion Ratio Lower Upper
107 100
121 49.8 34.8 52.2
122 82.8 60.0 90.0









#27
 Bis(2-Chloroethoxy)methane
 Concen: 24.221 ng/ul
 RT: 10.213 min Scan# 1186
 Delta R.T. 0.000 min
 Lab File: BM034145.D
 Acq: 08 Mar 2022 10:31




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020041


Tgt Ion: 93 Resp: 257196
 Ion Ratio Lower Upper
 93 100
 95 32.8 25.3 37.9
 123 12.8 7.9 11.9#



#29
 2,4-Dichlorophenol
 Concen: 22.853 ng/ul
 RT: 10.448 min Scan# 1226
 Delta R.T. 0.000 min
 Lab File: BM034145.D
 Acq: 08 Mar 2022 10:31

Tgt Ion:162 Resp: 223587
 Ion Ratio Lower Upper
 162 100
 164 64.5 50.0 75.0
 98 32.5 30.6 45.8









#30
Naphthalene
Concen: 23.428 ng/ul
RT: 10.860 min Scan# 1296
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31




Instrument :
BNA_M
ClientSampleId :
SSTD020041


Tgt Ion:128 Resp: 688354
Ion Ratio Lower Upper
128 100
129 10.4 9.1 13.7
127 13.1 10.2 15.4



#32
4-Chloroaniline
Concen: 24.881 ng/ul
RT: 10.960 min Scan# 1313
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31

Tgt Ion:127 Resp: 315987
Ion Ratio Lower Upper
127 100
129 34.1 25.4 38.0









#33
Hexachlorobutadiene
Concen: 22.083 ng/u1
RT: 11.154 min Scan# 1346
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31




Instrument :
BNA_M
ClientSampleId :
SSTD020041


Tgt Ion:225 Resp: 165944
Ion Ratio Lower Upper
225 100
223 64.7 52.2 78.2
227 64.5 52.8 79.2



#34
Caprolactam
Concen: 14.665 ng/u1
RT: 11.719 min Scan# 1442
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31

Tgt Ion:113 Resp: 38965
Ion Ratio Lower Upper
113 100
55 117.9 157.9 236.9#
56 99.5 131.8 197.6#









#35
 4-Chloro-3-methylphenol
 Concen: 24.304 ng/ul
 RT: 12.078 min Scan# 1503
 Delta R.T. 0.000 min
 Lab File: BM034145.D
 Acq: 08 Mar 2022 10:31




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020041


Tgt Ion:107 Resp: 222653
 Ion Ratio Lower Upper
 107 100
 144 28.5 17.1 25.7#
 142 85.2 49.4 74.2#



#36
 2-Methylnaphthalene
 Concen: 23.141 ng/ul
 RT: 12.466 min Scan# 1569
 Delta R.T. 0.000 min
 Lab File: BM034145.D
 Acq: 08 Mar 2022 10:31

Tgt Ion:142 Resp: 496316
 Ion Ratio Lower Upper
 142 100
 141 88.3 71.1 106.7





#37

1-Methylnaphthalene

Concen: 23.658 ng/ul

RT: 12.683 min Scan# 1606

Delta R.T. 0.000 min

Lab File: BM034145.D

Acq: 08 Mar 2022 10:31

Instrument :

BNA_M

ClientSampleId :

SSTD020041





Tgt Ion:142 Resp: 514355

Ion Ratio Lower Upper

142 100

141 90.9 75.5 113.3

116 4.0 4.0 6.0#



#39

1,2,4,5-Tetrachlorobenzene

Concen: 22.851 ng/ul

RT: 12.830 min Scan# 1631

Delta R.T. 0.000 min

Lab File: BM034145.D

Acq: 08 Mar 2022 10:31

Tgt Ion:216 Resp: 310624




Ion Ratio Lower Upper


216 100

214 78.1 61.3 91.9

179 21.1 18.0 27.0

108 15.3 16.6 25.0#









#40
Hexachlorocyclopentadiene
Concen: 21.968 ng/ul
RT: 12.819 min Scan# 1629
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31




Instrument :
BNA_M
ClientSampleId :
SSTD020041


Tgt Ion:237	Resp: 203641
Ion Ratio	Lower Upper
237 100	
235 62.9	50.6 76.0
272 13.4	9.6 14.4



#41
2,4,6-Trichlorophenol
Concen: 23.280 ng/ul
RT: 13.066 min Scan# 1671
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31

Tgt Ion:196	Resp: 193309
Ion Ratio	Lower Upper
196 100	
198 98.9	78.9 118.3
200 31.7	23.7 35.5









#42
2,4,5-Trichlorophenol
Concen: 23.601 ng/ul
RT: 13.136 min Scan# 1683
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31




Instrument :
BNA_M
ClientSampleId :
SSTD020041


Tgt Ion:196 Resp: 215536
Ion Ratio Lower Upper
196 100
198 98.0 82.2 123.2
200 31.5 27.9 41.9



#43
1,1'-Biphenyl
Concen: 23.148 ng/ul
RT: 13.466 min Scan# 1739
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31

Tgt Ion:154 Resp: 698144
Ion Ratio Lower Upper
154 100
153 39.6 33.1 49.7
76 9.7 10.7 16.1#





#44

2-Chloronaphthalene

Concen: 23.264 ng/ul

RT: 13.507 min Scan# 1746

Delta R.T. 0.000 min

Lab File: BM034145.D

Acq: 08 Mar 2022 10:31

Instrument :

BNA_M

ClientSampleId :

SSTD020041





Tgt Ion:162 Resp: 545134

Ion Ratio Lower Upper

162 100

164 34.5 26.2 39.4

127 37.8 33.7 50.5



#45

2-Nitroaniline

Concen: 24.777 ng/ul

RT: 13.701 min Scan# 1779

Delta R.T. 0.000 min

Lab File: BM034145.D

Acq: 08 Mar 2022 10:31




Tgt Ion: 65 Resp: 130701


Ion Ratio Lower Upper

65 100

92 71.3 47.8 71.6

138 119.3 61.0 91.6#









#47
Dimethylphthalate
Concen: 22.982 ng/ul
RT: 14.083 min Scan# 1844
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31




Instrument :
BNA_M
ClientSampleId :
SSTD020041


Tgt Ion:163 Resp: 688761
Ion Ratio Lower Upper
163 100
194 4.4 3.3 4.9
164 10.0 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 23.566 ng/ul
RT: 14.195 min Scan# 1863
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31

Tgt Ion:165 Resp: 138217
Ion Ratio Lower Upper
165 100
89 55.7 55.7 83.5
121 19.4 20.0 30.0#









#50
Acenaphthylene
Concen: 23.433 ng/ul
RT: 14.354 min Scan# 1890
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31




Instrument :
BNA_M
ClientSampleId :
SSTD020041


Tgt Ion:152 Resp: 867742
Ion Ratio Lower Upper
152 100
151 19.6 15.8 23.8
153 12.8 10.1 15.1



#51
3-Nitroaniline
Concen: 16.775 ng/ul
RT: 14.519 min Scan# 1918
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31

Tgt Ion:138 Resp: 95698
Ion Ratio Lower Upper
138 100
108 10.3 10.8 16.2#
92 109.5 110.6 166.0#





#52

Acenaphthene

Concen: 22.935 ng/ul

RT: 14.695 min Scan# 1948

Delta R.T. 0.000 min

Lab File: BM034145.D

Acq: 08 Mar 2022 10:31

Instrument :

BNA_M

ClientSampleId :

SSTD020041





Tgt Ion:153 Resp: 563626

Ion Ratio Lower Upper

153 100

152 46.6 39.4 59.2

154 89.0 72.7 109.1



#53

2,4-Dinitrophenol

Concen: 19.425 ng/ul

RT: 14.724 min Scan# 1953

Delta R.T. 0.000 min

Lab File: BM034145.D

Acq: 08 Mar 2022 10:31




Tgt Ion:184 Resp: 75582


Ion Ratio Lower Upper

184 100

63 55.5 73.8 110.8#

154 61.6 63.5 95.3#









#55
4-Nitrophenol
Concen: 22.706 ng/ul
RT: 14.819 min Scan# 1969
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31




Instrument :
BNA_M
ClientSampleId :
SSTD020041


Tgt Ion:109 Resp: 109599
Ion Ratio Lower Upper
109 100
139 117.4 77.6 116.4#
65 107.1 112.2 168.4#



#56
Dibenzofuran
Concen: 23.100 ng/ul
RT: 15.024 min Scan# 2004
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31

Tgt Ion:168 Resp: 833696
Ion Ratio Lower Upper
168 100
139 38.5 34.0 51.0









#57
2,4-Dinitrotoluene
Concen: 24.015 ng/ul
RT: 14.977 min Scan# 1996
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31




Instrument :
BNA_M
ClientSampleId :
SSTD020041


Tgt Ion:165 Resp: 204665
Ion Ratio Lower Upper
165 100
63 38.3 51.8 77.8#
182 3.0 2.6 4.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 23.651 ng/ul
RT: 15.248 min Scan# 2042
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31

Tgt Ion:232 Resp: 186907
Ion Ratio Lower Upper
232 100
131 39.9 41.4 62.2#
130 2.0 2.0 3.0#
166 30.6 27.4 41.0









#59
Diethylphthalate
Concen: 22.944 ng/ul
RT: 15.442 min Scan# 2075
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31




Instrument :
BNA_M
ClientSampleId :
SSTD020041


Tgt Ion:149 Resp: 686482
Ion Ratio Lower Upper
149 100
177 22.1 17.1 25.7
150 12.6 9.9 14.9



#61
Fluorene
Concen: 22.748 ng/ul
RT: 15.671 min Scan# 2114
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31

Tgt Ion:166 Resp: 684693
Ion Ratio Lower Upper
166 100
165 96.3 77.9 116.9
167 13.4 10.3 15.5









#62
4-Chlorophenyl-phenylether
Concen: 22.742 ng/ul
RT: 15.666 min Scan# 2113
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31




Instrument :
BNA_M
ClientSampleId :
SSTD020041


Tgt Ion:204 Resp: 366094
Ion Ratio Lower Upper
204 100
206 33.2 27.8 41.8
141 55.4 54.9 82.3



#63
4-Nitroaniline
Concen: 14.428 ng/ul
RT: 15.677 min Scan# 2115
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31

Tgt Ion:138 Resp: 84213
Ion Ratio Lower Upper
138 100
92 47.0 48.0 72.0#
108 82.0 89.5 134.3#









#66
 4,6-Dinitro-2-methylphenol
 Concen: 20.357 ng/ul
 RT: 15.742 min Scan# 2126
 Delta R.T. 0.000 min
 Lab File: BM034145.D
 Acq: 08 Mar 2022 10:31




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020041


Tgt Ion:198 Resp: 118199
 Ion Ratio Lower Upper
 198 100
 182 3.8 4.2 6.2#
 77 21.3 26.3 39.5#



#67
 N-Nitrosodiphenylamine
 Concen: 22.507 ng/ul
 RT: 15.877 min Scan# 2149
 Delta R.T. 0.000 min
 Lab File: BM034145.D
 Acq: 08 Mar 2022 10:31

Tgt Ion:169 Resp: 594744
 Ion Ratio Lower Upper
 169 100
 168 65.5 52.9 79.3
 167 35.9 29.9 44.9









#68
4-Bromophenyl-phenylether
Concen: 21.882 ng/ul
RT: 16.560 min Scan# 2172
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31




Instrument :
BNA_M
ClientSampleId :
SSTD020041


Tgt Ion:248 Resp: 217256
Ion Ratio Lower Upper
248 100
250 95.9 79.0 118.4
141 65.5 69.7 104.5#



#69
Hexachlorobenzene
Concen: 23.718 ng/ul
RT: 16.683 min Scan# 2286
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31

Tgt Ion:284 Resp: 285602
Ion Ratio Lower Upper
284 100
142 27.9 34.3 51.5#
249 29.8 26.1 39.1









#70
Atrazine
Concen: 21.631 ng/ul
RT: 16.824 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31




Instrument :
BNA_M
ClientSampleId :
SSTD020041


Tgt Ion:200 Resp: 234575
Ion Ratio Lower Upper
200 100
173 25.0 20.6 30.8
215 50.2 36.9 55.3



#71
Pentachlorophenol
Concen: 21.668 ng/ul
RT: 17.018 min Scan# 2343
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31

Tgt Ion:266 Resp: 169908
Ion Ratio Lower Upper
266 100
264 61.7 49.1 73.7
268 62.7 47.8 71.8





#72

Phenanthrene

Concen: 22.800 ng/u1

RT: 17.413 min Scan# 2410

Delta R.T. 0.000 min

Lab File: BM034145.D

Acq: 08 Mar 2022 10:31

Instrument :

BNA_M

ClientSampleId :

SSTD020041





Tgt Ion:178 Resp: 1125476

Ion Ratio Lower Upper

178 100

179 15.3 11.7 17.5

176 18.8 15.1 22.7



#74

Anthracene

Concen: 22.804 ng/u1

RT: 17.501 min Scan# 2425

Delta R.T. 0.000 min

Lab File: BM034145.D

Acq: 08 Mar 2022 10:31




Tgt Ion:178 Resp: 1139924


Ion Ratio Lower Upper

178 100

179 15.6 12.3 18.5

176 17.9 14.5 21.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 23.464 ng/uL

RT: 13.436 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM034145.D

Acq: 08 Mar 2022 10:31

Instrument :

BNA_M

ClientSampleId :

SSTD020041





Tgt Ion:216 Resp: 337064

Ion Ratio Lower Upper

216 100

214 77.8 65.4 98.0

218 47.6 38.8 58.2



#76

Pentachlorobenzene

Concen: 22.173 ng/uL

RT: 14.948 min Scan# 1991

Delta R.T. 0.000 min

Lab File: BM034145.D

Acq: 08 Mar 2022 10:31




Tgt Ion:250 Resp: 315072


Ion Ratio Lower Upper

250 100

248 64.7 51.6 77.4

252 64.7 52.0 78.0









#77
 Carbazole
 Concen: 23.172 ng/ul
 RT: 17.765 min Scan# 2470
 Delta R.T. 0.000 min
 Lab File: BM034145.D
 Acq: 08 Mar 2022 10:31




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020041


Tgt Ion:167 Resp: 1009501
 Ion Ratio Lower Upper
 167 100
 166 21.7 16.9 25.3
 139 12.6 10.5 15.7



#78
 Di-n-butylphthalate
 Concen: 22.328 ng/ul
 RT: 18.336 min Scan# 2567
 Delta R.T. 0.000 min
 Lab File: BM034145.D
 Acq: 08 Mar 2022 10:31

Tgt Ion:149 Resp: 1134067
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.0 10.6
 104 5.2 5.1 7.7





#80

Fluoranthene

Concen: 21.830 ng/ul

RT: 19.424 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM034145.D

Acq: 08 Mar 2022 10:31

Instrument :

BNA_M

ClientSampleId :

SSTD020041





Tgt Ion:202 Resp: 1414820

Ion Ratio Lower Upper

202 100

101 7.5 8.2 12.4#

100 5.7 6.9 10.3#



#82

Pyrene

Concen: 21.872 ng/ul

RT: 19.783 min Scan# 2813

Delta R.T. 0.000 min

Lab File: BM034145.D

Acq: 08 Mar 2022 10:31




Tgt Ion:202 Resp: 1464324


Ion Ratio Lower Upper

202 100

101 8.6 9.8 14.6#

100 7.4 8.6 13.0#









#83
 Butylbenzylphthalate
 Concen: 22.251 ng/ul
 RT: 20.677 min Scan# 2965
 Delta R.T. 0.000 min
 Lab File: BM034145.D
 Acq: 08 Mar 2022 10:31




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020041


Tgt Ion:149 Resp: 525280
 Ion Ratio Lower Upper
 149 100
 91 61.8 69.3 103.9#
 206 23.4 16.1 24.1



#84
 3,3'-Dichlorobenzidine
 Concen: 18.396 ng/ul
 RT: 21.453 min Scan# 3097
 Delta R.T. 0.000 min
 Lab File: BM034145.D
 Acq: 08 Mar 2022 10:31

Tgt Ion:252 Resp: 405777
 Ion Ratio Lower Upper
 252 100
 254 65.0 49.4 74.2
 126 9.2 8.2 12.4





#85

Benzo(a)anthracene

Concen: 23.668 ng/ul

RT: 21.524 min Scan# 3109

Delta R.T. 0.000 min

Lab File: BM034145.D

Acq: 08 Mar 2022 10:31

Instrument :

BNA_M

ClientSampleId :

SSTD020041





Tgt Ion:228 Resp: 1505434

Ion Ratio Lower Upper

228 100

229 19.4 15.0 22.4

226 26.5 21.4 32.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 21.480 ng/ul

RT: 21.453 min Scan# 3097

Delta R.T. 0.000 min

Lab File: BM034145.D

Acq: 08 Mar 2022 10:31




Tgt Ion:149 Resp: 781800


Ion Ratio Lower Upper

149 100

167 27.5 21.8 32.8

279 4.5 3.2 4.8





#87

Chrysene

Concen: 23.673 ng/ul

RT: 21.577 min Scan# 3118

Delta R.T. 0.000 min

Lab File: BM034145.D

Acq: 08 Mar 2022 10:31

Instrument :

BNA_M

ClientSampleId :

SSTD020041





Tgt Ion:228 Resp: 1478501

Ion Ratio Lower Upper

228 100

226 29.6 24.5 36.7

229 20.2 15.9 23.9



#89

Di-n-octyl phthalate

Concen: 20.384 ng/ul




RT: 22.395 min Scan# 3257


Delta R.T. 0.000 min

Lab File: BM034145.D

Acq: 08 Mar 2022 10:31

Tgt Ion:149 Resp: 1301296









#90
Benzo(b)fluoranthene
Concen: 24.016 ng/ul
RT: 23.242 min Scan# 3401
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31




Instrument :
BNA_M
ClientSampleId :
SSTD020041


Tgt Ion:252 Resp: 1579792
Ion Ratio Lower Upper
252 100
253 21.8 17.7 26.5
125 6.8 7.8 11.6#



#91
Benzo(k)fluoranthene
Concen: 23.640 ng/ul
RT: 23.289 min Scan# 3409
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31

Tgt Ion:252 Resp: 1463883
Ion Ratio Lower Upper
252 100
253 21.5 17.6 26.4
125 7.3 7.4 11.2#









#93
Benzo(a)pyrene
Concen: 0.146 ng/ul
RT: 24.036 min Scan# 3536
Delta R.T. 0.159 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31




Instrument :
BNA_M
ClientSampleId :
SSTD020041


Tgt Ion:252 Resp: 9196
Ion Ratio Lower Upper
252 100
253 23.8 17.4 26.0
125 10.5 7.8 11.6



#94
Indeno(1,2,3-cd)pyrene
Concen: 21.913 ng/ul
RT: 26.541 min Scan# 3962
Delta R.T. 0.000 min
Lab File: BM034145.D
Acq: 08 Mar 2022 10:31

Tgt Ion:276 Resp: 1535840
Ion Ratio Lower Upper
276 100
138 17.1 16.5 24.7
277 25.4 20.6 31.0





#95

Dibenzo(a,h)anthracene

Concen: 21.668 ng/ul

RT: 26.559 min Scan# 3965

Delta R.T. 0.000 min

Lab File: BM034145.D

Acq: 08 Mar 2022 10:31

Instrument :

BNA_M

ClientSampleId :

SSTD020041





Tgt Ion:278 Resp: 1314701

Ion Ratio Lower Upper

278 100

139 11.6 11.0 16.4

279 24.1 18.1 27.1



#96

Benzo(g,h,i)perylene

Concen: 21.567 ng/ul

RT: 27.324 min Scan# 4095

Delta R.T. 0.000 min

Lab File: BM034145.D

Acq: 08 Mar 2022 10:31

Tgt Ion:276 Resp: 1276636

Ion Ratio Lower Upper

276 100

138 16.1 15.6 23.4

277 24.1 18.7 28.1

