

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030822\
 Data File : BM034146.D
 Acq On : 08 Mar 2022 11:08
 Operator : CG/JU
 Sample : SSTD04042
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD040042

Quant Time: Mar 08 11:59:26 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM030822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 08 11:29:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.995	152	172826	20.000	ng/ul	0.00
20) Naphthalene-d8	10.807	136	711986	20.000	ng/ul	# 0.00
38) Acenaphthene-d10	14.630	164	485696	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.371	188	1056414	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.542	240	1025610	20.000	ng/ul	0.00
88) Perylene-d12	23.988	264	885568	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.372	96	56172	14.986	ng/uL	0.00
4) Pyridine-d5	3.790	84	410934	41.721	ng/ul	0.00
7) Phenol-d5	7.143	99	506419	41.028	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.313	67	275499	40.205	ng/ul	0.00
11) 2-Chlorophenol-d4	7.519	132	437118	41.079	ng/ul	0.00
15) 4-Methylphenol-d8	8.701	113	429333	40.865	ng/ul	0.00
21) Nitrobenzene-d5	9.154	128	217951	40.527	ng/ul	0.00
24) 2-Nitrophenol-d4	9.884	143	248117	40.588	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.425	165	486931	39.360	ng/ul	0.00
31) 4-Chloroaniline-d4	10.936	131	621931	39.864	ng/ul	0.00
46) Dimethylphthalate-d6	14.036	166	1430103	38.916	ng/ul	0.00
49) Acenaphthylene-d8	14.324	160	1820036	40.665	ng/ul	0.00
54) 4-Nitrophenol-d4	14.813	143	258571	41.079	ng/ul	0.00
60) Fluorene-d10	15.618	176	1259638	39.658	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.730	200	260296	37.741	ng/ul	0.00
73) Anthracene-d10	17.471	188	1997560	39.103	ng/ul	0.00
81) Pyrene-d10	19.753	212	2319561	41.954	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.830	264	1879490	41.646	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.407	88	59943	16.293	ng/uL#	74
5) Pyridine	3.813	79	410862	40.874	ng/ul#	65
6) Benzaldehyde	7.119	77	285568	40.740	ng/ul#	78
8) Phenol	7.172	94	510427	42.358	ng/ul#	79
10) Bis(2-Chloroethyl)ether	7.407	93	404005	42.562	ng/ul#	83
12) 2-Chlorophenol	7.554	128	449512	42.564	ng/ul#	80
13) 2-Methylphenol	8.431	108	402561	42.507	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.531	45	360603	34.814	ng/ul#	80
16) Acetophenone	8.819	105	678075	41.502	ng/ul#	79
17) N-Nitroso-di-n-propyla...	8.813	70	320028	42.943	ng/ul#	77
18) 4-Methylphenol	8.766	108	436711	42.519	ng/ul	94
19) Hexachloroethane	9.089	117	184920	41.814	ng/ul#	66
22) Nitrobenzene	9.201	77	482977	41.867	ng/ul#	85
23) Isophorone	9.731	82	895721	42.115	ng/ul#	91
25) 2-Nitrophenol	9.913	139	267012	42.307	ng/ul#	74
26) 2,4-Dimethylphenol	9.972	107	512629	40.555	ng/ul	89
27) Bis(2-Chloroethoxy)met...	10.213	93	546245	42.360	ng/ul#	97
29) 2,4-Dichlorophenol	10.454	162	471706	39.701	ng/ul	92
30) Naphthalene	10.860	128	1445952	40.525	ng/ul	99
32) 4-Chloroaniline	10.960	127	629845	40.838	ng/ul	98
33) Hexachlorobutadiene	11.154	225	353274	38.712	ng/ul	97
34) Caprolactam	11.730	113	80091	24.822	ng/ul#	53
35) 4-Chloro-3-methylphenol	12.083	107	471065	42.341	ng/ul#	74
36) 2-Methylnaphthalene	12.466	142	1047865	40.232	ng/ul	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030822\
 Data File : BM034146.D
 Acq On : 08 Mar 2022 11:08
 Operator : CG/JU
 Sample : SSTD04042
 Misc :
 ALS Vial : 5 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SSTD040042


Quant Time: Mar 08 11:59:26 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM030822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 08 11:29:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.683	142	1068273	40.461	ng/ul#	98
39) 1,2,4,5-Tetrachloroben...	12.836	216	650291	40.301	ng/ul#	96
40) Hexachlorocyclopentadiene	12.819	237	445827	40.515	ng/ul	99
41) 2,4,6-Trichlorophenol	13.066	196	420741	42.686	ng/ul	99
42) 2,4,5-Trichlorophenol	13.136	196	456069	42.070	ng/ul	95
43) 1,1'-Biphenyl	13.472	154	1456951	40.696	ng/ul#	97
44) 2-Chloronaphthalene	13.513	162	1138179	40.919	ng/ul	94
45) 2-Nitroaniline	13.707	65	280084	44.729	ng/ul#	59
47) Dimethylphthalate	14.089	163	1440131	40.481	ng/ul	99
48) 2,6-Dinitrotoluene	14.201	165	307229	44.129	ng/ul#	81
50) Acenaphthylene	14.354	152	1808424	41.141	ng/ul	99
51) 3-Nitroaniline	14.524	138	291997	43.120	ng/ul#	76
52) Acenaphthene	14.695	153	1188729	40.751	ng/ul	97
53) 2,4-Dinitrophenol	14.730	184	177270	38.381	ng/ul#	67
55) 4-Nitrophenol	14.830	109	222131	38.769	ng/ul#	71
56) Dibenzofuran	15.030	168	1720582	40.163	ng/ul	91
57) 2,4-Dinitrotoluene	14.983	165	439741	43.468	ng/ul#	66
58) 2,3,4,6-Tetrachlorophenol	15.254	232	396044	42.219	ng/ul#	86
59) Diethylphthalate	15.448	149	1429197	40.240	ng/ul	97
61) Fluorene	15.677	166	1432712	40.100	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.671	204	760533	39.801	ng/ul#	87
63) 4-Nitroaniline	15.683	138	283232	40.880	ng/ul#	77
66) 4,6-Dinitro-2-methylph...	15.748	198	263529	39.097	ng/ul#	80
67) N-Nitrosodiphenylamine	15.877	169	1225250	39.944	ng/ul	98
68) 4-Bromophenyl-phenylether	16.560	248	458758	39.803	ng/ul#	89
69) Hexachlorobenzene	16.683	284	580362	41.518	ng/ul#	85
70) Atrazine	16.830	200	489400	38.876	ng/ul	97
71) Pentachlorophenol	17.018	266	369749	40.620	ng/ul	98
72) Phenanthrene	17.412	178	2285902	39.892	ng/ul	99
74) Anthracene	17.507	178	2322968	40.032	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.436	216	710005	42.578	ng/uL	97
76) Pentachlorobenzene	14.954	250	656532	39.801	ng/uL	98
77) Carbazole	17.771	167	2021738	39.977	ng/ul	99
78) Di-n-butylphthalate	18.342	149	2386140	40.470	ng/ul	98
80) Fluoranthene	19.424	202	2768354	43.337	ng/ul#	93
82) Pyrene	19.783	202	2794813	42.352	ng/ul#	93
83) Butylbenzylphthalate	20.677	149	1012630	43.520	ng/ul#	79
84) 3,3'-Dichlorobenzidine	21.453	252	852648	39.217	ng/ul	97
85) Benzo(a)anthracene	21.530	228	2618310	41.763	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.459	149	1471551	41.019	ng/ul#	98
87) Chrysene	21.583	228	2534193	41.166	ng/ul	99
89) Di-n-octyl phthalate	22.394	149	2256831	41.409	ng/ul	100
90) Benzo(b)fluoranthene	23.241	252	2404542	42.818	ng/ul#	98
91) Benzo(k)fluoranthene	23.294	252	2329639	44.068	ng/ul#	98
93) Benzo(a)pyrene	23.883	252	2275598	42.325	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.547	276	2383164	39.829	ng/ul	96
95) Dibenzo(a,h)anthracene	26.565	278	2054676	39.667	ng/ul	96
96) Benzo(g,h,i)perylene	27.335	276	1995127	39.480	ng/ul#	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Mar 08 11:59:26 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM030822.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Mar 08 11:29:45 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 16.293 ng/uL
RT: 3.407 min Scan# 29
Delta R.T. -0.000 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion: 88 Resp: 59943
Ion Ratio Lower Upper
88 100
43 26.2 36.2 54.4#
58 62.7 68.5 102.7#



#5
Pyridine
Concen: 40.874 ng/uL
RT: 3.813 min Scan# 98
Delta R.T. -0.000 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08

Tgt Ion: 79 Resp: 410862
Ion Ratio Lower Upper
79 100
52 57.5 78.1 117.1#
51 31.1 36.0 54.0#









#6
Benzaldehyde
Concen: 40.740 ng/ul
RT: 7.119 min Scan# 660
Delta R.T. -0.000 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion: 77 Resp: 285568
Ion Ratio Lower Upper
77 100
105 99.0 65.6 98.4#
106 96.4 60.6 90.8#



#8
Phenol
Concen: 42.358 ng/ul
RT: 7.172 min Scan# 669
Delta R.T. 0.006 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08

Tgt Ion: 94 Resp: 510427
Ion Ratio Lower Upper
94 100
65 28.5 31.7 47.5#
66 40.1 45.4 68.0#









#10
 Bis(2-Chloroethyl)ether
 Concen: 42.562 ng/ul
 RT: 7.407 min Scan# 709
 Delta R.T. -0.000 min
 Lab File: BM034146.D
 Acq: 08 Mar 2022 11:08




Instrument :
 BNA_M
 ClientSampleId :
 SST040042


Tgt Ion: 93 Resp: 404005
 Ion Ratio Lower Upper
 93 100
 63 68.1 71.7 107.5#
 95 32.1 26.3 39.5



#12
 2-Chlorophenol
 Concen: 42.564 ng/ul
 RT: 7.554 min Scan# 734
 Delta R.T. -0.000 min
 Lab File: BM034146.D
 Acq: 08 Mar 2022 11:08

Tgt Ion:128 Resp: 449512
 Ion Ratio Lower Upper
 128 100
 64 39.4 49.4 74.2#
 130 31.7 26.8 40.2









#13
2-Methylphenol
Concen: 42.507 ng/ul
RT: 8.431 min Scan# 883
Delta R.T. -0.000 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:108 Resp: 402561
Ion Ratio Lower Upper
108 100
107 90.0 70.2 105.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 34.814 ng/ul
RT: 8.531 min Scan# 900
Delta R.T. 0.006 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08

Tgt Ion: 45 Resp: 360603
Ion Ratio Lower Upper
45 100
77 21.3 9.9 14.9#
79 16.7 8.3 12.5#









#16
Acetophenone
Concen: 41.502 ng/ul
RT: 8.819 min Scan# 948
Delta R.T. 0.006 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:105 Resp: 678075
Ion Ratio Lower Upper
105 100
77 73.0 73.1 109.7#
51 23.1 30.5 45.7#



#17
N-Nitroso-di-n-propylamine
Concen: 42.943 ng/ul
RT: 8.813 min Scan# 948
Delta R.T. 0.012 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08

Tgt Ion: 70 Resp: 320028
Ion Ratio Lower Upper
70 100
42 49.4 56.9 85.3#
101 10.7 6.9 10.3#
130 24.6 13.8 20.8#









#18
4-Methylphenol
Concen: 42.519 ng/ul
RT: 8.766 min Scan# 940
Delta R.T. 0.006 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:108 Resp: 436711
Ion Ratio Lower Upper
108 100
107 115.6 87.4 131.2



#19
Hexachloroethane
Concen: 41.814 ng/ul
RT: 9.089 min Scan# 995
Delta R.T. 0.006 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08

Tgt Ion:117 Resp: 184920
Ion Ratio Lower Upper
117 100
201 118.4 69.6 104.4#
199 74.2 40.6 60.8#









#22
Nitrobenzene
Concen: 41.867 ng/ul
RT: 9.201 min Scan# 1104
Delta R.T. 0.006 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion: 77 Resp: 482977
Ion Ratio Lower Upper
77 100
123 48.1 29.0 43.6#
65 13.5 11.4 17.2



#23
Isophorone
Concen: 42.115 ng/ul
RT: 9.731 min Scan# 1104
Delta R.T. 0.006 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08

Tgt Ion: 82 Resp: 895721
Ion Ratio Lower Upper
82 100
95 8.7 6.5 9.7
138 20.4 11.8 17.8#









#25
2-Nitrophenol
Concen: 42.307 ng/ul
RT: 9.913 min Scan# 1135
Delta R.T. -0.000 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:139 Resp: 267012
Ion Ratio Lower Upper
139 100
65 48.2 59.0 88.6#
109 32.9 37.2 55.8#



#26
2,4-Dimethylphenol
Concen: 40.555 ng/ul
RT: 9.972 min Scan# 1145
Delta R.T. -0.000 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08

Tgt Ion:107 Resp: 512629
Ion Ratio Lower Upper
107 100
121 49.3 34.8 52.2
122 85.8 60.0 90.0









#27
 Bis(2-Chloroethoxy)methane
 Concen: 42.360 ng/ul
 RT: 10.213 min Scan# 1186
 Delta R.T. -0.000 min
 Lab File: BM034146.D
 Acq: 08 Mar 2022 11:08




Instrument :
 BNA_M
 ClientSampleId :
 SSTD040042


Tgt Ion: 93 Resp: 546245
 Ion Ratio Lower Upper
 93 100
 95 32.2 25.3 37.9
 123 12.8 7.9 11.9#



#29
 2,4-Dichlorophenol
 Concen: 39.701 ng/ul
 RT: 10.454 min Scan# 1227
 Delta R.T. 0.006 min
 Lab File: BM034146.D
 Acq: 08 Mar 2022 11:08

Tgt Ion: 162 Resp: 471706
 Ion Ratio Lower Upper
 162 100
 164 66.6 50.0 75.0
 98 31.1 30.6 45.8









#30
Naphthalene
Concen: 40.525 ng/ul
RT: 10.860 min Scan# 1296
Delta R.T. -0.000 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:128 Resp: 1445952
Ion Ratio Lower Upper
128 100
129 10.9 9.1 13.7
127 13.3 10.2 15.4



#32
4-Chloroaniline
Concen: 40.838 ng/ul
RT: 10.960 min Scan# 1313
Delta R.T. -0.000 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08

Tgt Ion:127 Resp: 629845
Ion Ratio Lower Upper
127 100
129 32.5 25.4 38.0









#33
Hexachlorobutadiene
Concen: 38.712 ng/ul
RT: 11.154 min Scan# 111
Delta R.T. -0.000 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:225 Resp: 353274
Ion Ratio Lower Upper
225 100
223 62.8 52.2 78.2
227 63.9 52.8 79.2



#34
Caprolactam
Concen: 24.822 ng/ul
RT: 11.730 min Scan# 1444
Delta R.T. 0.012 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08

Tgt Ion:113 Resp: 80091
Ion Ratio Lower Upper
113 100
55 126.0 157.9 236.9#
56 102.1 131.8 197.6#









#35
4-Chloro-3-methylphenol
Concen: 42.341 ng/ul
RT: 12.083 min Scan# 1504
Delta R.T. 0.006 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:107 Resp: 471065
Ion Ratio Lower Upper
107 100
144 27.2 17.1 25.7#
142 85.3 49.4 74.2#



#36
2-Methylnaphthalene
Concen: 40.232 ng/ul
RT: 12.466 min Scan# 1569
Delta R.T. -0.000 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08

Tgt Ion:142 Resp: 1047865
Ion Ratio Lower Upper
142 100
141 88.5 71.1 106.7









#37
1-Methylnaphthalene
Concen: 40.461 ng/ul
RT: 12.683 min Scan# 1606
Delta R.T. -0.000 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:142 Resp: 1068273
Ion Ratio Lower Upper
142 100
141 92.2 75.5 113.3
116 4.0 4.0 6.0#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 40.301 ng/ul
RT: 12.836 min Scan# 1632
Delta R.T. 0.006 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08

Tgt Ion:216 Resp: 650291
Ion Ratio Lower Upper
216 100
214 78.2 61.3 91.9
179 20.1 18.0 27.0
108 14.5 16.6 25.0#









#40
Hexachlorocyclopentadiene
Concen: 40.515 ng/ul
RT: 12.819 min Scan# 1629
Delta R.T. -0.000 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:237 Resp: 445827
Ion Ratio Lower Upper
237 100
235 62.8 50.6 76.0
272 12.5 9.6 14.4



#41
2,4,6-Trichlorophenol
Concen: 42.686 ng/ul
RT: 13.066 min Scan# 1671
Delta R.T. -0.000 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08

Tgt Ion:196 Resp: 420741
Ion Ratio Lower Upper
196 100
198 97.9 78.9 118.3
200 30.9 23.7 35.5









#42
2,4,5-Trichlorophenol
Concen: 42.070 ng/ul
RT: 13.136 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:196 Resp: 456069
Ion Ratio Lower Upper
196 100
198 98.0 82.2 123.2
200 30.7 27.9 41.9



#43
1,1'-Biphenyl
Concen: 40.696 ng/ul
RT: 13.472 min Scan# 1740
Delta R.T. 0.006 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08

Tgt Ion:154 Resp: 1456951
Ion Ratio Lower Upper
154 100
153 40.5 33.1 49.7
76 9.6 10.7 16.1#









#44
2-Chloronaphthalene
Concen: 40.919 ng/ul
RT: 13.513 min Scan# 1746
Delta R.T. 0.006 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:162 Resp: 1138179
Ion Ratio Lower Upper
162 100
164 33.6 26.2 39.4
127 36.6 33.7 50.5



#45
2-Nitroaniline
Concen: 44.729 ng/ul
RT: 13.707 min Scan# 1780
Delta R.T. 0.006 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08

Tgt Ion: 65 Resp: 280084
Ion Ratio Lower Upper
65 100
92 74.2 47.8 71.6#
138 126.4 61.0 91.6#









#47
Dimethylphthalate
Concen: 40.481 ng/ul
RT: 14.089 min Scan# 1845
Delta R.T. 0.006 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:163 Resp: 1440131
Ion Ratio Lower Upper
163 100
194 4.4 3.3 4.9
164 9.6 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 44.129 ng/ul
RT: 14.201 min Scan# 1864
Delta R.T. 0.006 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08

Tgt Ion:165 Resp: 307229
Ion Ratio Lower Upper
165 100
89 51.9 55.7 83.5#
121 19.5 20.0 30.0#









#50
Acenaphthylene
Concen: 41.141 ng/ul
RT: 14.354 min Scan# 1890
Delta R.T. -0.000 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:152 Resp: 1808424
Ion Ratio Lower Upper
152 100
151 20.0 15.8 23.8
153 13.1 10.1 15.1



#51
3-Nitroaniline
Concen: 43.120 ng/ul
RT: 14.524 min Scan# 1919
Delta R.T. 0.006 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08

Tgt Ion:138 Resp: 291997
Ion Ratio Lower Upper
138 100
108 10.4 10.8 16.2#
92 107.2 110.6 166.0#









#52
Acenaphthene
Concen: 40.751 ng/ul
RT: 14.695 min Scan# 1948
Delta R.T. -0.000 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:153 Resp: 1188729
Ion Ratio Lower Upper
153 100
152 46.9 39.4 59.2
154 89.0 72.7 109.1



#53
2,4-Dinitrophenol
Concen: 38.381 ng/ul
RT: 14.730 min Scan# 1954
Delta R.T. 0.006 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08

Tgt Ion:184 Resp: 177270
Ion Ratio Lower Upper
184 100
63 51.3 73.8 110.8#
154 60.2 63.5 95.3#









#55
4-Nitrophenol
Concen: 38.769 ng/ul
RT: 14.830 min Scan# 1969
Delta R.T. 0.012 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:109 Resp: 222131
Ion Ratio Lower Upper
109 100
139 123.3 77.6 116.4#
65 103.3 112.2 168.4#



#56
Dibenzofuran
Concen: 40.163 ng/ul
RT: 15.030 min Scan# 2005
Delta R.T. 0.006 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08

Tgt Ion:168 Resp: 1720582
Ion Ratio Lower Upper
168 100
139 36.7 34.0 51.0









#57
2,4-Dinitrotoluene
Concen: 43.468 ng/ul
RT: 14.983 min Scan# 1997
Delta R.T. 0.006 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:165 Resp: 439741
Ion Ratio Lower Upper
165 100
63 36.4 51.8 77.8#
182 3.0 2.6 4.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 42.219 ng/ul
RT: 15.254 min Scan# 2043
Delta R.T. 0.006 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08

Tgt Ion:232 Resp: 396044
Ion Ratio Lower Upper
232 100
131 40.0 41.4 62.2#
130 2.2 2.0 3.0
166 28.4 27.4 41.0









#59
Diethylphthalate
Concen: 40.240 ng/ul
RT: 15.448 min Scan# 2075
Delta R.T. 0.006 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:149 Resp: 1429197
Ion Ratio Lower Upper
149 100
177 22.9 17.1 25.7
150 13.0 9.9 14.9



#61
Fluorene
Concen: 40.100 ng/ul
RT: 15.677 min Scan# 2115
Delta R.T. 0.006 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08

Tgt Ion:166 Resp: 1432712
Ion Ratio Lower Upper
166 100
165 97.5 77.9 116.9
167 13.4 10.3 15.5









#62
4-Chlorophenyl-phenylether
Concen: 39.801 ng/ul
RT: 15.671 min Scan# 2114
Delta R.T. 0.006 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:204 Resp: 760533
Ion Ratio Lower Upper
204 100
206 33.4 27.8 41.8
141 54.2 54.9 82.3#



#63
4-Nitroaniline
Concen: 40.880 ng/ul
RT: 15.683 min Scan# 2116
Delta R.T. 0.006 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08

Tgt Ion:138 Resp: 283232
Ion Ratio Lower Upper
138 100
92 45.8 48.0 72.0#
108 85.1 89.5 134.3#









#66
4,6-Dinitro-2-methylphenol
Concen: 39.097 ng/ul
RT: 15.748 min Scan# 2126
Delta R.T. 0.006 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:198 Resp: 263529
Ion Ratio Lower Upper
198 100
182 3.1 4.2 6.2#
77 20.1 26.3 39.5#



#67
N-Nitrosodiphenylamine
Concen: 39.944 ng/ul
RT: 15.877 min Scan# 2149
Delta R.T. -0.000 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08

Tgt Ion:169 Resp: 1225250
Ion Ratio Lower Upper
169 100
168 66.7 52.9 79.3
167 35.2 29.9 44.9









#68
4-Bromophenyl-phenylether
Concen: 39.803 ng/ul
RT: 16.560 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:248 Resp: 458758
Ion Ratio Lower Upper
248 100
250 97.4 79.0 118.4
141 67.4 69.7 104.5#



#69
Hexachlorobenzene
Concen: 41.518 ng/ul
RT: 16.683 min Scan# 2286
Delta R.T. -0.000 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08

Tgt Ion:284 Resp: 580362
Ion Ratio Lower Upper
284 100
142 28.6 34.3 51.5#
249 30.2 26.1 39.1









#70
 Atrazine
 Concen: 38.876 ng/ul
 RT: 16.830 min Scan# 21
 Delta R.T. 0.006 min
 Lab File: BM034146.D
 Acq: 08 Mar 2022 11:08




Instrument :
 BNA_M
 ClientSampleId :
 SST040042


Tgt Ion:200 Resp: 489400
 Ion Ratio Lower Upper
 200 100
 173 25.5 20.6 30.8
 215 49.5 36.9 55.3



#71
 Pentachlorophenol
 Concen: 40.620 ng/ul
 RT: 17.018 min Scan# 2343
 Delta R.T. -0.000 min
 Lab File: BM034146.D
 Acq: 08 Mar 2022 11:08

Tgt Ion:266 Resp: 369749
 Ion Ratio Lower Upper
 266 100
 264 62.6 49.1 73.7
 268 62.2 47.8 71.8









#72
Phenanthrene
Concen: 39.892 ng/ul
RT: 17.412 min Scan# 2410
Delta R.T. -0.000 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:178 Resp: 2285902
Ion Ratio Lower Upper
178 100
179 15.5 11.7 17.5
176 19.0 15.1 22.7



#74
Anthracene
Concen: 40.032 ng/ul
RT: 17.507 min Scan# 2426
Delta R.T. 0.006 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08

Tgt Ion:178 Resp: 2322968
Ion Ratio Lower Upper
178 100
179 15.4 12.3 18.5
176 18.3 14.5 21.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 42.578 ng/uL
RT: 13.436 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:216 Resp: 710005
Ion Ratio Lower Upper
216 100
214 77.0 65.4 98.0
218 48.3 38.8 58.2



#76
Pentachlorobenzene
Concen: 39.801 ng/uL
RT: 14.954 min Scan# 1992
Delta R.T. 0.006 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08

Tgt Ion:250 Resp: 656532
Ion Ratio Lower Upper
250 100
248 62.7 51.6 77.4
252 63.6 52.0 78.0









#77
 Carbazole
 Concen: 39.977 ng/ul
 RT: 17.771 min Scan# 2471
 Delta R.T. 0.006 min
 Lab File: BM034146.D
 Acq: 08 Mar 2022 11:08




Instrument :
 BNA_M
 ClientSampleId :
 SST040042


Tgt Ion:167 Resp: 2021738
 Ion Ratio Lower Upper
 167 100
 166 21.4 16.9 25.3
 139 12.3 10.5 15.7



#78
 Di-n-butylphthalate
 Concen: 40.470 ng/ul
 RT: 18.342 min Scan# 2568
 Delta R.T. 0.006 min
 Lab File: BM034146.D
 Acq: 08 Mar 2022 11:08

Tgt Ion:149 Resp: 2386140
 Ion Ratio Lower Upper
 149 100
 150 8.9 7.0 10.6
 104 5.2 5.1 7.7









#80
 Fluoranthene
 Concen: 43.337 ng/ul
 RT: 19.424 min Scan# 2752
 Delta R.T. -0.000 min
 Lab File: BM034146.D
 Acq: 08 Mar 2022 11:08




Instrument :
 BNA_M
 ClientSampleId :
 SSTD040042


Tgt Ion:202 Resp: 2768354
 Ion Ratio Lower Upper
 202 100
 101 7.7 8.2 12.4#
 100 6.1 6.9 10.3#



#82
 Pyrene
 Concen: 42.352 ng/ul
 RT: 19.783 min Scan# 2813
 Delta R.T. -0.000 min
 Lab File: BM034146.D
 Acq: 08 Mar 2022 11:08

Tgt Ion:202 Resp: 2794813
 Ion Ratio Lower Upper
 202 100
 101 9.5 9.8 14.6#
 100 8.1 8.6 13.0#









#83
Butylbenzylphthalate
Concen: 43.520 ng/ul
RT: 20.677 min Scan# 2965
Delta R.T. -0.000 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:149 Resp: 1012630
Ion Ratio Lower Upper
149 100
91 63.7 69.3 103.9#
206 23.2 16.1 24.1



#84
3,3'-Dichlorobenzidine
Concen: 39.217 ng/ul
RT: 21.453 min Scan# 3097
Delta R.T. -0.000 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08

Tgt Ion:252 Resp: 852648
Ion Ratio Lower Upper
252 100
254 63.8 49.4 74.2
126 8.9 8.2 12.4





#85

Benzo(a)anthracene

Concen: 41.763 ng/ul

RT: 21.530 min Scan# 3110

Delta R.T. 0.006 min

Lab File: BM034146.D

Acq: 08 Mar 2022 11:08

Instrument :

BNA_M

ClientSampleId :

SSTD040042





Tgt Ion:228 Resp: 2618310

Ion Ratio Lower Upper

228 100

229 19.6 15.0 22.4

226 26.5 21.4 32.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 41.019 ng/ul

RT: 21.459 min Scan# 3098

Delta R.T. 0.006 min

Lab File: BM034146.D

Acq: 08 Mar 2022 11:08




Tgt Ion:149 Resp: 1471551


Ion Ratio Lower Upper

149 100

167 28.3 21.8 32.8

279 5.3 3.2 4.8#









#87
Chrysene
Concen: 41.166 ng/ul
RT: 21.583 min Scan# 3118
Delta R.T. 0.006 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:228 Resp: 2534193
Ion Ratio Lower Upper
228 100
226 29.4 24.5 36.7
229 19.8 15.9 23.9



#89
Di-n-octyl phthalate
Concen: 41.409 ng/ul
RT: 22.394 min Scan# 3257
Delta R.T. -0.000 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08

Tgt Ion:149 Resp: 2256831









#90
Benzo(b)fluoranthene
Concen: 42.818 ng/ul
RT: 23.241 min Scan# 3401
Delta R.T. -0.000 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:252 Resp: 2404542
Ion Ratio Lower Upper
252 100
253 22.2 17.7 26.5
125 7.2 7.8 11.6#



#91
Benzo(k)fluoranthene
Concen: 44.068 ng/ul
RT: 23.294 min Scan# 3410
Delta R.T. 0.006 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08

Tgt Ion:252 Resp: 2329639
Ion Ratio Lower Upper
252 100
253 21.7 17.6 26.4
125 7.1 7.4 11.2#









#93
Benzo(a)pyrene
Concen: 42.325 ng/ul
RT: 23.883 min Scan# 3510
Delta R.T. 0.006 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:252 Resp: 2275598
Ion Ratio Lower Upper
252 100
253 21.9 17.4 26.0
125 8.0 7.8 11.6



#94
Indeno(1,2,3-cd)pyrene
Concen: 39.829 ng/ul
RT: 26.547 min Scan# 3963
Delta R.T. 0.006 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08

Tgt Ion:276 Resp: 2383164
Ion Ratio Lower Upper
276 100
138 17.3 16.5 24.7
277 24.8 20.6 31.0









#95
Dibenzo(a,h)anthracene
Concen: 39.667 ng/ul
RT: 26.565 min Scan# 3965
Delta R.T. 0.006 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08

Instrument :
BNA_M
ClientSampleId :
SSTD040042

Tgt Ion:278 Resp: 2054676
Ion Ratio Lower Upper
278 100
139 11.5 11.0 16.4
279 23.9 18.1 27.1



#96
Benzo(g,h,i)perylene
Concen: 39.480 ng/ul
RT: 27.335 min Scan# 4097
Delta R.T. 0.012 min
Lab File: BM034146.D
Acq: 08 Mar 2022 11:08

Tgt Ion:276 Resp: 1995127
Ion Ratio Lower Upper
276 100
138 14.9 15.6 23.4#
277 23.3 18.7 28.1

