


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030822\
 Data File : BM034142.D
 Acq On : 08 Mar 2022 08:43
 Operator : CG/JU
 Sample : DFTPP040
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP040

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM030822.M
 Title : SVOA CALIBRATION
 Last Update : Tue Mar 08 13:09:16 2022



AutoFind: Scans 2478, 2479, 2480; Background Corrected with Scan 2472

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	26.9	81683	PASS
68	69	0.00	2	1.8	1814	PASS
69	198	0.00	100	33.8	102944	PASS
70	69	0.00	2	0.6	583	PASS
127	198	10	80	43.5	132251	PASS
197	198	0.00	2	0.5	1640	PASS
198	198	100	100	100.0	304192	PASS
199	198	5	9	6.7	20309	PASS
275	198	10	60	32.6	99253	PASS
365	198	1	100	5.0	15178	PASS
441	198	0.01	100	14.5	44144	PASS
442	198	50	100	96.7	294059	PASS
443	442	15	24	18.8	55392	PASS

m/z	Abundance
39.90	672.0
43.90	650.0
48.90	241.0
51.00	157.0
73.00	2307.0
74.10	333.0
76.90	209.0
83.90	292.0
85.90	196.0
95.70	207.0
132.90	340.0
147.00	1200.0
148.00	189.0
149.00	199.0
177.00	256.0
190.90	454.0
193.10	272.0
207.00	4000.0
208.10	1029.0
208.90	659.0
221.10	1063.0
222.10	328.0
267.00	400.0
280.90	1345.0
282.00	421.0
282.80	213.0
295.00	319.0
341.10	379.0
342.00	155.0
355.00	597.0
356.00	222.0
429.20	350.0

Instrument :
BNA_M
ClientSampleId :
DFTPP040

m/z	Abundance
37.00	594.0
38.00	1506.0
39.00	7851.0
40.00	934.0
40.90	239.0
42.90	176.0
44.00	671.0
44.90	247.0
49.00	1166.0
50.00	26368.0
51.00	96584.0
52.00	5270.0
53.00	211.0
54.90	721.0
56.00	3002.0
57.00	6883.0
57.90	301.0
59.00	209.0
60.00	272.0
61.00	1383.0
62.00	1691.0
63.00	4311.0
64.00	730.0
65.00	2342.0
66.10	165.0
67.10	263.0
68.00	2238.0
69.00	120728.0
70.00	612.0
73.10	2139.0
74.00	12644.0
75.00	20088.0
76.10	6290.0
77.10	127496.0
78.00	8901.0
79.00	10008.0
80.00	7573.0
81.00	10179.0
82.00	2563.0
83.00	2358.0
83.90	354.0
85.00	1781.0
86.00	3082.0
87.00	1585.0
87.80	473.0
89.00	359.0
91.00	2343.0
92.00	2840.0
93.00	18008.0
94.00	1318.0
94.80	229.0
96.00	840.0
97.10	278.0
98.00	13891.0
99.00	11718.0
100.00	895.0
101.00	6234.0
101.90	211.0
103.00	1725.0
103.90	3534.0
105.00	4106.0
105.90	1179.0
107.00	44160.0
108.00	6831.0
109.10	1185.0
110.00	72432.0
111.00	12029.0
112.00	1594.0
112.90	405.0
114.80	223.0
116.00	2742.0
117.00	42264.0
118.00	2932.0
118.90	567.0
120.00	651.0
121.00	215.0
121.90	2919.0
123.00	4603.0
124.00	2373.0
125.00	2204.0
126.10	740.0
127.00	147328.0
128.00	11371.0
129.00	61152.0
130.00	5299.0
130.90	994.0
132.00	507.0
132.90	727.0
133.90	1842.0
135.00	5282.0
135.90	1590.0
137.00	3133.0

Instrument :
BNA_M
ClientSampleId :
DFTPP040

137.90	558.0
138.90	273.0
140.00	832.0
141.00	8390.0
142.00	2478.0
142.90	1967.0
144.00	630.0
145.00	416.0
146.00	1144.0
147.00	5054.0
148.00	10106.0
149.00	2414.0
150.00	576.0
151.10	1300.0
153.00	2730.0
153.90	1743.0
155.00	4484.0
156.00	6851.0
157.00	1263.0
157.90	1785.0
159.00	992.0
159.90	2886.0
161.00	4134.0
161.90	1260.0
162.90	352.0
164.10	464.0
165.00	3378.0
166.00	2219.0
167.00	16416.0
168.00	9648.0
169.00	1897.0
170.00	811.0
170.70	518.0
171.90	1828.0
172.90	1796.0
174.00	3322.0
175.00	6740.0
176.00	1914.0
177.00	3817.0
178.00	1047.0
179.00	13633.0
180.00	9356.0
181.00	3644.0
182.10	637.0
182.90	254.0
184.00	845.0
185.00	5302.0
186.00	45576.0
187.00	13364.0
188.00	1860.0
188.90	3374.0
190.00	705.0
191.00	2035.0
192.00	4192.0
193.00	4392.0
194.00	1242.0
194.90	491.0
196.00	9016.0
197.00	1610.0
197.90	309760.0
198.90	21440.0
200.00	1521.0
201.60	1619.0
202.90	2629.0
204.00	12523.0
205.00	21320.0
206.00	89680.0
207.00	16362.0
208.00	4013.0
209.00	1697.0
209.90	1323.0
211.00	3896.0
211.80	708.0
212.90	326.0
213.80	168.0
214.90	1117.0
216.00	2535.0
217.00	24792.0
218.00	3035.0
219.10	388.0
221.00	11922.0
221.80	3325.0
223.00	6007.0
224.00	50496.0
225.00	13064.0
226.00	1826.0
227.00	23352.0
228.00	3349.0
228.90	5184.0
229.90	551.0
230.90	1936.0
232.00	383.0
233.00	358.0
234.00	1472.0
235.00	1779.0
236.00	846.0

Instrument :
BNA_M
ClientSampleId :
DFTPP040

237.00	2551.0
238.00	286.0
239.10	664.0
239.90	848.0
240.90	1335.0
242.00	2699.0
243.00	2812.0
244.00	39960.0
245.00	5091.0
246.00	8612.0
246.90	1652.0
248.00	407.0
249.00	1637.0
249.90	403.0
251.00	392.0
252.10	244.0
253.10	1064.0
254.10	1520.0
255.00	207680.0
256.00	30904.0
257.00	2687.0
258.00	12432.0
259.00	2328.0
260.10	361.0
260.90	215.0
264.00	370.0
265.00	5394.0
265.90	754.0
266.90	390.0
269.90	303.0
271.10	763.0
272.00	496.0
273.00	5925.0
274.00	15560.0
275.00	94576.0
276.00	12549.0
277.00	8299.0
277.90	1565.0
278.90	232.0
281.00	1013.0
281.90	279.0
283.10	1023.0
284.10	545.0
285.10	1224.0
286.00	342.0
288.80	296.0
289.90	201.0
290.90	325.0
292.10	458.0
293.00	2008.0
294.10	684.0
295.00	674.0
296.00	28040.0
297.00	3989.0
297.90	308.0
301.10	264.0
302.00	280.0
303.00	3657.0
304.00	874.0
304.90	234.0
307.90	415.0
309.00	311.0
310.10	536.0
312.00	159.0
313.00	220.0
314.00	1503.0
314.90	2602.0
315.90	1222.0
317.00	171.0
319.90	178.0
321.00	1017.0
322.00	318.0
323.00	8362.0
324.00	1593.0
324.80	183.0
325.80	193.0
326.90	2095.0
328.00	841.0
332.10	581.0
333.00	706.0
334.00	5956.0
335.00	1309.0
335.80	289.0
338.90	201.0
341.00	1396.0
342.00	368.0
346.00	2179.0
346.90	499.0
352.00	3219.0
352.90	1699.0
354.10	2832.0
355.00	998.0
355.90	186.0
357.10	173.0
359.00	244.0
365.00	13209.0

Instrument :
BNA_M
ClientSampleId :
DFTPP040

366.00	1773.0
370.10	355.0
371.00	770.0
372.00	4495.0
373.00	1332.0
374.00	169.0
383.00	1327.0
383.90	349.0
385.00	173.0
389.90	483.0
391.00	450.0
392.10	194.0
400.90	281.0
402.00	1593.0
403.00	2722.0
404.00	987.0
404.80	171.0
421.00	2508.0
421.90	2117.0
423.00	13987.0
424.00	4830.0
424.90	629.0
439.90	170.0
441.10	35672.0
442.10	234048.0
443.10	45760.0
444.10	4294.0
444.90	296.0

Instrument :
BNA_M
ClientSampleId :
DFTPP040

m/z	Abundance
37.00	595.0
38.00	1352.0
39.00	7043.0
40.00	755.0
41.10	302.0
44.00	648.0
49.10	835.0
50.10	26936.0
51.00	95096.0
52.10	4482.0
53.00	174.0
54.90	573.0
56.00	3410.0
57.00	6496.0
58.00	375.0
59.00	152.0
61.00	1593.0
62.00	1461.0
63.00	4532.0
63.90	498.0
65.00	2549.0
66.10	236.0
68.00	2023.0
69.00	121472.0
70.00	748.0
73.00	1766.0
74.00	12818.0
75.00	20328.0
76.00	6444.0
77.00	128560.0
78.00	8549.0
79.00	10397.0
80.00	7458.0
81.00	10657.0
82.00	2515.0
83.00	2358.0
83.80	397.0
85.00	1776.0
86.00	3493.0
87.00	1298.0
87.80	456.0
88.90	280.0
91.00	2194.0
92.00	2981.0
93.00	18304.0
94.00	1311.0
94.90	199.0
95.90	780.0
98.00	14184.0
99.00	11418.0
99.90	976.0
101.00	5754.0
102.10	474.0
103.00	1689.0
104.00	3174.0
105.00	3902.0
106.10	1118.0
107.00	47560.0
108.00	7153.0
109.00	1290.0
110.00	78648.0
111.00	12605.0
111.90	1544.0
112.80	507.0
114.90	308.0
116.00	3224.0
117.00	44336.0
118.00	3254.0
118.90	391.0
120.10	753.0
120.80	239.0
122.00	3830.0
123.00	4997.0
123.90	1936.0
125.00	2227.0
127.00	155200.0
128.00	12128.0
129.00	64200.0
130.00	5676.0
131.00	1380.0
132.10	603.0
132.90	678.0
134.00	1628.0
135.00	5016.0
136.00	2288.0
137.00	2836.0
137.80	745.0
139.00	594.0
139.90	1136.0
141.00	9639.0
142.00	2686.0
142.90	2517.0

Instrument :
BNA_M
ClientSampleId :
DFTPP040

143.90	565.0
144.90	511.0
146.00	1812.0
147.00	4873.0
148.00	12191.0
149.00	2615.0
150.10	688.0
151.00	907.0
151.70	829.0
152.90	2925.0
154.00	2240.0
155.00	5572.0
156.00	7987.0
157.10	1217.0
157.90	1746.0
159.00	1422.0
160.00	2791.0
161.00	4627.0
162.00	1172.0
163.00	448.0
164.00	311.0
165.00	3751.0
166.00	3144.0
167.00	18576.0
168.00	10199.0
169.00	1579.0
170.00	685.0
170.80	587.0
171.10	553.0
171.90	1544.0
173.00	1947.0
173.90	4223.0
175.00	7980.0
176.00	2201.0
177.00	4217.0
178.00	1123.0
178.90	14727.0
180.00	10907.0
181.00	4460.0
182.00	699.0
182.80	412.0
183.90	1186.0
185.00	7181.0
186.00	52680.0
187.00	15647.0
188.00	1903.0
188.90	3862.0
189.90	476.0
191.00	2620.0
192.00	4698.0
193.00	6313.0
193.90	1237.0
195.10	522.0
196.00	11244.0
197.00	1905.0
197.90	368640.0
199.00	24736.0
200.00	1757.0
201.20	1231.0
201.70	1230.0
202.90	2826.0
204.00	14473.0
205.00	26008.0
206.00	103128.0
207.00	17432.0
208.00	4648.0
209.00	1848.0
210.20	1745.0
211.00	3356.0
211.80	849.0
212.90	313.0
214.90	1008.0
216.10	2511.0
217.00	29032.0
218.00	4192.0
219.00	382.0
221.00	13221.0
221.80	3730.0
223.00	6466.0
224.00	61256.0
225.00	14277.0
225.90	1737.0
227.00	27632.0
228.00	3834.0
229.00	5093.0
230.00	721.0
231.00	2200.0
231.90	331.0
232.90	369.0
233.90	1667.0
235.00	2114.0
236.00	1594.0
237.00	2516.0
237.90	340.0
239.00	1070.0
239.90	1018.0

Instrument :
BNA_M
ClientSampleId :
DFTPP040

241.00	1456.0
242.00	3253.0
243.00	3083.0
244.00	47248.0
245.00	5956.0
246.00	10900.0
246.90	1881.0
247.90	442.0
248.90	1727.0
249.90	336.0
250.90	570.0
252.00	435.0
252.90	1125.0
254.00	1674.0
255.00	256320.0
256.00	38160.0
257.00	2789.0
258.00	15466.0
259.00	2048.0
260.10	403.0
260.90	430.0
264.10	383.0
265.00	6263.0
265.90	973.0
266.80	254.0
267.90	217.0
269.90	197.0
271.00	718.0
272.00	1007.0
273.00	8345.0
274.00	21800.0
275.00	118792.0
276.00	15665.0
277.00	10397.0
277.90	2010.0
279.00	323.0
281.00	1094.0
282.00	547.0
283.00	1250.0
284.00	713.0
285.00	1880.0
286.10	455.0
288.90	429.0
289.90	301.0
290.90	206.0
292.00	544.0
293.00	2497.0
293.90	678.0
295.00	873.0
296.00	36360.0
297.00	5343.0
297.90	464.0
301.00	510.0
302.00	694.0
303.00	4600.0
304.00	1056.0
305.00	178.0
308.00	341.0
309.00	286.0
310.10	462.0
313.10	251.0
314.00	1823.0
315.00	3995.0
316.00	1905.0
317.10	346.0
320.00	182.0
321.00	857.0
322.10	690.0
323.00	11353.0
324.00	1999.0
324.90	274.0
325.80	196.0
326.90	2291.0
328.00	910.0
332.00	777.0
333.00	1320.0
334.00	8396.0
335.00	1950.0
336.20	285.0
339.00	339.0
341.10	1427.0
342.00	449.0
346.00	2502.0
347.00	560.0
352.00	3982.0
353.00	2615.0
354.10	3745.0
355.10	956.0
359.00	167.0
364.00	171.0
365.00	17992.0
366.00	2327.0
370.00	300.0
371.00	1030.0
372.10	6730.0
373.10	1306.0

Instrument :
BNA_M
ClientSampleId :
DFTPP040

377.10	155.0
383.00	1695.0
384.00	425.0
390.00	880.0
391.00	595.0
392.00	435.0
401.00	414.0
402.00	3028.0
403.10	3476.0
404.00	1471.0
405.00	288.0
414.90	215.0
421.00	2924.0
422.00	2741.0
423.10	21272.0
424.00	6023.0
425.00	738.0
437.90	165.0
439.80	219.0
441.10	53760.0
442.10	354176.0
443.10	66464.0
444.00	5753.0
445.10	462.0

Instrument :
BNA_M
ClientSampleId :
DFTPP040

m/z	Abundance
37.00	222.0
38.00	762.0
39.00	4031.0
40.00	847.0
44.00	438.0
49.10	638.0
50.00	15229.0
51.00	53840.0
52.00	2535.0
55.00	194.0
56.00	1888.0
57.00	3777.0
58.00	171.0
61.00	676.0
62.00	737.0
63.00	2632.0
64.10	359.0
65.00	1222.0
66.10	156.0
68.00	1181.0
69.00	66632.0
70.00	388.0
73.00	1271.0
74.00	7557.0
75.00	11585.0
76.00	3413.0
77.00	72624.0
78.10	4822.0
79.00	5727.0
80.00	4028.0
81.00	6166.0
82.00	1486.0
83.00	1169.0
84.00	378.0
85.00	1135.0
86.00	1896.0
86.90	737.0
87.90	300.0
91.00	1461.0
92.00	1733.0
93.00	10772.0
94.00	1086.0
94.90	175.0
95.90	613.0
97.00	255.0
98.00	7757.0
99.00	6593.0
100.00	734.0
101.00	3421.0
102.00	429.0
102.90	1085.0
104.00	2244.0
105.00	2362.0
106.00	805.0
107.00	28272.0
107.90	3734.0
109.10	767.0
110.00	42712.0
111.00	7187.0
112.00	1056.0
112.90	292.0
114.90	247.0
116.00	1682.0
117.00	26336.0
118.00	1816.0
119.00	411.0
119.90	380.0
120.80	171.0
122.00	2366.0
123.00	2669.0
124.00	1560.0
125.00	1431.0
126.10	531.0
127.00	94224.0
128.00	6774.0
129.00	38776.0
130.00	3098.0
131.00	760.0
132.10	353.0
132.90	528.0
134.00	1282.0
135.00	3143.0
136.00	1273.0
137.00	1785.0
138.00	306.0
139.00	180.0
140.10	448.0
141.00	5893.0
142.00	1623.0
142.90	1219.0
143.80	318.0
144.90	409.0

Instrument :
BNA_M
ClientSampleId :
DFTPP040

145.90	1043.0
147.00	3869.0
148.00	7340.0
149.00	1537.0
150.00	257.0
151.20	810.0
152.00	383.0
152.90	2425.0
154.00	1433.0
155.00	3016.0
156.00	4564.0
157.00	650.0
157.90	1254.0
158.90	683.0
160.00	2069.0
161.00	2313.0
161.90	689.0
162.90	402.0
163.80	245.0
164.90	2355.0
166.00	1252.0
167.00	12247.0
168.00	6014.0
168.90	1048.0
170.10	249.0
171.00	626.0
172.00	1090.0
173.00	1487.0
174.00	2255.0
175.00	4367.0
176.10	1280.0
176.90	2359.0
177.90	846.0
178.90	9307.0
180.00	6831.0
180.90	2991.0
182.00	347.0
183.00	293.0
184.00	856.0
185.00	4506.0
186.00	33896.0
187.00	9403.0
188.00	1014.0
189.00	2513.0
190.10	519.0
190.90	1586.0
192.00	3064.0
193.00	3939.0
193.90	776.0
195.00	359.0
196.00	6636.0
197.00	1404.0
197.90	234176.0
198.90	14751.0
200.00	1189.0
201.30	848.0
203.00	2006.0
204.00	9956.0
205.00	15446.0
206.00	65568.0
207.00	12028.0
208.00	2799.0
209.00	1273.0
210.10	1157.0
211.00	2749.0
211.80	578.0
212.90	220.0
215.00	898.0
216.00	1446.0
217.00	18744.0
217.90	2368.0
218.80	267.0
220.00	303.0
221.00	9985.0
221.80	2627.0
223.00	4008.0
224.00	39264.0
225.00	10282.0
226.10	990.0
227.00	19208.0
228.00	2660.0
229.00	3760.0
229.80	268.0
231.00	1722.0
232.00	248.0
232.90	243.0
233.90	1181.0
234.90	956.0
236.00	656.0
237.00	1390.0
237.80	206.0
239.00	679.0
239.90	593.0
241.00	739.0
242.00	1916.0
243.10	2164.0

Instrument :
BNA_M
ClientSampleId :
DFTPP040

244.10	29184.0
245.00	4686.0
246.00	7009.0
247.00	1579.0
247.80	240.0
248.90	1256.0
249.90	274.0
251.00	358.0
251.90	451.0
252.90	720.0
254.10	1055.0
255.00	172096.0
256.00	24608.0
257.00	2346.0
258.00	9853.0
259.00	1985.0
259.80	387.0
260.80	273.0
264.00	301.0
265.00	4535.0
265.90	636.0
267.00	175.0
269.90	285.0
270.80	283.0
271.90	624.0
273.00	5574.0
274.00	13457.0
275.00	84392.0
276.00	11433.0
277.00	7465.0
278.00	1235.0
279.00	404.0
281.00	1207.0
282.10	446.0
282.90	1095.0
283.90	598.0
285.00	1169.0
285.90	185.0
288.90	488.0
290.00	232.0
290.90	184.0
291.90	308.0
293.00	1768.0
294.00	647.0
295.00	841.0
296.00	27080.0
297.00	3308.0
297.90	220.0
300.90	415.0
302.00	639.0
303.00	2677.0
303.90	754.0
307.90	552.0
309.00	327.0
310.00	342.0
313.00	359.0
313.90	1396.0
315.00	2553.0
316.00	1612.0
316.90	317.0
321.00	898.0
322.00	275.0
323.00	8779.0
324.00	1713.0
325.90	297.0
326.90	2011.0
328.00	1015.0
328.90	184.0
332.00	603.0
332.90	764.0
334.00	5765.0
335.00	1304.0
336.10	274.0
338.90	306.0
340.10	236.0
341.00	1137.0
341.90	304.0
342.90	206.0
345.90	1791.0
347.00	312.0
350.90	187.0
352.00	2477.0
353.00	1862.0
354.10	2749.0
355.00	903.0
356.00	174.0
358.90	236.0
365.00	14332.0
365.90	1895.0
370.10	241.0
371.00	823.0
372.00	4670.0
373.10	1124.0
383.00	1544.0
384.00	323.0
385.00	179.0

Instrument :
BNA_M
ClientSampleId :
DFTPP040

390.10	524.0
391.00	349.0
392.00	373.0
401.00	344.0
402.00	2165.0
403.00	3106.0
403.90	1007.0
405.00	225.0
421.10	2304.0
422.00	2339.0
423.10	16928.0
424.00	4737.0
425.10	359.0
441.10	43000.0
442.10	293952.0
443.10	53952.0
444.00	4779.0
445.00	396.0

Instrument :

BNA_M

ClientSampleId :

DFTPP040