

DDT Breakdown

Date	Instrument Name	DFTPP Data File
3/17/2023	BNA_M	<u>BM039028.D</u>
Compound Name	Response	Retention Time
DDT	2103891	20.833
DDD	36517	20.392
DDE	3987	19.886
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
40504	2144395	1.89