




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031723\
Data File : BM039028.D
Acq On : 17 Mar 2023 09:12
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 18 03:20:42 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM030823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 17 05:25:34 2023
Response via : Initial Calibration



TIC: BM039028.D\data.ms

(70) Pentachlorophenol (C)




17.004min (-0.006) 157472.19 ng

response 745392

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.70	63.92
264.00	62.90	62.33
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031723\
Data File : BM039028.D
Acq On : 17 Mar 2023 09:12
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 18 03:20:42 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM030823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 17 05:25:34 2023
Response via : Initial Calibration



TIC: BM039028.D\data.ms

(77) Benzidine

19.598min (+ 0.000) 16052.26 ng

response 3914350

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.30	14.93
183.00	11.00	11.32
0.00	0.00	0.00